

### Invalid Samples in Maricopa County, AZ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
33497	VOID	843	5/15/03	FIELD/FINAL VAC = 0	TO-15

### Invalid Samples in South Phoenix, AZ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32805	VOID	041	4/3/03	VAC = 0	TO-15
35031	VOID	844	8/1/03	PRESSURE TOO LOW TO ANALYZE	TO-15

### Invalid Samples in St. John, AZ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32924	D1 - VOID	171	4/27/03	FINAL PRESSURE = 0"	TO-15
33397	VOID	3956	5/27/03	FIELD/LAB PRESSURE = 0	TO-15
33675	D1 - VOID	ER066	6/14/03	FINAL PRESS = 0	TO-15
33676	D2 - VOID	565	6/14/03	FINAL PRESS = 0	TO-15
33845	VOID	626	6/20/03	AMBIENT PRESS = 0	TO-15
35012	VOID	1478	8/13/03	CANISTER LEAK	TO-15
35013	VOID	1911	8/13/03	CANISTER LEAK	TO-15
35015	D2 - VOID	171	8/13/03	CANISTER LEAK	TO-15
37076	D1 - VOID	982	11/5/03	RECEIVED BEYOND HOLD TIME	TO-15
37077	D2 - VOID	2165	11/5/03	RECEIVED BEYOND 30 DAYS	TO-15

### Invalid Samples in Supersite, Phoenix, AZ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31529	VOID	1473	1/3/03	VAC = 0 DID NOT RUN	TO-15

### Invalid Samples in Denver, CO (Site 1 - DECO)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32989	VOID	1910	4/27/03	VOID BY LAB - VACUUM WENT TO AMBIENT	TO-15
34378	903334 - VOID		1/15/03	NO REASON GIVEN	METALS
34379	V00079 - VOID		1/21/03	NO REASON GIVEN	METALS
34421	V00105 - VOID		4/27/03	NO REASON GIVEN	METALS

### Invalid Samples in Denver, CO (Site 2 - WECO)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31717	D1 - VOID	1912	1/27/03	VAC = 20" TOO HIGH TO ANALYZE	TO-15
31718	D2 - VOID	1886	1/27/03	VAC = 20" TOO HIGH TO ANALYZE	TO-15
31719	D1 - VOID		1/27/03	INVALID CANISTER	TO-11A
31720	D2 - VOID		1/27/03	INVALID CANISTER	TO-11A
34064	903371		10/29/02	NO REASON GIVEN	METALS
34396	903410 - VOID		2/2/03	NO REASON GIVEN	METALS

### Invalid Samples in Fort Lauderdale, FL

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31540	VOID		1/15/03	DID NOT RUN	TO-11A

### Invalid Samples in Orlando, FL

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
36728	D1 - VOID		10/24/03	OUT OF HOLD TIME	TO-11A
36729	D2 - VOID		10/24/03	OUT OF HOLD TIME	TO-11A

### Invalid Samples in Tampa (Gandy), FL

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31828	KEEP FB AS VALID SAMPLE		2/2/03	POWER FAILURE	TO-11A

### Invalid Samples in Tampa (Lullwater), FL

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32361	VOID		3/16/03		TO-11A
32853	VOID		4/21/03	VOIDED BY FIELD	TO-11A

### Invalid Samples in Northbrook, IL

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
36895	VOID	22293	10/18/03	RECEIVED BEYOND 30 DAYS	TO-15
36896	VOID	21131	10/24/03	RECEIVED BEYOND 30 DAYS	TO-15
36897	VOID	21067	10/30/03	RECEIVED BEYOND 30 DAYS	TO-15
36978	VOID	21037	11/5/03	RECEIVED BEYOND 30 DAYS	TO-15
37380	VOID	21044	12/5/03	RECEIVED BEYOND HOLD TIME	TO-15
37381	VOID	21141	12/11/03	RECEIVED BEYOND HOLD TIME	TO-15
37385	VOID	21042	11/29/03	RECEIVED BEYOND HOLD TIME	TO-15

### Invalid Samples in Schiller Park, IL

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
36713	VOID	21052	9/30/03	RECEIVED BEYOND 30 DAYS	TO-15
36976	VOID	21029	10/12/03	RECEIVED BEYOND 30 DAYS	TO-15

### Invalid Samples in Detroit (Dearborn), MI

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32072	D1 - VOID	163	2/20/03	VOIDED IN FIELD	TO-15
32073	D2 - VOID	642	2/20/03	VOIDED IN FIELD	TO-15
32230	VOID	1895	3/4/03	VALVE OPEN VAC = 0	TO-15
34221	D1 - VOID	723	7/2/03	VAC = 30" - TOO HIGH TO ANALYZE	TO-15
34222	D2 - VOID	642	7/2/03	VAC = 30" - TOO HIGH TO ANALYZE	TO-15
34663	D1 - VOID	639	7/26/03	NO REASON GIVEN	TO-15
34665	D2 - VOID	ER054	7/26/03	NO REASON GIVEN	TO-15
35904	D1 - VOID	799	9/24/03	NO REASON GIVEN	TO-15
35906	D2 - VOID	089	9/24/03	NO REASON GIVEN	TO-15
36298	D1 - VOID	3952	10/18/03	VALVE OPEN VAC = 0	TO-15
36299	D2 - VOID	ER039	10/18/03	VALVE OPEN VAC = 0	TO-15
36465	D1 - VOID	658	10/30/03	OPERATOR ERROR	TO-15
36467	D2 - VOID	090	10/30/03	OPERATOR ERROR	TO-15

### Invalid Samples in E7MI, MI

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
33762	VOID	071	6/8/03	VAC = 26" - TOO HIGH TO ANALYZE	TO-15
34223	VOID	809	7/2/03	VAC = 30" - TOO HIGH TO ANALYZE	TO-15

### Invalid Samples in Houghton Lake, MI

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31442	VOID	702	1/3/03	NOT ANALYZED BY LAB	TO-15
31907	VOID	ER096	2/8/03	VOID BY FIELD	TO-15
32234	VOID	1479	3/4/03	NO VAC AT SITE	TO-15
33955	VOID	163	6/20/03	FINAL VAC = 0	TO-15
36477	VOID	607	10/30/03	CANISTER LEAK	TO-15
36636	VOID	ER015	11/11/03	SAMPLER ERROR	TO-15
37029	VOID	ER113	12/5/03	FINAL FIELD PRESSURE = 0	TO-15

### Invalid Samples in Intertribal Council, MI

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
33835	VOID	MI001	6/20/03	SAMPLER ERROR	TO-15
34255	VOID	MI008	7/8/03	VAC = 19" - TOO HIGH TO ANALYZE	TO-15
36479	VOID	4034	10/30/03	OPERATOR ERROR	TO-15
37311	VOID	698	12/29/03	VAC = 0	TO-15
36422	FB - VOID		10/24/03	RECEIVED LATE	TO-13A
36423	VOID		10/24/03	RECEIVED LATE	TO-13A
37312	VOID		12/29/03	OUT OF HOLD TIME	TO-13A

### Invalid Samples in Bonne Terre, MO

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31523	VOID	R5	1/12/03	SYSTEM NOT SWITCHED TO RUN MODE	SNMOC
32639	VOID	ER062	4/9/03	TIMER NOT PUT IN RUN MODE	SNMOC
34128	D2 - VOID	1886	6/26/03	CANISTER LOST VACUUM	SNMOC
36847	VOID	184	11/23/03	TIMER = 20.4 HRS	SNMOC
32636	VOID		4/9/03	TIMER NOT PUT IN RUN MODE	TO-11A
36848	FB - VOID		11/23/03	TIMER = 20.4 HRS	TO-11A
36849	VOID		11/23/03	TIMER = 20.4 HRS	TO-11A

### Invalid Samples in St. Louis, MO - Grant School

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
34315	VOID	1904	7/8/03	VAC = 0	SNMOC
36058	VOID	407	9/30/03	SAMPLER ERROR	SNMOC
33521	VOID		6/2/03	ONLY SCRUBBER SCENT	TO-11A
36059	VOID		9/30/03	SAMPLER ERROR	TO-11A

### Invalid Samples in Gulf Port, MS

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
34473	D2 - VOID	676	7/14/03	VAC = 30" - TOO HIGH TO ANALYZE	TO-15
36629	D1 - VOID	ER120	11/11/03	VAC = 20" - TOO HIGH TO ANALYZE	TO-15
36631	D2 - VOID	072	11/11/03	VAC = 20" - TOO HIGH TO ANALYZE	TO-15
36630	D1 - VOID		11/11/03	VAC = 20" - TOO HIGH TO ANALYZE	TO-11A
36632	D2 - VOID		11/11/03	VAC = 20" - TOO HIGH TO ANALYZE	TO-11A

### Invalid Samples in Jackson, MS

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32673	VOID	035	4/9/03	NO VAC IN CANISTER	TO-15
34475	D1 - VOID	632	7/14/03	SAMPLE COLLECTED FOR 2 DAYS	TO-15
34477	D2 - VOID	40790	7/14/03	SAMPLE COLLECTED FOR 2 DAYS	TO-15
34895	VOID	1898	8/7/03	VAC = 0	TO-15
32674	VOID		4/9/03	NO VAC IN CANISTER	TO-11A
34476	D1 - VOID		7/14/03	SAMPLE COLLECTED FOR 2 DAYS	TO-11A
34478	D2 - VOID		7/14/03	SAMPLE COLLECTED FOR 2 DAYS	TO-11A

### Invalid Samples in Pascagoula, MS

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
34479	D1 - VOID	770	7/14/03	VAC = 0	TO-15
34481	D2 - VOID	631	7/14/03	VAC = 0	TO-15
34691	VOID	1886	7/26/03	CANISTER LOST VACUUM	TO-15
34903	VOID	3506	8/7/03	VAC = 0	TO-15

### Invalid Samples in Tupelo, MS

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
33382	VOID	ER096	5/27/03	LOGIN PRESSURE = 0	TO-15

### Invalid Samples in Candor, NC

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
34441	VOID		7/2/03	OPERATOR LEFT CARB TUBE IN CAR	TO-11A

### Invalid Samples in Beulah, ND

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32165	D1 - VOID	4005	2/26/03	VALVE LEFT OPEN	TO-15
32166	D2 - VOID	1908	2/26/03	VALVE LEFT OPEN ON DUP	TO-15

### Invalid Samples in Lincoln, NE (Site 1 - LINE)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31490	VOID	ER023	1/9/03	CARB TUBE NOT SENT TO SITE	TO-15
31728	D2 - VOID		1/27/03	INVALID CANISTER. KEEP D1 AS VALID SAMPLE	TO-11A

### Invalid Samples in Lincoln, NE (LONE)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31726	D2 - VOID	1904	1/27/03	NO VAC IN CAN - KEEP D1	TO-15

### Invalid Samples in Belvidere, NJ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32351	VOID	135	3/13/03	VAC = 30" TOO HIGH TO ANALYZE	TO-15/MEOH
33360	C1R1 - D1 VOID	1887	5/20/03	VALVE OPEN 0 VAC	TO-15/MEOH

### Invalid Samples in Camden, NJ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32669	VOID	723	4/9/03	NO VAC IN CANISTER	TO-15
32951	VOID	ER090	4/27/03	VAC = 0"	TO-15
35048	VOID	1482	8/13/03	VAC = 0	TO-15
32670	VOID		4/9/03	NO VAC IN CANISTER	TO-11A
32874	VOID		4/21/03	ELAPSED TIME NOT LONG ENOUGH	TO-11A
32952	VOID		4/27/03	VAC = 0"	TO-11A
35049	VOID - FB		8/13/03	VAC = 0	TO-11A
35050	VOID		8/13/03	VAC = 0	TO-11A



### Invalid Samples in Chester, NJ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31468	VOID	ER060	1/3/03	NOT ANALYZED BY LAB	TO-15
31699	D1 - VOID	115	1/27/03	CANISTER INSTALLED INCORRECTLY	TO-15
31700	D2 - VOID	ER095	1/27/03	CANISTER INSTALLED INCORRECTLY	TO-15
32264	VOID	ER119	3/10/03	FIELD FINAL VAC = 0"	TO-15
34174	VOID	ER113	7/2/03	FINAL VAC = 0	TO-15
31701	D1 - VOID		1/27/03	INVALID CANISTER	TO-11A
31702	D2 - VOID		1/27/03	INVALID CANISTER	TO-11A
32266	VOID		3/10/03	INVALID CANISTER	TO-11A
32265	FB - VOID		3/10/03	INVALID CANISTER	TO-11A
34175	VOID		7/2/03	FINAL VAC = 0	TO-11A

### Invalid Samples in Elizabeth, NJ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31987	VOID	ER020	2/14/03	VAC = 22" - TOO HIGH TO ANALYZE	TO-15
32066	VOID	575	2/20/03	VACUUM INSIDE CAN TOO HIGH	TO-15
32135	D1 - VOID	1914	2/26/03	VAC = 16" - TOO HIGH TO ANALYZE	TO-15
32137	D2 - VOID	1801	2/26/03	VAC = 16" - TOO HIGH TO ANALYZE	TO-15
32632	VOID	189	4/3/03	POWER FAILURE	TO-15
31988	VOID		2/14/03	INVALID CANISTER	TO-11A
32068	VOID		2/20/03	CAN VOIDED	TO-11A
32136	D1 - VOID		2/26/03	INVALID CANISTER	TO-11A
32138	D2 - VOID		2/26/03	INVALID CANISTER	TO-11A
32633	VOID		4/3/03	POWER FAILURE	TO-11A
32067	FB - VOID		2/20/03	CAN VOIDED	TO-11A

### Invalid Samples in New Brunswick, NJ

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32183	VOID	3947	3/4/03	VACUUM TOO HIGH	TO-15
32280	VOID	1878	3/10/03	VAC TOO HIGH TO ANALYZE	TO-15
32345	VOID	702	3/16/03	VAC = 30" - TOO HIGH TO ANALYZE	TO-15
32438	VOID	089	3/22/03	VAC = 30" - TOO HIGH TO ANALYZE	TO-15
33269	D1 - VOID	129	5/15/03	VAC = 0	TO-15
33271	D2 - VOID	ER043	5/15/03	VAC = 0	TO-15
33668	VOID	ER094	6/14/03	FIELD/LAB FINAL PRESS = 0	TO-15
34333	D1 - VOID	799	7/14/03	VAC = 0	TO-15
34335	D2 - VOID	652	7/14/03	VAC = 0"	TO-15
35980	VOID	4041	9/24/03	VALVE OPEN VAC = 0	TO-15
37111	VOID	ER096	12/14/03	VALVE OPEN VAC = 0	TO-15
32184	VOID		3/4/03	VACUUM TOO HIGH TO ANALYZE	TO-11A
32281	FB - VOID		3/10/03	INVALID CANISTER	TO-11A
32282	VOID		3/10/03	INVALID CANISTER	TO-11A
32346	FB - VOID		3/16/03	CANISTER VOIDED	TO-11A
32347	VOID		3/16/03	CANISTER VOIDED	TO-11A
32439	VOID		3/22/03	INVALID CANISTER	TO-11A
33270	D1 - VOID		5/15/03	VAC = 0	TO-11A
33272	D2 - VOID		5/15/03	VAC = 0	TO-11A
33669	FB - VOID		6/14/03	INVALID CANISTER	TO-11A
33670	VOID		6/14/03	INVALID CANISTER	TO-11A
34334	D1 - VOID		7/14/03	VAC = 0	TO-11A
34336	D2 - VOID		7/14/03	VAC = 0"	TO-11A
35981	VOID		9/24/03	VALVE OPEN VAC = 0	TO-11A
37112	VOID		12/14/03	VALVE OPEN VAC = 0	TO-11A
36933	VOID		11/23/03	RECEIVED BEYOND EXTRACTION TIME	TO-13A
37189	VOID		12/5/03	RECEIVED BEYOND EXTRACTION TIME	TO-13
37190	VOID		12/11/03	RECEIVED BEYOND EXTRACTION TIME	TO-13

### Invalid Samples in Barceloneta, PR

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31750	D1 - VOID	ER024	1/27/03	VAC = 14" TOO HIGH TO ANALYZE	TO-15
31751	D2 - VOID	ER090	1/27/03	VAC = 14" TOO HIGH TO ANALYZE	TO-15
31864	VOID	4039	2/8/03	VAC = 14" TOO HIGH TO ANALYZE	TO-15
31752	D1 - VOID		1/27/03	INVALID CANISTER	TO-11A
31753	D2 - VOID		1/27/03	INVALID CANISTER	TO-11A

### Invalid Samples in San Juan, PR

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31458	VOID	001	1/3/03	NOT ANALYZED BY LAB	TO-15
33376	VOID	3947	5/27/03	VAC = 0	TO-15
33627		663	6/8/03		TO-15
34010		115	6/20/03		TO-15
34226		084	7/2/03		TO-15
34339		ER076	7/14/03		TO-15
34737		3500	7/26/03		TO-15
34907		1883	8/7/03		TO-15

### Invalid Samples in Sioux Falls, SD

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31426	VOID	692	1/3/03	NO CARBS SAMPLED	TO-15
31758	D1 - VOID	3947	1/27/03	ARRIVED AT SITE WITH NO VAC.	TO-15
31759	D2 - VOID	1895	1/27/03	ARRIVED AT SITE WITH NO VAC.	TO-15
32101	D1 - VOID	662	2/20/03	VACUUM INSIDE CAN TOO HIGH	TO-15
32109	D2 - VOID	080	2/26/03	VACUUM TOO HIGH TO ANALYZE	TO-15
32491	D1 - VOID	400	3/28/03	VAC = 18" - TOO HIGH TO ANALYZE	TO-15
32493	D2 - VOID	678	3/28/03	VAC = 18" - TOO HIGH TO ANALYZE	TO-15
33303	D1 - VOID	723	5/15/03	VAC = 22" - TOO HIGH TO ANALYZE	TO-15
33305	D2 - VOID	807	5/15/03	VAC = 16" - TOO HIGH TO ANALYZE	TO-15
36345	D1 - VOID	4022	10/24/03	MFC SETTING INCORRECT	TO-15
36347	D2 - VOID		10/24/03	MFC SETTING INCORRECT	TO-15
37298	VOID	1401	12/23/03	VAC LIMIT EXCEEDED	TO-15
37300	VOID	137	12/29/03	VAC LIMIT EXCEEDED	TO-15
31760	D1 - VOID		1/27/03	INVALID CANISTER	TO-11A
31761	D2 - VOID		1/27/03	INVALID CANISTER	TO-11A
32108	D1 - VOID		2/26/03	VACUUM TOO HIGH TO ANALYZE	TO-11A
32110	D2 - VOID		2/26/03	VACUUM TOO HIGH TO ANALYZE	TO-11A
32492	D1 - VOID		3/28/03	INVALID CANISTER	TO-11A
32494	D2 - VOID		3/28/03	INVALID CANISTER	TO-11A
33304	D1 - VOID		5/15/03	VAC = 22" - TOO HIGH TO ANALYZE	TO-11A
33306	D2 - VOID		5/15/03	VAC = 16" - TOO HIGH TO ANALYZE	TO-11A
36346	D1 - VOID		10/24/03	MFC SETTING I CORRECT	TO-11A
36348	D2 - VOID		10/24/03	MFC SETTING INCORRECT	TO-11A
37299	VOID		12/23/03	INVALID CANISTER	TO-11A
37301	FB - VOID		12/29/03	INVALID CANISTER	TO-11A
37302	VOID		12/29/03	INVALID CANISTER	TO-11A

### Invalid Samples in Kingsport, TN

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31691	VOID	TNAPC-4	1/21/03	POWER FAILURE	TO-15
32270	D1 - VOID	TNAPC-10	2/26/03	COULD NOT OPEN VALVE IN FIELD	TO-15
33758	VOID	628	6/14/03	SAMPLER MALFUNCTION	TO-15
31503	D2 VOID		1/9/03	CARTRIDGE WAS AN OZONE SCRUBBER	TO-11A
31692	VOID		1/21/03	POWER FAILURE	TO-11A
31693	FB - VOID		1/21/03	POWER FAILURE	TO-11A
32271	D1 - VOID		3/4/03	INVALID CANISTER	TO-11A
33759	FB - VOID		6/14/03	SAMPLER MALFUNCTION	TO-11A
33760	VOID		6/14/03	SAMPLER MALFUNCTION	TO-11A

### Invalid Samples in Nashville, TN (Site 1 - EATN)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31422	VOID	TN004	1/3/03	POWER FAILURE AT SITE	TO-15
31711	D1 - VOID	TN002	1/27/03	POWER FAILURE AT SITE	TO-15
31712	D2 - VOID	TN004	1/27/03	POWER FAILURE AT SITE	TO-15
34188	VOID	TN010	7/2/03	SAMPLE DID NOT RUN - SAMPLER ERROR	TO-15
34430	VOID	TN007	7/14/03	SAMPLER ERROR	TO-15
35421	VOID	TN007	8/31/03	SAMPLER ERROR CAN P LIMIT EXCEEDED	TO-15
36304	VOID	162	10/18/03	SAMPLER ERROR	TO-15
37002	VOID	607	12/5/03	CANISTER VAC AT = 0	TO-15
31423	VOID		1/3/03	POWER FAILURE AT SITE	TO-11A
31713	D1 - VOID		1/27/03	POWER FAILURE AT SITE	TO-11A
31714	D2 - VOID		1/27/03	POWER FAILURE AT SITE	TO-11A
34189	VOID		7/2/03	SAMPLE DID NOT RUN - SAMPLER ERROR	TO-11A
34431	VOID		7/14/03	SAMPLER ERROR	TO-11A
35422	VOID -FB		8/31/03	SAMPLER ERROR	TO-11A
35423	VOID		8/31/03	SAMPLER ERROR	TO-11A
36305	VOID		10/18/03	SAMPLER ERROR	TO-11A
37003	VOID		12/5/03	VAC = 0	TO-11A
32210	7040830 - VOID		1/27/03	POWER OUTAGE	METALS
32211	7040831- VOID		1/27/03	POWER OUTAGE	METALS
36306	VOID		10/18/03	SAMPLER ERROR	METALS

## Invalid Samples in Nashville, TN (Site 2 - LOTN)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
31695	D1 - VOID	TN005	1/27/03	POST FLOW LIMIT EXCEEDED	TO-15
31696	D2 - VOID	TN016	1/27/03	POST FLOW LIMIT EXCEEDED	TO-15
33559	VOID	TN001	6/8/03	VOIDED BY FIELD PRESSURE LIMIT EXCEEDED	TO-15
33906	VOID	TN008	6/20/03	USED CAN PRESSURE LIMIT EXCEEDED	TO-15
34433	VOID	TN014	7/14/03	SAMPLER ERROR	TO-15
34714	VOID	TN015	7/26/03	SAMPLER ERROR	TO-15
35417	VOID	TN006	8/31/03	SAMPLER ERROR CAN PRESSURE LIMIT EXCEEDED	TO-15
35783	D1 - VOID	TN002	9/12/03	INVALID RUN	TO-15
35785	D2 - VOID	TN006	9/12/03	INVALID SAMPLE	TO-15
36302	VOID	551	10/18/03	SAMPLER ERROR	TO-15
36425	VOID	TN014	10/30/03	SAMPLER ERROR	TO-15
37210	VOID	501	12/17/03	SAMPLE MALFUNCTION	TO-15
37354	VOID	1408	12/29/03	SAMPLER ERROR	TO-15
31697	D1 - VOID		1/27/03	INVALID CANISTER	TO-11A
31698	D2 - VOID		1/27/03	INVALID CANISTER	TO-11A
33560	VOID		6/8/03	INVALID CANISTER	TO-11A
33907	FB - VOID		6/20/03	INVALID CANISTER	TO-11A
33908	VOID		6/20/03	INVALID CANISTER	TO-11A
34434	VOID		7/14/03		TO-11A
34715	VOID - FB		7/26/03	SAMPLER ERROR	TO-11A
34716	VOID		7/26/03	SAMPLER ERROR	TO-11A
35418	VOID -FB		8/31/03	SAMPLER ERROR	TO-11A
35419	VOID		8/31/03		TO-11A
35784	D1 - VOID		9/12/03	INVALID SAMPLE	TO-11A
35786	D2 - VOID		9/12/03	INVALID SAMPLE	TO-11A
36303	VOID		10/18/03	SAMPLER ERROR	TO-11A
36426	FB - VOID		10/30/03	SAMPLER ERROR	TO-11A
36427	VOID		10/30/03	SAMPLER ERROR	TO-11A
37211	VOID		12/17/03	SAMPLER MALFUNCTION	TO-11A
37355	FB - VOID		12/29/03	SAMPLER ERROR	TO-11A
37356	VOID		12/29/03	SAMPLER ERROR	TO-11A

### Invalid Samples in Bountiful, UT (Site 1 - BOUT)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
32615	VOID	844	4/3/03	VOID AT FIELD-FLOW CONTROLLER SET TO 0	TO-15/SNMOC
32616	VOID		4/3/03	VOID AT FIELD-FLOW CONTROLLER SET TO 0	TO-11A

### Invalid Samples in Bountiful, UT (Site 2- BTUT)

ERG ID	Comments	Canister #	Date Sampled	Invalid Reason	Analysis Method
35915	VOID	151	8/13/03	RECEIVED OUT OF HOLD TIME	TO-15/SNMOC
36143	VOID	1432	10/6/03	VAC = 0	TO-15/SNMOC
36501	VOID	ER106	9/30/03	RECEIVED BEYOND 30 DAYS	TO-15/SNMOC
37019	VOID	ER043	11/29/03	ELAPSED TIME = 0	TO-15/SNMOC
37253	VOID	680	7/26/03	OUT OF HOLD TIME	TO-15/SNMOC
37254	VOID	046	12/17/03	SAMPLER ERROR	TO-15/SNMOC
35916	VOID		8/13/03	RECEIVED OUT OF HOLD TIME	TO-11A
36144	VOID		10/6/03	INVALID CANISTER	TO-11A
36502	FB - VOID		9/30/03	RECEIVED BEYOND 30 DAYS	TO-11A
36503	VOID		9/30/03	RECEIVED BEYOND 30 DAYS	TO-11A
37020	VOID		11/29/03	ELAPSED TIME = 0	TO-11A
37252	VOID		12/17/03	INVALID CANISTER	TO-11A