

TECHNICAL SUPPORT SECTION EFFICACY REVIEW - I

Disinfectants Branch

IN 11-14-85

OUT 11-18-85

Reviewed By Dorothy M. Portner *DP* *11/18/85* Date 11-18-85

EPA Reg. No. 46781-1 & 46781-2

EPA Petition or EUP No. None

Date Division Received 10-17-85

Type Product Disinfecting/Sterilizing Solution

Data Accession No(s).

Product Manager PM-31 (Lee)

Product Name Metricide MX-1400 & Metricide MX-2800

Company Name Metrex Reserach Corporation

Submission Purpose Resubmission of a revised proposed re-use
protocol

Type Formulation Liquid concentrate to be activated and used
undiluted

Active Ingredient(s): 8

Glutaraldehyde.....2.00

Technical Support Section Efficacy Review-II

Disinfectants Branch

EPA.Reg. No.or File Symbol 46781-1 & 46781-2

Data Division Received 10-17-85

Data Accession No(s). _____

Product Manager No. PM 31 (Lee)

Product Name Metricide MX-1400 & Metricide MX-2800

Company Name Metrex Research Corporation

The revised proposed re-use test protocol for a manual system, submitted with your letter of October 17, 1985, incorporates the minimal specifications required for conducting a simulated re-use study that were indicated in our letter of October 7, 1985. This proposed re-use test protocol is acceptable as a general guidance outline. The specific testing details actually employed in each study conducted in accordance with the protocol outline should also accompany the derived data submitted for evaluation.