

TECHNICAL SUPPORT SECTION EFFICACY REVIEW: INCOMPLETE SUBMISSION

Disinfectants Branch

IN 05-23-88

OUT 07-28-88

EPA Reg. No. or File Symbol 46781-G

Product Name MetriCide Plus 14™

Company Name Metrex Research Corporation

Date Application Received 05-19-88

Submission Purpose Resubmission to register "MetriCide Plus 14™" based on Metrex
Research Corporation's "MetriCide MX-1400", EPA Reg.No.46781-1

Type Product Hospital Disinfectant/Stelilant

Product Manager 31(Lee)

Active Ingredients 8

Glutaraldehyde.....3.2

Recommendations:

Confirmatory Efficacy Data Not required

The subject product is identical in formulation to the referenced product with the exception of increase in the concentration of the active ingredient. Therefore, no confirmatory efficacy data are required to be submitted.

Additional Data Required to Support Efficacy

To fulfill the tuberculocidal data requirements, validation data must be developed on one reused sample from manual systems (14 days in bucket) at 10 minutes and 20°C exposure by a laboratory other than the one which developed the original data (refer to the attached TB Data Call-In Notice) or change recommended exposure temperature from 20°C to 25°C.

Labeling:

In lieu of tuberculocidal validation data change exposure temperature to 25°C.

No other adverse comments.

Reviewed By Srinival Gowda
Srinivas Gowda

WEC
7/29/88

Date 07-28-88