

PM 32

675-15

10.11.90

46

ACCEPTED  
with COMMENTS  
in EPA Letter Dated:

FEB 12 1990

Front/Back Panel

Under the Federal Insecticide,  
Fungicide, and Rodenticide Act  
as amended, for the pesticide  
registered under EPA Reg. No.

675-15

**LEHN & FINK  
INSTRUMENT GERMICIDE  
FOR COLD DISINFECTION OF SURGICAL AND DENTAL INSTRUMENTS  
(Contains rust inhibitor)**

**GERMICIDAL\*** . Staphylocidal . Pseudomonacidal . Tuberculocidal  
**FUNGICIDAL\*** (Against pathogenic fungi)  
**VIROCIDAL\*** (Against Influenza A<sub>2</sub> (Japan), Herpes Simplex Type 1, Adenovirus  
Type 2, Vaccinia, and HIV-1 (AIDS Virus))

\*On Environmental Surfaces

**ACTIVE INGREDIENTS:** o-Phenylphenol 0.518%, o-Benzyl-p-chlorophenol  
0.250%.

**INERT INGREDIENTS:** 99.232%  
Contains sodium nitrite


EPA REGISTRATION NO. 675-15  
EPA EST. NO. 777-OH-1

**KEEP OUT OF REACH OF CHILDREN**

**CAUTION:** Harmful if swallowed. Avoid contact with skin and eyes.  
Avoid contamination of food. See side panel for additional precautionary  
statements.

**NET CONTENTS ONE GALLON (3.79 liters)**

**NATIONAL LABORATORIES  
LEHN & FINK PRODUCTS GROUP  
Montvale, N.J. 07645**



Side Panel (left)

FEB 12 1990

**HOW TO USE  
LEHN & FINK  
INSTRUMENT  
GERMICIDE**

Under the Federal Insecticide,  
Fungicide, and Rodenticide Act  
as amended, for the pesticide  
registered under EPA Reg. No.

675-15

**DIRECTIONS FOR USE:** It is a violation of federal law to use this product in a manner inconsistent with its labeling. Lehn & Fink Instrument Germicide is recommended for use in hospitals, nursing homes, dental offices, and other institutional facilities.

**CLEANING SURGICAL AND DENTAL INSTRUMENTS** Instruments must be cleaned prior to disinfection to remove all adhering blood and serous exudates, body fluids, and human wastes.

**DISINFECTING SURGICAL AND DENTAL INSTRUMENTS** Immerse instruments in Lehn & Fink Instrument Germicide for 15 minutes to wet all surfaces thoroughly. Flush and fill the lumen of hollow instruments/equipment, avoiding entrapment of air bubbles. If instruments are to remain in Lehn & Fink Instrument Germicide for substantially longer periods of time, add 1 gram of sodium bicarbonate to each quart of solution; stir to dissolve completely. Lehn & Fink Instrument Germicide is used for cold disinfection of medical/surgical and dental instruments. Fresh solution must be prepared at least daily if the solution becomes diluted or soiled.

**FOR TUBERCULOCIDAL EFFECTIVENESS** Thoroughly clean instruments to remove excess organic debris. Rinse and rough dry before immersing in Lehn & Fink Instrument Germicide for 10 minutes at 20°C (68°F).

**HYPODERMIC NEEDLES AND SYRINGES** Flush with Lehn & Fink Instrument Germicide, immerse for 15 minutes, then flush with sterile distilled water.

**RUBBER GLOVES, TUBES AND OTHER RUBBER ARTICLES** Immerse for not over 15 minutes. Remove, then immerse in 70% alcohol for one minute.

**EYE AND DELICATE LENS INSTRUMENTS** To disinfect instruments with optical systems, rinse off mucus and exudates and place in Lehn & Fink Instrument Germicide for 15 minutes but not longer. Assure thorough contact by flushing. On removal flush and rinse with sterile water. When not using instrument immediately, place in sterile wrap for future use.

If rapid drying is required, immerse for a few seconds in alcohol after removing from the Lehn & Fink Instrument Germicide.

**NOTE:** Before using Lehn & Fink Instrument Germicide, consult the instrument manufacturer's instructions for care and disinfection of their instrument.

Lehn & Fink Instrument Germicide may cause damage to certain types of plastic trays. Refer to manufacturer's instructions for proper use.

Side Panel

ACCEPTED  
with COMMENTS  
to EPA Letter Dated:


FEB 12 1990

Under the Federal Insecticide,  
Fungicide, and Rodenticide Act  
as amended, for the pesticide  
registered under EPA Reg. No

675-15

FOR COLD DISINFECTION  
OF SURGICAL  
AND DENTAL  
INSTRUMENTS

Contains Rust Inhibitor



KILLS HUMAN IMMUNODEFICIENCY VIRUS TYPE 1 (HIV-1) (AIDS VIRUS) ON  
PRECLEANED ENVIRONMENTAL SURFACES/OBJECTS PREVIOUSLY SOILED WITH BLOOD/BODY  
FLUIDS.

SPECIAL INSTRUCTIONS FOR CLEANING AND DECONTAMINATION AGAINST HIV-1 OF  
SURFACES/OBJECTS SOILED WITH BLOOD/BODY FLUIDS.

PERSONAL PROTECTION: When handling items soiled with blood or body fluids  
use disposable latex gloves, gowns, masks, or eye coverings.

CLEANING PROCEDURES: Blood and other body fluids must be thoroughly  
cleaned from surfaces and objects before application of Lehn & Fink  
Instrument Germicide.

CONTACT TIME: Allow surface to remain wet for 10 minutes.


DISPOSAL OF INFECTIOUS MATERIALS: Blood and other body fluids should be  
autoclaved and disposed of according to local regulations for infectious  
waste disposal.

ACCEPTED  
with COMMENTS  
to EPA Letter Dated:

FEB 12 1990

Under the Federal Insecticide,  
Fungicide, and Rodenticide Act  
as amended, for the pesticide  
registered under EPA Reg. No.

675-15



ACCEPTED  
with COMMENTS  
in EPA Letter Dated:

FEB 12 1990

Under the Federal Insecticide,  
Fungicide, and Rodenticide Act  
as amended, for the pesticide  
registered under EPA Reg. No.

67575

**FOR COLD DISINFECTION  
OF SURGICAL  
AND DENTAL  
INSTRUMENTS**

**Contains Rust Inhibitor**

**PRECAUTIONARY STATEMENTS: HAZARDS TO HUMANS AND DOMESTIC ANIMALS**

**CAUTION:** Harmful if swallowed. May cause eye and skin irritation. Avoid contact with eyes or skin. Avoid contamination of food.


**STATEMENT OF PRACTICAL TREATMENT**

In case of eye contact, IMMEDIATELY flush eyes thoroughly with plenty of water. Remove any contact lenses and continue to flush with plenty of water for at least 15 minutes. Get medical attention. If on skin, wash with plenty of soap and water. If irritation persists, get medical attention. If swallowed, do not induce vomiting. Drink a large amount of water. Call a physician.

**STORAGE AND DISPOSAL**

**STORAGE:** Store in original container in areas inaccessible to small children. Keep securely closed.

**DISPOSAL:** Do not reuse empty container. Wrap and discard in trash.



FIFRA Labeling Information

Ingredient Statement - General

- (A) C-1 An ingredient statement must declare the name and percentage of each active ingredient and the total percentage of all inert ingredients. The following form of ingredient statement complies with the Federal Insecticide, Fungicide, and Rodenticide Act, as amended by the Federal Environmental Pesticide Control Act of 1972:

ACTIVE INGREDIENTS

Name of ingredient . . . . .	_____	wt
Name of ingredient . . . . .	_____	wt
Name of ingredient . . . . .	_____	wt

INERT INGREDIENTS . . . . . \_\_\_\_\_ wt

Products containing arsenic in any form must also declare the percentages of total and water soluble arsenic, calculated as elementary arsenic.

The Confidential Statement of Formula (EPA Form 8570-4), must be submitted for the Registration Division's files, giving the names and percentage by weight of each ingredient, both active and inert.

- (A) 2. The words, "ACTIVE INGREDIENT(S)" and "INERT INGREDIENT(S)," must be printed in type of the same size and be given equal prominence. The words "ACTIVE INGREDIENT(S)" and "INERT INGREDIENT(S)" must be aligned to the same margin.

