

U.S. ENVIRONMENTAL PROTECTION AGENCY

Office of Pesticide Programs Antimicrobials Division (7510C) 1200 Pennsylvania Avenue NW Washington, D.C. 20460

NOTICE OF PESTICIDE:

x Registration
Reregistration

(under FIFRA, as amended)

EPA Reg.

Number:

Date of

111N 2 9 20101

Term of Issuance:

Conditional

1043-125

Name of Pesticide Product:

HASTe-SSD Component A

Name and Address of Registrant (include ZIP Code):

Steris Corporation 7501 Page Avenue St Louis, MO 63133

Note (Changes in labeling differing in substance from that accepted in connection with this registration must be submitted to and accepted by the Registration Division prior to use of the label in commerce; linkany correspondence on this product always refer to the above EPA registration number.

On the basis of information furnished by the registrant, the above named pesticide is hereby registered/reregistered under the Federal Insecticide, Fungicide and Rodenticide Act.

Registration is in no way to be construed as an endorsement or recommendation of this product by the Agency. In order to protect health and the environment, the Administrator, on his motion, may at any time suspend or cancel the registration of a pesticide in accordance with the Act. The acceptance of any name in connection with the registration of a product under this Act is not to be construed as giving the registrant a right to exclusive use of the name or to its use if it has been covered by others.

This product is conditionally registered in accordance with FIFRA sec 3(c)(7)(A) provided that you:

- 1. Submit and/or cite all data required for registration of your product under FIFRA sec. 3(c)(5) when the Agency requires all registrants of similar products to submit such data; and submit acceptable responses required for re-registration of your product under FIFRA section 4.
 - 2. Make the labeling changes listed below before you release the product for shipment:
 - a. Revise the "EPA Registration Number to read, "EPA Reg. No."1-43-125".

Signature of Approving Offical:

Marshall Swindell

Product Manager Team-33

Regulatory Management Branch I Antimicrobials Division (7510P) Date:

JUN 2 9 2010

Page 2 EPA Registration No. 1043-125

3. Submit two (2) copies of your final printed labeling before distributing or selling the product bearing the revised labeling.

If these conditions are not complied with, the registration will be subject to cancellation in accordance with FIFRA sec. 6(e). Your release for shipment of the product constitutes acceptance of these conditions.

A stamped copy of the label is enclosed for your records.

If you have any questions concerning this Notice of Registration, please contact Martha Terry at (703) 308-6217.

Sincerely,

Marshall Swindell Product Manager 33

Regulatory Management Branch I Antimicrobials Division (7510P)

Enclosure: (Stamped Label)

HASTe-SSD - Component A

For Use Only with HASTe-SSD - Component B

Use this product on washable, hard, non-porous environmental surfaces: floors, walls, counters in healthcare facilities, hospitals, medical and dental offices.

One Step Cleaner Disinfectant

Sterilant/Sporicidal, Bactericidal, Tuberculocidal, Fungicidal, *Virucidal

Active Ingredients:

Tetraacetyl ethylenediamine

61.6%

Inert Ingredients:

<u>38.4%</u>

Total

100.0%

KEEP OUT OF REACH OF CHILDREN

CAUTION

See side panel for precautionary statements

FIRST AID	
	•Hold eye open and rinse slowly and gently with water for 15-20
IF IN EYES	minutes.
	•Remove contact lenses, if present, after the first 5 minutes, then
	continue rinsing the eye.
	Call a poison control center or doctor for treatment advice.
	Take off contaminated clothing.
IF ON SKIN OR ON CLOTHING	•Rinse skin immediately with plenty of water for 15-20 minutes.
	Call a poison control center or doctor for treatment.
	Call poison control center or doctor immediately for treatment
IF SWALLOWED	advice.
	•Have person sip a glass of water if able to swallow.
	•Do not induce vomiting unless told to do so by the poison control
	center or doctor.
	•Do not give anything by mouth to an unconscious person.
HOT LINE NUMBER	
Have the product container or label with you when calling a poison control center or doctor, or going for	

Have the product container or label with you when calling a poison control center or doctor, or going for treatment. You may also contact Chemtrec at 800-424-9300 for emergency medical treatment information.

NET CONTENTS:

STERIS Corporation 7501 Page Avenue St. Louis, MO 63133 U.S.A. EPA Reg. No. 1043-XXX EPA Est. No.

ACCEPTED with COMMENTS
EPA Letter Dated:

JUN 2 9 2010

Under the Federal Insecticide, Fungicide, and Rodenticide Act as amended, for the pesticide, registered under EPA Reg. No.

Page 1 of 4 Draft 8/20/09

PRECAUTIONARY STATEMENTS

Hazard to Humans and Domestic Animals: CAUTION: Causes moderate eye irritation. Harmful if swallowed or absorbed through skin. Do not get in eyes, on skin, or on clothing. Wash hands before eating, drinking, chewing gum, using tobacco or using the toilet. Remove contaminated clothing and wash hands before reuse

For the end use solution, causes moderate eye irritation. Do not get in eyes, on skin, or on clothing. Avoid breathing vapor or mist. Use in a well-ventilated area. Wash hands before eating, drinking, chewing gum, using tobacco or using the toilet. Remove contaminated clothing and wash hands before reuse.

PRODUCT CLAIMS	
Sterilant/Sporicidal [<i>B. subtilis</i> (ATCC 19659), <i>C. sporogenes</i> (ATCC 3584)]	1 hour 20°C
Mouse parvovirus (Isolate, Yale University) in the presence of 5% organic soil	10 minutes 20°C
Clostridium difficile spores (ATCC 43599) in the presence of 5% organic soil	5 minutes 20°C
Hospital Disinfectant [S. aureus (ATCC 6538), S. enterica (ATCC 10708), Ps. aeruginosa (ATCC 15442)] in the presence of 5% organic soil	2 minutes 20°C
The following organisms were all tested in the presence of 5%	1 minute 20°C
organic soil:	ACCEPTED
Tuberculocidal (<i>Mycobacterium bovis</i> BCG) (Clinical Isolate)	with COMMENTS
*HIV-1** (Clinical Isolate) Tricophyton mentagrophytes (ATCC 9533)	EPA Letter Dated:
Candida albicans (ATCC 10231)	0 2010:
MRSA (ATCC 33591)	JUN 2 9 2010
VRE (ATCC 51299)	Under the Federal insections,
ESBL (ATCC)	Pungicide, and Rodenticide Act
Klebisellla pneumoniae (ATCC)	amended, for the pesticide,
Acinetobacter baumannii (ATCC 19606)	registered under EPA Reg No.

**KILLS HIV- 1 ON PRE-CLEANED ENVIRONMENTAL SURFACES/OBJECTS PREVIOUSLY SOILED WITH BLOOD/BODILY FLUIDS in health care settings or other settings in which there is an expected likelihood of soiling inanimate surfaces/objects with blood or body fluids, and in which the surfaces/objects likely to be soiled with blood or body fluids can be associated with the potential for transmission of HIV-1 (Human Immunodeficiency Virus Type 1 - associated with AIDS).

SPECIAL INSTRUCTIONS FOR CLEANING AND DECONTAMINATION AGAINST HIV OF SURFACES/OBJECTS SOILED WITH BLOOD/BODY FLUIDS:

PERSONAL PROTECTION: Specific barrier protection items to be used when handling items soiled with blood or body fluids are disposable latex gloves, gowns, masks and eye coverings.

CLEANING PROCEDURE: Blood and other body fluids must be thoroughly cleaned from surfaces and objects before the application of the sterilant.

Clostridium difficile: Use only on hard, non-porous surfaces.

Personal Protection: Wear appropriate barrier protection such as gloves, gowns, masks or eye covering.

Cleaning Procedure: Fecal matter/waste must be thoroughly cleaned from surfaces/objects before disinfection by application with a clean cloth, mop, and/or sponge saturated with the disinfectant product. Cleaning is to include vigorous wiping and/or scrubbing, until all visible soil is removed. Special attention is needed for high-touch surfaces. Surfaces in patient rooms are to be cleaned in an appropriate manner, such as from right to left or left to right, on horizontal surfaces, and top to bottom, on vertical surfaces, to minimize spreading of the spores. Restrooms are to be cleaned last. Do not reuse soiled cloths.

Infectious Materials Disposal: Materials used in the cleaning process that may contain feces/wastes are to be disposed of immediately in accordance with local regulations for infectious materials disposal.

Contact Time: Leave surfaces wet with undiluted product for a minimum of 5 minutes.

Mycobacterium tuberculosis BCG: Remove any obvious debris or organic material from the surface to be treated. This can often be accomplished by rinsing with water or by detergent cleaning followed by a water rinse. Apply product to hard, non-porous surfaces, thoroughly wetting surfaces with a clean cloth, mop, sponge, or by immersion. Treated surfaces must remain wet for 1 minute. Wipe dry with a cloth, sponge or mop or allow to air dry.

DIRECTIONS FOR One-Step Cleaner Disinfection, Hospital Disinfection and as a Fungicide: For disinfection of items such as external surfaces of medical equipment, counters, bathroom fixtures, etc. Remove any obvious debris or organic material from the surface to be treated. Apply product to hard, non-porous surfaces, thoroughly wetting surfaces with a clean cloth, mop, sponge or by immersion. Treated surfaces must remain wet for 2 minutes. Wipe dry with a clean cloth, sponge or mop or allow to air dry. This product may be used to preclean or decontaminate critical or semi-critical medical devices prior to sterilization or high level disinfection.

USES

Primarily intended for sterilization, or disinfection of: High-Touch Hard, Non-porous Surfaces such as bedrails, doorknobs, bathroom fixtures, toilet handle, toilet seat, Floors, and Walls

Can be used on the following materials: Stainless Steel, Plastic Items and surfaces, Glass, Formica, Vinyl

ACCEPTED with COMMENTS EPA Letter Dated:

Under the Federal Insecticide, Fungicide, and Rocerutide Act as amended, for the perficide, registered under SPA Reg. IVo.

1043-125 Reg. Wo.

Page 4 of 4 Draft 8/20/09

Under the R

DISPOSAL OF INFECTIOUS MATERIALS: Blood and other body fluids should be autoclaved and disposed of according to federal, state and local regulations for infectious waste disposal.

CONTACT TIME Leave surfaces wet with undiluted product for a minimum of 1 minute.

STORAGE AND DISPOSAL

Do not contaminate water, food or feed by storage or disposal.

Pesticide Storage

Store in original closed container.

Pesticide Disposal

registered under EPA Reg. N Wastes resulting from the use of this product may be disposed of on-site in a sanitary sewer or at an approved waste disposal facility. Pesticide wastes are acutely hazardous. Improper disposal of excess pesticide, spray mixture, or rinsate is a violation of Federal Law. If these wastes cannot be disposed of by use according to label instructions, contact your State Pesticide or Environmental Control Agency. or the Hazardous Waste Representative at the nearest EPA Regional Office for Guidance.

Container Disposal

Nonrefillable container. Do not reuse or refill this container. Offer for recycling, if available.

DIRECTIONS FOR USE

It is a violation of Federal law to use this product in a manner inconsistent with its labeling.

Preparation of End-Use Product: Component A and Component B are mixed together to create an end-use solution containing approximately 0.085% peroxyacetic acid and 1.0% hydrogen peroxide.

To activate the end-use solution, tear open the protective pouch of Component A and remove the activator dose pack. Open Component B and insert the dose pack into bottle, then recap and seal tightly. Shake thoroughly and let sit for 10 minutes to allow reaction to occur. Apply per directions.

Prepare daily a fresh end-use dilution. As solution becomes dirty, discard and replace with new end-use solution.

DIRECTIONS FOR STERILIZATION: This product is not to be used as a terminal sterilant on any critical/semi-critical medical device. Remove any obvious debris or organic material from the surface to be sterilized. This can often be accomplished by rinsing with water or by detergent cleaning followed by a water rinse. Immerse the item to be sterilized in a sufficient volume of undiluted product to cover the item and fill all passages requiring sterilization. Hold in the sterilizing solution for a minimum of 1 hour at 20°C temperature (68°F). Remove items after 1 hour and rinse with deionized water.