

PM 31 7078-4
 UNITED STATES ENVIRONMENTAL PROTECTION AGENCY

10415

Johnson & Johnson Medical, Inc.
 2500 Arbrook Boulevard
 P.O. Box 90120
 Arlington, TX 76004-3130

MAY 24 1994

Attention: T. M. Wendt, Ph.D., Director
 Regulatory Affairs

Paul J. Kruger, Agent
 for Johnson & Johnson Medical, Inc.

Subject: CIDEX FORMULA 7* Long-Life
 Activated Dialdehyde Solution
 EPA Registration No. 7078-4
 Your Amendment Dated April 8, 1994

The amendment referred to above, submitted in connection with registration under the Federal Insecticide, Fungicide, and Rodenticide Act, as amended, to revise the product labeling to include a package insert as per FDA 510(k) labeling requirements, has been reviewed by the Agency and found to be acceptable, provided you make the following change.

Amend the submitted product label with the following statement prior to releasing the product for shipment. Immediately under "DIRECTIONS FOR USE" include: "It is a violation of Federal Law to use this product in a manner inconsistent with its labeling." Your release for shipment of the product constitutes acceptance of this condition.

The submitted package insert is appropriate, acceptable, and complies with FDA 510(k) labeling requirements for liquid chemical germicides, as per 21 CFR Section 801.5.

A stamped copy of the product label and package insert is enclosed for your records.

If you have any questions concerning this letter, please contact Doreen Aviado at (703) 305-6554.

Sincerely,



Marshall Swindell
 Acting Product Manager-31
 Antimicrobial Program Branch
 Registration Division (7505C)

CONCURRENCES

SYMBOL	Enclosures 7505C						
SURNAME	D. Aviado						
DATE	5/28/94						

INDICATIONS FOR USE

STERILIZATION: This solution should be used for the sterilization of heat sensitive medical equipment for which alternative methods of sterilization are not suitable. Medical equipment which should always be sterilized is that which is categorized as critical (i.e., used in procedures in which contact will be made with tissues that are normally considered sterile).

HIGH LEVEL DISINFECTANT: This solution should be used for the high level disinfection of heat sensitive medical equipment for which sterilization is not practical. Medical equipment which should always be subjected to high level disinfection or sterilization is that which will be used in procedures categorized as semi-critical (i.e., used in procedures in which contact will be made with mucous membranes or other body surfaces which are not normally considered to be sterile).

INTERMEDIATE LEVEL DISINFECTANT: This solution should be used for the disinfection of medical equipment for which a risk of cross-contamination exists. Medical equipment which should always be disinfected is that which is to be used in procedures categorized as non-critical (i.e., used in procedures in which contact will only be made with intact skin).

DIRECTIONS FOR USE

ACTIVATION: Activate the CIDEX FORMULA 7 Solution by mixing the entire contents of the solution vial which is attached to the CIDEX FORMULA 7 Solution container. Shake well. When activated, solution changes color to green. Remove the date of activation on the indicated space below, in a log book or a label affixed to any secondary container used for the activated solution. See package insert for additional instructions and information regarding activated solution.

CLEANING/DECONTAMINATION: Glass and other body fluids must be thoroughly cleaned from the surfaces of medical instruments and equipment, thoroughly clean, rinse and rough dry objects before immersing in CIDEX FORMULA 7 Solution. Cleanse and rinse the lenses of optical instruments before filling with CIDEX FORMULA 7 Solution. See package insert for additional cleaning/decontamination instructions.

STERILIZATION: Immerse medical instrument/equipment completely in CIDEX FORMULA 7 Solution for a minimum of 30 hours at 20°C - 25°C. Remove instrument from the solution using aseptic technique and rinse thoroughly with sterile water. See package insert for complete instructions/information on sterilization.

HIGH LEVEL DISINFECTANT: Immerse medical instrument/equipment completely in CIDEX FORMULA 7 Solution for a minimum of 30 minutes at 20°C - 25°C. Remove from the solution and rinse thoroughly with sterile water or with potable tap water. The quality of this water used is dependent upon the intended use of the instrument/equipment. See package insert for complete instructions/information on high level disinfection.

INTERMEDIATE LEVEL DISINFECTANT: Immerse medical instrument/equipment completely in CIDEX FORMULA 7 Solution for a minimum of 30 minutes at 20°C - 25°C to destroy all vegetative bacteria except for large numbers of spore-forming tubercle bacilli, but including Pseudomonas aeruginosa, pathogenic fungi and spore-forming yeasts, as indicated in the package insert.

A 30 minute immersion at 20°C will kill 99.9% *Sporobacterium thymophilus* (Charalike 70 Method).
A 30 minute immersion at 20°C will kill 99.9% *Sporobacterium thymophilus* (Charalike 70 Method).
Remove instrument/equipment from the solution and thoroughly clean with sterile water or potable tap water. The quality of the rinse water used is dependent upon the intended use of the instrument/equipment. See package insert for complete instructions/information on intermediate level disinfection.

Reorder No.

2745

REUSABLE

Sterilizing and Disinfecting Solution

Sporicidal Fungicidal Bactericidal
Virucidal Tuberculocidal Pseudomonocidal

Cidex 7

long-life activated dialdehyde solution

Active Ingredient	
Gluconaldehyde	2.5%
Inert Ingredients	97.5%
Total	100.0%

NOTE - Contents of attached vial must be added to solution before this product is effective.
See Directions for Use - Activation.
Contains Sodium Nitrite as a corrosion inhibitor.
EPA REG. No. 7078-4
EPA EST. No. 38126-PP-1

Keep out of reach of children.

DANGER: See right side panel for additional precautionary statements.

*TRADEMARK

CONTENTS: 0.946 L (ONE QT.)

Johnson & Johnson
MEDICAL INC.

PRECAUTIONARY STATEMENTS

WARNING TO HUMANS AND DOMESTIC ANIMALS:

DANGER: Keep Out of Reach of Children.

Direct contact is corrosive to exposed skin, eyes, nose and mucous membranes. Avoid contact with clothing. Wear protective clothing. Avoid contamination of food.

STATEMENT OF PRACTICAL TREATMENT:

of water for at least 15 minutes. For quantities of water and call a physician for

NOTE TO PHYSICIAN: Possible mucous

STORAGE AND DISPOSAL: Store at room

Possible Hazard: Observe restricted use

Container Disposal: Do not reuse empty

Refer to package insert for detailed use

CONTROL NUMBER AND EXPIRATION DATE

NOTE: Do not use beyond 30 days after activation date.
CIDEX FORMULA 7 Solution
EXPIRATION DATE FOR

John
M

APL INC.

For technical safety and TRADEMARK

SCALE: 100%

ACCEPTED
with COMMENTS
in EPA Letter Dated:

MAY 24 1994

Under the Federal Insecticide,
Fungicide, and Rodenticide Act as
amended, for the pesticide
registered under EPA Reg. No.

7078-4

BEST AVAILABLE COPY

4
7
1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
66
67
68
69
70
71
72
73
74
75
76
77
78
79
80
81
82
83
84
85
86
87
88
89
90
91
92
93
94
95
96
97
98
99
100
101
102
103
104
105
106
107
108
109
110
111
112
113
114
115
116
117
118
119
120
121
122
123
124
125
126
127
128
129
130
131
132
133
134
135
136
137
138
139
140
141
142
143
144
145
146
147
148
149
150
151
152
153
154
155
156
157
158
159
160
161
162
163
164
165
166
167
168
169
170
171
172
173
174
175
176
177
178
179
180
181
182
183
184
185
186
187
188
189
190
191
192
193
194
195
196
197
198
199
200
201
202
203
204
205
206
207
208
209
210
211
212
213
214
215
216
217
218
219
220
221
222
223
224
225
226
227
228
229
230
231
232
233
234
235
236
237
238
239
240
241
242
243
244
245
246
247
248
249
250
251
252
253
254
255
256
257
258
259
260
261
262
263
264
265
266
267
268
269
270
271
272
273
274
275
276
277
278
279
280
281
282
283
284
285
286
287
288
289
290
291
292
293
294
295
296
297
298
299
300
301
302
303
304
305
306
307
308
309
310
311
312
313
314
315
316
317
318
319
320
321
322
323
324
325
326
327
328
329
330
331
332
333
334
335
336
337
338
339
340
341
342
343
344
345
346
347
348
349
350
351
352
353
354
355
356
357
358
359
360
361
362
363
364
365
366
367
368
369
370
371
372
373
374
375
376
377
378
379
380
381
382
383
384
385
386
387
388
389
390
391
392
393
394
395
396
397
398
399
400
401
402
403
404
405
406
407
408
409
410
411
412
413
414
415
416
417
418
419
420
421
422
423
424
425
426
427
428
429
430
431
432
433
434
435
436
437
438
439
440
441
442
443
444
445
446
447
448
449
450
451
452
453
454
455
456
457
458
459
460
461
462
463
464
465
466
467
468
469
470
471
472
473
474
475
476
477
478
479
480
481
482
483
484
485
486
487
488
489
490
491
492
493
494
495
496
497
498
499
500
501
502
503
504
505
506
507
508
509
510
511
512
513
514
515
516
517
518
519
520
521
522
523
524
525
526
527
528
529
530
531
532
533
534
535
536
537
538
539
540
541
542
543
544
545
546
547
548
549
550
551
552
553
554
555
556
557
558
559
560
561
562
563
564
565
566
567
568
569
570
571
572
573
574
575
576
577
578
579
580
581
582
583
584
585
586
587
588
589
590
591
592
593
594
595
596
597
598
599
600
601
602
603
604
605
606
607
608
609
610
611
612
613
614
615
616
617
618
619
620
621
622
623
624
625
626
627
628
629
630
631
632
633
634
635
636
637
638
639
640
641
642
643
644
645
646
647
648
649
650
651
652
653
654
655
656
657
658
659
660
661
662
663
664
665
666
667
668
669
670
671
672
673
674
675
676
677
678
679
680
681
682
683
684
685
686
687
688
689
690
691
692
693
694
695
696
697
698
699
700
701
702
703
704
705
706
707
708
709
710
711
712
713
714
715
716
717
718
719
720
721
722
723
724
725
726
727
728
729
730
731
732
733
734
735
736
737
738
739
740
741
742
743
744
745
746
747
748
749
750
751
752
753
754
755
756
757
758
759
760
761
762
763
764
765
766
767
768
769
770
771
772
773
774
775
776
777
778
779
780
781
782
783
784
785
786
787
788
789
790
791
792
793
794
795
796
797
798
799
800
801
802
803
804
805
806
807
808
809
810
811
812
813
814
815
816
817
818
819
820
821
822
823
824
825
826
827
828
829
830
831
832
833
834
835
836
837
838
839
840
841
842
843
844
845
846
847
848
849
850
851
852
853
854
855
856
857
858
859
860
861
862
863
864
865
866
867
868
869
870
871
872
873
874
875
876
877
878
879
880
881
882
883
884
885
886
887
888
889
890
891
892
893
894
895
896
897
898
899
900
901
902
903
904
905
906
907
908
909
910
911
912
913
914
915
916
917
918
919
920
921
922
923
924
925
926
927
928
929
930
931
932
933
934
935
936
937
938
939
940
941
942
943
944
945
946
947
948
949
950
951
952
953
954
955
956
957
958
959
960
961
962
963
964
965
966
967
968
969
970
971
972
973
974
975
976
977
978
979
980
981
982
983
984
985
986
987
988
989
990
991
992
993
994
995
996
997
998
999
1000

MAY 24 1994

Under the Federal Insecticide,
Fungicide, and Rodenticide Act
amended for the pesticide
registered under EPA Reg. No.

7073-4

CIDEX FORMULA 7* Solution is a liquid chemical sterilant and a high or intermediate level disinfectant, when used according to the Directions for Use.

Sterilant: CIDEX FORMULA 7 Solution is a sterilant when used or reused, according to Directions for Use, at full strength for a maximum of 28 days at 20°C to 25°C with an immersion time of at least 10 hours.

High Level Disinfectant: CIDEX FORMULA 7 Solution is a high level disinfectant when used or reused, according to Directions for Use, at full strength for a maximum of 28 days at 25°C with an immersion time of at least 90 minutes.

Intermediate Level Disinfectant: CIDEX FORMULA 7 Solution is an intermediate level disinfectant when used or reused, according to Directions for Use, at full strength for a maximum of 28 days.

A 10 minute immersion at 20°C to 25°C will destroy all vegetative bacteria, except for large numbers of *Mycobacterium tuberculosis*, but including *Pseudomonas aeruginosa*, pathenogenic fungi and specified viruses, as indicated in Section E - Directions for Use - Item 3(c).

A 10 minute immersion at 20°C will kill 89.4% *Mycobacterium tuberculosis* (Quantitative TB Method).

A 10 minute immersion at 25°C will kill 98.4% *Mycobacterium tuberculosis* (Quantitative TB Method).

2) Reuse Period

BEST AVAILABLE COPY

CIDEX FORMULA 7 Solution has demonstrated efficacy in the presence of 2% organic soil and a simulated amount of microbiological burden during reuse. CIDEX FORMULA 7 Solution can be reused for a period not to exceed 28 days provided the required conditions of glutaraldehyde concentration, pH, and temperature exist based upon monitoring described in Directions for Use. **DO NOT** rely solely on days in use. **Efficacy of this product during its use-life must be verified by the CIDEX FORMULA 7* Solution Test Strip to determine that at least the minimum effective concentration (MEC) of 1.8% glutaraldehyde is present.**

*Trademark

3) **General Information on Selection and Use of Germicides for Medical Device Reprocessing**

Choose a germicide with the level of microbial activity that is appropriate for the reusable medical device or equipment surface. Follow the reusable device labeling and standard institutional practices. In the absence of complete instructions, use the following process:

First, for patient contacting devices, determine whether the reusable device to be reprocessed is a critical, semi-critical, or noncritical device.

- A **critical device** presents a high risk of infection if not sterile. Critical devices routinely penetrate the skin or mucous membranes during use or are otherwise used in normally sterile tissue of the body.
- A **semi-critical device** makes contact with mucous membranes but does not ordinarily penetrate normally sterile areas of the body.
- A **noncritical device** contacts only intact skin during routine use.

Second, determine the level of germicidal activity that is needed for the reusable device.

Critical Device	Sterilization required (e.g.: products that enter sterile tissue or the vascular system, such as laparoscopes and microsurgical instruments)
Semi-critical Device	Sterilization recommended whenever practical, otherwise High Level Disinfection acceptable (e.g.: GI endoscopes, anesthesia equipment for the airway, diaphragm-fitting rings, etc.)
Noncritical Device	Intermediate Level Disinfection recommended when there is a risk of cross-contamination, otherwise Low Level Disinfection acceptable (e.g.: bed pans, blood pressure cuffs, linens, etc.)
Medical Equipment Surface	Same as for noncritical device (e.g.: crutches, oxygen cylinders, etc.)

Third, select a germicide that is labeled for the appropriate germicidal level and is compatible with the reusable device. Follow directions for the germicide.

BEST AVAILABLE COPY

4) **Microbial Activity**

The following table indicates the spectrum of activity as demonstrated by testing of CIDEX FORMULA 7 Solution**:

<u>BACTERIA</u>		<u>FUNGI</u>	<u>VIRUSES</u>	
SPORES	VEGETATIVE ORGANISMS		NON-ENVELOPED	ENVELOPED
Bacillus subtilis	Staphylococcus aureus	Trichophyton mentagrophytes	Poliavirus Type 1	Coronavirus
Clostridium sporogenes	Salmonella choleraesuis		Rhinovirus Type 14	Cytomegalovirus
	Pseudomonas aeruginosa		Adenovirus Type 2	Influenza virus Type A [WS/33]
	Mycobacterium tuberculosis		Vaccinia	HIV-1 (AIDS Virus)
			Coxsackievirus B-1	Herpes simplex Type 1, 2

** Testing was done after 28 days of simulated rescue using prescribed testing methods.

BEST AVAILABLE COPY

5) **Material Compatibility**

CIDEX FORMULA 7 Solution is recommended for use with medical devices made from the materials shown below. Care must be taken with medical and dental equipment such as anesthesia and respiratory therapy tubing, dental mirrors and burs. These devices may be damaged when cleaned with a highly alkaline detergent, poorly rinsed after disinfection, stored wet or dried at temperatures exceeding 105°F.

METALS

- Chrome plate¹
- Aluminum⁴
- Anodized aluminum⁵
- Brass⁵
- Carbon steel⁷
- Stainless steel⁷

PLASTICS

- Polyethylene terephthalate (Polyester)²
- Polyvinylchloride (PVC)²
- Polystyrene²
- Polymethylmethacrylate (Acrylic)²
- Polycarbonate^{4,8}
- Polypropylene⁷
- Polyethylene⁷
- Acrylonitrile-butadiene-styrene (ABS)⁷
- Nylon⁷

ELASTOMERS

- Black natural rubber³
- Red natural rubber³
- Polychloroprene rubber (Neoprene)⁶
- Silicone rubber (Silastic)⁷

NOTE:

- 1 Represents 13 hours of total contact with CIDEX FORMULA 7 Solution over 40 disinfection cycles.
- 2 Represents 20 hours of total contact with CIDEX FORMULA 7 Solution over 20 disinfection cycles.
- 3 Represents 35 hours of total contact with CIDEX FORMULA 7 Solution over 104 disinfection cycles.
- 4 Represents 40 hours of total contact with CIDEX FORMULA 7 Solution over 40 disinfection cycles.
- 5 Represents 144 hours of continuous contact with CIDEX FORMULA 7 Solution.
- 6 Represents 240 hours of total contact with CIDEX FORMULA 7 Solution over 10 sterilization cycles.
- 7 Represents 336 hours or greater of continuous contact with CIDEX FORMULA 7 Solution.
- 8 Triple rinse equipment and dry at room temperature or below 104°F (40°C) to minimize possibility of stress cracking.

*Trademark

7/2/15

6) **Precleaning Agent Compatibility**

CIDEX FORMULA 7 Solution is compatible with enzymatic detergents (e.g.: ENZOL* Enzymatic Detergent) which are mild in pH, low foaming, and easily rinsed from equipment. Detergents that are either highly acid or alkaline are contraindicated as precleaning agents since improper rinsing could affect the efficacy of the CIDEX FORMULA 7 Solution by altering its pH.

B) **Contraindications**

1) **Sterilant Usage**

Routine biological monitoring is not feasible with CIDEX FORMULA 7 Solution, and therefore, CIDEX FORMULA 7 Solution should NOT be used to sterilize reusable medical devices that are compatible with other available methods of sterilization that can be biologically monitored, e.g.: heat, ethylene oxide, or peroxide gas plasma.

CIDEX FORMULA 7 Solution should not be used for sterilization of critical devices intended for single use (e.g.: catheters).

2) **High Level Disinfectant Usage**

CIDEX FORMULA 7 Solution should NOT be used to high level disinfect a semi-critical device when sterilization is practical.

3) **Endoscope Usage**

CIDEX FORMULA 7 Solution is not the method of choice for sterilization of rigid endoscopes which the device manufacturer indicates are compatible with steam sterilization. In general, glutaraldehyde solutions that do not contain surfactants (e.g.: CIDEX* Activated Diakdehyde Solution) are more appropriate for flexible endoscopes, since glutaraldehyde solutions containing surfactants (e.g.: CIDEX FORMULA 7 Solution or CIDEX PLUS* 28 Day Solution) are more difficult to rinse from the devices. However, these surfactant containing disinfectants may be used for reprocessing of flexible endoscopes if a validated protocol for rinsing and leak testing is employed.

BEST AVAILABLE COPY

C) Warnings

CIDEX FORMULA 7 SOLUTION IS HAZARDOUS TO HUMANS AND DOMESTIC ANIMALS

**DANGER: Keep Out of Reach of Children
Contains Glutaraldehyde**

- 1) Direct contact is corrosive to exposed tissue, causing eye damage and skin irritation/damage. Do not get into eyes, on skin or on clothing.**
- 2) Avoid contamination of food.**
- 3) Use in well ventilated area in closed containers.**

In case of contact, immediately flush eyes or skin with plenty of water for at least 15 minutes. For eyes, get medical attention.

Harmful if swallowed. Drink large quantities of water and call a physician immediately.

Probable mucosal damage from oral exposure may contraindicate the use of gastric lavage.

Emergency, safety, or technical information about CIDEX FORMULA 7 Solution can be obtained from Johnson & Johnson Medical, Inc. Customer Relations at 1-800-423-5850, or by contacting your local Johnson & Johnson Medical, Inc. sales representative.

D) Precautions

- 1) Disposable latex gloves, eye protection, face masks, and liquid-proof gowns should be worn when cleaning and sterilizing/disinfecting soiled devices and equipment.
- 2) Contaminated, reusable devices **MUST BE THOROUGHLY CLEANED** prior to disinfection or sterilization, since residual contamination will decrease effectiveness of the germicide.
- 3) The user **MUST** adhere to the **Directions for Use** since any modification will affect the safety and effectiveness of the germicide.
- 4) The reusable device manufacturer should provide the user with a validated reprocessing procedure for that device using CIDEX FORMULA 7 Solution.
- 5) The use of CIDEX FORMULA 7 Solution in automated endoscope washers **must be part of a validated reprocessing procedure provided by the washer manufacturer**. Contact the manufacturer of the endoscope washer for instructions on the maximum number of reprocessing cycles which may be used before refilling with fresh CIDEX FORMULA 7 Solution. Use CIDEX FORMULA 7 Solution Test Strips to monitor glutaraldehyde concentration before each cycle to detect unexpected dilution.

E) **Directions for Use**

1) **Activation**

Activate the CIDEX FORMULA 7 Solution by adding the entire contents of the Activator Vial which is attached to the CIDEX FORMULA 7 Solution Container. Shake well. Activated solution immediately changes color to green, thereby indicating solution is ready to use. CIDEX FORMULA 7 Solution is intended for use in manual (bucket and tray) systems made from polypropylene, ABS, polyethylene, glass-filled polypropylene or specially molded polycarbonate plastics. Record the date of activation (mixing date) and expiration date on the CIDEX FORMULA 7 Solution container label in the space provided, in a log book or on a label affixed to any secondary container for the activated solution.

2) **Cleaning/Decontamination**

Blood and other body fluids must be thoroughly cleaned from surfaces, lumens, and objects before application of the disinfectant or sterilant. Blood and other body fluids should be autoclaved and disposed of according to all applicable federal, state and local regulations for infectious waste disposal.

For complete disinfection or sterilization of medical instruments and equipment, thoroughly clean, rinse and rough dry objects before immersing in CIDEX FORMULA 7 Solution. Cleanse and rinse the lumens of hollow instruments before filling with CIDEX FORMULA 7 Solution. Refer to the reusable device manufacturer's labeling for additional instructions on disassembly, decontamination, cleaning and leak testing of their equipment.

3) **Usage**

It is a violation of Federal Law to use this product in a manner inconsistent with its labeling.

a) **Sterilization (Bucket/Tray Manual System)**

Immerse medical equipment/device completely in CIDEX FORMULA 7 Solution for a minimum of 10 hours at 20°C to 25°C to eliminate all microorganisms including *Clostridium sporogenes* and *Bacillus subtilis* spores. Remove equipment from the solution using sterile technique and rinse thoroughly with sterile water following the rinsing instructions below.

b) **High Level Disinfection (Bucket/Tray Manual System)**

Immerse medical equipment/device completely in CIDEX FORMULA 7 Solution for a minimum of 90 minutes at 25°C to destroy all pathenogenic microorganisms, except for large numbers of bacterial endospores, but including *Mycobacterium tuberculosis* (Quantitative TB Method). Remove devices and equipment from the solution and rinse thoroughly following the rinsing instructions below.

c) **Intermediate Level Disinfection (Bucket/Tray Manual System)**

Immerse medical equipment/device completely in CIDEX FORMULA 7 Solution for a minimum of 10 minutes at 20°C to 25°C to destroy all vegetative bacteria, except for large numbers of *Mycobacterium tuberculosis*, but including *Pseudomonas aeruginosa*, pathogenic fungi, and viruses (Poliovirus Type 1; Adenovirus Type 2; Herpes simplex Type 1,2; HIV-1 (AIDS virus); Influenza Type A [WS/33]; Vaccinia; Coronavirus; Cytomegalovirus; Rhinovirus Type 14) on inanimate surfaces.

A 10 minute immersion at 20°C will kill 89.4% *Mycobacterium tuberculosis* (Quantitative TB Method).

A 10 minute immersion at 25°C will kill 98.4% *Mycobacterium tuberculosis* (Quantitative TB Method).

Remove devices and equipment from the solution and rinse thoroughly following the rinsing instructions below.



d) Rinsing Instructions

Following immersion in CIDEX FORMULA 7 Solution, thoroughly rinse the equipment or medical device by immersing it completely in three separate copious volumes of water. Each rinse should be a minimum of 1 minute in duration unless otherwise noted by the device or equipment manufacturer. Use fresh portions of water for each rinse. Discard the water following each rinse. Do not reuse the water for rinsing or any other purpose as it will be contaminated with glutaraldehyde.

Refer to the reusable device/equipment manufacturer's labeling for additional rinsing instructions.

STERILE WATER RINSE

The following devices should be rinsed with sterile water, using sterile technique when rinsing and handling:

1. Devices intended for use in normally sterile areas of the body;
2. Devices intended for use in known immunocompromised patients, or potentially immunocompromised patients based on institutional procedures (e.g.: high risk population served) and;
3. When practicable, bronchoscopes, due to a risk of atypical Mycobacteria contamination from potable water supply.

POTABLE WATER RINSE

BEST AVAILABLE COPY

For all other devices a sterile water rinse is recommended when practicable, otherwise a high quality potable tap water rinse is acceptable. A high quality potable water is one that meets Federal Clean Water Standards at the point of use.

When using potable water for rinsing, the user should be aware of the increased risk of recontaminating the device or medical equipment with **Pseudomonads** and **atypical (fast growing) Mycobacteria** often present in potable water supplies. A device (e.g.: colonoscope) that is not completely dried provides an ideal situation for rapid colonization of bacteria. Additionally, **Mycobacteria** are highly resistant to drying, therefore, rapid drying will avoid possible colonization but may not result in a device free from **atypical Mycobacteria**. Although these bacteria are not normally pathogenic in patients with healthy immune systems, AIDS patients or other immunocompromised individuals may be placed at high risk of infection by these opportunistic microorganisms. A final rinse using a 70% isopropyl alcohol solution is useful to speed the drying process and reduce the numbers of any organism present as a result of rinsing with potable water.

e) **Reusage**

CIDEX FORMULA 7 Solution has also demonstrated efficacy in the presence of 2% organic soil contamination and a simulated amount of microbiological burden during reuse. This solution may be used and reused within the limitations indicated above for up to 28 days after activation. Do not use activated solution beyond 28 days. Efficacy of this product during its use-life must be verified by the CIDEX FORMULA 7 Solution Test Strip to determine that the minimum effective concentration (MEC) of 1.8% is present.

4) **Monitoring of Germicide to Ensure Specifications are Met**

During the usage of CIDEX FORMULA 7 Solution, as a high or intermediate level disinfectant and/or sterilant, it is recommended that a thermometer and timer be utilized to ensure that the optimum usage conditions are met. In addition, it is recommended that the CIDEX FORMULA 7 Solution be tested with the CIDEX FORMULA 7 Solution Test Strip prior to each usage. This is to insure that the appropriate concentration of glutaraldehyde is present and to guard against a dilution which may lower the effectiveness of the solution below its MEC. The pH of the activated solution may also be periodically checked to verify that the pH of the solution is between 7.5 and 8.1.

5) **Post-Processing Handling and Storage of Reusable Devices**

Disinfected and/or sterilized reusable devices are either to be immediately used or stored in a manner to minimize recontamination. Note that only terminal sterilization (sterilization in a suitable wrap) provides maximum assurance against recontamination. Refer to the reusable device-equipment manufacturer's labeling for additional storage and/or handling instructions.

140415

F) Storage Conditions and Expiration Date

- 1) Prior to activation, CIDEX FORMULA 7 Solution should be stored in its original sealed container at controlled room temperature 15°-30°C (59°-86°F).

Once the CIDEX FORMULA 7 Solution has been activated, it should be stored in the original container until transferred to the closed containers in which the immersion for disinfection or sterilization is to take place. Containers should be stored in a well ventilated, low traffic area at controlled room temperature.

- 2) The expiration date of the inactivated CIDEX FORMULA 7 Solution will be found on the bottom of the immediate container.
- 3) The use period for activated CIDEX FORMULA 7 Solution is for no longer than 28 days following activation or as indicated by the CIDEX FORMULA 7 Solution Test Strip. Once activated, the solution requires no further dilution prior to its usage.

G) Emergency and Technical Product Information

Emergency, safety, or technical information about CIDEX FORMULA 7 Solution can be obtained from Johnson & Johnson Medical, Inc. Customer Relations at 1-800-423-5850, or by contacting your local Johnson & Johnson Medical, Inc. sales representative.

H) User Proficiency

It is a violation of Federal Law to use this product in a manner inconsistent with its labeling. The user should be adequately trained in the decontamination and disinfection or sterilization of medical devices and the handling of toxic substances such as liquid chemical germicides. Additional information about CIDEX FORMULA 7 Solution can be obtained from Johnson & Johnson Medical, Inc. Customer Relations at 1-800-423-5850, or by contacting your local Johnson & Johnson Medical, Inc. sales representative.

I) Disposal Information

Germicide Disposal

Discard residual solution in drain. Flush thoroughly with water.

Container Disposal

- **0.946 L (One Quart) and 3.785 L (One Gallon) Size Containers**
Do not reuse empty container. Wrap container and put in trash.
- **9.462 L (2 1/2 Gallon) Size Container**
Triple rinse empty container with water and dispose of in an incinerator or landfill approved for pesticide containers.

J) How Supplied

<u>Reorder</u>	<u>Description</u>	<u>Case Contains</u>
2745	0.946 L (One Quart) Container	4 x 0.946 L (4 qts)/box; 4 boxes/case; 16 x 0.946 L (16qts)/case
2750	3.785 L (One Gallon) Container	4 x 3.785 L (4 gals) /case
2753	9.462 L (2 1/2 Gallon) Container	2 x 9.462 L (2 x 2 1/2 gals)/case
2923	CIDEX FORMULA 7 Solution Test Strips	60 strips/btl; 2 btls/case
2928	CIDEX FORMULA 7 Solution Test Strips	15 strips/btl; 2 btls/case

References supplied upon request.

*Trademark