DIRECTIONS FOR USE

For complete sterilization or disinfection of medical instruments and equipment thoroughly clean, rinse and roughdry objects before immersing in full strength solution. Clean and rinse the lumens of any hollow instruments before filling with solution. Avoid trapped air bubbles.

Disinfection. Immerse equipment completely for a minimum of 10 minutes at 20°C to destroy vegetative pathogenic bacteria including tuberculosis, Pseudomonas, fungi, pathogenic yeast, and viruses including HBV, HIV, Type 1 Herpes Simplex, Type 2, Coxsackievirus B5, Echovirus 9, Poliovirus Type 2, or Rotavirus Simian. Remove equipment from METRICIDE Solution and rinse thoroughly.

Sterilization. Immerse equipment completely for a minimum of 10 hours at 20°C to destroy resistant pathogenic spores as represented by Clostridium sporogenes and Bacillus subtilis. Remove equipment from METRICIDE Solution using sterile technique and rinse thoroughly with sterile water.

DIRECTIONS FOR ACTIVATING

A 3.5 oz. ampule is attached at the top of this container of METRICIDE Solution. Shake. Solution is ready to use. When activated, solution changes color to green. This solution may be used and reused for up to 21 days after activation, so long as the activated solution is beyond 14 days.

General Disinfection — For use in hospitals, Dental Offices, Nursing Homes and Health Care Institutions. For repeated use in human (habitat and pet) environments.

It is a violation of Federal Law to use this product in a manner inconsistent with its labeling.

Product No. **MX-1400**

Active Ingredient	
Glutaraldehyde	2.0%
Inert Ingredient	98.0%
Total	100.0%

MIXING DATE

EXPIRATION DATE

REUSABLE
Sterilizing and Disinfecting Solution

Rapid-acting • Non-staining • Non-rusting

METRICIDE™

Activated Dialdehyde Solution

Sporicidal Fungicidal Bactericidal
Virucidal* Tuberculocidal Pseudomonacidal

DANGER: Keep out of reach of children.
See right side panel for precautions.

Manufactured by

EPA Reg. No. 46781
EPA Est. No. 46781-CO-1

46781-1

BEST AVAILABLE COPY