

TUBERCULOCIDE • VIRUCIDE • BACTERICIDE • FUNGICIDE

VIRAHOL®

HOSPITAL DISINFECTANT / CLEANER & INSTRUMENT PRESOAK*

READY TO USE • DISINFECTS AND DEODORIZES

See Right Panel for Additional Instructions

U.S. PATENT 5,441,723
U.S. PATENT 5,145,663

KEEP OUT OF REACH OF CHILDREN

WARNING

SEE RIGHT PANEL FOR ADDITIONAL PRECAUTIONARY STATEMENTS



© 1998 - ERDIEN CORPORATION

ACTIVE INGREDIENT - Isopropanol ...	70%
INERT INGREDIENTS	30%
TOTAL	100%

NET CONTENTS - 1 U.S. GALLON - (128 Fluid Ounces) - 3.785 Liters

ACCEPTED

JUN 26 1998

Under the Federal Insecticide, Fungicide, and
Rodenticide Act as amended, for the
pesticide, registered under
EPA Reg. No. 68142-1

PM 33
68142-1
6/26/98
Page 1 & 2

ACCEPTED

JUN 26 1998

Under the Federal Insecticide, Fungicide, and
Herbicide Act as amended, for the
pesticide, registered under
EPA Reg. No. 60142-1

PRECAUTIONARY STATEMENTS

**HAZARDS TO HUMANS &
DOMESTIC ANIMALS**

WARNING

Causes eye irritation. Do not get in eyes. Harmful if
swallowed.

STATEMENT OF PRACTICAL TREATMENT: In case of eye
contact, immediately flush eyes with plenty of water. Get
medical attention if irritation persists.

PHYSICAL OR CHEMICAL HAZARDS: Flammable, keep
away from heat & open flame.

STORAGE AND DISPOSAL

STORAGE: Store in a cool, dry place.

CONTAINER DISPOSAL: Triple rinse (or equivalent). Then
offer for recycling or reconditioning, or puncture and dispose
of in a sanitary landfill, or incineration, or, if allowed by state
and local authorities, by burning. If burned, stay out of
smoke.

PESTICIDE DISPOSAL: Do not contaminate water, food or
feed by storage or disposal. Wastes resulting from use of this
product may be disposed of on site or at an appropriate
waste disposal facility.



VERIDIEN CORPORATION
11800 28th St. N.
St. Petersburg, FL 33716
1-800-345-5444

EPA Reg. No. 60142-1
EPA Est. No. 60142-FL-1

EFFECTIVE AGAINST A BROAD RANGE OF PATHOGENS

HIV-1 (associated with AIDS)	Mycobacterium tuberculosis	Pseudomonas aeruginosa
Staphylococcus aureus	Salmonella choleraesuis	Listeria monocytogenes
Poliovirus type 1	Tetrahymena thermophila	

COMPATIBLE IN CONTACT WITH THESE MATERIALS

Aluminum	Buna N	Cast Bronze
Epoxy	Natural Rubber	Naugahyde ***
Neoprene	Nylon	Polypropylene
Rigid PVC	Silicone	Stainless Steel (304,316,440)
Teflon *	Titanium	Tygon **
Viton ***		

* a registered trademark of E.I. du Pont de Nemours ** a registered trademark of Norton Co. *** a registered trademark of Uniflow Engineering Products, Inc.

DIRECTIONS FOR USE

It is a violation of Federal Law to use this product in a manner inconsistent with its labeling. VIRAHOL® Hospital Disinfectant/Cleaner & Instrument Presoak is designed for effective disinfection, cleaning and deodorizing of hard inanimate surfaces such as walls, sink tops, tables, chairs, telephones and bed frames, and for the decontamination of instruments prior to sterilization. When using strong cleaners and disinfecting agents, it is always advisable to first test in an inconspicuous area.

TO DISINFECT CONTAMINATED SURFACES AND OBJECTS:

For heavily soiled areas, or surfaces of objects soiled with blood or body fluids, a precleaning step is required. Use a clean sponge or lint-free cloth. On precleaned surfaces, apply VIRAHOL® Hospital Disinfectant/Cleaner & Instrument Presoak with a mop, cloth, sponge or mechanical sprayer so as to thoroughly wet precleaned surfaces. DO NOT DILUTE. Allow surfaces to remain wet for 10 minutes.

***INSTRUMENT PRESOAK - TO DECONTAMINATE INSTRUMENTS PRIOR TO STERILIZATION:**

Completely immerse instruments for 10 minutes prior to preparation for sterilization. This product is not to be used as a terminal sterilant/high level disinfectant on any surface or instrument that (1) is introduced directly into the human body, either into or in contact with the bloodstream or normally sterile areas of the body, or (2) contacts intact mucous membranes but which does not ordinarily penetrate the blood barrier or otherwise enter normally sterile areas of the body. This product may be used to preclean or decontaminate critical or semi-critical medical devices prior to sterilization or high-level disinfection.

Inactivates HIV-1 on precleaned environmental surfaces and objects previously soiled with blood or body fluids in healthcare settings or other settings in which there is an expected likelihood of soiling of inanimate surfaces or objects with blood or body fluids, and in which the surfaces & objects likely to be soiled with blood or body fluids can be associated with the potential for transmission of Human Immunodeficiency Virus (HIV-1) (associated with AIDS).

SPECIAL INSTRUCTIONS FOR CLEANING AND DECONTAMINATION AGAINST HIV-1 ON HARD INANIMATE SURFACES AND / OR OBJECTS SOILED WITH BLOOD OR BODY FLUIDS:

Personal Protection: When handling items soiled with blood or body fluids use disposable latex gloves, gown, mask and / or eye coverings.

Cleaning Procedures: Blood and body fluids must be thoroughly cleaned from surfaces and objects before application of Virahol® Hospital Disinfectant/Cleaner & Instrument Presoak.

Disposal of Infectious Materials: Blood and other body fluids should be autoclaved and disposed of according to federal, state, and local regulations for infectious waste disposal.

Contact Time: Apply Virahol® Hospital Disinfectant/Cleaner & Instrument Presoak with a mop, cloth, sponge or mechanical sprayer so as to thoroughly wet precleaned surfaces. DO NOT DILUTE. Allow surfaces to remain wet for one minute to ensure total inactivation of HIV virus. This contact time will not control other common types of viruses and bacteria. For example, a ten-minute contact time is required to control Mycobacterium tuberculosis at 20°C.

No warranty, either expressed or implied, is given to usage contrary to the instructions given herein.

2 of 2