



U.S. ENVIRONMENTAL PROTECTION AGENCY

Office of Pesticide Programs Antimicrobials Division (7510P) 1200 Pennsylvania Avenue NW Washington, D.C. 20460

JO	TI	CF	OF	PFS	TICI	DF.	

x Registration
Reregistration

EPA Reg.

Date of Issuance:

Number:

88691-1

APR -5 2012

Term of Issuance:

Unconditional

Name of Pesticide Product:

Clear Bath® Algae Inhibitor

(under FIFRA, as amended)

Name and Address of Registrant (include ZIP Code): Laboratory Safety Inc.

1090 E Desert Inn Road

Las Vegas, NV 89109

Note: Changes in labeling differing in substance from that accepted in connection with this registration must be submitted to and accepted by the Antimicrobials Division prior to use of the label in commerce. In any correspondence on this product always refer to the above EPA registration number.

On the basis of information furnished by the registrant, the above named pesticide is hereby registered/reregistered under the Federal Insecticide, Fungicide and Rodenticide Act.

Registration is in no way to be construed as an endorsement or recommendation of this product by the Agency. In order to protect health and the environment, the Administrator, on his motion, may at any time suspend or cancel the registration of a pesticide in accordance with the Act. The acceptance of any name in connection with the registration of a product under this Act is not to be construed as giving the registrant a right to exclusive use of the name or to its use if it has been covered by others.

This product (OPP Decision No. 455829 is unconditionally registered in accordance with FIFRA sec 3(c)(7)(A) provided that you:

- 1. Submit and/or cite all data required for registration of your product under FIFRA sec. 3(c)(5) when the Agency requires all registrants of similar products to submit such data; and submit acceptable responses required for re-registration of your product under FIFRA section 4.
- 2. Make the labeling changes listed below before you release the product for shipment:
- a. Revise the "EPA Registration Symbol" to read, "EPA Reg. No 88691-1".
- b. Revise the precautionary statement on the label to read "... Wash thoroughly with soap and water after handling and before eating, drinking, chewing gum, using tobacco or using the toilet".

Signature of Approving Official:

Product Manager Team-31

Regulatory Management Branch Antimicrobials Division (7510P) Date:

APR - 5 2012

- c. Continuous feeding of this product into laboratory incubators is inappropriate. Revise the "Laboratory incubators" section to read exactly as follows: "Remove contents of incubator and clean walls and shelves with 70% ethanol and 10% bleach, remove racks and autoclave them. Treat water tray by replacing water with a solution of 1 ml of this product per 1 gallon of water. Replace solution every 3 months or when solution becomes visibly dirty or discolored".
- d. Submit three (3) copies of your final printed labeling before distributing or selling the product bearing the revised labeling.

If these conditions are not complied with, the registration will be subject to cancellation in accordance with FIFRA sec. 6(e). Your release for shipment of the product constitutes acceptance of these conditions. Should you have any questions concerning this letter, please contact Velma Noble at (703) 308-623 or Emilia Oiguenblik at (703) 347-0199.

A stamped copy of the label is enclosed for your records.

Sincerely,

Velma Noble

Product Manager (31)

Regulatory Management Branch Antimicrobials Division (7510P)

Enclosure: Stamped label

Clear Bath® Algae Inhibitor

[ABN: Clear Bath®]

Helps to effectively control algae and fungi in laboratory water baths, laboratory incubators, laboratory humidifiers, and on stainless steel surfaces

ACTIVE	INGREDIENTS:

Alkyl (50% C14, 40% C12, 10% C16) dimethyl, benzyl ammonium chloride	50.00%
OTHER INGREDIENTS.	50.00%
TOTAL	100.00%

KEEP OUT OF REACH OF CHILDREN

DANGER

	FIRST AID
IF IN EYES	 Hold eye open and rinse slowly and gently with water for 15-20 minutes. Remove contact lenses, if present, after the first 5 minutes, then continue rinsing eye. Call a poison control center or doctor for treatment advice.
IF ON SKIN OR CLOTHING	 Take off contaminated clothing. Rinse skin immediately with plenty of water for 15-20 minutes. Call a poison control center or doctor for treatment advice.
IF SWALLOWED	 Have person sip a glass of water if able to swallow. Do not induce vomiting unless told to do so by a poison control center or doctor. Do not give anything by mouth to an unconscious person. Call a poison control center or doctor immediately for treatment advice.
circulatory Shock,	Probable mucosal damage may contraindicate the use of gastric lavage. Measures against respiratory depression, and convulsion may be needed. ontainer or label with you when calling a poison control center or doctor, or going for

See [back] [side] panel for additional precautionary statements.

Manufactured for: Laboratory Safety Inc. 1090 E Desert Inn Road Las Vegas, NV 89109

EPA Reg. No. 88691-NEW EPA Est. No. 88691-NV-1

Net Contents: (see container label)

ACCEPTED
with COMMENTS
in EPA Letter Dated:

APR 5 2012

Under the Federal Insecticide, Fungicide, and Rodenticide Act as amended, for the pesticide, registered under EPA Reg. No. 88691

Batch Code # / Lot #: (see container label)

PRECAUTIONARY STATEMENTS HAZARD TO HUMANS AND DOMESTIC ANIMALS

DANGER: CORROSIVE. Causes irreversible eye damage and skin burns. May be fatal if inhaled, swallowed or absorbed through the skin. Do not get in eyes, on skin or on clothing. Do not breathe spray. Wear a dust/mist filtering respirator (MSHA/NIOSH approval number prefix TC-21C). Wear protective eyewear (goggles or face shield), protective clothing and rubber gloves. Wash thoroughly with soap and water after handling and before eating, drinking, or using tobacco. Remove contaminated clothing and wash clothing before reuse.

ENVIRONMENTAL HAZARDS

This pesticide is toxic to fish and aquatic organisms.

PHYSICAL OR CHEMICAL HAZARDS

Do not use or store near heat or flame.

DIRECTIONS FOR USE

It is a violation of federal law to use this product in a manner inconsistent with its labeling. Read entire label and use strictly in accordance with precautionary statements and directions.

Dilution Instructions:

Add 1ml of this product per gallon of water.

Helps to effectively control algae, bacteria, and fungi in laboratory water baths, laboratory incubators, humidifiers, and cleans stainless steel surfaces.

CIRCULATING LABORATORY WATER BATHS: Formulated for use in circulating water baths. Thoroughly clean the hard non-porous surfaces of the water tank and filters. Add 1ml of this product per gallon of water. Product may be fed continuously until organisms are under control and on an intermittent basis to control organism growth as needed. Replace solution every 3 months or when solution becomes visibly dirty or discolored.

LABORATORY INCUBATORS: Remove contents of incubator and clean walls and shelves with 70% ethanol and 10% bleach, remove racks and autoclave them. Water tray: replace water; add 1ml of this product per gallon of water. Product may be fed continuously until organisms are under control and on an intermittent basis to control organism growth as needed. Replace solution every 3 months or when solution becomes visibly dirty or discolored.

LABORATORY HUMIDIFIER DIRECTIONS: Formulated for use in portable humidifiers. Thoroughly clean the hard non-porous surfaces of the water tank and filters (before each heating season, or sooner if necessary). Add 1ml of this product to per gallon of water. Product may be fed continuously until organisms are under control and on an intermittent basis to control organism growth as needed. Replace solution every 3 months or when solution becomes visibly dirty or discolored.

ACCEPTED with COMMENTS in EPA Letter Dated:

Clear Bath® Algae Inhibitor (EPA Reg. No: 88691-R)

56

STAINLESS STEEL SURFACES CLEANER DIRECTIONS: Formulated for use on stainless steel surfaces, including laboratory bench-tops and laboratory table-tops. Add 1ml of this product to per gallon of water and use to wipe surfaces clean with cloth, paper towel, or sponge.

NOTE: This product should not be used or compounded with any reducing or oxidizing agents (such as calcium hypochlorite, solid perchlorate, or nitric acid) since such mixtures may be explosive. Do not use in conjunction with soap or any anionic wetting agent.

STORAGE AND DISPOSAL


Do not contaminate water, food, or feed by storage and disposal.

PESTICIDE STORAGE: Store in original container and place in locked storage area. Keep away from freezing.

PESTICIDE DISPOSAL: Pesticide wastes are acutely hazardous. Improper disposal of excess pesticide, spray mixture, or rinsate is a violation of Federal Law. If these wastes cannot be disposed of by use according to label instructions, contact your State Pesticide or Environmental Control Agency, or the Hazardous Waste Representative at the nearest EPA Regional Office for guidance.

CONTAINER HANDLING: Nonrefillable container: Do not refill or reuse container. Triple rinse as follows: Fill container ¼ full with water and recap. Shake for 10 seconds. Follow Pesticide Disposal instructions for rinsate disposal. Drain for 10 seconds after the flow begins to drip. Repeat procedure two more times. Then offer for recycling or reconditioning. If not available, puncture and dispose of in a sanitary landfill.

NOTE: Buyer assumes all responsibility for safety and use not in accordance with directions.



66

{Optional Marketing Statements, shown in []}

[Clear Bath® Algae Inhibitor] (or) [This product] is the efficient and economical way to improve the appearance and cleanliness of your laboratory water baths. It requires no special equipment for treating water and may be added directly to the laboratory water bath or to the laboratory water circulating equipment.

[Clear Bath® Algae Inhibitor keeps laboratory water baths and other equipment clean.]

[Water treated with Clear Bath® Algae Inhibitor remains clear for weeks.]

[Humidifier Bacteria-Algae treatment is a bacteriostat which can control the build-up of bacteria and algae in water tanks of portable, (evaporative) humidifiers that are located in laboratories.]

[For Laboratory Use Only Not For Clinical Use]

