FORM 7
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: STL BURLINGTON Contract: 24000

Lab Code: STLVT Case No.: 24000 SAS No.: SDG No.: 100226

Instrument ID: V Calibration Date: 06/08/04 Time: 1332

Lab File ID: EF010 Init. Calib. Date(s): 06/08/04 06/08/04

Heated Purge: (Y/N) N Init. Calib. Times: 1201 1541

GC Column: RTX-624 ID: 0.32 (mm)

COMPOUND	$\overline{RRF}$	RRF10	MIN RRF	%D	MAX %D
Dichlorodifluoromethane	2.314	2.173	0.01	6.1	30.0
Chloromethane	0.634	0.625	0.01	1.4	30.0
Vinyl Chloride	0.816	0.758	0.01	7.1	30.0
Bromomethane	0.864	0.885	0.01	2.4	30.0
Chloroethane	0.433	0.443	0.01	2.3	30.0
Trichlorofluoromethane	2.346	2.377	0.01	1.3	30.0
Freon TF	1.679	1.733	0.01	3.2	30.0
1,1-Dichloroethene	0.730	0.741	0.01	1.5	30.0
Methylene Chloride	1.059	0.970	0.01	8.4	30.0
1,1-Dichloroethane	1.511	1.567	0.1	3.7	30.0
cis-1,2-Dichloroethene	0.897	0.879	0.01	2.0	30.0
Chloroform	2.037	2.035	0.01	0.1	30.0
1,1,1-Trichloroethane	0.451	0.451	0.01	0.0	30.0
Carbon Tetrachloride	0.515	0.525	0.01	1.9	30.0
Benzene	0.554	0.552	0.01	0.4	30.0
1,2-Dichloroethane	0.305	0.304	0.01	0.3	30.0
Trichloroethene	0.272	0.272	0.01	0.0	30.0
1,2-Dichloropropane	0.213	0.217	0.01	1.9	30.0
cis-1,3-Dichloropropene	0.399	0.394	0.01	1.2	30.0
Toluene	0.442	0.441	0.01	0.2	30.0
trans-1,3-Dichloropropene	0.360	0.350	0.01	2.8	30.0
1,1,2-Trichloroethane	0.250	0.256	0.01	2.4	30.0
Tetrachloroethene	0.325	0.337	0.01	3.7	30.0
Chlorobenzene	0.569	0.591	0.3	3.9	30.0
Ethylbenzene	0.955	0.982	0.01	2.8	30.0
Xylene (m,p)	0.323	0.342	0.01	5.9	30.0
Styrene	0.555	0.595	0.01	7.2	30.0
Xylene (o)	0.328	0.348	0.01	6.1	30.0
1,1,2,2-Tetrachloroethane	0.607	0.640	0.01	5.4	30.0
1,3-Dichlorobenzene	0.528	0.546	0.01	3.4	30.0
1,4-Dichlorobenzene	0.513	0.530	0.01	3.3	30.0
1,2-Dichlorobenzene	0.463	0.482	0.01	4.1	30.0
1,2,4-Trichlorobenzene	0.228	0.241	0.01	5.7	30.0
Hexachlorobutadiene	0.216	0.218	0.01	0.9	30.0
1,3,5-Trimethylbenzene	0.931	0.911	0.01	2.1	30.0
1,2,4-Trimethylbenzene	0.830	0.863	0.01	4.0	30.0
1,2-Dichlorotetrafluoroethan	2.238	2.304	0.01	2.9	30.0

FORM 7
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: STL BURLINGTON Contract: 24000

Lab Code: STLVT Case No.: 24000 SAS No.: SDG No.: 100226

Instrument ID: V Calibration Date: 06/08/04 Time: 1332

Lab File ID: EF010 Init. Calib. Date(s): 06/08/04 06/08/04

Heated Purge: (Y/N) N Init. Calib. Times: 1201 1541

GC Column: RTX-624 ID: 0.32 (mm)

COMPOUND	$\overline{RRF}$	RRF10	MIN RRF	%D	MAX %D
1,2-Dibromoethane	0.529	0.540	0.01	2.1	30.0
1,3-Butadiene	0.514	0.498	0.01	3.1	30.0
Carbon Disulfide	2.572	2.572	0.01	0.0	30.0
Acetone	1.266	1.258	0.01	0.6	30.0
Isopropyl Alcohol	1.439	1.482	0.01	3.0	30.0
Methyl tert-Butyl Ether	2.415	2.459	0.01	1.8	30.0
Cyclohexane	0.237	0.235	0.01	0.8	30.0
Dibromochloromethane	0.640	0.653	0.01	2.0	30.0
Methyl Ethyl Ketone	0.463	0.460	0.01	0.6	30.0
Methyl Isobutyl Ketone	0.531	0.523	0.01	1.5	30.0
Methyl Butyl Ketone	0.569	0.567	0.01	0.4	30.0
Bromodichloromethane	0.524	0.533	0.01	1.7	30.0
trans-1,2-Dichloroethene	1.228	1.248	0.01	1.6	30.0
4-Ethyltoluene	1.086	1.199	0.01	10.4	30.0
2-Chlorotoluene	0.849	0.892	0.01	5.1	30.0
n-Hexane	1.272	1.191	0.01	6.4	30.0
Tetrahydrofuran	0.241	0.247	0.01	2.5	30.0
n-Heptane	0.293	0.296	0.01	1.0	30.0

FORM 7
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: STL BURLINGTON Contract: 24000

Lab Code: STLVT Case No.: 24000 SAS No.: SDG No.: 100226

Instrument ID: V Calibration Date: 06/09/04 Time: 1121

Lab File ID: EF010A Init. Calib. Date(s): 06/08/04 06/08/04

Heated Purge: (Y/N) N Init. Calib. Times: 1201 1541

GC Column: RTX-624 ID: 0.32 (mm)

COMPOUND	$\overline{RRF}$	RRF10	MIN RRF	%D	MAX %D
Dichlorodifluoromethane	2.314	2.437	0.01	5.3	30.0
Chloromethane	0.634	0.646	0.01	1.9	30.0
Vinyl Chloride	0.816	0.757	0.01	7.2	30.0
Bromomethane	0.864	0.899	0.01	4.0	30.0
Chloroethane	0.433	0.429	0.01	0.9	30.0
Trichlorofluoromethane	2.346	2.611	0.01	11.3	30.0
Freon TF	1.679	1.800	0.01	7.2	30.0
1,1-Dichloroethene	0.730	0.744	0.01	1.9	30.0
Methylene Chloride	1.059	0.994	0.01	6.1	30.0
1,1-Dichloroethane	1.511	1.588	0.1	5.1	30.0
cis-1,2-Dichloroethene	0.897	0.880	0.01	1.9	30.0
Chloroform	2.037	2.168	0.01	6.4	30.0
1,1,1-Trichloroethane	0.451	0.490	0.01	8.6	30.0
Carbon Tetrachloride	0.515	0.567	0.01	10.1	30.0
Benzene	0.554	0.557	0.01	0.5	30.0
1,2-Dichloroethane	0.305	0.336	0.01	10.2	30.0
Trichloroethene	0.272	0.277	0.01	1.8	30.0
1,2-Dichloropropane	0.213	0.208	0.01	2.3	30.0
cis-1,3-Dichloropropene	0.399	0.406	0.01	1.8	30.0
Toluene	0.442	0.454	0.01	2.7	30.0
trans-1,3-Dichloropropene	0.360	0.373	0.01	3.6	30.0
1,1,2-Trichloroethane	0.250	0.256	0.01	2.4	30.0
Tetrachloroethene	0.325	0.355	0.01	9.2	30.0
Chlorobenzene	0.569	0.606	0.3	6.5	30.0
Ethylbenzene	0.955	1.026	0.01	7.4	30.0
Xylene (m,p)	0.323	0.360	0.01	11.4	30.0
Styrene	0.555	0.611	0.01	10.1	30.0
Xylene (o)	0.328	0.360	0.01	9.8	30.0
1,1,2,2-Tetrachloroethane	0.607	0.650	0.01	7.1	30.0
1,3-Dichlorobenzene	0.528	0.578	0.01	9.5	30.0
1,4-Dichlorobenzene	0.513	0.572	0.01	11.5	30.0
1,2-Dichlorobenzene	0.463	0.504	0.01	8.8	30.0
1,2,4-Trichlorobenzene	0.228	0.197	0.01	13.6	30.0
Hexachlorobutadiene	0.216	0.200	0.01	7.4	30.0
1,3,5-Trimethylbenzene	0.931	1.189	0.01	27.7	30.0
1,2,4-Trimethylbenzene	0.830	0.929	0.01	11.9	30.0
1,2-Dichlorotetrafluoroethan	2.238	2.482	0.01	10.9	30.0

FORM 7
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: STL BURLINGTON Contract: 24000

Lab Code: STLVT Case No.: 24000 SAS No.: SDG No.: 100226

Instrument ID: V Calibration Date: 06/09/04 Time: 1121

Lab File ID: EF010A Init. Calib. Date(s): 06/08/04 06/08/04

Heated Purge: (Y/N) N Init. Calib. Times: 1201 1541

GC Column: RTX-624 ID: 0.32 (mm)

COMPOUND	$\overline{RRF}$	RRF10	MIN RRF	%D	MAX %D
1,2-Dibromoethane	0.529	0.550	0.01	4.0	30.0
1,3-Butadiene	0.514	0.509	0.01	1.0	30.0
Carbon Disulfide	2.572	2.553	0.01	0.7	30.0
Acetone	1.266	1.383	0.01	9.2	30.0
Isopropyl Alcohol	1.439	1.538	0.01	6.9	30.0
Methyl tert-Butyl Ether	2.415	2.570	0.01	6.4	30.0
Cyclohexane	0.237	0.235	0.01	0.8	30.0
Dibromochloromethane	0.640	0.694	0.01	8.4	30.0
Methyl Ethyl Ketone	0.463	0.461	0.01	0.4	30.0
Methyl Isobutyl Ketone	0.531	0.555	0.01	4.5	30.0
Methyl Butyl Ketone	0.569	0.604	0.01	6.2	30.0
Bromodichloromethane	0.524	0.564	0.01	7.6	30.0
trans-1,2-Dichloroethene	1.228	1.287	0.01	4.8	30.0
4-Ethyltoluene	1.086	1.112	0.01	2.4	30.0
2-Chlorotoluene	0.849	0.994	0.01	17.1	30.0
n-Hexane	1.272	1.178	0.01	7.4	30.0
Tetrahydrofuran	0.241	0.258	0.01	7.0	30.0
n-Heptane	0.293	0.312	0.01	6.5	30.0

FORM 8
VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: STL BURLINGTON Contract: 24000

Lab Code: STLVT Case No.: 24000 SAS No.: SDG No.: 100226

Lab File ID (Standard): EF010 Date Analyzed: 06/08/04

Instrument ID: V Time Analyzed: 1332

GC Column: RTX-624 ID: 0.32 (mm) Heated Purge: (Y/N) N

		IS1(BCM) AREA #	RT #	IS2(CBZ) AREA #	RT #	IS3(DFB) AREA #	RT #
	12 HOUR STD	187037	9.63	738513	12.68	820225	10.42
	UPPER LIMIT	261852	9.96	1033918	13.01	1148315	10.75
	LOWER LIMIT	112222	9.30	443108	12.35	492135	10.09
	CLIENT SAMPLE NO.						
01	W9ICVLCS	144055	9.63	575032	12.69	636393	10.42
02	W9ICVLCSD	166591	9.63	671205	12.69	750674	10.42
03	ABLKW9	153550	9.65	609932	12.69	683418	10.42
04	SG-ABC ROOFI	149619	9.65	576235	12.69	644740	10.42
05	SG-VINING	141770	9.65	545749	12.69	612310	10.42
06	SG-GANGLIANI	136014	9.65	540419	12.70	600797	10.42
07	SG-SACCO_A	142380	9.64	558484	12.70	621204	10.43
08	SG-GRAPHIQUE	139633	9.64	543752	12.70	606540	10.43
09	SG-SACCO_B	133297	9.64	528699	12.70	586268	10.42
10	SG-PACER	137637	9.64	544965	12.70	604191	10.42
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

IS1 (BCM) = Bromochloromethane
IS2 (CBZ) = Chlorobenzene-d5
IS3 (DFB) = 1,4-Difluorobenzene

AREA UPPER LIMIT = + 40% of internal standard area
AREA LOWER LIMIT = - 40% of internal standard area
RT UPPER LIMIT = + 0.33 minutes of internal standard RT
RT LOWER LIMIT = - 0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

FORM VIII VOA

FORM 8
VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: STL BURLINGTON Contract: 24000

Lab Code: STLVT Case No.: 24000 SAS No.: SDG No.: 100226

Lab File ID (Standard): EF010A Date Analyzed: 06/09/04

Instrument ID: V Time Analyzed: 1121

GC Column: RTX-624 ID: 0.32 (mm) Heated Purge: (Y/N) N

		IS1(BCM) AREA #	RT #	IS2(CBZ) AREA #	RT #	IS3(DFB) AREA #	RT #
	12 HOUR STD	144111	9.64	571110	12.70	639200	10.43
	UPPER LIMIT	201755	9.97	799554	13.03	894880	10.76
	LOWER LIMIT	86467	9.31	342666	12.37	383520	10.10
	CLIENT SAMPLE NO.						
01	X1LCS	145527	9.65	585191	12.70	645059	10.44
02	X1LCSD	149881	9.64	576547	12.70	642079	10.43
03	ABLKX1	147027	9.65	574208	12.70	629444	10.44
04	SG-ABC ROOFI	164002	9.64	632959	12.70	717411	10.42
05	SG-PACERDL	160288	9.65	624046	12.68	700060	10.42
06							
07							
08							
09							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

IS1 (BCM) = Bromochloromethane
IS2 (CBZ) = Chlorobenzene-d5
IS3 (DFB) = 1,4-Difluorobenzene

AREA UPPER LIMIT = + 40% of internal standard area
AREA LOWER LIMIT = - 40% of internal standard area
RT UPPER LIMIT = + 0.33 minutes of internal standard RT
RT LOWER LIMIT = - 0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

FORM VIII VOA

STL Burlington
Colchester, Vermont

Sample Data Summary Package

SDG: 100862

SEVERN TRENT STL

STL Burlington
208 South Park Drive, Suite 1
Colchester, VT 05446

Tel: 802 655 1203 Fax: 802 655 1248
www.stl-inc.com

July 7, 2004

Mr. Chris Milone
Roux Associates
25 Corporate Drive
Suite 230
Burlington, MA 01803

Re: Laboratory Project No. 24000
Case 24000; SDG: 100862

Dear Mr. Milone:

Enclosed are the analytical results for samples received by STL Burlington on June 19, 2004. This report is sequentially numbered starting with page 0001 and ending with page 0558.

Laboratory numbers have been assigned and designated as follows:

Lab ID	Client Sample ID	Sample Date	Sample Matrix
	Received: 06/19/04 ETR No: 100862		
576334	SG-TERADYNE	06/18/04	Air
576335	SG-TERADYNE-DUP	06/18/04	Air
576336	TRIP BLANK		Air
576337	MODELING CLAY	06/18/04	Solid

Documentation of the condition of the samples at the time their receipt and any exceptions to the laboratory's Sample Acceptance Policy is included in the Sample Handling section of this submittal.

In order to accommodate field length limitations in processing the data summary forms, the laboratory did, in certain instances, abbreviate the sample identifiers. The electronically formatted data provides the full sample identifier.

EPA SW846 Method 8260B – Volatile Organics:
The analysis of the method blank identified as VBLKW8 exhibited the presence of the target compound Naphthalene. However the concentrations detected in the method blank sample were below the method reporting limit. This compound was not detected in the field samples of this delivery group. All associated results have been identified with the qualifier "B".

The analysis of the blank spike sample NXXI LCS and the associated blank spike duplicate sample NXXI LCSD exhibited percent recoveries of select target compounds that marginally exceeded the control limits. These exceedences are presented on the analytical form 3s.

The responses for the target compound Acrolein in a select continuing calibration check acquisition exceeded the maximum percent difference criterion. This compound was not detected in the samples of this delivery group.

Method TO-15 – Volatile Organics:
The analyses of the blank spike samples B1LCS and B3LCS the associated blank spike duplicate samples exhibited percent recoveries of select target compounds that exceeded the control limits. These exceedences are presented on the analytical form 3s.

The responses for the target compounds 1,3,5-Trimethylbenzene and 2-CHlorotoluene in select continuing calibration check acquisitions exceeded the maximum percent difference criterion. These compounds were not detected in the samples of this delivery group.

The original analysis of the sample SG-TERADYNE exhibited select internal standard area responses that were outside the control criterion. The laboratory suspects that this anomaly is due to the nature of the sample matrix. This sample was re-analyzed at a dilution in order to reduce the matrix effect, but yielded in similar results. Both sets of data have been presented in this case submittal.

The original analysis of the sample SG-TERADYNE-DUP exhibited select internal standard area responses that were outside the control criterion. The laboratory suspects that this anomaly is due to the nature of the sample matrix. This sample was re-analyzed and yielded results that were within the control limits. The results of the re-analysis have been presented in this case submittal.

The analytical results presented in this data report were generated under a quality system that adheres to the requirements specified in the NELAC standard. This report shall not be reproduced, except in full, without the written approval of the laboratory. The release of the data in this report is authorized by the Laboratory Director or his designee, as verified by the following signature.

If there are any questions regarding this submittal, please contact Ron Pentkowski at (802) 655-1203.

Sincerely,

Ron Pentkowski for

Michael F. Wheeler, Ph.D.
Laboratory Director

Enclosure

TO-14/15
Result Summary

CLIENT SAMPLE NO.

SG-TERADYNE

ab Name: STL Burlington

DG Number: 100862

ase Number:

ample Matrix: Air

Lab Sample No.: 576334

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
ichlorodifluoromethane	75-71-8	0.59		0.50	2.9		2.5
hloromethane	74-87-3	0.58		0.50	1.2		1.0
inyl Chloride	75-01-4	0.50	U	0.50	1.3	U	1.3
romomethane	74-83-9	0.50	U	0.50	1.9	U	1.9
hloroethane	75-00-3	0.58		0.50	1.5		1.3
richlorofluoromethane	75-69-4	7.9		0.50	44		2.8
reon TF	76-13-1	0.50	U	0.50	3.8	U	3.8
,1-Dichloroethene	75-35-4	0.50	U	0.50	2.0	U	2.0
lethylene Chloride	75-09-2	0.50	U	0.50	1.7	U	1.7
1-Dichloroethane	75-34-3	0.50	U	0.50	2.0	U	2.0
s-1,2-Dichloroethene	156-59-2	0.50	U	0.50	2.0	U	2.0
hloroform	67-66-3	0.50	U	0.50	2.4	U	2.4
1,1-Trichloroethane	71-55-6	0.50	U	0.50	2.7	U	2.7
arbon Tetrachloride	56-23-5	0.50	U	0.50	3.1	U	3.1
enzene	71-43-2	0.91		0.50	2.9		1.6
2-Dichloroethane	107-06-2	0.50	U	0.50	2.0	U	2.0
ichloroethene	79-01-6	5.1		0.50	27		2.7
2-Dichloropropane	78-87-5	0.50	U	0.50	2.3	U	2.3
s-1,3-Dichloropropene	10061-01-5	0.50	U	0.50	2.3	U	2.3
oluene	108-88-3	7.1		0.50	27		1.9
ans-1,3-Dichloropropene	10061-02-6	0.50	U	0.50	2.3	U	2.3
1,2-Trichloroethane	79-00-5	0.50	U	0.50	2.7	U	2.7
etrachloroethene	127-18-4	0.50	U	0.50	3.4	U	3.4
hlorobenzene	108-90-7	0.50	U	0.50	2.3	U	2.3
hylbenzene	100-41-4	1.3		0.50	5.6		2.2
lene (m,p)	1330-20-7	4.7		0.50	20		2.2
yrene	100-42-5	1.8		0.50	7.7		2.1
lene (o)	95-47-6	1.4		0.50	6.1		2.2
1,2,2-Tetrachloroethane	79-34-5	0.50	U	0.50	3.4	U	3.4
3-Dichlorobenzene	541-73-1	0.50	U	0.50	3.0	U	3.0
4-Dichlorobenzene	106-46-7	0.50	U	0.50	3.0	U	3.0
2-Dichlorobenzene	95-50-1	0.50	U	0.50	3.0	U	3.0
2,4-Trichlorobenzene	120-82-1	0.50	U	0.50	3.7	U	3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

SG-TERADYNE

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: Air

Lab Sample No.: 576334

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	0.50	U	0.50	5.3	U	5.3
1,3,5-Trimethylbenzene	108-67-8	0.50	U	0.50	2.5	U	2.5
1,2,4-Trimethylbenzene	95-63-6	0.67		0.50	3.3		2.5
1,2-Dichlorotetrafluoroethane	76-14-2	0.50	U	0.50	3.5	U	3.5
1,2-Dibromoethane	106-93-4	0.50	U	0.50	3.8	U	3.8
1,3-Butadiene	106-99-0	0.50	U	0.50	1.1	U	1.1
Carbon Disulfide	75-15-0	2.0		0.50	6.2		1.6
Acetone	67-64-1	34		5.0	81		12
Isopropyl Alcohol	67-63-0	29		5.0	71		12
Methyl tert-Butyl Ether	1634-04-4	0.50	U	0.50	1.8	U	1.8
Cyclohexane	110-82-7	1.7		0.50	5.9		1.7
Dibromochloromethane	124-48-1	0.50	U	0.50	4.3	U	4.3
Methyl Ethyl Ketone	78-93-3	2.0		0.50	5.9		1.5
Methyl Isobutyl Ketone	108-10-1	0.50	U	0.50	2.0	U	2.0
Methyl Butyl Ketone	591-78-6	0.50	U	0.50	2.0	U	2.0
Bromodichloromethane	75-27-4	0.50	U	0.50	3.4	U	3.4
trans-1,2-Dichloroethene	156-60-5	0.50	U	0.50	2.0	U	2.0
4-Ethyltoluene	622-96-8	0.52		0.50	2.6		2.5
2-Chlorotoluene	95-49-8	0.50	U	0.50	2.6	U	2.6
n-Hexane	110-54-3	1.8		0.50	6.3		1.8
Tetrahydrofuran	109-99-9	5.0	U	5.0	15	U	15
n-Heptane	142-82-5	7.5		0.50	31		2.0

TO-14/15
Result Summary

CLIENT SAMPLE NO.

SG-TERADYNEDL

.ab Name: STL Burlington

;DG Number: 100862

:ase Number:

.ample Matrix: Air

Lab Sample No.: 576334D1

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
)ichlorodifluoromethane	75-71-8	1.0	U	1.0	4.9	U	4.9
:hloromethane	74-87-3	1.0	U	1.0	2.1	U	2.1
'inyl Chloride	75-01-4	1.0	U	1.0	2.6	U	2.6
·romomethane	74-83-9	1.0	U	1.0	3.9	U	3.9
:hloroethane	75-00-3	1.0	U	1.0	2.6	U	2.6
richlorofluoromethane	75-69-4	8.5	D	1.0	48	D	5.6
reon TF	76-13-1	1.0	U	1.0	7.7	U	7.7
,1-Dichloroethene	75-35-4	1.0	U	1.0	4.0	U	4.0
lethylene Chloride	75-09-2	1.0	U	1.0	3.5	U	3.5
,1-Dichloroethane	75-34-3	1.0	U	1.0	4.0	U	4.0
s-1,2-Dichloroethene	156-59-2	1.0	U	1.0	4.0	U	4.0
hloroform	67-66-3	1.0	U	1.0	4.9	U	4.9
,1,1-Trichloroethane	71-55-6	1.0	U	1.0	5.5	U	5.5
arbon Tetrachloride	56-23-5	1.0	U	1.0	6.3	U	6.3
enzene	71-43-2	1.1	D	1.0	3.5	D	3.2
,2-Dichloroethane	107-06-2	1.0	U	1.0	4.0	U	4.0
richloroethene	79-01-6	5.3	D	1.0	28	D	5.4
.2-Dichloropropane	78-87-5	1.0	U	1.0	4.6	U	4.6
s-1,3-Dichloropropene	10061-01-5	1.0	U	1.0	4.5	U	4.5
oluene	108-88-3	7.3	D	1.0	28	D	3.8
ans-1,3-Dichloropropene	10061-02-6	1.0	U	1.0	4.5	U	4.5
1,2-Trichloroethane	79-00-5	1.0	U	1.0	5.5	U	5.5
etrachloroethene	127-18-4	1.0	U	1.0	6.8	U	6.8
hlorobenzene	108-90-7	1.0	U	1.0	4.6	U	4.6
:hylbenzene	100-41-4	1.4	D	1.0	6.1	D	4.3
/lene (m,p)	1330-20-7	4.9	D	1.0	21	D	4.3
yrene	100-42-5	1.8	D	1.0	7.7	D	4.3
/lene (o)	95-47-6	1.5	D	1.0	6.5	D	4.3
1,2,2-Tetrachloroethane	79-34-5	1.0	U	1.0	6.9	U	6.9
3-Dichlorobenzene	541-73-1	1.0	U	1.0	6.0	U	6.0
4-Dichlorobenzene	106-46-7	1.0	U	1.0	6.0	U	6.0
2-Dichlorobenzene	95-50-1	1.0	U	1.0	6.0	U	6.0
2,4-Trichlorobenzene	120-82-1	1.0	U	1.0	7.4	U	7.4

TO-14/15
Result Summary

CLIENT SAMPLE NO.

SG-TERADYNEDL

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: Air

Lab Sample No.: 576334D1

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	1.0	U	1.0	11	U	11
1,3,5-Trimethylbenzene	108-67-8	1.0	U	1.0	4.9	U	4.9
1,2,4-Trimethylbenzene	95-63-6	1.0	U	1.0	4.9	U	4.9
1,2-Dichlorotetrafluoroethane	76-14-2	1.0	U	1.0	7.0	U	7.0
1,2-Dibromoethane	106-93-4	1.0	U	1.0	7.7	U	7.7
1,3-Butadiene	106-99-0	1.0	U	1.0	2.2	U	2.2
Carbon Disulfide	75-15-0	2.2	D	1.0	6.9	D	3.1
Acetone	67-64-1	39	D	10	93	D	24
Isopropyl Alcohol	67-63-0	33	D	10	81	D	25
Methyl tert-Butyl Ether	1634-04-4	1.0	U	1.0	3.6	U	3.6
Cyclohexane	110-82-7	1.8	D	1.0	6.2	D	3.4
Dibromochloromethane	124-48-1	1.0	U	1.0	8.5	U	8.5
Methyl Ethyl Ketone	78-93-3	2.0	D	1.0	5.9	D	2.9
Methyl Isobutyl Ketone	108-10-1	1.1	D	1.0	4.5	D	4.1
Methyl Butyl Ketone	591-78-6	2.4	D	1.0	9.8	D	4.1
Bromodichloromethane	75-27-4	1.0	U	1.0	6.7	U	6.7
trans-1,2-Dichloroethene	156-60-5	1.0	U	1.0	4.0	U	4.0
4-Ethyltoluene	622-96-8	1.0	U	1.0	4.9	U	4.9
2-Chlorotoluene	95-49-8	1.0	U	1.0	5.2	U	5.2
n-Hexane	110-54-3	1.4	D	1.0	4.9	D	3.5
Tetrahydrofuran	109-99-9	10	U	10	29	U	29
n-Heptane	142-82-5	8.2	D	1.0	34	D	4.1

TO-14/15
Result Summary

CLIENT SAMPLE NO.

SG-TERADYNE-DUP

ab Name: STL Burlington

DG Number: 100862

ase Number:

ample Matrix: Air

Lab Sample No.: 576335

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
ichlorodifluoromethane	75-71-8	0.62		0.50	3.1		2.5
hloromethane	74-87-3	0.69		0.50	1.4		1.0
inyl Chloride	75-01-4	0.50	U	0.50	1.3	U	1.3
romomethane	74-83-9	0.50	U	0.50	1.9	U	1.9
hloroethane	75-00-3	0.61		0.50	1.6		1.3
richlorofluoromethane	75-69-4	8.6		0.50	48		2.8
reon TF	76-13-1	0.50	U	0.50	3.8	U	3.8
,1-Dichloroethene	75-35-4	0.50	U	0.50	2.0	U	2.0
lethylene Chloride	75-09-2	0.51		0.50	1.8		1.7
,1-Dichloroethane	75-34-3	0.50	U	0.50	2.0	U	2.0
is-1,2-Dichloroethene	156-59-2	0.50	U	0.50	2.0	U	2.0
hloroform	67-66-3	0.50	U	0.50	2.4	U	2.4
,1,1-Trichloroethane	71-55-6	0.50	U	0.50	2.7	U	2.7
arbon Tetrachloride	56-23-5	0.50	U	0.50	3.1	U	3.1
enzene	71-43-2	0.73		0.50	2.3		1.6
,2-Dichloroethane	107-06-2	0.50	U	0.50	2.0	U	2.0
richloroethene	79-01-6	4.1		0.50	22		2.7
,2-Dichloropropane	78-87-5	0.50	U	0.50	2.3	U	2.3
s-1,3-Dichloropropene	10061-01-5	0.50	U	0.50	2.3	U	2.3
oluene	108-88-3	6.5		0.50	24		1.9
ans-1,3-Dichloropropene	10061-02-6	0.50	U	0.50	2.3	U	2.3
,1,2-Trichloroethane	79-00-5	0.50	U	0.50	2.7	U	2.7
etrachloroethene	127-18-4	0.50	U	0.50	3.4	U	3.4
hlorobenzene	108-90-7	0.50	U	0.50	2.3	U	2.3
thylbenzene	100-41-4	1.1		0.50	4.8		2.2
ylene (m,p)	1330-20-7	4.5		0.50	20		2.2
tyrene	100-42-5	1.4		0.50	6.0		2.1
ylene (o)	95-47-6	1.4		0.50	6.1		2.2
1,2,2-Tetrachloroethane	79-34-5	0.50	U	0.50	3.4	U	3.4
3-Dichlorobenzene	541-73-1	0.50	U	0.50	3.0	U	3.0
4-Dichlorobenzene	106-46-7	0.50	U	0.50	3.0	U	3.0
2-Dichlorobenzene	95-50-1	0.50	U	0.50	3.0	U	3.0
2,4-Trichlorobenzene	120-82-1	0.50	U	0.50	3.7	U	3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

SG-TERADYNE-DUP

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: Air

Lab Sample No.: 576335

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	0.50	U	0.50	5.3	U	5.3
1,3,5-Trimethylbenzene	108-67-8	0.50	U	0.50	2.5	U	2.5
1,2,4-Trimethylbenzene	95-63-6	0.70		0.50	3.4		2.5
1,2-Dichlorotetrafluoroethane	76-14-2	0.50	U	0.50	3.5	U	3.5
1,2-Dibromoethane	106-93-4	0.50	U	0.50	3.8	U	3.8
1,3-Butadiene	106-99-0	0.50	U	0.50	1.1	U	1.1
Carbon Disulfide	75-15-0	0.98		0.50	3.1		1.6
Acetone	67-64-1	32		5.0	76		12
Isopropyl Alcohol	67-63-0	32		5.0	79		12
Methyl tert-Butyl Ether	1634-04-4	0.50	U	0.50	1.8	U	1.8
Cyclohexane	110-82-7	1.7		0.50	5.9		1.7
Dibromochloromethane	124-48-1	0.50	U	0.50	4.3	U	4.3
Methyl Ethyl Ketone	78-93-3	2.0		0.50	5.9		1.5
Methyl Isobutyl Ketone	108-10-1	0.50	U	0.50	2.0	U	2.0
Methyl Butyl Ketone	591-78-6	0.50	U	0.50	2.0	U	2.0
Bromodichloromethane	75-27-4	0.50	U	0.50	3.4	U	3.4
trans-1,2-Dichloroethene	156-60-5	0.50	U	0.50	2.0	U	2.0
4-Ethyltoluene	622-96-8	0.52		0.50	2.6		2.5
2-Chlorotoluene	95-49-8	0.50	U	0.50	2.6	U	2.6
n-Hexane	110-54-3	1.9		0.50	6.7		1.8
Tetrahydrofuran	109-99-9	5.0	U	5.0	15	U	15
n-Heptane	142-82-5	8.4		0.50	34		2.0

TO-14/15
Result Summary

CLIENT SAMPLE NO.

TRIP BLANK

ab Name: STL Burlington

;DG Number: 100862

:ase Number:

;ample Matrix: Air

Lab Sample No.: 576336

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
)ichlorodifluoromethane	75-71-8	0.50	U	0.50	2.5	U	2.5
:hloromethane	74-87-3	0.50	U	0.50	1.0	U	1.0
/inyl Chloride	75-01-4	0.50	U	0.50	1.3	U	1.3
3romomethane	74-83-9	0.50	U	0.50	1.9	U	1.9
:hloroethane	75-00-3	0.50	U	0.50	1.3	U	1.3
'richlorofluoromethane	75-69-4	0.50	U	0.50	2.8	U	2.8
:reon TF	76-13-1	0.50	U	0.50	3.8	U	3.8
,1-Dichloroethene	75-35-4	0.50	U	0.50	2.0	U	2.0
/lethylene Chloride	75-09-2	0.50	U	0.50	1.7	U	1.7
,1-Dichloroethane	75-34-3	0.50	U	0.50	2.0	U	2.0
:is-1,2-Dichloroethene	156-59-2	0.50	U	0.50	2.0	U	2.0
:hloroform	67-66-3	0.50	U	0.50	2.4	U	2.4
,1,1-Trichloroethane	71-55-6	0.50	U	0.50	2.7	U	2.7
:arbon Tetrachloride	56-23-5	0.50	U	0.50	3.1	U	3.1
3enzene	71-43-2	0.50	U	0.50	1.6	U	1.6
,2-Dichloroethane	107-06-2	0.50	U	0.50	2.0	U	2.0
'richloroethene	79-01-6	0.50	U	0.50	2.7	U	2.7
,2-Dichloropropane	78-87-5	0.50	U	0.50	2.3	U	2.3
:is-1,3-Dichloropropene	10061-01-5	0.50	U	0.50	2.3	U	2.3
'oluene	108-88-3	0.50	U	0.50	1.9	U	1.9
rans-1,3-Dichloropropene	10061-02-6	0.50	U	0.50	2.3	U	2.3
,1,2-Trichloroethane	79-00-5	0.50	U	0.50	2.7	U	2.7
'etrachloroethene	127-18-4	0.50	U	0.50	3.4	U	3.4
:hlorobenzene	108-90-7	0.50	U	0.50	2.3	U	2.3
:thylbenzene	100-41-4	0.50	U	0.50	2.2	U	2.2
(ylene (m,p)	1330-20-7	0.50	U	0.50	2.2	U	2.2
3tyrene	100-42-5	0.50	U	0.50	2.1	U	2.1
(ylene (o)	95-47-6	0.50	U	0.50	2.2	U	2.2
,1,2,2-Tetrachloroethane	79-34-5	0.50	U	0.50	3.4	U	3.4
,3-Dichlorobenzene	541-73-1	0.50	U	0.50	3.0	U	3.0
,4-Dichlorobenzene	106-46-7	0.50	U	0.50	3.0	U	3.0
,2-Dichlorobenzene	95-50-1	0.50	U	0.50	3.0	U	3.0
,2,4-Trichlorobenzene	120-82-1	0.50	U	0.50	3.7	U	3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

TRIP BLANK

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: Air

Lab Sample No.: 576336

Date Analyzed: 07/02/2004

Date Received: 06/19/2004

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	0.50	U	0.50	5.3	U	5.3
1,3,5-Trimethylbenzene	108-67-8	0.50	U	0.50	2.5	U	2.5
1,2,4-Trimethylbenzene	95-63-6	0.50	U	0.50	2.5	U	2.5
1,2-Dichlorotetrafluoroethane	76-14-2	0.50	U	0.50	3.5	U	3.5
1,2-Dibromoethane	106-93-4	0.50	U	0.50	3.8	U	3.8
1,3-Butadiene	106-99-0	0.50	U	0.50	1.1	U	1.1
Carbon Disulfide	75-15-0	0.50	U	0.50	1.6	U	1.6
Acetone	67-64-1	5.0	U	5.0	12	U	12
Isopropyl Alcohol	67-63-0	5.0	U	5.0	12	U	12
Methyl tert-Butyl Ether	1634-04-4	0.50	U	0.50	1.8	U	1.8
Cyclohexane	110-82-7	0.50	U	0.50	1.7	U	1.7
Dibromochloromethane	124-48-1	0.50	U	0.50	4.3	U	4.3
Methyl Ethyl Ketone	78-93-3	0.50	U	0.50	1.5	U	1.5
Methyl Isobutyl Ketone	108-10-1	0.50	U	0.50	2.0	U	2.0
Methyl Butyl Ketone	591-78-6	0.50	U	0.50	2.0	U	2.0
Bromodichloromethane	75-27-4	0.50	U	0.50	3.4	U	3.4
trans-1,2-Dichloroethene	156-60-5	0.50	U	0.50	2.0	U	2.0
4-Ethyltoluene	622-96-8	0.50	U	0.50	2.5	U	2.5
2-Chlorotoluene	95-49-8	0.50	U	0.50	2.6	U	2.6
n-Hexane	110-54-3	0.50	U	0.50	1.8	U	1.8
Tetrahydrofuran	109-99-9	5.0	U	5.0	15	U	15
n-Heptane	142-82-5	0.50	U	0.50	2.0	U	2.0

TO-14/15
Result Summary

CLIENT SAMPLE NO.

ABLKB1

ab Name: STL Burlington

DG Number: 100862

ase Number:

ample Matrix: AIR

Lab Sample No.: ABLKB1

Date Analyzed: 07/01/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
ichlorodifluoromethane	75-71-8	0.50	U	0.50	2.5	U	2.5
hloromethane	74-87-3	0.50	U	0.50	1.0	U	1.0
inyl Chloride	75-01-4	0.50	U	0.50	1.3	U	1.3
romomethane	74-83-9	0.50	U	0.50	1.9	U	1.9
hloroethane	75-00-3	0.50	U	0.50	1.3	U	1.3
richlorofluoromethane	75-69-4	0.50	U	0.50	2.8	U	2.8
reon TF	76-13-1	0.50	U	0.50	3.8	U	3.8
,1-Dichloroethene	75-35-4	0.50	U	0.50	2.0	U	2.0
lethylene Chloride	75-09-2	0.50	U	0.50	1.7	U	1.7
,1-Dichloroethane	75-34-3	0.50	U	0.50	2.0	U	2.0
is-1,2-Dichloroethene	156-59-2	0.50	U	0.50	2.0	U	2.0
hloroform	67-66-3	0.50	U	0.50	2.4	U	2.4
,1,1-Trichloroethane	71-55-6	0.50	U	0.50	2.7	U	2.7
arbon Tetrachloride	56-23-5	0.50	U	0.50	3.1	U	3.1
enzene	71-43-2	0.50	U	0.50	1.6	U	1.6
,2-Dichloroethane	107-06-2	0.50	U	0.50	2.0	U	2.0
richloroethene	79-01-6	0.50	U	0.50	2.7	U	2.7
,2-Dichloropropane	78-87-5	0.50	U	0.50	2.3	U	2.3
is-1,3-Dichloropropene	10061-01-5	0.50	U	0.50	2.3	U	2.3
oluene	108-88-3	0.50	U	0.50	1.9	U	1.9
ans-1,3-Dichloropropene	10061-02-6	0.50	U	0.50	2.3	U	2.3
,1,2-Trichloroethane	79-00-5	0.50	U	0.50	2.7	U	2.7
etrachloroethene	127-18-4	0.50	U	0.50	3.4	U	3.4
hlorobenzene	108-90-7	0.50	U	0.50	2.3	U	2.3
thylbenzene	100-41-4	0.50	U	0.50	2.2	U	2.2
ylene (m,p)	1330-20-7	0.50	U	0.50	2.2	U	2.2
tyrene	100-42-5	0.50	U	0.50	2.1	U	2.1
ylene (o)	95-47-6	0.50	U	0.50	2.2	U	2.2
,1,2,2-Tetrachloroethane	79-34-5	0.50	U	0.50	3.4	U	3.4
,3-Dichlorobenzene	541-73-1	0.50	U	0.50	3.0	U	3.0
,4-Dichlorobenzene	106-46-7	0.50	U	0.50	3.0	U	3.0
,2-Dichlorobenzene	95-50-1	0.50	U	0.50	3.0	U	3.0
2,4-Trichlorobenzene	120-82-1	0.50	U	0.50	3.7	U	3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

ABLKB1

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: ABLKB1

Date Analyzed: 07/01/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	0.50	U	0.50	5.3	U	5.3
1,3,5-Trimethylbenzene	108-67-8	0.50	U	0.50	2.5	U	2.5
1,2,4-Trimethylbenzene	95-63-6	0.50	U	0.50	2.5	U	2.5
1,2-Dichlorotetrafluoroethane	76-14-2	0.50	U	0.50	3.5	U	3.5
1,2-Dibromoethane	106-93-4	0.50	U	0.50	3.8	U	3.8
1,3-Butadiene	106-99-0	0.50	U	0.50	1.1	U	1.1
Carbon Disulfide	75-15-0	0.50	U	0.50	1.6	U	1.6
Acetone	67-64-1	5.0	U	5.0	12	U	12
Isopropyl Alcohol	67-63-0	5.0	U	5.0	12	U	12
Methyl tert-Butyl Ether	1634-04-4	0.50	U	0.50	1.8	U	1.8
Cyclohexane	110-82-7	0.50	U	0.50	1.7	U	1.7
Dibromochloromethane	124-48-1	0.50	U	0.50	4.3	U	4.3
Methyl Ethyl Ketone	78-93-3	0.50	U	0.50	1.5	U	1.5
1,4-Dioxane	123-91-1	5.0	U	5.0	18	U	18
Methyl Isobutyl Ketone	108-10-1	0.50	U	0.50	2.0	U	2.0
Methyl Butyl Ketone	591-78-6	0.50	U	0.50	2.0	U	2.0
Bromoform	75-25-2	0.50	U	0.50	5.2	U	5.2
Bromodichloromethane	75-27-4	0.50	U	0.50	3.4	U	3.4
trans-1,2-Dichloroethene	156-60-5	0.50	U	0.50	2.0	U	2.0
4-Ethyltoluene	622-96-8	0.50	U	0.50	2.5	U	2.5
3-Chloropropene	107-05-1	0.50	U	0.50	1.6	U	1.6
2,2,4-Trimethylpentane	540-84-1	0.50	U	0.50	2.3	U	2.3
Bromoethene	593-60-2	0.50	U	0.50	2.2	U	2.2
2-Chlorotoluene	95-49-8	0.50	U	0.50	2.6	U	2.6
n-Hexane	110-54-3	0.50	U	0.50	1.8	U	1.8
Tetrahydrofuran	109-99-9	5.0	U	5.0	15	U	15
n-Heptane	142-82-5	0.50	U	0.50	2.0	U	2.0
Methyl Methacrylate	80-62-6	0.50	U	0.50	2.0	U	2.0
1,2-Dichloroethene (total)	540-59-0	0.50	U	0.50	2.0	U	2.0
Xylene (total)	1330-20-7	0.50	U	0.50	2.2	U	2.2
tert-Butyl Alcohol	75-65-0	5.0	U	5.0	15	U	15

TO-14/15
Result Summary

CLIENT SAMPLE NO.

ABLKB3

.ab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: ABLKB3

Date Analyzed: 07/02/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Dichlorodifluoromethane	75-71-8	0.50	U	0.50	2.5	U	2.5
Chloromethane	74-87-3	0.50	U	0.50	1.0	U	1.0
Vinyl Chloride	75-01-4	0.50	U	0.50	1.3	U	1.3
Bromomethane	74-83-9	0.50	U	0.50	1.9	U	1.9
Chloroethane	75-00-3	0.50	U	0.50	1.3	U	1.3
Trichlorofluoromethane	75-69-4	0.50	U	0.50	2.8	U	2.8
Freon TF	76-13-1	0.50	U	0.50	3.8	U	3.8
,1-Dichloroethene	75-35-4	0.50	U	0.50	2.0	U	2.0
Methylene Chloride	75-09-2	0.50	U	0.50	1.7	U	1.7
,1-Dichloroethane	75-34-3	0.50	U	0.50	2.0	U	2.0
is-1,2-Dichloroethene	156-59-2	0.50	U	0.50	2.0	U	2.0
Chloroform	67-66-3	0.50	U	0.50	2.4	U	2.4
,1,1-Trichloroethane	71-55-6	0.50	U	0.50	2.7	U	2.7
Carbon Tetrachloride	56-23-5	0.50	U	0.50	3.1	U	3.1
Benzene	71-43-2	0.50	U	0.50	1.6	U	1.6
,2-Dichloroethane	107-06-2	0.50	U	0.50	2.0	U	2.0
Trichloroethene	79-01-6	0.50	U	0.50	2.7	U	2.7
,2-Dichloropropane	78-87-5	0.50	U	0.50	2.3	U	2.3
is-1,3-Dichloropropene	10061-01-5	0.50	U	0.50	2.3	U	2.3
oluene	108-88-3	0.50	U	0.50	1.9	U	1.9
ans-1,3-Dichloropropene	10061-02-6	0.50	U	0.50	2.3	U	2.3
,1,2-Trichloroethane	79-00-5	0.50	U	0.50	2.7	U	2.7
etrachloroethene	127-18-4	0.50	U	0.50	3.4	U	3.4
hlorobenzene	108-90-7	0.50	U	0.50	2.3	U	2.3
thylbenzene	100-41-4	0.50	U	0.50	2.2	U	2.2
ylene (m,p)	1330-20-7	0.50	U	0.50	2.2	U	2.2
tyrene	100-42-5	0.50	U	0.50	2.1	U	2.1
ylene (o)	95-47-6	0.50	U	0.50	2.2	U	2.2
1,2,2-Tetrachloroethane	79-34-5	0.50	U	0.50	3.4	U	3.4
3-Dichlorobenzene	541-73-1	0.50	U	0.50	3.0	U	3.0
4-Dichlorobenzene	106-46-7	0.50	U	0.50	3.0	U	3.0
2-Dichlorobenzene	95-50-1	0.50	U	0.50	3.0	U	3.0
2,4-Trichlorobenzene	120-82-1	0.50	U	0.50	3.7	U	3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

ABLKB3

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: ABLKB3

Date Analyzed: 07/02/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	0.50	U	0.50	5.3	U	5.3
1,3,5-Trimethylbenzene	108-67-8	0.50	U	0.50	2.5	U	2.5
1,2,4-Trimethylbenzene	95-63-6	0.50	U	0.50	2.5	U	2.5
1,2-Dichlorotetrafluoroethane	76-14-2	0.50	U	0.50	3.5	U	3.5
1,2-Dibromoethane	106-93-4	0.50	U	0.50	3.8	U	3.8
1,3-Butadiene	106-99-0	0.50	U	0.50	1.1	U	1.1
Carbon Disulfide	75-15-0	0.50	U	0.50	1.6	U	1.6
Acetone	67-64-1	5.0	U	5.0	12	U	12
Isopropyl Alcohol	67-63-0	5.0	U	5.0	12	U	12
Methyl tert-Butyl Ether	1634-04-4	0.50	U	0.50	1.8	U	1.8
Cyclohexane	110-82-7	0.50	U	0.50	1.7	U	1.7
Dibromochloromethane	124-48-1	0.50	U	0.50	4.3	U	4.3
Methyl Ethyl Ketone	78-93-3	0.50	U	0.50	1.5	U	1.5
Methyl Isobutyl Ketone	108-10-1	0.50	U	0.50	2.0	U	2.0
Methyl Butyl Ketone	591-78-6	0.50	U	0.50	2.0	U	2.0
Bromodichloromethane	75-27-4	0.50	U	0.50	3.4	U	3.4
trans-1,2-Dichloroethene	156-60-5	0.50	U	0.50	2.0	U	2.0
4-Ethyltoluene	622-96-8	0.50	U	0.50	2.5	U	2.5
2-Chlorotoluene	95-49-8	0.50	U	0.50	2.6	U	2.6
n-Hexane	110-54-3	0.50	U	0.50	1.8	U	1.8
Tetrahydrofuran	109-99-9	5.0	U	5.0	15	U	15
n-Heptane	142-82-5	0.50	U	0.50	2.0	U	2.0

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B1LCS

.ab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: B1LCS

Date Analyzed: 07/01/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Dichlorodifluoromethane	75-71-8	11		0.50	54		2.5
Chloromethane	74-87-3	11		0.50	23		1.0
Vinyl Chloride	75-01-4	11		0.50	28		1.3
Bromomethane	74-83-9	12		0.50	47		1.9
Chloroethane	75-00-3	12		0.50	32		1.3
Trichlorofluoromethane	75-69-4	12		0.50	67		2.8
Freon TF	76-13-1	11		0.50	84		3.8
,1-Dichloroethene	75-35-4	11		0.50	44		2.0
Methylene Chloride	75-09-2	10		0.50	35		1.7
,1-Dichloroethane	75-34-3	11		0.50	45		2.0
is-1,2-Dichloroethene	156-59-2	10		0.50	40		2.0
Chloroform	67-66-3	11		0.50	54		2.4
,1,1-Trichloroethane	71-55-6	11		0.50	60		2.7
Carbon Tetrachloride	56-23-5	11		0.50	69		3.1
Benzene	71-43-2	11		0.50	35		1.6
,2-Dichloroethane	107-06-2	12		0.50	49		2.0
richloroethene	79-01-6	10		0.50	54		2.7
,2-Dichloropropane	78-87-5	10		0.50	46		2.3
is-1,3-Dichloropropene	10061-01-5	10		0.50	45		2.3
oluene	108-88-3	11		0.50	41		1.9
ans-1,3-Dichloropropene	10061-02-6	10		0.50	45		2.3
,1,2-Trichloroethane	79-00-5	11		0.50	60		2.7
etrachloroethene	127-18-4	11		0.50	75		3.4
hlorobenzene	108-90-7	11		0.50	51		2.3
thylbenzene	100-41-4	11		0.50	48		2.2
ylene (m,p)	1330-20-7	24		0.50	100		2.2
tyrene	100-42-5	12		0.50	51		2.1
ylene (o)	95-47-6	12		0.50	52		2.2
1,2,2-Tetrachloroethane	79-34-5	12		0.50	82		3.4
3-Dichlorobenzene	541-73-1	12		0.50	72		3.0
4-Dichlorobenzene	106-46-7	12		0.50	72		3.0
2-Dichlorobenzene	95-50-1	12		0.50	72		3.0
2,4-Trichlorobenzene	120-82-1	18		0.50	130		3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B1LCS

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: B1LCS

Date Analyzed: 07/01/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	14		0.50	150		5.3
1,3,5-Trimethylbenzene	108-67-8	14		0.50	69		2.5
1,2,4-Trimethylbenzene	95-63-6	12		0.50	59		2.5
1,2-Dichlorotetrafluoroethane	76-14-2	11		0.50	77		3.5
1,2-Dibromoethane	106-93-4	11		0.50	85		3.8
1,3-Butadiene	106-99-0	12		0.50	27		1.1
Carbon Disulfide	75-15-0	11		0.50	34		1.6
Acetone	67-64-1	12		5.0	29		12
Isopropyl Alcohol	67-63-0	9.8		5.0	24		12
Methyl tert-Butyl Ether	1634-04-4	11		0.50	40		1.8
Cyclohexane	110-82-7	11		0.50	38		1.7
Dibromochloromethane	124-48-1	12		0.50	100		4.3
Methyl Ethyl Ketone	78-93-3	10		0.50	29		1.5
1,4-Dioxane	123-91-1	9.1		5.0	33		18
Methyl Isobutyl Ketone	108-10-1	9.8		0.50	40		2.0
Methyl Butyl Ketone	591-78-6	9.8		0.50	40		2.0
Bromoform	75-25-2	12		0.50	120		5.2
Bromodichloromethane	75-27-4	11		0.50	74		3.4
trans-1,2-Dichloroethene	156-60-5	11		0.50	44		2.0
4-Ethyltoluene	622-96-8	10		0.50	49		2.5
3-Chloropropene	107-05-1	10		0.50	31		1.6
2,2,4-Trimethylpentane	540-84-1	11		0.50	51		2.3
Bromoethene	593-60-2	12		0.50	52		2.2
2-Chlorotoluene	95-49-8	12		0.50	62		2.6
n-Hexane	110-54-3	11		0.50	39		1.8
Tetrahydrofuran	109-99-9	10		5.0	29		15
n-Heptane	142-82-5	12		0.50	49		2.0
Methyl Methacrylate	80-62-6	10		0.50	41		2.0
1,2-Dichloroethene (total)	540-59-0	22		0.50	87		2.0
Xylene (total)	1330-20-7	32		0.50	140		2.2
tert-Butyl Alcohol	75-65-0	11		5.0	33		15

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B1LCSD

.ab Name: STL Burlington

;DG Number: 100862

:ase Number:

ample Matrix: AIR

Lab Sample No.: B1LCSD

Date Analyzed: 07/01/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
ichlorodifluoromethane	75-71-8	12		0.50	59		2.5
hloromethane	74-87-3	12		0.50	25		1.0
inyl Chloride	75-01-4	12		0.50	31		1.3
romomethane	74-83-9	12		0.50	47		1.9
hloroethane	75-00-3	12		0.50	32		1.3
richlorofluoromethane	75-69-4	12		0.50	67		2.8
reon TF	76-13-1	11		0.50	84		3.8
,1-Dichloroethene	75-35-4	11		0.50	44		2.0
lethylene Chloride	75-09-2	10		0.50	35		1.7
,1-Dichloroethane	75-34-3	11		0.50	45		2.0
s-1,2-Dichloroethene	156-59-2	10		0.50	40		2.0
hloroform	67-66-3	11		0.50	54		2.4
1,1-Trichloroethane	71-55-6	12		0.50	65		2.7
arbon Tetrachloride	56-23-5	11		0.50	69		3.1
enzene	71-43-2	11		0.50	35		1.6
2-Dichloroethane	107-06-2	12		0.50	49		2.0
ichloroethene	79-01-6	10		0.50	54		2.7
2-Dichloropropane	78-87-5	11		0.50	51		2.3
s-1,3-Dichloropropene	10061-01-5	10		0.50	45		2.3
oluene	108-88-3	11		0.50	41		1.9
ans-1,3-Dichloropropene	10061-02-6	11		0.50	50		2.3
1,2-Trichloroethane	79-00-5	11		0.50	60		2.7
etrachloroethene	127-18-4	11		0.50	75		3.4
hlorobenzene	108-90-7	11		0.50	51		2.3
hylbenzene	100-41-4	12		0.50	52		2.2
lene (m,p)	1330-20-7	25		0.50	110		2.2
yrene	100-42-5	12		0.50	51		2.1
lene (o)	95-47-6	12		0.50	52		2.2
1,2,2-Tetrachloroethane	79-34-5	12		0.50	82		3.4
3-Dichlorobenzene	541-73-1	12		0.50	72		3.0
4-Dichlorobenzene	106-46-7	12		0.50	72		3.0
2-Dichlorobenzene	95-50-1	13		0.50	78		3.0
2,4-Trichlorobenzene	120-82-1	20		0.50	150		3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B1LCSD

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: B1LCSD

Date Analyzed: 07/01/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	16		0.50	170		5.3
1,3,5-Trimethylbenzene	108-67-8	14		0.50	69		2.5
1,2,4-Trimethylbenzene	95-63-6	13		0.50	64		2.5
1,2-Dichlorotetrafluoroethane	76-14-2	12		0.50	84		3.5
1,2-Dibromoethane	106-93-4	11		0.50	85		3.8
1,3-Butadiene	106-99-0	12		0.50	27		1.1
Carbon Disulfide	75-15-0	11		0.50	34		1.6
Acetone	67-64-1	12		5.0	29		12
Isopropyl Alcohol	67-63-0	9.9		5.0	24		12
Methyl tert-Butyl Ether	1634-04-4	12		0.50	43		1.8
Cyclohexane	110-82-7	11		0.50	38		1.7
Dibromochloromethane	124-48-1	12		0.50	100		4.3
Methyl Ethyl Ketone	78-93-3	10		0.50	29		1.5
1,4-Dioxane	123-91-1	8.8		5.0	32		18
Methyl Isobutyl Ketone	108-10-1	9.8		0.50	40		2.0
Methyl Butyl Ketone	591-78-6	10		0.50	41		2.0
Bromoform	75-25-2	12		0.50	120		5.2
Bromodichloromethane	75-27-4	11		0.50	74		3.4
trans-1,2-Dichloroethene	156-60-5	11		0.50	44		2.0
4-Ethyltoluene	622-96-8	12		0.50	59		2.5
3-Chloropropene	107-05-1	10		0.50	31		1.6
2,2,4-Trimethylpentane	540-84-1	11		0.50	51		2.3
Bromoethene	593-60-2	12		0.50	52		2.2
2-Chlorotoluene	95-49-8	13		0.50	67		2.6
n-Hexane	110-54-3	11		0.50	39		1.8
Tetrahydrofuran	109-99-9	11		5.0	32		15
n-Heptane	142-82-5	12		0.50	49		2.0
Methyl Methacrylate	80-62-6	10		0.50	41		2.0
1,2-Dichloroethene (total)	540-59-0	22		0.50	87		2.0
Xylene (total)	1330-20-7	33		0.50	140		2.2
tert-Butyl Alcohol	75-65-0	11		5.0	33		15

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B3LCS

.ab Name: STL Burlington

;DG Number: 100862

:ase Number:

·ample Matrix: AIR

Lab Sample No.: B3LCS

Date Analyzed: 07/02/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
)ichlorodifluoromethane	75-71-8	9.8		0.50	48		2.5
:hloromethane	74-87-3	10		0.50	21		1.0
'inyl Chloride	75-01-4	10		0.50	26		1.3
ıromomethane	74-83-9	11		0.50	43		1.9
:hloroethane	75-00-3	10		0.50	26		1.3
'richlorofluoromethane	75-69-4	10		0.50	56		2.8
'reon TF	76-13-1	10		0.50	77		3.8
,1-Dichloroethene	75-35-4	11		0.50	44		2.0
'lethylene Chloride	75-09-2	9.2		0.50	32		1.7
,1-Dichloroethane	75-34-3	10		0.50	40		2.0
is-1,2-Dichloroethene	156-59-2	10		0.50	40		2.0
:hloroform	67-66-3	10		0.50	49		2.4
,1,1-Trichloroethane	71-55-6	10		0.50	55		2.7
:arbon Tetrachloride	56-23-5	10		0.50	63		3.1
;enzene	71-43-2	11		0.50	35		1.6
,2-Dichloroethane	107-06-2	11		0.50	45		2.0
richloroethene	79-01-6	10		0.50	54		2.7
,2-Dichloropropane	78-87-5	10		0.50	46		2.3
is-1,3-Dichloropropene	10061-01-5	10		0.50	45		2.3
oluene	108-88-3	11		0.50	41		1.9
ans-1,3-Dichloropropene	10061-02-6	10		0.50	45		2.3
,1,2-Trichloroethane	79-00-5	10		0.50	55		2.7
etrachloroethene	127-18-4	11		0.50	75		3.4
hlorobenzene	108-90-7	11		0.50	51		2.3
thylbenzene	100-41-4	11		0.50	48		2.2
ylene (m,p)	1330-20-7	23		0.50	100		2.2
tyrene	100-42-5	11		0.50	47		2.1
ylene (o)	95-47-6	11		0.50	48		2.2
,1,2,2-Tetrachloroethane	79-34-5	11		0.50	76		3.4
,3-Dichlorobenzene	541-73-1	11		0.50	66		3.0
4-Dichlorobenzene	106-46-7	11		0.50	66		3.0
2-Dichlorobenzene	95-50-1	11		0.50	66		3.0
2,4-Trichlorobenzene	120-82-1	17		0.50	130		3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B3LCS

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: B3LCS

Date Analyzed: 07/02/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	13		0.50	140		5.3
1,3,5-Trimethylbenzene	108-67-8	11		0.50	54		2.5
1,2,4-Trimethylbenzene	95-63-6	11		0.50	54		2.5
1,2-Dichlorotetrafluoroethane	76-14-2	10		0.50	70		3.5
1,2-Dibromoethane	106-93-4	10		0.50	77		3.8
1,3-Butadiene	106-99-0	10		0.50	22		1.1
Carbon Disulfide	75-15-0	10		0.50	31		1.6
Acetone	67-64-1	10		5.0	24		12
Isopropyl Alcohol	67-63-0	9.1		5.0	22		12
Methyl tert-Butyl Ether	1634-04-4	10		0.50	36		1.8
Cyclohexane	110-82-7	11		0.50	38		1.7
Dibromochloromethane	124-48-1	11		0.50	94		4.3
Methyl Ethyl Ketone	78-93-3	9.8		0.50	29		1.5
Methyl Isobutyl Ketone	108-10-1	9.8		0.50	40		2.0
Methyl Butyl Ketone	591-78-6	9.3		0.50	38		2.0
Bromodichloromethane	75-27-4	10		0.50	67		3.4
trans-1,2-Dichloroethene	156-60-5	10		0.50	40		2.0
4-Ethyltoluene	622-96-8	11		0.50	54		2.5
2-Chlorotoluene	95-49-8	11		0.50	57		2.6
n-Hexane	110-54-3	9.7		0.50	34		1.8
Tetrahydrofuran	109-99-9	10		5.0	29		15
n-Heptane	142-82-5	11		0.50	45		2.0

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B3LCSD

.ab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: B3LCSD

Date Analyzed: 07/02/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Dichlorodifluoromethane	75-71-8	11		0.50	54		2.5
Chloromethane	74-87-3	11		0.50	23		1.0
Vinyl Chloride	75-01-4	11		0.50	28		1.3
Bromomethane	74-83-9	11		0.50	43		1.9
Chloroethane	75-00-3	11		0.50	29		1.3
Trichlorofluoromethane	75-69-4	11		0.50	62		2.8
Freon TF	76-13-1	11		0.50	84		3.8
,1-Dichloroethene	75-35-4	11		0.50	44		2.0
Methylene Chloride	75-09-2	9.6		0.50	33		1.7
,1-Dichloroethane	75-34-3	10		0.50	40		2.0
is-1,2-Dichloroethene	156-59-2	10		0.50	40		2.0
Chloroform	67-66-3	10		0.50	49		2.4
,1,1-Trichloroethane	71-55-6	11		0.50	60		2.7
Carbon Tetrachloride	56-23-5	11		0.50	69		3.1
Benzene	71-43-2	11		0.50	35		1.6
,2-Dichloroethane	107-06-2	11		0.50	45		2.0
Trichloroethene	79-01-6	10		0.50	54		2.7
,2-Dichloropropane	78-87-5	10		0.50	46		2.3
is-1,3-Dichloropropene	10061-01-5	10		0.50	45		2.3
Toluene	108-88-3	10		0.50	38		1.9
rans-1,3-Dichloropropene	10061-02-6	10		0.50	45		2.3
,1,2-Trichloroethane	79-00-5	11		0.50	60		2.7
etrachloroethene	127-18-4	11		0.50	75		3.4
Chlorobenzene	108-90-7	11		0.50	51		2.3
thylbenzene	100-41-4	12		0.50	52		2.2
ylene (m,p)	1330-20-7	25		0.50	110		2.2
tyrene	100-42-5	12		0.50	51		2.1
ylene (o)	95-47-6	12		0.50	52		2.2
,1,2,2-Tetrachloroethane	79-34-5	12		0.50	82		3.4
,3-Dichlorobenzene	541-73-1	12		0.50	72		3.0
,4-Dichlorobenzene	106-46-7	12		0.50	72		3.0
,2-Dichlorobenzene	95-50-1	12		0.50	72		3.0
2,4-Trichlorobenzene	120-82-1	16		0.50	120		3.7

TO-14/15
Result Summary

CLIENT SAMPLE NO.

B3LCSD

Lab Name: STL Burlington

SDG Number: 100862

Case Number:

Sample Matrix: AIR

Lab Sample No.: B3LCSD

Date Analyzed: 07/02/2004

Date Received: / /

Target Compound	CAS Number	Results in ppbv	Q	RL in ppbv	Results in ug/m3	Q	RL in ug/m3
Hexachlorobutadiene	87-68-3	14		0.50	150		5.3
1,3,5-Trimethylbenzene	108-67-8	15		0.50	74		2.5
1,2,4-Trimethylbenzene	95-63-6	12		0.50	59		2.5
1,2-Dichlorotetrafluoroethane	76-14-2	11		0.50	77		3.5
1,2-Dibromoethane	106-93-4	11		0.50	85		3.8
1,3-Butadiene	106-99-0	11		0.50	24		1.1
Carbon Disulfide	75-15-0	11		0.50	34		1.6
Acetone	67-64-1	10		5.0	24		12
Isopropyl Alcohol	67-63-0	9.4		5.0	23		12
Methyl tert-Butyl Ether	1634-04-4	11		0.50	40		1.8
Cyclohexane	110-82-7	11		0.50	38		1.7
Dibromochloromethane	124-48-1	12		0.50	100		4.3
Methyl Ethyl Ketone	78-93-3	10		0.50	29		1.5
Methyl Isobutyl Ketone	108-10-1	9.7		0.50	40		2.0
Methyl Butyl Ketone	591-78-6	10		0.50	41		2.0
Bromodichloromethane	75-27-4	11		0.50	74		3.4
trans-1,2-Dichloroethene	156-60-5	11		0.50	44		2.0
4-Ethyltoluene	622-96-8	10		0.50	49		2.5
2-Chlorotoluene	95-49-8	12		0.50	62		2.6
n-Hexane	110-54-3	11		0.50	39		1.8
Tetrahydrofuran	109-99-9	10		5.0	29		15
n-Heptane	142-82-5	11		0.50	45		2.0

SEVERN TRENT LABORATORIES, INC. 208 South Park Drive, Suite 1
Colchester, VT 05446 Tel 802 655 1203

CHAIN OF CUSTODY RECORD

Report to:
Company: ROUX ASSOCIATES
Address: 25 Corporate Dr.
Burlington MA 01803
Contact: Larry McTiernan
Phone: 781-270-6600
Fax: 781-270-9060
Contract/ Quote:

Invoice to:
Company: SAME AS
Address: REPORT TO:
Contact:
Phone:
Fax:

Lab Use Only
Due Date:

Temp. of coolers when received (C°): 1 | 2 | 3 | 4 | 5

Custody Seal N / Y
Intact N / Y

Screened For Radioactivity ☐

Sampler's Name: C. Milone
Sampler's Signature: C. Milone

Proj. No.: 11940/M
Project Name: Solutia-WHP/Vining

ANALYSIS REQUESTED: TO15

Matrix[1]	Date	Time	Comp	Grab	Identifying Marks of Sample(s)	No/Type of Containers[2] VOA	A/G 1 Lt.	250 ml	P/O	TO15	Lab/Sample ID (Lab Use Only)
A	6/12/04	6:42		X	SG - Teradyne				1	X	
A	6/12/04	6:45		X	SG - Teradyne-DUP				1	X	
A					TRIP BLANK				1	X	

Relinquished by: (Signature)	Date	Time	Received by: (Signature)	Date	Time
C. Milone	6/12/04	16:00	Angela Patnaud	06.19.04	1030
Relinquished by: (Signature)	Date	Time	Received by: (Signature	Date	Time
Relinquished by: (Signature)	Date	Time	Received by: (Signature	Date	Time

Remarks: Fedex # 3456 3284 3060

Client's delivery of samples constitutes acceptance of Severn Trent Laboratories terms and conditions contained in the Price Schedule.

[1]Matrix WW - Wastewater W - Water S - Soil L - Liquid A - Air bag C - Charcoal Tube SL - Sludge O - Oil
[2]Container VOA - 40 ml vial A/G - Amber / Or Glass 1 Liter 250 ml - Glass wide mouth P/O - Plastic or other

STL cannot accept verbal changes. Please Fax written changes to (802) 655-1248

STL8234-200 (12/02)

CLICK TO CONTINUE

TRENT STL
SEVERN TRENT LABORATORIES, INC.

STL Burlington
208 South Park Drive, Suite 1
Colchester, VT 05446 Tel 802 655 1203

CHAIN OF CUSTODY RECORD

Report to:
Company: ROUX ASSOCIATES
Address: 25 Corporate Dr. Suite 230 Burlington, MA 01803
Contact:
Phone: 781-270-6600
Fax: 781-270-9066
Contract/ Quote:

Invoice to:
Company: SAME AS REPORT TO
Address:
Contact:
Phone:
Fax:

Sampler's Name: C. Milone
Sampler's Signature: C. Milone

Proj. No.: 11940/M
Project Name: Solutia - WHP/Vinity

ANALYSIS REQUESTED: VOC (8260B)

Lab Use Only
Due Date:
Temp. of coolers when received (C°): 1 | 2 | 3 | 4 | 5
Custody Seal N / Y
Intact N / Y
Screened For Radioactivity ☐

Matrix[1]	Date	Time	Comp	Grab	Identifying Marks of Sample(s)	No/Type of Containers[2] VOA	A/G 1 Lt.	250 ml	P/O	VOC (8260B)	Lab/Sample ID (Lab Use Only)
	6/18/04	9:00			Modeling Clay.	2				X	
	6/18/04	CJM			TRIP BLANK CJM	2				X	CJM CJM

Relinquished by: (Signature)	Date	Time	Received by: (Signature)	Date	Time
C. Milone	6/18/04	16:00	Angela Patnoad	06.19.04	1030
Relinquished by: (Signature)	Date	Time	Received by: (Signature	Date	Time
Relinquished by: (Signature)	Date	Time	Received by: (Signature	Date	Time

Remarks: FedEx # 8456 3284 3070

Client's delivery of samples constitutes acceptance of Severn Trent Laboratories terms and conditions contained in the Price Schedule.

[1]Matrix WW - Wastewater W - Water S - Soil L - Liquid A - Air bag C - Charcoal Tube SL - Sludge O - Oil
[2]Container VOA - 40 ml vial A/G - Amber / Or Glass 1 Liter 250 ml - Glass wide mouth P/O - Plastic or other

STL cannot accept verbal changes.
Please Fax written changes to

STL8234-200 (12/02)