



Tier II Data Validation Report Summary

Client: Chevron Environmental Management Company (EMC) Cincinnati	Laboratory: Lancaster Laboratories, Inc.
Project Name: Barrier Wall Monitoring Network Sampling	Sample Matrix: Groundwater
Project Number: 500-017-012	Sample Start Date: December 8, 2009
Date Validated: February 2, 2010	Sample End Date: December 8, 2009
Parameters Included: Volatile Organic Compounds (VOC) by Solid Waste-846 (SW-846) Method 8260B; Carbon Dioxide by SW-846 Method 8000B; Methane by SW-846 Modified Method 8015B; Total and Dissolved Metals by SW-846 Method 6010B; Ferric Iron by SW-846 Modified Method 6010B; Chloride and Sulfate by Environmental Protection Agency (EPA) Method 300.0; Kjeldahl Nitrogen by EPA Method 351.2; Nitrate Nitrogen and Nitrite Nitrogen by EPA Method 353.2; Total Carbon (TC), Total Inorganic Carbon (TIC), and Total Organic Carbon (TOC) by EPA Method 415.1; Alkalinity by Method Standard Method 20 th Edition (SM20) 2320B; Ferrous Iron by Modified Method SM20 3500 Fe B; Sulfide by Method SM20 4500 S2 D; and Ammonia Nitrogen by Modified Method SM20 4500NH3 B/C	
Laboratory Project ID: 1174348	
Data Validator: Tim Gunn, CHMM	

DATA EVALUATION CRITERIA SUMMARY

A Tier II Data Validation was performed by Trihydro Corporation's Chemical Data Evaluation Services group on the analytical data report package generated by Lancaster Laboratories evaluating samples from the Chevron EMC site located in Cincinnati, Ohio.

Precision, accuracy, method compliance, and completeness of this data package were assessed during this data review. Precision was determined by evaluating the calculated relative percent difference (RPD) values of samples from laboratory duplicate pairs. Laboratory accuracy was established by reviewing the demonstrated percent recoveries of matrix spike (MS) and matrix spike duplicate (MSD) samples, and of laboratory control samples (LCS) and laboratory control sample duplicates (LCSD) to verify that none of the data were biased. Additionally, field accuracy was established by collecting a trip blank to monitor for possible ambient or cross contamination during sampling. Method compliance was established by reviewing holding times, detection limits, surrogate recoveries, method blanks, and the LCS and LCSD percent recoveries against method specific requirements. Completeness was evaluated by determining the overall ratio of the number of samples planned versus the number of samples with valid analyses. Determination of completeness included a review of the chain-of-custody, laboratory analytical methods, and any other necessary documents associated with this analytical data set.

Data were evaluated in general accordance with validation criteria set forth in the USEPA Contract Laboratory Program (CLP) National Functional Guidelines for Superfund Organic Methods Data Review, document number USEPA-540-R-08-01, June 2008 with additional reference to USEPA Contract Laboratory Program (CLP) National Functional Guidelines for Organic Data Review, document number EPA 540/R-99-008 of October 1999 and the USEPA CLP National Functional Guidelines for Inorganic Data Review, document number EPA 540R-04-004, October 2004. Review of duplicates is conducted in accordance with USEPA Region 1 Laboratory Data Validation Function Guidelines for Evaluation of Organic Analysis, December 1996.



Tier II Data Validation Report Summary

SAMPLE NUMBERS TABLE

Client Sample ID	Laboratory Sample Number
MW-135D, 120809	5858865
MW-135D, 120809 Filtered	5858866
MW-135I, 120809	5858867
MW-135I, 120809 Filtered	5858868
MW-135S, 120809	5858869
MW-135S, 120809 Filtered	5858870
BMW-1D, 120809	5858871
BMW-1D, 120809 Filtered	5858872
BMW-1I, 120809	5858873
BMW-1I, 120809 Filtered	5858874
BMW-1S, 120809	5858875
BMW-1S, 120809 Filtered	5858876
MW-136D, 120809	5858877
MW-136D, 120809 Filtered	5858878
Trip_Blank, 120809	5858879



Tier II Data Validation Report

The samples were analyzed for client-specified analytes. Chain-of-custody (COC) completeness is included in Section #3. The laboratory data were reviewed to evaluate compliance with the required methods and the quality of the reported data. A leading check mark (✓) indicates that the referenced data were deemed acceptable. A preceding crossed circle (⊗) signifies problems with the referenced data that may have warranted attaching qualifiers to the data.

- ✓ Data Completeness
- ✓ COC Documentation
- ⊗ Holding Times and Preservation
- ✓ Laboratory Blanks
- ✓ System Monitoring Compounds (i.e. Surrogates)
- ✓ Laboratory Control Samples/Laboratory Control Sample Duplicates (LCS/LCSD)
- ✓ Matrix Spike/Matrix Spike Duplicates (MS/MSD)
- ✓ Laboratory Duplicates
- ✓ Trip Blank

OVERALL DATA PACKAGE ASSESSMENT

Based on a data validation review, the data are acceptable as delivered. Data qualified by the laboratory are discussed in Section #2.

The purpose of validating data and assigning qualifiers is to assist in proper data interpretation. Data which are not qualified meet the site data quality objectives. If values are assigned qualifiers other than an R, the data may be used for site evaluation, with the reasons for qualification being given consideration when interpreting sample concentrations. Data points which are assigned an R qualifier should not be used for any site evaluation purposes. Data were qualified with J data flags by the laboratory if the result was greater than or equal to the method detection limit (MDL) but less than the limit of quantitation (LOQ). Laboratory J flags were preserved in the data and included in the Data Qualification Summary table at the end of this report.

Data qualifiers used during this validation included:

- J – Estimated concentration

Data Completeness

The analyses appeared to be performed as requested on the chain-of-custody records. The associated samples were received by the laboratory and appeared to be analyzed properly. No data points were rejected. The data completeness measure for this data package is 100% and is acceptable.

VALIDATION CRITERIA CHECKLIST	
1. Was the report free of any non-conformances related to the analytical data identified by the laboratory?	Yes
Comments: The laboratory did not note any non-conformances related to the analytical data with the following exception. For methane in sample MW-136D, the laboratory reported that due to interfering peaks on the chromatograms, the value reported represent the lowest reporting limit attainable.	
2. Were data qualification flags or any other notes used by the laboratory? If yes, define.	Yes
Comments: The laboratory noted that the samples were filtered in the field for dissolved metals. The laboratory used the following data qualification flags with this data set. J – Estimated value (1) The result for one or both determinations was less than five times the limit of quantitation (LOQ). (2) The unspiked result was more than four times the spike added.	
3. Were sample COC forms complete?	Yes
Comments: The COC form was complete from the field to the laboratory. Custody was maintained as evidenced by proper signatures, dates, and times of receipt.	
4. Were detection limits in accordance with the QAPP, permit, or method?	Yes
Comments: As indicated by the Tier I data validator, the detection limits were acceptable. Dilutions were not required for the sample analyses.	
5. Were the requested analytical methods in compliance with the QAPP, permit, or COC?	Yes
Comments: The requested analytical methods were in compliance with the COC and the attached analyte list, Analytical Requests for Groundwater.	
6. Were samples received in good condition within method specified requirements?	Yes
Comments: The samples were received in good condition and within the recommended temperature range of 4°C +/- 2°C at 2.9°- 3.5° C. Custody seals were present and intact.	
7. Were samples analyzed within method specified or technical holding times?	No
Comments: The samples were extracted or analyzed within method specified holding times with the following exception. The ferrous iron analysis was performed past the immediate recommended analysis time. The modified Method SM20 3500 Fe B states that holding time is 24 hours but the procedure can also be used in the laboratory if it is understood that normal sample exposure to air during shipment may result in precipitation of iron. As a result, the data were accepted with qualification of J for detections.	
8. Were reported units appropriate for the associated sample matrix/matrices and method(s) of analyses?	Yes
Comments: Sample results were reported in µg/L or mg/L, which are appropriate units for the requested analyses and the water matrix.	
9. Do the laboratory reports include all constituents requested to be reported?	Yes
Comments: The laboratory report included the requested constituents.	
10. Was there indication from the laboratory that the initial or continuing calibration verification results were within acceptable limits?	N/A
Comments: Initial and continuing calibration data were not included as part of this data set; however, these data are assumed to be acceptable as the laboratory did not note that any calibration verification results were outside acceptable limits.	
11. Was the total number of method blank samples prepared equal to at least 5% of the total number of samples, or analyzed as required by the method?	Yes
Comments: The total number of method blanks prepared was greater than 5% of the total number of samples.	

VALIDATION CRITERIA CHECKLIST	
12. Were method blank samples free of analyte contamination?	Yes
Comments: There were no detections of the requested analytes reported in the method blank samples.	
13. Was the total number of matrix spike samples prepared equal to at least 5% of the total number of samples, or analyzed as required by the method?	Yes
Comments: Matrix spike samples were prepared on at least a 5% basis for the total number of samples. Matrix spike samples for Methane batch 093450000A, chloride/sulfate batch 09349196602B, and Alkalinity pH 4.5 batch 09349020201A were prepared from sample MW-136D, 120809. The remaining matrix spikes were prepared from samples not associated with this data set.	
14. Were MS/MSD percent recoveries and MS/MSD RPD values within data validation or laboratory quality control (QC) limits?	Yes
Comments: The project specific MS/MSD recoveries were within laboratory-specified limits or were greater than four times the spiked concentration. In QC batch 09350106101A, the analyte nitrate nitrogen (30%; acceptable range 90-110%) was recovered outside of the laboratory QC limits. The project specific MS and MSD percent recoveries and MS/MSD RPD value were within data validation QC limits. MS and MSD spike recoveries for non-project samples were considered but data were not qualified since matrix similarity to project samples could not be guaranteed.	
15. Was the total number of LCSs analyzed equal to at least 5% of the total number of samples, or analyzed as required by the method?	Yes
Comments: Laboratory control samples were prepared on at least a 5% basis for the total number of samples.	
16. Were LCS/LCSD percent recoveries and LCS/LCSD RPD values within laboratory QC limits?	Yes
Comments: The LCS/LCSD percent recoveries and LCS/LCSD RPD values were within laboratory QC limits.	
17. Were surrogate recoveries within laboratory control limits?	Yes
Comments: Surrogate recoveries were within laboratory control limits.	
18. Was the number of equipment, trip, or field blanks collected equal to at least 10% of the total number of samples, or as required by the project guidelines, QAPP, SAP, or permit, or as indicated by the Tier I validator?	Yes
Comments: There was one trip blank (Trip Blank, 120809) collected with the samples of this data set, which is greater than 10% the total number of samples.	
19. Were the trip blank, field blank, and/or equipment blank samples free of analyte contamination?	Yes
Comments: There were no detections of the requested analytes in the sample Trip Blank, 120809.	
20. Were the field duplicates collected equal to at least 10% of the total number of samples, or as required by the project guidelines, QAPP, SAP, or permit, or as indicated by the Tier I validator?	N/A
Comments: Field duplicates were not collected with the samples of this data set.	
21. Were field duplicate RPD values within data validation QC limits (soil 0-50%, water 0-30%, or air 0-25%)?	N/A
Comments: Field duplicates were not collected with the samples of this data set.	

VALIDATION CRITERIA CHECKLIST

22. Were laboratory duplicate RPD values within laboratory-specified limits?

Yes

Comments: Laboratory duplicates were prepared for inorganic analyses including metals, nitrite-nitrogen, total organic carbon, kjeldahl nitrogen, chloride/sulfate, nitrate-nitrogen, total carbon, ferrous iron, ammonia nitrogen, sulfide, and alkalinity. Laboratory duplicates for chloride/sulfate batch 09349196602B, and Alkalinity pH 4.5 batch 09349020201A were prepared from sample MW-136D, 120809.

The project specific laboratory duplicate RPD value was within the data validation QC limits or were qualified by the laboratory with (1) indicating that the result for one or both determinations was less than five times the LOQ. Laboratory duplicate RPDs for non-project samples were considered but data were not qualified since matrix similarity to project samples could not be guaranteed.



DATA QUALIFICATION SUMMARY

Analyte	Field Sample ID	Lab Sample ID	Result	Reviewer Qualifier	Reviewer Qualifier Reason
CO2 by Headspace	MW-136D,120809	5858877	8,700 µg/L	J	Flagged by the Lab: Result between MDL and RL.
Total Iron	MW-136D,120809	5858877	0.0545 mg/L	J	Flagged by the Lab: Result between MDL and RL.
Ferrous Iron	MW-136D,120809	5858877	0.12 mg/L	J	Sample Extracted Outside of Acceptable Holding Time