

ANALYTICAL RESULTS

Prepared for:

Chevron
5000 State Route 128
HOOVEN OH 45033

Prepared by:

Lancaster Laboratories
2425 New Holland Pike
Lancaster, PA 17605-2425

January 07, 2010

Project: Hooven Cincinnati Final Remedy

Samples arrived at the laboratory on Tuesday, December 29, 2009. The PO# for this group is 0015039270 and the release number is 50008931. The group number for this submittal is 1176742.

<u>Client Sample Description</u>	<u>Lancaster Labs (LLI) #</u>
MW-95S,122809 Grab Water	5873647
MW-95S,122809 Filtered Grab Water	5873648
MW-113,122809 Grab Water	5873649
MW-113,122809 Filtered Grab Water	5873650
Trip Blank,122809 Water	5873651

The specific methodologies used in obtaining the enclosed analytical results are indicated on the Laboratory Sample Analysis Record.

ELECTRONIC COPY TO	Trihydro Corporation	Attn: Trihydro Database
ELECTRONIC COPY TO	Trihydro Corporation	Attn: Tim Gunn
ELECTRONIC COPY TO	Trihydro Corporation	Attn: Matthew Mitchell

Questions? Contact your Client Services Representative
Katherine A Klinefelter at (717) 656-2300

Respectfully Submitted,



Max E. Snavelly
Senior Specialist



Analysis Report

2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

Sample Description: MW-95S,122809 Grab Water
2nd Semi-Annual 2009-Cincinnati Final Remedy

LLI Sample # WW 5873647
LLI Group # 1176742
OH

Project Name: Hooven Cincinnati Final Remedy

Collected: 12/28/2009 13:05 by DB

Account Number: 11494

Submitted: 12/29/2009 09:10

Chevron

Reported: 01/07/2010 at 07:20

5000 State Route 128

Discard: 03/09/2010

HOOVEN OH 45033

SE95S SDG#: HVQ38-01

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Method Detection Limit	Dilution Factor
GC/MS	Volatiles	SW-846 8260B	ug/l	ug/l	
10903	Benzene	71-43-2	N.D.	0.5	1
10903	Chlorobenzene	108-90-7	N.D.	0.8	1
10903	Ethylbenzene	100-41-4	N.D.	0.8	1
10903	Toluene	108-88-3	N.D.	0.7	1
10903	Xylene (Total)	1330-20-7	N.D.	0.8	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01163	GC/MS VOA Water Prep	SW-846 5030B	1	W093641AA	12/30/2009 21:46	Sara E Johnson	1
10903	8260 BTEX + Chlorobenzene	SW-846 8260B	1	W093641AA	12/30/2009 21:46	Sara E Johnson	1



Analysis Report

2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

**Sample Description: MW-95S,122809 Filtered Grab Water
2nd Semi-Annual 2009-Cincinnati Final Remedy**

**LLI Sample # WW 5873648
LLI Group # 1176742
OH**

Project Name: Hooven Cincinnati Final Remedy

Collected: 12/28/2009 13:05 by DB

Account Number: 11494

Submitted: 12/29/2009 09:10

Chevron

Reported: 01/07/2010 at 07:20

5000 State Route 128

Discard: 03/09/2010

HOOVEN OH 45033

F-95S SDG#: HVQ38-02

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Method Detection Limit	Dilution Factor
Metals Dissolved			mg/l	mg/l	
07035	Arsenic	7440-38-2	N.D.	0.0072	1
07055	Lead	7439-92-1	N.D.	0.0069	1

General Sample Comments

This sample was field filtered for metals.

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
07035	Arsenic	SW-846 6010B	1	093651848009	01/05/2010 11:25	Eric L Eby	1
07055	Lead	SW-846 6010B	1	093651848009	01/05/2010 11:25	Eric L Eby	1
01848	WW SW846 ICP Digest (tot rec)	SW-846 3005A	1	093651848009	01/04/2010 10:12	Denise K Connors	1



Analysis Report

2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

Sample Description: MW-113,122809 Grab Water
2nd Semi-Annual 2009-Cincinnati Final Remedy

LLI Sample # WW 5873649
LLI Group # 1176742
OH

Project Name: Hooven Cincinnati Final Remedy

Collected: 12/28/2009 13:25 by DB Account Number: 11494

Submitted: 12/29/2009 09:10 Chevron
Reported: 01/07/2010 at 07:20 5000 State Route 128
Discard: 03/09/2010 HOOVEN OH 45033

SE113 SDG#: HVQ38-03

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Method Detection Limit	Dilution Factor
GC/MS	Volatiles	SW-846 8260B	ug/l	ug/l	
10903	Benzene	71-43-2	N.D.	0.5	1
10903	Chlorobenzene	108-90-7	N.D.	0.8	1
10903	Ethylbenzene	100-41-4	N.D.	0.8	1
10903	Toluene	108-88-3	N.D.	0.7	1
10903	Xylene (Total)	1330-20-7	N.D.	0.8	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01163	GC/MS VOA Water Prep	SW-846 5030B	1	W093641AA	12/30/2009 22:09	Sara E Johnson	1
10903	8260 BTEX + Chlorobenzene	SW-846 8260B	1	W093641AA	12/30/2009 22:09	Sara E Johnson	1



Analysis Report

2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

**Sample Description: MW-113,122809 Filtered Grab Water
2nd Semi-Annual 2009-Cincinnati Final Remedy**

**LLI Sample # WW 5873650
LLI Group # 1176742
OH**

Project Name: Hooven Cincinnati Final Remedy

Collected: 12/28/2009 13:25 by DB

Account Number: 11494

Submitted: 12/29/2009 09:10

Chevron

Reported: 01/07/2010 at 07:20

5000 State Route 128

Discard: 03/09/2010

HOOVEN OH 45033

F-113 SDG#: HVQ38-04

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Method Detection Limit	Dilution Factor
Metals Dissolved			mg/l	mg/l	
07035	Arsenic	7440-38-2	N.D.	0.0072	1
07055	Lead	7439-92-1	N.D.	0.0069	1

General Sample Comments

This sample was field filtered for metals.

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
07035	Arsenic	SW-846 6010B	1	093651848009	01/05/2010 11:28	Eric L Eby	1
07055	Lead	SW-846 6010B	1	093651848009	01/05/2010 11:28	Eric L Eby	1
01848	WW SW846 ICP Digest (tot rec)	SW-846 3005A	1	093651848009	01/04/2010 10:12	Denise K Connors	1



Analysis Report

2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

Sample Description: Trip Blank, 122809 Water
2nd Semi-Annual 2009-Cincinnati Final Remedy

LLI Sample # WW 5873651
LLI Group # 1176742
OH

Project Name: Hooven Cincinnati Final Remedy

Collected: 12/28/2009

Account Number: 11494

Submitted: 12/29/2009 09:10

Chevron

Reported: 01/07/2010 at 07:20

5000 State Route 128

Discard: 03/09/2010

HOOVEN OH 45033

TBSEM SDG#: HVQ38-05TB*

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Method Detection Limit	Dilution Factor
GC/MS	Volatiles	SW-846 8260B	ug/l	ug/l	
10903	Benzene	71-43-2	N.D.	0.5	1
10903	Chlorobenzene	108-90-7	N.D.	0.8	1
10903	Ethylbenzene	100-41-4	N.D.	0.8	1
10903	Toluene	108-88-3	N.D.	0.7	1
10903	Xylene (Total)	1330-20-7	N.D.	0.8	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01163	GC/MS VOA Water Prep	SW-846 5030B	1	W093641AA	12/30/2009 21:23	Sara E Johnson	1
10903	8260 BTEX + Chlorobenzene	SW-846 8260B	1	W093641AA	12/30/2009 21:23	Sara E Johnson	1

Quality Control Summary

 Client Name: Chevron
 Reported: 01/07/10 at 07:20 AM

Group Number: 1176742

Matrix QC may not be reported if site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD was performed, unless otherwise specified in the method.

Laboratory Compliance Quality Control

<u>Analysis Name</u>	<u>Blank Result</u>	<u>Blank MDL</u>	<u>Report Units</u>	<u>LCS %REC</u>	<u>LCSD %REC</u>	<u>LCS/LCSD Limits</u>	<u>RPD</u>	<u>RPD Max</u>
Batch number: W093641AA	Sample number(s): 5873647,5873649,5873651							
Benzene	N.D.	0.5	ug/l	104	101	79-120	2	30
Chlorobenzene	N.D.	0.8	ug/l	102	101	80-120	2	30
Ethylbenzene	N.D.	0.8	ug/l	101	101	79-120	0	30
Toluene	N.D.	0.7	ug/l	101	102	79-120	1	30
Xylene (Total)	N.D.	0.8	ug/l	101	101	80-120	0	30
Batch number: 093651848009	Sample number(s): 5873648,5873650							
Arsenic	N.D.	0.0072	mg/l	105		89-115		
Lead	N.D.	0.0069	mg/l	104		80-120		

Sample Matrix Quality Control

Unspiked (UNSPK) = the sample used in conjunction with the matrix spike
 Background (BKG) = the sample used in conjunction with the duplicate

<u>Analysis Name</u>	<u>MS %REC</u>	<u>MSD %REC</u>	<u>MS/MSD Limits</u>	<u>RPD</u>	<u>RPD MAX</u>	<u>BKG Conc</u>	<u>DUP Conc</u>	<u>DUP RPD</u>	<u>Dup RPD Max</u>
Batch number: W093641AA	Sample number(s): 5873647,5873649,5873651 UNSPK: 5873647								
Benzene	109		80-126						
Chlorobenzene	104		87-124						
Ethylbenzene	104		71-134						
Toluene	105		80-125						
Xylene (Total)	105		79-125						
Batch number: 093651848009	Sample number(s): 5873648,5873650 UNSPK: P872412 BKG: P872412								
Arsenic	106	105	75-125	1	20	N.D.	N.D.	0 (1)	20
Lead	99	100	75-125	1	20	N.D.	N.D.	0 (1)	20

Surrogate Quality Control

Surrogate recoveries which are outside of the QC window are confirmed unless attributed to dilution or otherwise noted on the Analysis Report.

 Analysis Name: 8260 Std. Water Master
 Batch number: W093641AA

	Dibromofluoromethane	1,2-Dichloroethane-d4	Toluene-d8	4-Bromofluorobenzene
5873647	90	89	91	84
5873649	92	86	91	84
5873651	92	89	90	84

*- Outside of specification

- (1) The result for one or both determinations was less than five times the LOQ.
- (2) The unspiked result was more than four times the spike added.

Quality Control Summary

Client Name: Chevron
Reported: 01/07/10 at 07:20 AM

Group Number: 1176742

Surrogate Quality Control

Blank	91	88	90	85
LCS	92	89	91	87
LCSD	92	88	91	88
MS	94	84	90	87
Limits:	80-116	77-113	80-113	78-113

*- Outside of specification

- (1) The result for one or both determinations was less than five times the LOQ.
- (2) The unspiked result was more than four times the spike added.

Analysis Request/ Environmental Services Chain of Custody



For Lancaster Laboratories use only

Acct. # 11494 Group# 1176742 Sample # 5873647-51 **COC #** 225013

Please print. Instructions on reverse side correspond with circled numbers.

1 Client: <u>Chevron</u> Acct. #: <u>11494</u> Project Name/#: <u>2nd Semi-Annual 2009</u> PWSID #: <u>MWREN17000M</u> Project Manager: <u>Doug Lam</u> P.O.#: _____ Sampler: <u>Dale Barrett</u> Quote #: _____ Name of state where samples were collected: <u>Ohio</u>				4		5 Analyses Requested					For Lab Use Only								
				Matrix		Preservation Codes					FSC: _____ SCR#: _____								
2 Sample Identification				3 Date Collected Time Collected		Origin Composite		Soil Water Other		Total # of Containers		H N VOCs PAH 8260 Pb/As D-SOLVED METALS TPH GRO TPX BRO MNA					Preservation Codes H=HCl T=Thiosulfate N=HNO ₃ B=NaOH S=H ₂ SO ₄ O=Other		6 Temperature of samples upon receipt (if requested)
												Remarks							
MW-95, 122809 MW-113, 122809 TRIP Blank, 122809				12/28/09 1305 12/28/09 1325 12/28/09 —		X X X		X X X		4 4 2		See Attached Analytic List Dissolved metals are field filtered QC Summary Data Package					(Diagonal line through this section)		

7 Turnaround Time Requested (TAT) (please circle): Normal Rush
 (Rush TAT is subject to Lancaster Laboratories approval and surcharge.)
 Date results are needed: _____
 Rush results requested by (please circle): Phone Fax E-mail
 Phone #: 513-353-1323 Fax #: 513-353-4664
 E-mail address: mm.tckill@trihydro.com

8 Data Package Options (please circle if required)

Type I (validation/NJ Reg)	TX TRRP-13	SDG Complete? <u>Yes</u> No
Type II (Tier II)	MA MCP CT RCP	
Type III (Reduced NJ)	Site-specific QC (MS/MSD/Dup)? Yes No	
Type IV (CLP SOW)	(If yes, indicate QC sample and submit triplicate volume.)	
Type VI (Raw Data Only)	Internal COC Required? Yes / No	

Relinquished by: <u>[Signature]</u>	Date: <u>12/28/09</u>	Time: <u>1400</u>	Received by: _____	Date: _____	Time: _____
Relinquished by: _____	Date: _____	Time: _____	Received by: _____	Date: _____	Time: _____
Relinquished by: _____	Date: _____	Time: _____	Received by: _____	Date: _____	Time: _____
Relinquished by: _____	Date: _____	Time: _____	Received by: _____	Date: _____	Time: _____
Relinquished by: _____	Date: _____	Time: _____	Received by: <u>[Signature]</u>	Date: <u>12/28/09</u>	Time: <u>910</u>

Analytical Requests for Groundwater
Chevron Cincinnati Facility, Hooven, Ohio

Volatile Organics

Benzene
Chlorobenzene
Ethylbenzene
Toluene
Xylenes (total)

Dissolved Metals- field filtered

Arsenic
Lead

TPH

GRO
DRO

Monitored Natural Attenuation

Alkalinity
Calcium
Chemical Oxygen Demand
Chloride
Iron (II) and Iron (III)
Dissolved and Total Manganese
Methane
Nitrate Nitrogen
Nitrite Nitrogen
Ammonia Nitrogen
Total Kjeldahl Nitrogen
Potassium
Sodium
Sulfate
Sulfide
Total Organic Carbon



Environmental Sample Administration Receipt Documentation Log

Client/Project: Cheriton
 Date of Receipt: 12/29/09
 Time of Receipt: 9:10
 Source Code: 501
 Unpacker Emp. No.: 2316

Shipping Container Sealed: YES NO
 Custody Seal Present * : YES NO
 * Custody seal was intact unless otherwise noted in the discrepancy section
 Package: Chilled Not Chilled

Temperature of Shipping Containers							
Cooler #	Thermometer ID	Temperature (°C)	Temp Bottle (TB) or Surface Temp (ST)	Wet Ice (WI) or Dry Ice (DI) or Ice Packs (IP)	Ice Present? Y/N	Loose (L) Bagged Ice (B) or NA	Comments
1	D121465	0.82	TB	WI	Y	L	
2							
3							
4							
5							
6							

Number of Trip Blanks received NOT listed on chain of custody: 0

Paperwork Discrepancy/Unpacking Problems:

Sample Administration Internal Chain of Custody			
Name	Date	Time	Reason for Transfer
<i>Mary Beth Reed</i>	12/29/09	1038	Unpacking / Storage
<i>Mary Beth Reed</i>	12/29/09	1041	Place in Storage or <input checked="" type="radio"/> Entry
			Entry
			Entry

Lancaster Laboratories Explanation of Symbols and Abbreviations

The following defines common symbols and abbreviations used in reporting technical data:

N.D.	none detected	BMQL	Below Minimum Quantitation Level
TNTC	Too Numerous To Count	MPN	Most Probable Number
IU	International Units	CP Units	cobalt-chloroplatinate units
umhos/cm	micromhos/cm	NTU	nephelometric turbidity units
C	degrees Celsius	F	degrees Fahrenheit
Cal	(diet) calories	lb.	pound(s)
meq	milliequivalents	kg	kilogram(s)
g	gram(s)	mg	milligram(s)
ug	microgram(s)	l	liter(s)
ml	milliliter(s)	ul	microliter(s)
m3	cubic meter(s)	fib >5 um/ml	fibers greater than 5 microns in length per ml
<	less than – The number following the sign is the <u>limit of quantitation</u> , the smallest amount of analyte which can be reliably determined using this specific test.		
>	greater than		
ppm	parts per million – One ppm is equivalent to one milligram per kilogram (mg/kg), or one gram per million grams. For aqueous liquids, ppm is usually taken to be equivalent to milligrams per liter (mg/l), because one liter of water has a weight very close to a kilogram. For gases or vapors, one ppm is equivalent to one microliter of gas per liter of gas.		
ppb	parts per billion		
Dry weight basis	Results printed under this heading have been adjusted for moisture content. This increases the analyte weight concentration to approximate the value present in a similar sample without moisture.		

U.S. EPA data qualifiers:

Organic Qualifiers

A	TIC is a possible aldol-condensation product
B	Analyte was also detected in the blank
C	Pesticide result confirmed by GC/MS
D	Compound quantitated on a diluted sample
E	Concentration exceeds the calibration range of the instrument
J	Estimated value
N	Presumptive evidence of a compound (TICs only)
P	Concentration difference between primary and confirmation columns >25%
U	Compound was not detected
X,Y,Z	Defined in case narrative

Inorganic Qualifiers

B	Value is <CRDL, but ≥IDL
E	Estimated due to interference
M	Duplicate injection precision not met
N	Spike amount not within control limits
S	Method of standard additions (MSA) used for calculation
U	Compound was not detected
W	Post digestion spike out of control limits
*	Duplicate analysis not within control limits
+	Correlation coefficient for MSA <0.995

Analytical test results for methods listed on the laboratories' accreditation scope meet all requirements of NELAC unless otherwise noted under the individual analysis.

Tests results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY – In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. THE FOREGOING EXPRESS WARRANTY IS EXCLUSIVE AND IS GIVEN IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED. WE DISCLAIM ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING A WARRANTY OF FITNESS FOR PARTICULAR PURPOSE AND WARRANTY OF MERCHANTABILITY. IN NO EVENT SHALL LANCASTER LABORATORIES BE LIABLE FOR INDIRECT, SPECIAL, CONSEQUENTIAL, OR INCIDENTAL DAMAGES INCLUDING, BUT NOT LIMITED TO, DAMAGES FOR LOSS OF PROFIT OR GOODWILL REGARDLESS OF (A) THE NEGLIGENCE (EITHER SOLE OR CONCURRENT) OF LANCASTER LABORATORIES AND (B) WHETHER LANCASTER LABORATORIES HAS BEEN INFORMED OF THE POSSIBILITY OF SUCH DAMAGES. We accept no legal responsibility for the purposes for which the client uses the test results. No purchase order or other order for work shall be accepted by Lancaster Laboratories which includes any conditions that vary from the Standard Terms and Conditions of Lancaster Laboratories and we hereby object to any conflicting terms contained in any acceptance or order submitted by client.