



# Quality Assurance Site Survey Report for Closet World (Quemetco)

Last updated May, 2015



AQS ID	ARB Number	Site Start Date	Reporting Agency and Agency Code
060371404	70043	10/03/2008	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
500 S. 7th Ave. City of Industry, CA 91746	Los Angeles	South Coast	34° 01' 34"N	117° 58' 54"W	89 m



## Detailed Site Information

Local site name	Closet World (Quemetco)			
AQS ID	060371404			
GPS coordinates (decimal degrees)	Latitude: 34° 01' 34" Longitude: 117° 58' 54"			
Street Address	720 S 7th Ave. City of Industry, CA 91746			
County	Los Angeles			
Distance to roadways (meters)	30			
Traffic count (AADT, year)	20,000 / 2012			
Groundcover (e.g. asphalt, dirt, sand)	Asphalt			
Representative statistical area name (i.e. MSA, CBSA, other)	31080-Los Angeles-Long Beach-Anaheim, MSA			
Pollutant, POC	Lead, 1			
Parameter code	14129			
Basic monitoring objective(s)	NAAQS			
Site type(s)	Source Oriented			
Monitor (type)	SLAMS			
Instrument manufacturer and model	GMW 1200 TSP			
Method code	110			
FRM/FEM/ARM/ other	FRM			
Collecting Agency	SCAQMD			
Analytical Lab (i.e. weigh lab, toxics lab, other)	SCAQMD			
Reporting Agency	SCAQMD			
Spatial scale (e.g. micro, neighborhood)	Micro			
Monitoring start date (MM/DD/YYYY)	10/03/2008			
Current sampling frequency (e.g. 1:3, continuous)	1:6			
Calculated sampling frequency (e.g. 1:3/1:1)	1:6			
Sampling season (MM/DD-MM/DD)	01/01-12/31			
Probe height (meters)	2.6			
Distance from supporting structure (meters)	1			
Distance from obstructions on roof (meters)	N/A			
Distance from obstructions not on roof (meters)	N/A			

Distance from trees (meters)	N/A			
Distance to furnace or incinerator flue (meters)	N/A			
Distance between collocated monitors (meters)	N/A			
Unrestricted airflow (degrees)	360°			
Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A			
Residence time for reactive gases (seconds)	N/A			
Will there be changes within the next 18 months? (Y/N)	No			
Is it suitable for comparison against the annual PM2.5? (Y/N)	N/A			
Frequency of flow rate verification for manual PM samplers	Monthly			
Frequency of flow rate verification for automated PM analyzers	N/A			
Frequency of one-point QC check for gaseous instruments	N/A			
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A			
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	05/28/2014, 12/05/2014			

**Quemetco – Closet World  
Site Photos**



**Looking North from the probe**



**Looking East from the probe.**



**Looking South toward the probe.**



**Looking West from the probe**