


# Quality Assurance Site Survey Report for SA Recycling

*Last updated May, 2015*



AQS ID	ARB Number	Site Start Date	Reporting Agency and Agency Code
Unavailable	Unavailable	6/2012	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
8822 Etiwanda Ave. , Rancho Cucamonga, CA, 91739	San Bernardino	South Coast	34° 05' 35"N	117° 31' 41"W	351 m



## Detailed Site Information

Local site name	SA Recycling		
AQS ID	Unavailable		
GPS coordinates (decimal degrees)	Latitude: 34° 05' 35"N Longitude: 117° 31' 41"W		
Street Address	8822 Etiwanda Ave. , Rancho Cucamonga,CA,91739		
County	San Bernardino		
Distance to roadways (meters)	400 m		
Traffic count (AADT, year)	Unavailable		
Groundcover (e.g. asphalt, dirt, sand)	Asphalt		
Representative statistical area name (i.e. MSA, CBSA, other)	40140-Riverside-San Bernardino-Ontario, CA MSA		
Pollutant, POC	Lead, 1	Metals, CR6, 1	
Parameter code	14129	See Table 26	
Basic monitoring objective(s)	NAAQS	NAAQS	
Site type(s)	Source Oriented	Highest Concentration	
Monitor (type)	SLAMS	NATTS	
Instrument manufacturer and model	GMW 1200 TSP	RM Env. 924, A Sampler	
Method code	110	See Table 26	
FRM/FEM/ARM/ other	FRM	Other	
Collecting Agency	SCAQMD	SCAQMD	
Analytical Lab (i.e. weigh lab, toxics lab, other)	SCAQMD	SCAQMD	
Reporting Agency	SCAQMD	SCAQMD	
Spatial scale (e.g. micro, neighborhood)	Micro	Neighborhood	
Monitoring start date (MM/DD/YYYY)	6/26/12	7/19/12	
Current sampling frequency (e.g. 1:3, continuous)	1:6	1:3	
Calculated sampling frequency (e.g. 1:3/1:1)	1:6	No CFR mandated sampling schedule.	
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	
Probe height (meters)	2.6	3	
Distance from supporting structure (meters)	1	1.6	
Distance from obstructions on roof (meters)	N/A	N/A	
Distance from obstructions not on roof (meters)	N/A	N/A	

Distance from trees (meters)	N/A	N/A		
Distance to furnace or incinerator flue (meters)	N/A	N/A		
Distance between collocated monitors (meters)	N/A	N/A		
Unrestricted airflow (degrees)	360°	360°		
Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A		
Residence time for reactive gases (seconds)	N/A	N/A		
Will there be changes within the next 18 months? (Y/N)	No	No		
Is it suitable for comparison against the annual PM2.5? (Y/N)	N/A	N/A		
Frequency of flow rate verification for manual PM samplers	Monthly	Monthly		
Frequency of flow rate verification for automated PM analyzers	N/A	N/A		
Frequency of one-point QC check for gaseous instruments	N/A	N/A		
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A		
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	unavailable	N/A		