

### Measurement Quality Objectives - Parameter NO<sub>2</sub> (Chemiluminescence)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Standard Reporting Units</b>	All data	ppm	40 CFR, Pt 50.11	
<b>Shelter Temperature</b> Temperature range Temperature control	Daily Daily	20 to 30° C # ± 2° C	40 CFR, Pt. 53.20 Vol II, S 7.1 <sup>1</sup> / <sub>2</sub> Vol II, MS 2.3.2	Instruments designated as reference or equivalent have been tested over this temperature range. Maintain shelter temperature above sample dewpoint. Shelter should have a 24- hour temperature recorder. Flag all data for which temperature range or fluctuations are outside acceptance criteria.
<b>Equipment</b> NO <sub>2</sub> analyzer Air flow controllers Flowmeters	Purchase specification	Reference or equivalent method Flow rate regulated to ± 2 % Accuracy ± 2 %	40 CFR, Pt 53.9 40 CFR, Pt 50, App F, S 2.2 EPA-600/4-75-003	
<b>Detection</b> Noise Lower detectable level	Purchase specification	0.005 ppm 0.01 ppm	40 CFR, Pt 53.20 & 23 “	Instruments designated as reference or equivalent have been determined to meet these acceptance criteria
<b>Completeness</b> Hourly Data	Quarterly	75 %	40 CFR, Pt 50.11	
<b>Compressed Gases</b> Dilution gas (zero air)  Gaseous standards	Purchase specification  Purchase specification	Free of contaminants  NIST Traceable (e.g., EPA Protocol Gas)	EPA-600/4-75-003  40 CFR, Pt 50, App F, S 1.3 EPA-600/R-97/121	Return cylinder to supplier.  Nitric oxide in nitrogen EPA Protocol Gases have a 24-month certification period and must be recertified to extend the certification.
<b>Calibration</b> Multipoint calibration (at least 5 points)  Convertor efficiency  Zero/span check- level 1    Flowmeters	≥ 1/6 months., after failure of QC check or after maintenance  During multipoint calibrations 1/ 2 weeks    1/3 months	Residence time ≤ 2 min Dynam. parameter ≥ 2.75 ppm-min All points within ± 2 % of full scale of best-fit straight line  \$ 96 % Zero drift # ± 20 to 30 ppb Span drift # ± 20 to 25 %  Zero drift # ± 10 to 15 ppb Span drift # ± 15 %  Accuracy ± 2 %	40 CFR, Pt 50, App F, S 1 Vol II, S 12.6 Vol II, MS 2.3.2  40 CFR, Pt. 50, App F Vol II, MS.2.3.2 Vol II, S 12.6 Vol II, MS 2.3.2  Vol II, S 12.6 Vol II, MS 2.3.2  Vol II, App 12	Zero gas and at least four upscale calibration points. Points outside acceptance criterion are repeated. If still outside consult manufacturers manual and invalidate data to last acceptable multipoint calibration or zero/span check .  Replace or service converter. <b>If calibration factors are updated after each zero/span,</b> invalidate data to last acceptable zero/span check, adjust analyzer, and perform multipoint calibration. <b>If fixed calibration factors are used to calculate data,</b> invalidate data to last acceptable zero/span check, adjust analyzer, and perform multipoint calibration. Flowmeter calibration should be traceable to NIST standards.

**Measurement Quality Objectives - Parameter NO<sub>2</sub> (Chemiluminescence)**

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Performance Evaluation</b> (NPAP)  State audits	1/year at selected sites 1/year	Mean absolute difference # 15 %  State requirements	NPAP QAPP  Vol II, App 15, S 3	Use information to inform reporting agency for corrective action and technical systems audits.
<b>Precision</b> Single analyzer Reporting organization	1/ 2 weeks 1/3 months	None 95 % Confidence Interval # ± 15 %	40 CFR, Pt 58, App A EPA-600/4-83-023 Vol II, App 15, S 6	Concentration. = 0.08-0.10 ppm.
<b>Accuracy</b> Single analyzer Reporting organization	25 % of sites quarterly (all sites yearly)	None 95% Confidence Interval # ± 20%	40 CFR, Pt 58, App A EPA-600/4-83-023 Vol II, App 15, S 3	Four concentration ranges. If failure, recalibrate analyzer and reanalyze samples. Repeated failure requires corrective action.

<sup>1/</sup> - reference refers to the QA Handbook for Air Pollution Measurement Systems, Volume II . The use of “S” refers to sections within Part 1 of Volume II. The use of “MS” refers to meth specific sections in Volume II.

### Measurement Quality Objectives - Parameter O<sub>3</sub> (Ultraviolet Photometric)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Standard Reporting Units</b>	All data	ppm	40 CFR, Pt 50.9	
<b>Shelter Temperature</b> Temperature range Temperature control	Daily Daily	20 to 30° C. # ± 2° C	40 CFR, Pt. 53.20 Vol II, S 7.1 <sup>U</sup> <i>Determination of Ozone by Ultraviolet Analysis (draft)</i>	Instruments designated as reference or equivalent have been tested over this temperature range. Maintain shelter temperature above sample dewpoint. Shelter should have a 24- hour temperature recorder. Flag all data for which temperature range or fluctuations are outside acceptance criteria.
<b>Equipment</b> O <sub>3</sub> analyzer	Purchase specification	Reference or equivalent method	40 CFR, Pt 53.9 EPA-600/4-79-057	Air flow controllers must be capable of regulating air flows as necessary to meet the output stability and photometer precision requirements. The photometric measurement of absorption is not directly related to flow rate, but may be indirectly related due to thermal or other effects.
<b>Detection</b> Noise Lower detectable level	Purchase specification	0.005 ppm 0.01 ppm	40 CFR, Pt. 53.20 & 23 “	Instruments designated as reference or equivalent have been determined to meet these acceptance criteria.
<b>Completeness</b> (seasonal) Maximum 1-hour concentration	Daily	75% values from 9:01 AM to 9:00 PM (LST)	40 CFR, Pt 50, App H, S 3	A missing daily maximum ozone value may be assumed to be less than the standard if valid daily maxima on the preceding and following days do not exceed 75 percent of the standard.
<b>Transfer standard</b> Qualification and certification Recertification to local primary standard	Upon receipt of transfer standard 1/3 months (if at a fixed site)	±4% or ±4 ppb (whichever greater) RSD of six slopes # 3.7% Std. dev. of six intercepts #1.5% New slope = ±0.05 of previous	EPA-600/4-79-056 EPA-600/4-79-057 “ “	6 comparison runs that include, at minimum, 6 concentrations per comparison run including 0 and 90 ± 5% of upper range. A single six-point comparison run.
<b>Local primary standard</b> Certification/recertification to Standard Photometer (if recertified via a transfer standard)	1/year "	Difference # ±5 % (preferably ± 3%) Regression slopes = 1.00 ± 0.03 and two intercepts are 0 ± 3 ppb	<i>Determination of Ozone by Ultraviolet Analysis (draft)</i> "	The local primary standard is a standard in its own right, but it must be repaired and recertified if the acceptance criterion is exceeded.
<b>EPA Standard Reference Photometer recertification</b>	1/year	Regression slope = 1.00 ± 0.01 and intercept < 3 ppb	Protocol for Recertification of Standard Reference Photometers... (TRC Environmental Document)	9 replicate analysis over 12 conc. ranges. Disagreement must be resolved. EPA Standard Reference Photometer rechecked with NIST. If OK Network STANDARD REFERENCE PHOTOMETER must be repaired.
<b>Zero air</b>	Purchase specification	Free of O <sub>3</sub> or any substance that might react with O <sub>3</sub> (e.g., NO, NO <sub>2</sub> , hydrocarbons, and particulates)	EPA-600/4-79-057	Return cylinder to supplier

### Measurement Quality Objectives - Parameter O<sub>3</sub> (Ultraviolet Photometric)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Ozone analyzer calibration</b> Zero/span check -level 1  Multipoint calibration (at least 5 points)	1/ 2 weeks  Upon receipt, adjustment, or 1/ 6 months	Zero drift # ± 20 to 30 ppb Span drift # ± 20 to 25 %  Zero drift # ± 10 to 15 ppb Span drift # ± 15%  Linearity error <5%	Vol II, S 12.6 “  Vol II, S 12.6 “  40 CFR, Pt 50, App D, S 5.2.3 EPA-600/4-79-057 S.5 Vol II, S 12.2	<b>If calibration updated at each zero/span</b> , Invalidate data to last acceptable check, adjust analyzer, perform multipoint calibration. <b>If fixed calibration used to calculate data</b> , Invalidate data to last acceptable check, adjust analyzer, perform multipoint calibration. Zero gas and at least four upscale calibration points. Check verify accuracy of flow dilution. Redo analysis. If failure persists corrective action required.
<b>Performance Evaluation (NPAP)</b>  State audits	1/year at selected sites 1/year	Mean absolute difference # 15%  State requirements	Vol II, S 16.3  Vol II, App 15, S 3	Use information to inform reporting agency for corrective action and technical systems audits.
<b>Precision</b> Single analyzer Reporting organization	1/ 2 weeks 1/3 months	None 95% CI < ± 15%	40 CFR, Pt 58, App A EPA-600/4-83-023 Vol II, App 15, S 6	Concentration = 0.08-0.10 ppm.
<b>Accuracy</b> Single analyzer Annual accuracy	25 % of sites quarterly (all sites yearly)	None 95% CI # ± 20%	40 CFR, Pt 58, App A EPA-600/4-83-023 Vol II, App 15, S 6	Four concentration ranges. If failure, recalibrate and reanalyze. Repeated failure requires corrective action.

<sup>1/</sup> - reference refers to the QA Handbook for Air Pollution Measurement Systems, Volume II. The use of “S” refers to sections within Part 1 of Volume II. The use of “MS” refers to method specific sections in Volume II.

### Measurement Quality Objectives - Parameter Lead (Atomic Absorption Spectroscopy)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Reporting Units</b>	All data	: g/m <sup>3</sup>	40 CFR, Pt 50.12	
<b>Filter Checks</b>				
Visual defect check	All filters	See reference	Vol II, MS 2.2.4	Discard any defective filters
Filter Integrity				
Collection efficiency	Purchase specification	99%	40 CFR, Pt 50, App B, S 7.1	Measure using DOP test (ASTM-2988). Reject shipment
Integrity		2.4 mg max weight loss	"	
pH		6 to 10	"	
<b>Equipment</b>				
Sampler	Purchase specification	Reference or equivalent method	40 CFR, Pt 53.9	
Flow rate transfer standard	Purchase specification	0.02 std. m <sup>3</sup> /min	40 CFR, Pt 50, App B, S 7 "	
<b>Detection Limit</b>				
LDL	Not applicable	0.07 µg/m <sup>3</sup>	40 CFR, Pt 50, App G, S 2	This value is based on a collaborative test of the method. Assumed air volume of 2,400 m <sup>3</sup> .
<b>Completeness</b>	Quarterly	75%		
<b>Sampler calibration</b>				
Orifice calibration unit (flow rate transfer standard)	On receipt and yearly	Indicated flow rate within ±2% of actual flow rate	Vol II, MS 2.8.1 Vol II, MS 2.2.2	Adopt a new calibration curve. A rotary-type, gas displacement meter is the recommended NIST-traceable reference standard. Adjust or replace meter
Elapsed time meter	On receipt and 1/6 months	± 2 min/24 hours	"	
On/Off Timer	On receipt and 1/3 months	± 30 min/24 hour	Vol II, MS 2.2.2	Checked against elapsed time meter. Adjust or repair.
Sampler flow rate	On receipt, if audit deviation > 7 %, after maintenance	All points within ± 5 % of full scale of best-fit straight line	"	Rerun points outside limits until acceptable.
<b>Analytical calibration</b>				
Reproducibility test	On receipt	# 5%	Vol II, MS 2.8.1	Reproducibility = 100 ([high response-low response]/average response). Responses should be corrected for the blank level. If acceptance criterion is exceeded, instrument should be checked by a service rep or qualified operator.
Calibration stability	Before first sample, after every tenth sample, after last sample	# ± 5 % deviation from calibration curve.	Vol II, MS 2.8.5	Alternate between two control standards with concentrations # : g/mL or 1 to # 10 : g/mL. Take corrective action and repeat the previous ten analyses.

### Measurement Quality Objectives - Parameter Lead (Atomic Absorption Spectroscopy)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Performance Evaluation</b> (NPAP) Sampler performance Audit (flow rate)	1/year at selected sites 1/3 months	Mean absolute difference# 15% Percentage difference # $\pm 7\%$	Vol II, S 16.3 40 CFR, Pt 58, App A Vol II, MS 2.2.8	Use information to inform reporting agency for corrective action and technical systems audits Recalibrate before any additional sampling
<b>Precision</b> Single analyzer Reporting organization	1/6 days 1/ 3 months	None 95% CI $< \pm 15\%$	40 CFR, Pt 58, App A, S 5.3 40 CFR, Pt 58, App A, S 5.3	Both lead values must be $> 0.15 \text{ : g/m}^3$
<b>Accuracy</b> Single analyzer Reporting organization	25 % of sites quarterly	Percentage difference # $\pm 16\%$ 95% CI # $\pm 20\%$	Vol II, MS 2.8.8 40 CFR, Pt 58, App A, S 3.4 EPA-600/4-83-023	Analyze three audit samples in each of the two concentration ranges. The audit samples shall be distributed as much as possible over the entire calendar quarter.

U - reference refers to the QA Handbook for Air Pollution Measurement Systems, Volume II . The use of “S” refers to sections within Part 1 of Volume II. The use of “MS” refers to meth specific sections in Volume II.

### Measurement Quality Objectives - Parameter PM10 (Dichotomous Sampler)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Reporting Units</b>	All data	: g/m <sup>3</sup>	40 CFR, Pt 50.7	
<b>Filter Checks</b>				
Visual defect check	All filters	See reference	Vol II, MS 2.10.4	Discard any defective filters
Filter Integrity				
Collection efficiency	Purchase specification	99 %	40 CFR, Pt 50, App M, S 7.2	As measure by DOP test (ASTM-2988). Reject shipment.
Integrity		± 5 : g/m <sup>3</sup>	“	
Alkalinity		< 25.0 microequivalents/gram	“	Following 2 months storage at ambient temp and relative humidity. Reject filters
Filter Conditioning				
Equilibration time	All Filters	at least 24 hours	40 CFR, Pt 50, App M, S 9.3	Repeat equilibration
Temperature range	“	15 to 30° C	40 CFR, Pt 50, App M, S 7.4	Keep thermometer in balance room and record temperature daily.
Temperature control	“	± 3° C	“	
Humidity range	“	20 to 45 % relative humidity	“	Keep hygrometer in the balance room and record humidity daily.
Humidity control	“	± 5 % relative humidity	“	
<b>Equipment</b>				
Sampler	Purchase specification	Reference or equivalent method	40 CFR, Pt 53.9	
Flow rate transfer standard	Purchase specification	± 2 % accuracy (NIST traceable)	40 CFR, Pt 50, App M, S 7.3	
Analytical balance	Purchase specification	Sensitivity = 0.1 mg	40 CFR, Pt 50, App M, S 7.5 Vol II, MS 2.10.4	This acceptance criterion is inconsistent with other acceptance criteria for balance that are in the quality assurance handbook.
Mass reference standards	Purchase specification	NIST traceable (e.g., ANSI/ASTM Class 2)	Vol II, MS 2.10.4	
<b>Detection Limit</b>				
LDL	Not applicable	Not applicable	40 CFR, Pt 50, App M, S 3.1	The lower limit of the mass concentration is determined by the repeatability of filter tare weights, assuming the nominal air sample volume for the sampler.
<b>Completeness</b>	quarterly	75%	40 CFR, Pt 50, App K, S 2.3	
<b>Sampler Calibration</b>				
Flow control device	On installation, after repairs, after out-of-limits flow check	<4% difference from manufacturers spec and actual	40 CFR, Pt 50, App M, S 7.1 Vol II, MS 2.10.2	Adopt new calibration curve if no evidence of damage, otherwise replace.
Elapsed time meter		± 15 min	40 CFR, Pt 50, App M, S 7.1 Vol II, MS 2.10.1	Adjust or replace.
Flow-rate transfer Standard	On receipt and 1/6 months  Periodically	±2% over the expected range of ambient conditions	40 CFR, Pt 50, App M, S 8.2 Vol II, MS 2.10.1	Checked against NIST-traceable primary standard.

Measurement Quality Objectives - Parameter PM10 (Dichotomous Sampler)				
Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Balance Calibration</b>	1/year		Vol II, MS 2.10.4	Calibrate and maintain according to the manufacturer's recommendations.
<b>Performance Evaluation (NPAP)</b>	1/year at selected sites	Mean absolute difference # 15%	Vol II, S 16.3	Use information to inform reporting agency for corrective action and technical systems audits
<b>Precision</b> Single analyzer	1/6 days	# 5 µg/m <sup>3</sup> for conc. #80 : g/m <sup>3</sup>	40 CFR, Pt 50, App M, S 4.1	Both PM10 values must be > 20 : g/m <sup>3</sup> .
Reporting organization	1/ 3 months	7% for conc. >80 : g/m <sup>3</sup> 95% CI < ± 15%	40 CFR, Pt 58, App A, S 5.3 EPA-600/4-83-023	
<b>Accuracy</b> Single analyzer Annual accuracy	25 % of sites quarterly (all sites yearly)	None 95% CI # ± 20%	40 CFR, Pt 58, App A EPA-600/4-83-023 Vol II, App 15, S 6	Transfer standards different then those used in calibration. Recalibrate before any additional sampling. Invalidate data to last acceptable flow check if difference ≥ 10%.
<b>QC Checks</b> Field calibration flow check	1/month	Percentage difference # ±7 % from sampler's indicated flow rate or # ± 10 % from design condition flow rate	40 CFR, Pt 50, App M, S 8.2 Vol II, MS 2.10.3	Trouble shoot and recalibrate sampler.
"Standard" filter weighing	at beginning of weighing day	± 20 µg of original weight	Vol II, S 2.10.4	Trouble shoot and reweigh.
Reweighing filters	5 exposed and 5 unexposed/day	± 20 µg of original weight	Vol II, S 2.10.4	Trouble shoot and reweigh.
Balance zero and calibration check	every fifth filter	± 4 µg at zero ± 2 µg at 10 mg	Vol II, S 2.10.4	Trouble shoot and reweigh.

<sup>1/</sup> - reference refers to the QA Handbook for Air Pollution Measurement Systems, Volume II . The use of "S" refers to sections within Part 1 of Volume II. The use of "MS" refers to meth specific sections in Volume II.



### Measurement Quality Objectives - Parameter SO<sub>2</sub> (Ultraviolet Fluorescence)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Standard Reporting Units</b>	All data	ppm	40 CFR, Pt 50.4	
<b>Shelter Temperature</b> Temperature range Temperature control	Daily Daily	20 to 30° C # ± 2° C	40 CFR Pt. 53.20 Vol II, S 7.1 <sup>1</sup> / <sub>2</sub> Vol II, MS 2.9	Instruments designated as reference or equivalent have been tested over this temperature range. Maintain temperature above sample dewpoint. Shelter should have a 24- hour temperature recorder. Flag all data for which temperature range or fluctuations are outside acceptance criteria.
<b>Equipment</b> SO <sub>2</sub> analyzer Air flow controllers Flowmeters	Purchase specification	Reference or equivalent method Flow rate regulated to ± 2 % Accuracy ± 2 %	Vol II, MS 2.9 " "	
<b>Detection</b> Noise Lower detectable level	Purchase specification	.005 ppm .01 ppm	40 CFR, Pt 53.20 & 23 “	Instruments designated as reference or equivalent have been determined to meet these acceptance criteria.
<b>Completeness</b> Annual standard 24-hour standard 3-hour standard	Quarterly 24 hours 3 hours	75% 75% 75%	40 CFR, Pt 50.43 “ “	
<b>Compressed Gases</b> Dilution gas (zero air)  Gaseous standards	Purchase specification  Purchase specification	SO <sub>2</sub> free, 21 % O <sub>2</sub> /78 % N <sub>2</sub> , 300 to 400 ppm CO <sub>2</sub> , # 0.1 ppm aromatics NIST Traceable (e.g., permeation tube or EPA Protocol Gas	Vol II, MS 2.9.2  EPA-600/R97/121	Return cylinder to supplier. It is recommended that a clean air system be used instead of compressed air cylinders. Sulfur dioxide in nitrogen EPA Protocol Gases have a 24-month certification period for concentrations between 40 and 499 ppm and a 36-month certification period for higher concentrations.
<b>Calibration</b> Multipoint calibration (at least 4 points)  Zero/span check -level 1   Flowmeters	Upon receipt, adjustment, or 1/ 6 months 1/ 2 weeks   1/3 months	All points within + 2% of full scale of best-fit straight line  Zero drift # ± 20 to 30 ppb Span drift # ± 20 to 25 %  Zero drift # ± 10 to 15 ppb Span drift # ± 15%  Accuracy ± 2 %	Vol II, S 12.6 Vol II, MS 2.9.2  Vol II, S 12.6 “  Vol II, S 12.6 “  Vol II, App 12	Zero gas and at least three upscale points. Note: two pages from Section 2.4 (Calibration Procedures) of Vol II, MS 2.9.2 are missing from the 1994 reprinting of the QA Handbook. <b>If calibration updated at each zero/span-</b> Invalidate data to last acceptable check, adjust analyzer, perform multipoint calibration <b>If fixed calibration used to calculate data.</b> Invalidate data to last acceptable check, adjust analyzer, perform multipoint calibration Flowmeter calibration should be traceable to NIST standards

### Measurement Quality Objectives - Parameter SO<sub>2</sub> (Ultraviolet Fluorescence)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Performance Evaluation</b> (NPAP)  State audits	1/year at selected sites  1/year	Mean absolute difference # 15%  State requirements	Vol II, S 16.3  Vol II, App 15, S 3	Use information to inform reporting agency for corrective action and technical systems audits.
<b>Precision</b> Single analyzer Reporting organization	1/2 weeks  1/3 months	None 95% CI < ± 15%	40 CFR, Pt 58, App EPA-600/4-83-023 Vol II, S 16, S2	Concentration = 0.08-0.10 ppm.
<b>Accuracy</b> Annual accuracy check- Reporting organization	25 % of sites quarterly (all sites yearly)	None 95% CI # ± 20%	40 CFR, Pt 58, App A EPA-600/4-83-023 Vol II, S 16	Four concentration ranges. If failure, recalibrate and reanalyze. Repeated failure requires corrective action.

<sup>1/</sup> - reference refers to the QA Handbook for Air Pollution Measurement Systems, Volume II. The use of "S" refers to sections within Part 1 of Volume II. The use of "MS" refers to meth specific sections in Volume II.

**Measurement Quality Objectives - Parameter CO (Nondispersive Infrared Photometry)**

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Standard Reporting Units</b>	All data	ppm	40 CFR, Pt 50.8	
<b>Shelter Temperature</b> <b>Temperature range</b> <b>Temperature control</b>	Daily Daily	20 to 30° C. < ± 2° C	40 CFR, Pt. 53.20 Vol II, S 7.1 <sup>1/</sup>	Instruments designated as reference or equivalent have been tested over this temperature range. Maintain shelter temperature above sample dewpoint. Shelter should have a 24- hour temperature recorder. Flag all data for which temperature range or fluctuations are outside acceptance criteria.
<b>Equipment</b> CO analyzer Flow controllers Flowmeters	Purchase specification	Reference or equivalent method Flow rate regulated to ± 1% Accuracy ± 2%	40 CFR, Pt 50, App C " "	
<b>Detection Limit</b> Noise Lower detectable level	Purchase specification	0.5 ppm 1.0 ppm	40 CFR, Pt 53.20 & 23 "	Instruments designated as reference or equivalent have been determined to meet these acceptance criteria.
<b>Completeness</b> 8-hour average	hourly	\$75 % of hourly averages for the 8-hour period	40 CFR, Pt 50.8	
<b>Compressed Gases</b> Dilution gas (zero air)  Gaseous standards	Purchase specification  Purchase specification	< 0.1 ppm CO  NIST Traceable (e.g., EPA Protocol Gas)	40 CFR, Pt 50, App C " EPA-600/R97/12	Return cylinder to supplier.  Carbon monoxide in nitrogen or air EPA Protocol Gases have a 36-month certification period and must be recertified to extend the certification.
<b>Calibration</b> Multipoint calibration (at least 5 points)  Zero/span check-level 1   Flowmeters	Upon receipt, adjustment, or 1/ 6 months  1/ 2 weeks  1/3 months	All points within ±2% of full scale of best-fit straight line  Zero drift # ± 2 to 3 ppm Span drift # ± 20 to 25 %  Zero drift # ± 1 to 1.5 ppm Span drift # ± 15%  Accuracy ± 2 %	Vol II, S 12.6 Vol II, MS.2.6.1  Vol II, S 12.6 " Vol II, S 12.6 " Vol II, App 12	Zero gas and at least four upscale calibration points. Points outside acceptance criterion are repeated. If still outside criterion, consult manufacturers manual and invalidate data to last acceptable calibration. <b>If calibration updated at each zero/span</b> , invalidate data to last acceptable check, adjust analyzer, perform multipoint calibration. <b>If fixed calibration used to calculate data</b> , invalidate data to last acceptable check, adjust analyzer, perform multipoint calibration. Flowmeter calibration should be traceable to NIST standards.

<b>Measurement Quality Objectives - Parameter CO (Nondispersive Infrared Photometry)</b>				
<b>Requirement</b>	<b>Frequency</b>	<b>Acceptance Criteria</b>	<b>Reference</b>	<b>Information/Action</b>
<b>Performance Evaluation</b> (NPAP)  State audits	1/year at selected sites  1 /year	Mean absolute difference # 15%  State requirements	Vol II, S 16.3  Vol II, pp 15, S 3	Use information to inform reporting agency for corrective action and technical systems audits
<b>Precision</b> Single analyzer Reporting organization	1/2 weeks 1/3 months	None 95% CI # ± 15%	40 CFR, Pt 58, App A EPA-600/4-83-023 Vol II, App 15, S 5	Concentration = 8 to 10 ppm. Aggregation of a quarters measured precision values.
<b>Accuracy</b> <b>Single analyzer</b> <b>Reporting organization</b>	25 % of sites quarterly (all sites yearly)	None 95% CI # ± 20%	40 CFR, Pt 58, App A	Four concentration ranges. If failure, recalibrate and reanalyze. Repeated failure requires corrective action.

<sup>u</sup> - reference refers to the QA Handbook for Air Pollution Measurement Systems, Volume II . The use of “S” refers to sections within Part 1 of Volume II. The use of “MS” refers to meth specific sections in Volume II.

<b>Measurement Quality Objectives- Parameter PM<sub>2.5</sub></b>				
<b>Requirement</b>	<b>Frequency</b>	<b>Acceptance Criteria</b>	<b>40 CFR Reference</b>	<b>QA Guidance Document 2.12 Reference</b>
<b>Filter Holding Times</b> Pre-sampling	all filters	< 30 days before sampling	Part 50, App.L Sec 8.3	Sec. 7.9
Post-sampling Weighing	“	< 10 days at 25° C from sample end date < 30 days at 4°C from sample end date	“ “	Sec. 7.11 “
<b>Sampling Period</b>	All data	1380-1500 minutes or value if < 1380 and exceedance of NAAQS	Part 50, App.L Sec 3.3	
<b>Reporting Units</b>	All data	µg/m <sup>3</sup>	Part 50.3	Sec. 11.1
<b>Detection Limit</b> Lower DL Upper Conc. Limit	All data All data	2 µg/m <sup>3</sup> 200 µg/m <sup>3</sup>	Part 50, App.L Sec 3.1 Part 50, App.L Sec 3.2	
<b>Sampling Instrument</b> Flow Rate Filter Temp Sensor	every 24 hours of op “ “	≤ 5% of 16.67 ≤ 2% CV measured ≤ 5% average for < 5 min. ≤ 5° C of ambient for <30min	Part 50, App.L Sec 7.4 “ “	
<b>Data Completeness</b>	quarterly	75%	Part 50, App. N, Sec. 2.1	
<b>Filter</b> Visual Defect Check Filter Conditioning Environment Equilibration Temp. Range Temp. Control Humidity Range Humidity Control Pre/post sampling RH Balance	All Filters All filters “ “ “ “ “ “	See reference 24 hours minimum 20-23° C ±2° C SD over 24 hr 30% - 40% RH or ±5% sampling RH but >20%RH ± 5% SD over 24 hr. ± 5% RH located in filter conditioning environment	Part 50, App.L Sec 6.0 Part 50, App.L Sec 8.2 “ “ “ “ Part 50, App.L Sec 8.3.3 “8.3.2	Sec 7.5 Sec. 7.6 " " " "
<b>Filter Checks</b> Lot Blanks	3 filters per lot	less than 15 µg change between weighings	not described	Sec. 7.7
Exposure Lot Blanks	3 filters per lot	less than 15 µg change between weighings	not described	Sec. 7.7

<b>Measurement Quality Objectives- Parameter PM<sub>2.5</sub></b>				
<b>Requirement</b>	<b>Frequency</b>	<b>Acceptance Criteria</b>	<b>40 CFR Reference</b>	<b>QA Guidance Document 2.12 Reference</b>
<b>Lab QC Checks</b>				
Field Filter Blank	10% or 1 per weighing session	±30 µg change between weighings	Part 50, App.L Sec 8.3	Sec. 7.7
Lab Filter Blank	10% or 1 per weighing session	±15 µg change between weighings	Part 50, App.L Sec 8.3	“
Balance Check	beginning, every 10th sample, end	≤3 µg	not described	Sec. 7.9
Duplicate Filter Weighing	1 per weighing session	±15 µg change between weighings	not described	Sec 7.11
<b>Calibration/Verification</b>				
Flow Rate (FR) Calibration	If multi-point failure	± 2% of transfer standard	Part 50, App.L, Sec 9.2	Sec 6.3
FR multi-point verification	1/yr	± 2% of transfer standard	Part 50, App.L, Sec 9.2.5	Sec 6.3 & 8.4
One point FR verification	1/4 weeks	± 4% of transfer standard	Part 50, App.L, Sec 9.2	Sec 8.4
External Leak Check	every 5 sampling events	80 mL/min	Part 50, App.L, Sec 7.4	Sec. 6.6 & 8.4
Internal Leak Check	every 5 sampling events	80 mL/min	“	Sec. 6.6 & 8.4
Temperature Calibration	If multi-point failure	± 2% of standard	Part 50, App.L, Sec 9.3	Sec. 6.4
Temp M-point Verification	on installation, then 1/yr	± 2°C of standard	Part 50, App.L, Sec 9.3	Sec. 6.4 and 8.4
One-point temp Verification	1/4 weeks	± 4°C of standard	“	Sec. 6.4 and 8.4
Pressure Calibration	on installation, then 1/yr	±10 mm Hg	“	Sec. 6.5
Pressure Verification	1/4 weeks	±10 mm Hg	“	Sec. 8.2
Clock/timer Verification	1/ 4 weeks	1 min/mo	Part 50, App.L, Sec 7.4	not described
<b>Accuracy</b>				
FRM Performance Evaluation	25% of sites 4/yr	± 10%	Part 58, App A, Sec 3.5	Sec 10.2
External Leak Check	4/yr	< 80 mL/min	not described	Sec. 10.2
Internal Leak Check	4/yr	< 80 mL/min	not described	“
Temperature Audit	4/yr	± 2°C	not described	“
Pressure Audit	4/yr (?)	±10 mm Hg	not described	“
Balance Audit	1/yr	Manufacturers specs	not described	“
<b>Accuracy</b>				
Flow Rate Audit	1/2wk (automated) 4/yr (manual)	± 4% of audit standard	Part 58, App A, Sec 3.5	Sec. 10.2
<b>Precision</b>				
Collocated samples	every 6 days for 25% of sites	CV ≤ 10%	Part 58, App.A, Sec 3.5 and 5.5	Sec. 10.2
Single analyzer	1/3 mo.	CV ≤ 10%	not described	not described
Single Analyzer	1/ yr	CV ≤ 10%	not described	not described
Reporting Org.	1/ 3 mo.	CV ≤ 10%	not described	not described

**Measurement Quality Objectives- Parameter PM<sub>2.5</sub>**

Requirement	Frequency	Acceptance Criteria	40 CFR Reference	QA Guidance Document 2.12 Reference
<b>Calibration &amp; Check Standards</b>				
Flow Rate Transfer Std.	1/yr	±2% of NIST-traceable Std.	Part 50, App.L Sec 9.1 & 9.2	Sec. 6.3
Field Thermometer	1/yr	± 0.1° C resolution	not described	Sec 4.2 & 6.4
		± 0.5° C accuracy		“
Field Barometer	1/yr	± 1 mm Hg resolution	not described	“
		± 5 mm Hg accuracy		“
Working Mass Stds.	3-6 mo.	0.025 mg	not described	Sec 4.3 and 7.3
Primary Mass Stds.	1/yr	0.025 mg	not described	"

### Measurement Quality Objectives - Parameter PAMS Volatile Organic Compounds (VOC)

Requirement	Frequency	Acceptance Criteria	Reference	Information/Action
<b>Standard Reporting Units</b>	All data	ppbC	TAD, July 1997	
<b>Shelter Temperature Temperature range</b>	Daily	20 to 30° C.	Vol II, S 7.1 <sup>1/</sup>	Instruments designated as reference or equivalent have been tested over this temperature range. Maintain shelter temperature above sample dewpoint. Shelter should have a 24- hour temperature recorder. Flag all data for which temperature range or fluctuations are outside acceptance criteria.
<b>Detection Limit</b> System detection limit		1 ppbC	TAD Sect 2.8 2.3	Calculation based on multiple manual or automated analysis and 40 CFR recommendations
<b>Completeness</b> (sesonal)	annually	85 %	TAD 2.8.1	
<b>Calibration</b> Multipoint retention time calibration standard	Start of analytical season	correlation coefficient $\geq 0.995$	TAD 2.8.2.3	Triplicate analysis of multiple level propane standards over the expected sample concentration range (a minimum of three levels)
<b>Performance Evaluation</b> NPAP	prior to start of sampling season and twice during monitoring season	In absence of specified objectives within 25%	TAD 2.8.2.3	Useful for informing reporting agency for corrective actions and technical systems audits.
<b>Precision</b> Duplicate samples	once/2 weeks automated 10% -manual	$\pm 25\%$ RSD or RPD	TAD 2.8.2.1.1	Comparison of duplicate field samples, or replicate sample analysis using manual or automated field devices.
<b>QC Checks</b> Retention time (RT) calibration check  Canister cleaning	Weekly	Response Factor within 10% RPD of calibration curve  < 10 ppbC total	TAD 2.8.2.3	Retention time checked versus annual PAMS retention time cylinder provided to each site in the program.  Canister cleaning per approved methodology
Background/carryover	weekly and after calibration & RT	< 20 ppbC for both columns or <10 ppbC per column	TAD 2.8.2.3	Background testing according to TAD