

8-Hour Ozone Data Quality Objectives *lite!*


SAMWG

10/16/03

What are we here for & why now?

- What
 - Basic understanding of the DQO process
 - Agreement on proposed **measurement quality objectives** for precision and bias
- Why Now
 - Monitoring Strategy
 - Opportunity for review and revision of CFR (Nov. 03 draft for potential late 04 early 05 promulgation)
 - Did not rethink the QA requirements back in 97
 - DQO Endorsement –provides a rationale for the quality system requirements “Now we see why your asking us to do this”
 - Improved data quality
 - It’s been awhile- the monitoring technology is better and the data quality is showing this. (*“The +/- 20% calibration tolerance encourages sloppy practices by instrumentation technicians” -Mims*)

What is a performance curve?




Development of the 8-hour Ozone DQO

1. Aggregated 3 years (99-2001) of ozone data from ~1,125 sites
2. Reviewed data to identify the appropriate population distribution and population input parameters--Developed the ozone main simulation model
3. Reviewed measurement data quality indicators (precision, bias, completeness) and developed within-day module that adjusts the main simulation model
4. Calculate various decision performance curves depending on changes to input parameters

Population Input Parameters


- Seasonality Ratio - Ratio of high to low point on sine curve
- Population CV - Variability of the population distribution
- Autocorrelation – Similarity in concentration between successive days
- Shift – The time at which ozone concentrations peak during the year

Typical time series of daily maximum 8-hour average ozone concentrations



Measurement Input Parameters

- **Precision** – Repeatability of the measurement system
- **Bias** – Systematic deviation from the true value
- **Completeness**- acceptable data loss on a sampling day (6 hours) and annual basis (75%)
 - completeness values (75% and 95%) were used



Selecting Input Parameters- Using the Conservative Approach


We selected input parameters that create the more extreme but realistic performance curves.

Percentiles of the Estimated Population and Measurement Parameters from the 3-Year Ozone Data Set

Percentile	Population Input Parameters				Measurement Parameters	
	Ratio of High to Low Season	Population CV	Auto-Correlation	Expected date of peak ozone	Sampler Bias	Hourly Precision CV
10	1.33	36% (43%)	0.17 (0.15)	7/27	1.23%	1.32%
20	1.49	34%	0.27	7/12	1.64%	1.69%
30	1.62	32%	0.34	7/6	1.98%	2.04%
40	1.76	30%	0.39	7/1	2.38%	2.38%
50	1.94	29%	0.44	6/27	2.72%	2.74%
60	2.14	27%	0.51	6/24	3.06%	3.10%
70	2.39	26%	0.55	6/20	3.51%	3.52%
80	2.79	25%	0.60	6/14	4.04%	4.00%
90	3.46 (3.5)	22%	0.67	6/3	5.00%	4.76%


For the performance curves the population input parameters were held constant, the measurement parameters are varied.

Decision Performance Curves for 75% Completeness



- 20% CV & Bias
- 10% CV & Bias
- 7.5% CV & Bias
- 5.0% CV & Bias
- - - 5% Limit
- - - 95% Limit
- · - Action Limit

Decision Performance Curve for Ozone



— Conservative Proposal
 (.078-099 ppm)

... More Typical Curve
 (.080 - .090 ppm)


Bias 7.5
 CV 7.5
 Comp 75%
 10/90 -Pop param.

Bias 3.5
 CV 3.5
 Comp 95%
 70/30- Pop. Param

DQOs


Summary

- Initial planning- can be used to help understand data quality impacts on:
 - The level of the standard
 - The form of the standard
 - Optimize network designs (number of sites, sampling frequency etc.)
- Ongoing monitoring implementation
 - Provides an excellent assessment tool for achievement of data quality
 - Provides a way to focus on quality system improvements as well as site specific improvements




Steps to Data Assessment

- 1) Site level
 - a) Individual value
 - b) Trend




- 2) Site/Reporting Organization
 - a) Yearly
 - b) 3 year



Ozone CY 99-01 Data Quality Objective Parameters

State	Rep. Org	Siteid	Ave Conc. ppm	Season Ratio	Pop CV (%)	Autocor.	Complete.	Bias CL	Precision CL	99-01 Gray Zones	Gray Zone Within DQO
AK	15	20200018	0.066	1.33	25	0.27	0.90	2.7	7	(.078 - .093)	Yes
AK	15	20200044	0.074	1.37	23	0.34	0.94	-0.4	4.5	(.0789- .090)	Yes
AK	15	20900010	0.076	1.62	33	0.60	0.83	4.2	3.7	(.081 - .089)	Yes
AK	15	21100004	0.080	1.76	23	0.25	0.89	2.6	6.2	(.078 - .093)	Yes
AK	15	21700014	0.084	2.14	21	0.34	0.85	1.4	3.2	(.0789- .090)	Yes
AK	15	21700008	0.089	1.67	34	0.52	0.88	1.7	2.5	(.081 - .089)	Yes
AK	15	22900003	0.092	2.39	36	0.34	0.76	-5.6	4.7	(.078 - .093)	No
AK	15	21100005	0.096	2.19	22	0.25	0.89	4.6	6.3	(.0789- .090)	Yes
AK	15	21700007	0.101	1.24	32	0.34	0.85	4.8	5.2	(.081 - .089)	Yes
AK	15	21700022	0.104	2.22	12	0.52	0.88	4	4.4	(.078 - .093)	Yes
AK	15	22900004	0.108	3.46	33	0.34	0.76	3.7	5.8	(.0789- .090)	Yes

3) Assessments related to the DQOs



DQOs

Summary

- Are not used to invalidate data
 - Precision and bias QC checks can fail the MQOs without invalidation
- Identifies the probability of decision errors not the fact that these errors have occurred.
- Proposed
 - 5% decision error rate
 - 7% precision and bias MQO
 - 75% Completeness


Performance Curve Gray Zones

Measurement Input Parameters				Gray Zone (ppm)			
Daily Completeness	Minimum Hours per day	Hourly Measurement Precision (CV)	Measurement bias	Lower end point	Upper end point		
95%	18	7.5%	7.5%	0.077	0.095		
		10%	5%	0.078	0.093		
			10%	0.075	0.098		
			20%	0.069	0.110		
		15%	5%	0.078	0.092		
			10%	0.074	0.097		
			20%	0.068	0.110		
		20%	5%	0.076	0.091		
			10%	0.073	0.096		
			20%	0.066	0.108		
		75%	18	7.5	7.5	0.078	0.099
				10%	5%	0.080	0.096
10%	0.076				0.102		
20%	0.070				0.114		
15%	5%			0.079	0.096		
	10%			0.075	0.101		
	20%			0.069	0.114		
20%	5%			0.078	0.095		
	10%			0.074	0.100		
	20%			0.068	0.112		

Measurement Input Parameters

- **Precision** – Repeatability of the measurement system
- **Bias** – Systematic deviation from the true value
- **Completeness**- acceptable data loss on a sampling day (6 hours) and annual basis (75%)
 - completeness values (75% and 95%) were used

Relationship Between Bias and Imprecision on Data Quality





Why do we need DQO's?

Decision Maker

I want the right answer!
You have 1 million and 1 year

A man in a dark suit and glasses is sitting at a desk, writing on a document with a pen.

Technical Manager

I'll use the very best equipment for the best precision and accuracy

A woman in a red jacket is sitting at a desk with a computer monitor and keyboard, looking at the screen.

Reality

A balance scale with two pans. The left pan is lower and contains the text 'Data Quality'. The right pan is higher and contains the text 'Time Resources'.

Data Quality

Time Resources

QA Manager

I'll try to ensure that the precision and bias are near zero

A man in a dark suit and blue tie is holding a tablet computer and looking at it.

Statistician

Give me all the samples you can.
I don't want to make any assumptions

A man in a blue shirt is standing in front of a blackboard, writing on it with a piece of chalk.

DQO Process Promotes Communication/Collaboration




Problem
Decision
Error Tolerance
Resources
Time



Methods
Expertise
Logistics
Implementation



DQOs
Quality system development
Measurement quality



Uncertainties
Statistical sample design
Achievable error limits

DQOs

A process for ensuring that environmental data will be adequate for their intended use


Problem



Resource Effective
Data Collection Design

- State the **Problem**
- Identify the **Decision**
- Identify **Inputs** to decision
- Define the **Boundaries**
- Develop the Decision **Rule**
- Specify **Tolerable** limits to **Decision Error**
- **Optimize** the design

Understanding and Controlling Uncertainty in Order to Minimize **Decision Errors**



OZONE OPERATIONAL CRITERIA TABLE

Criteria	Acceptable Range	Frequency
Shelter Temperature		
<u>Temperature range</u>	20 to 30° C or Instrument must be operated per manufacturers specifications (see note D Below)	<u>Daily (hourly value)</u>
Temperature control	# 2°C SD over 24 hours	Daily
Accuracy		
<u>Single analyzer</u>	None (See note C)	25 % of sites
<u>Annual accuracy</u>	95% Confidence Interval# ± 20%	<u>quarterly (all sites year)</u>
Compressed Gases		
Zero air	Free of any substance that might react with ozone	
Calibration		
Multipoint calibration (at least 5 points)	Linearity error <5%	Upon receipt, adjust or 1/year if cont. zero/s 1/ 6 months if man zero/span
Precision		
<u>Reporting organization</u>	95 % Confidence Interval # ± 15 %	<u>1/3 months</u>
Performance Evaluation		
<u>Federal audits</u>	Mean absolute difference # 15%	<u>1/year at selected sites</u>
State audits	State requirements (see note E below)	1/year