

Suggested and Implemented Changes to the Validation Template. Posted on AMTIC 7/2014

Template Title	Criteria	Field changed	Original Value	Revised Value	Importance	Date of Change	Comment
PM _{2.5} Filter Based Local Conditions Validation Template	Collocated Samples	Information/ Action	App A Sec 2.3.1.3	App A Sec 2.3.1.1	Low		Initial reference for acceptance value was for PM10-2.5 DQO which was incorrect and should be PM2.5.
Pb High Volume TSP Validation Template		Template Title	Pb High Volume TSP Validation Template	Pb High Volume TSP Local Conditions Validation Template	Low		
NO ₂ , NO _x , NO Validation Template	Zero/span check	Acceptance criteria	Zero drift < + 1.5 ppb	Zero drift $\leq \pm 3.0$ ppb (24 hr) $\leq \pm 5.0$ ppb (>24hr-14 day)	High	6/3/2014	See Technical Guidance at http://www.epa.gov/ttn/amtic/cpreldoc.html
O3 Validation Template	Zero/span check	Acceptance criteria	Zero drift < + 1.5 ppb	Zero drift $\leq \pm 3.0$ ppb (24 hr) $\leq \pm 5.0$ ppb (>24hr-14 day)	High	6/3/2014	See Technical Guidance at http://www.epa.gov/ttn/amtic/cpreldoc.html
SO2 Validation Template	Zero/span check	Acceptance criteria	Zero drift < + 1.5 ppb	Zero drift $\leq \pm 3.0$ ppb (24 hr) $\leq \pm 5.0$ ppb (>24hr-14 day)	High	6/3/2014	See Technical Guidance at http://www.epa.gov/ttn/amtic/cpreldoc.html
CO Validation Template	Zero/span check	Acceptance criteria	Zero drift < + 0.03 ppm	Zero drift $\leq \pm 0.4$ ppm (24 hr) $\leq \pm 0.6$ ppm (>24hr-14 day)	High	6/3/2014	See Technical Guidance at http://www.epa.gov/ttn/amtic/cpreldoc.html
Continuous PM10 STP Conditions Validation Template		Template Title	Continuous PM10 STP Conditions Validation Template	Continuous PM10 STP Conditions Validation Template	Low		misspelling
Number of Low Volume PM		In Comment #1 sections	Comment about < 80	Revised comment to use	Low		Bottom of Template explaining the use of external