

Annual Data Summary Report for the Chemical Speciation of PM_{2.5} Filter Samples Project

January 1 through December 31, 2005

**Prepared for:
U.S. Environmental Protection Agency
Office of Air Quality Planning and Standards
Research Triangle Park, NC 27711**

EPA Contract No. 68-D-03-038

**Prepared by
RTI
P.O. Box 12194
3040 Cornwallis Road
Research Triangle Park, NC 27709**



This page was intentionally left blank.

TABLE OF CONTENTS

| <u>Section</u> | <u>Page</u> |
|--|--------------------|
| Executive Summary | ES-1 |
| 1.0 Introduction..... | 1-1 |
| 1.1 Program Overview | 1-1 |
| 1.2 Project/Task Description..... | 1-1 |
| 1.3 Major Laboratory Operational Areas | 1-2 |
| 2.0 Quality Issues and Corrective Actions..... | 2-1 |
| 2.1 Data Quality | 2-1 |
| 2.2 Summary of Data Completeness..... | 2-2 |
| 2.3 Corrective Actions | 2-3 |
| 2.3.1 Gravimetric Mass..... | 2-3 |
| 2.3.2 Elemental Analysis | 2-4 |
| 2.3.3 Ion Analysis | 2-4 |
| 2.3.4 OC/EC Analysis..... | 2-4 |
| 2.3.5 Sample Handling and Archiving Laboratory (SHAL)..... | 2-4 |
| 2.3.6 Data Processing..... | 2-4 |
| 2.4 Other Quality Issues..... | 2-4 |
| 3.0 Laboratory Quality Control Summaries..... | 3-1 |
| 3.1 Gravimetric Laboratory | 3-1 |
| 3.1.1 Quality Issues and Corrective Actions..... | 3-1 |
| 3.1.2 Description of QC Checks Applied | 3-1 |
| 3.1.3 Summary of QC Results | 3-6 |
| 3.1.4 Assessment of Between-instrument Comparability | 3-6 |
| 3.1.5 Determination of Uncertainties and Method Detection Limits..... | 3-6 |
| 3.1.6 Audits, Performance Evaluations, Training, and Accreditations..... | 3-7 |
| 3.2 Ions Analysis Laboratory | 3-8 |
| 3.2.1 Quality Issues and Corrective Actions..... | 3-8 |
| 3.2.2 Description of QC Checks Applied | 3-8 |
| 3.2.3 Summary of QC Results | 3-11 |
| 3.2.4 Assessment of Between-instrument Comparability | 3-13 |
| 3.2.5 Determination of Uncertainties and MDLs..... | 3-14 |
| 3.2.6 Audits, PEs, Training, and Accreditations..... | 3-14 |
| 3.3 Organic Carbon/Elemental Carbon Laboratory | 3-14 |
| 3.3.1 Quality Issues and Corrective Actions..... | 3-14 |
| 3.3.2 Description of QC Checks Applied | 3-14 |
| 3.3.3 Summary of QC Results | 3-16 |
| 3.3.4 Assessment of Between-Instrument Comparability..... | 3-21 |
| 3.3.5 Determination of Uncertainties and MDLs..... | 3-21 |
| 3.3.6 Audits, Performance Evaluations, Training, and Accreditations..... | 3-23 |
| 3.4 X-ray Fluorescence Laboratories | 3-25 |
| 3.4.1 Chester LabNet X-Ray Fluorescence Laboratory | 3-25 |
| 3.4.2 RTI International XRF Laboratory | 3-30 |

TABLE OF CONTENTS (Continued)

| <u>Section</u> | <u>Page</u> |
|---|--------------------|
| 3.5 Sample Handling and Archiving Laboratory | 3-39 |
| 3.5.1 Quality Issues and Corrective Actions | 3-39 |
| 3.5.2 Description of QC Checks Applied | 3-40 |
| 3.5.3 Summary of QC Results | 3-40 |
| 3.5.4 Summary of Scheduling Problems | 3-41 |
| 3.5.5 Support Activities for Site Operators and Data Users | 3-42 |
| 3.5.6 Audits, PEs, Training, and Accreditations | 3-43 |
| 3.6 Denuder Refurbishment Laboratory | 3-43 |
| 3.6.1 Quality Issues and Corrective Actions | 3-44 |
| 3.6.2 Operational Discussion | 3-44 |
| 3.6.3 Description of QC Checks Applied and Results | 3-45 |
| 4.0 Data Processing | 4-1 |
| 4.1 Quality Issues and Corrective Actions | 4-1 |
| 4.2 Operational Summary | 4-1 |
| 4.3 Operational Changes and Improvements | 4-1 |
| 4.4 Monthly Data Postings to Web Site | 4-2 |
| 4.5 Postings to AQS | 4-3 |
| 4.6 Data User Support Activities | 4-4 |
| 4.6.1 Data Change Requests | 4-4 |
| 4.6.2 Requests for Old Data | 4-4 |
| 4.6.3 SDVAT Support | 4-4 |
| 4.6.4 Data User Communications | 4-5 |
| 5.0 Quality Assurance and Data Validation | 5-1 |
| 5.1 QA Activities | 5-1 |
| 5.1.1 QAPP Updates | 5-1 |
| 5.1.2 SOP Updates | 5-1 |
| 5.1.3 Internal Surveillance Activities | 5-2 |
| 5.1.4 Data User Support Activities | 5-2 |
| 5.2 Data Validation and Review | 5-3 |
| 5.2.1 Review of Monthly Data Reports to STN Web Site | 5-3 |
| 5.2.2 Review of Monthly Data Packages to AQS | 5-3 |
| 5.3 Analysis of Collocated Data | 5-7 |
| 5.3.1 Precision | 5-7 |
| 5.3.2 Bias | 5-13 |
| 5.4 Analysis of Trip and Field Blanks | 5-13 |
| 5.4.1 Outliers By Sample Type | 5-16 |
| 5.4.2 Trends and Offsets in Blank Data | 5-16 |
| 6.0 External Audits | 6-1 |
| 6.1 Performance Evaluation Audit Results | 6-1 |
| 6.2 System Audit Results | 6-1 |
| 6.3 Synoptic Summary of 2005 Speciation Trends and IMPROVE Network Audits | 6-1 |
| 7.0 List of References | 7-1 |
| 7.1 List of Current STN Documents | 7-1 |
| 7.2 Special Reports Issued During the Reporting Period | 7-2 |

List of Appendices

- A. Method Detection Limits
- B. Data Completeness Summary
- C. Gravimetric Inter-Laboratory Comparison Study
- D. Experimental Inter-Comparison of Speciation Laboratories
- E. Systems Audit Report
- F. Synoptic Summary of 2005 Speciation Trends and IMPROVE Network Audits

LIST OF TABLES

| <u>Table</u> | <u>Page</u> |
|--|--------------------|
| 2-1 Gravimetry Laboratory - Corrective Actions in Response to Facility Problems | 2-2 |
| 3-1 Summary of QC Checks Applied in the Gravimetry Laboratory | 3-2 |
| 3-2 Description of Audits, PEs, Training, and Accreditations | 3-7 |
| 3-3 Description of Ion Chromatographic Systems Used for Analysis of PM _{2.5} Filter Samples | 3-8 |
| 3-4 Ion Analysis of PM _{2.5} – QA/QC Checks | 3-9 |
| 3-5 MDL* for Each Instrument and Analyte..... | 3-10 |
| 3-6 Definitions and Specifications for QA/QC Samples | 3-10 |
| 3-7 Average Percent Recovery for QA and QC Samples | 3-11 |
| 3-8 Average Percent Recovery for Spikes | 3-12 |
| 3-9 Filter Blank (N) and Reagent Blank Values (ppm) for all Analytes | 3-12 |
| 3-10 Between-Instrument Comparability | 3-13 |
| 3-11 OC/EC Laboratory QC Checks, Acceptance Criteria, and Corrective Actions | 3-15 |
| 3-12 OC/EC Laboratory-Assigned Data Flags | 3-16 |
| 3-13 OC/EC Instrument Blank Statistics | 3-16 |
| 3-14 OC/EC Three-Point Calibration Statistics..... | 3-18 |
| 3-15 OC/EC Daily Calibration Check Statistics..... | 3-19 |
| 3-16 Duplicate OC/EC Analysis Statistics..... | 3-20 |
| 3-17 Estimated Uncertainties for OC/EC Carbon Fractions | 3-22 |
| 3-18 Estimated Uncertainties for OC/EC Carbon Fractions | 3-23 |
| 3-19 QC Procedures Performed in Support of XRF Elemental Analysis..... | 3-26 |
| 3-20A Summary of Chester XRF 770 Laboratory QC Precision Data 1/1/2005 through 10/4/2005 ... | 3-27 |
| 3-20B Summary of Chester XRF 771 Laboratory QC Precision Data 1/1/2005 through 10/4/2005 ... | 3-27 |
| 3-21A Recovery Determined from Analysis of NIST SRMs 1832, 1833, and 2708 for Chester XRF 770 -- 1/1/2005 through 12/31/2005..... | 3-28 |
| 3-21B Recovery Determined from Analysis of NIST SRMs 1832, 1833 and 2708 for Chester XRF 771 -- 1/1/2005 through 12/31/2005..... | 3-28 |
| 3-22 Replicate Data for Chester XRF 770 and 771 | 3-29 |
| 3-23 Three-sigma MDLs ^a for Chester 770 and 771 Instruments..... | 3-31 |
| 3-24 QC Procedures Performed in RTI XRF Elemental Analysis Laboratory | 3-31 |
| 3-25A Summary of RTI XRF 1 Laboratory QC Precision Data, ug/cm ² 1/1/2005 through 10/4/2005 | 3-32 |
| 3-25B Summary of RTI XRF 1 Laboratory QC Precision Data, ug/cm ² 10/5/2005 through 12/31/2005 | 3-32 |
| 3-25C Summary of RTI XRF 2 Laboratory QC Precision Data, ug/cm ² 1/1/2005 through 12/31/2005 | 3-33 |
| 3-26A Recovery Determined from Analysis of NIST SRMs 1832 and 2783 for RTI XRF 1 -- 1/1/2005 through 12/31/2005 | 3-33 |
| 3-26B Recovery Determined from Analysis of NIST SRMs 1832 and 2783 for RTI XRF 2 -- 1/1/2005 through 12/31/2005 | 3-34 |
| 3-27 Replicates | 3-34 |
| 3-28 Counts of Round Robin Filter Analyses During 2005..... | 3-35 |
| 3-29A Regression Results for 15 Elements - RTI XRF Instruments | 3-36 |
| 3-29B Regression Results for 15 Elements - Chester XRF Instruments..... | 3-36 |

LIST OF TABLES (Continued)

| <u>Table</u> | <u>Page</u> |
|---|--------------------|
| 3-30 Average Laboratory Uncertainties | 3-37 |
| 3-31 MDL Values for XRF 1 and XRF 2 | 3-39 |
| 3-32 Sites with Less than 95% of Filters Run on Intended Sampling Date | 3-42 |
| 3-33 Summary of SHAL Communications With Site Operators and Data Users | 3-42 |
| 3-34 Review of SHAL Technician Processing Filter Modules | 3-43 |
| 3-35 Denuder Refurbishments January 1 through December 31, 2005 | 3-44 |
| 4-1 Events Posted To Web Site | 4-2 |
| 4-2 Records Posted To Web Site | 4-2 |
| 4-3 Events Posted To AQS | 4-3 |
| 4-4 Records Posted to AQS | 4-3 |
| 4-5 Change Requests Per Report Batch | 4-4 |
| 5-1 Summary of Validity Status Codes by Delivery Batch Number | 5-4 |
| 5-2 Summary of Null Value Codes by Delivery Batch Number | 5-5 |
| 5-3 RTI-assigned Flags (not reported to AQS) by Delivery Batch Number | 5-6 |
| 5-4 Collocated sites in the STN | 5-7 |
| 5-5 Trace Elements by XRF | 5-10 |
| 5-6 Anions and Cations by IC | 5-11 |
| 5-7 Organic and Elemental Carbon | 5-11 |
| 5-8 Particulate Matter (Gravimetry) | 5-12 |
| 5-9 Concentration Percentiles for Combined Trip and Field Blanks Reported During 2005 | 5-14 |
| 5-10 Outliers by Sampler Type for Selected Analytes | 5-17 |
| 6-1 Systems Audit Findings and Responses | 6-2 |

LIST OF FIGURES

| <u>Figure</u> | <u>Page</u> |
|--|--------------------|
| 3-1A Late Coolers By Sites | 3-41 |
| 3-1B Late Coolers By Federal Express | 3-41 |
| 5-1 Examples of the comparisons for mass, sulfate, nitrate, sulfur, organic carbon, and elemental carbon | 5-8 |

EXECUTIVE SUMMARY

Introduction

The U.S. Environmental Protection Agency (EPA) established a PM2.5 Chemical Speciation Trends Network (STN) in 1999. The STN included a core set of 54 trends analysis sites and some 160 other sites. RTI is assisting in the PM2.5 STN by shipping ready-to-use filter packs and denuders to all the field sites and by conducting gravimetric and chemical analyses of several types of filters used in the samplers. RTI staff performed an extensive array of quality assurance/quality control (QA/QC) activities to ensure that the data provided to EPA and the States are of the highest quality. The laboratory QA activities in terms of accuracy, precision, data completion, and any corrective actions taken on the chemical speciation of samples from the STN sites from January 1 to December 31, 2005, are described in this report.

Data Quality

Analytical completeness exceeded 95%, and laboratory accuracy and precision were under control as demonstrated by routine QC samples, laboratory audits, and instrument intercomparison. The Research Triangle Institute (RTI) laboratories were audited by EPA personnel during July 2005, and also received performance audit samples. Except for one gravimetric value (which may have been affected by the manufacturer's debris) all results were within acceptable limits, as shown in Appendices D and E.

Laboratory Performance

Section 3.0 of this report provides the details of accuracy, precision, and other measures of laboratory performance. The laboratories consistently met their QC goals of routine analyses, which are detailed in Sections 3.1 (Gravimetry Laboratory), 3.2 (Ion Analysis), 3.3 (Organic and Elemental Carbon [OC/EC]), and 3.4 (X-ray Fluorescence).

Problems with the weighing chamber environmental controls in the Gravimetry Laboratory (Section 3.1) were dealt with aggressively so that a minimum of data had to be flagged as outside holding time or environmental criteria. The other issue affecting the gravimetric data was a filter debris problem originating with a particular manufacturer's lot. RTI arranged for the exchange of 5,000 filters from this lot. The Standard Operating Procedure (SOP) for gravimetric analysis has been updated to increase the frequency of re-weighing in the laboratory in order to quickly recognize and correct future instances (Section 2.3). Systems and performance audit results (Section 3.1.6) for gravimetric mass were found to be satisfactory.

Minimal problems with laboratory operations and filter media were reported by the Ion and OC/EC laboratories during 2005. Systems and performance audit results for these laboratories were satisfactory (Sections 3.2.6 and 3.3.6).

The X-ray fluorescence (XRF) laboratories operated by RTI and subcontractor Chester LabNet (CLN) generally met the prescribed QC criteria for analysis (Sections 3.4.1 and 3.4.2). Both laboratories had equipment downtime, which affected sample analysis logistics, but this had no effect on data quality. The RTI and CLN laboratories participate in an intercomparison (round-robin) program described in Section 3.4.2.4. Systems and performance audit results (Section 3.4.6) were satisfactory.

Operations in the Sampling Handling and Archiving Laboratory (SHAL) proceeded normally during 2005, with a few rare problems such as switching the paperwork for coolers shipped to two different sites (Section 3.5.1). A small number of samples were missed due to late return of coolers from the field sites. Systems and performance audit results (Section 3.5.6) were satisfactory. No significant quality issues were reported by the denuder refurbishment laboratory (Section 3.6).

No significant quality issues were reported by the data processing and data validation functions during 2005 (Sections 4.0 and 5.0). Data continues to be reviewed and posted to a secure website on a monthly basis for review. Finalized data are posted to the EPA AQS database approximately 60 days after initial posting (Section 4.0). A number of data users contacted SHAL, data processing, and quality assurance (QA) personnel with questions about specific data items, or to request explanations about apparent discrepancies.

Estimation of MDLs and Uncertainties

Method Detection Limits (MDLs) for all laboratory methods are provided in Appendix A. Uncertainties are estimated based on laboratory QC data, augmented by a 5% concentration-proportional term to account for field handling and sample volume uncertainties. Results from collocated samplers (Section 5.3) indicate that this uncertainty model is reasonable.

Quality Issues

Aside from the specific issues discussed in the two Corrective Action Requests (CARs) issued during 2005 (Section 2.1), there are some ongoing issues that have not been assigned CARs because there was no specific action that RTI could take, or because they required input and cooperation from others outside RTI. These issues are summarized in the following table.

| CAR Number | Lab | Description | Response | Effect on Data |
|------------|-------|--|---|--|
| 9 | Grav. | Whatman Teflon Filter - Debris problem. Loose debris causes net filter weights to be occasionally too low. | 5,000 filters were returned to manufacturer and replaced; mass balance outlier flag applied; SOPs for data screening and replicate analysis were updated. | Increased incidence of mass balance outliers (AQS '5' code). Data users should be aware of isolated outliers in gravimetric mass data during 2005. |
| 10 | SHAL | Shipping error for 12/11/04 Paducah and Perkinstown coolers | Corrected data in database. Increased supervision and training of SHAL workers. | None. |
| none | SHAL | Late-arriving coolers | DOPO and others are notified whenever coolers are received late from the field. | Data are flagged as missing. |
| none | XRF | Harmonize XRF uncertainty calculations | RTI is consulting with recognized experts to identify the correct and consistent methods for calculation for uncertainty. | More accurate uncertainty values in the future could assist modelers and other data users achieve more accurate results. |
| none | All | Investigate sampler-dependent background levels | RTI is examining historical data for evidence of systematic contamination with different sampler types (in progress). | Potential to inform users of increased incidence of certain types of outliers. |

1.0 Introduction

1.1 Program Overview

In 1997, the U.S. Environmental Protection Agency (EPA) promulgated the new National Ambient Air Quality Standards (NAAQS) for particulate matter. The regulations (given in 40 CFR Parts 50, 53, and 58) apply to the mass concentrations ($\mu\text{g}/\text{meter}^3$ of air) of particles with aerodynamic diameters less than 10 micrometers (the PM₁₀ standard) and less than 2.5 micrometers (the PM_{2.5} standard). Currently, a 1500-site mass measurements network and a 214-site chemical speciation monitoring network have been established.

The ambient air data from the first network, which measures solely the mass of particulate matter, will be used principally for NAAQS comparison purposes in identifying areas that meet or do not meet the NAAQS criteria and in supporting designation of an area as attainment or non-attainment.

The smaller chemical Speciation Trends Network (STN) included a core set of 54 trends analysis sites and some 160 other sites from State and local agencies supported by RTI. This data summary report covers the quality assurance (QA) aspects of the collection and chemical speciation of samples from these sites from January 1 through December 31, 2005. Chemical speciation data will be used to support development of emission mitigation approaches to reduce ambient PM_{2.5} concentration levels. Such needs include emission inventory establishment, air quality model evaluations, and source attribution analysis. Other uses of the data sets will be regional haze assessments, estimating personal exposure to PM_{2.5} and its components, and evaluating potential linkages to health effects.

RTI is supporting the PM_{2.5} STN by shipping ready-to-use filter packs and denuders to the field sites and by conducting gravimetric and chemical analyses of the several types of filters used in the samplers. The details of the QA activities being performed are described in the RTI QA Project Plan (QAPP) for this project. The QAPP focuses on the QA activities associated with RTI's role in performing these analyses, as well as in validating and reporting the data, and should be considered a companion document to this annual QA report.

1.2 Project/Task Description

The STN laboratory contract involves four broad areas:

1. Supplying each site or state with sample collection media (loaded filter packs, denuders, and absorbent cartridges) and field data documentation forms. RTI ships the collection media to monitoring agencies on a schedule specified by the Delivery Order Project Officer (DOPO).

2. Receiving the samples from the field sites and analyzing the sample media for mass and for an array of chemical constituents including elements (by EDXRF), soluble anions and cations (by ion chromatography), and carbonaceous species (using the Sunset thermal degradation/laser transmittance system). Analysis of semi-volatile organic compounds and examination of particles by electron or optical microscopy have been performed on a very limited basis.
3. Assembling validated sets of data from the analyses, preparing data reports for EPA management and the states, and entering data into the Air Quality System (AQS) data bank 60 days after initial data reports are first submitted to the DOPO and the states.
4. Establishing and applying a comprehensive quality assurance/quality control (QA/QC) system. RTI's Quality Management Plan, QAPP, and associated Standard Operating Procedures (SOPs) provide the documentation for RTI's quality system.

1.3 Major Laboratory Operational Areas

This report addresses the operation of the Sample Handling and Archiving Laboratory (SHAL) and QA/QC for the four major analytical areas active during the time period of January 1 through December 31, 2005. These analytical areas are the: (1) gravimetric determination of particulate mass on Teflon® filters; (2) determination of 48 elements on Teflon® filters using X-ray fluorescence spectrometry; (3) determination of nitrate, sulfate, sodium, ammonium and potassium on nylon or Teflon filters using ion chromatography; and (4) determination of organic carbon, elemental carbon, total carbon, and five other peaks (PK1C, PK2C, PK3C, PK4C, and PyrolC) on quartz filters using thermal optical transmittance. Also addressed is denuder refurbishment, data processing, and QA and data validation.

2.0 Quality Issues and Corrective Actions

2.1 Data Quality

RTI staff perform an extensive array of QA/QC activities to ensure that the data provided to EPA and the States are of the highest quality. Further, RTI makes every effort to provide data that can serve as the basis for making important decisions.

Data quality for the STN has several dimensions, but the primary goal should be usefulness to data users and understanding of the data set's characteristics. There are several metrics that are typically considered in assessing the quality of the STN dataset:

- Accuracy - All analyses standardized to reference values that are traceable to NIST.
- Precision - Measured both as laboratory and whole-system through regular QC replicates and results from samplers collocated at the same site.
- Completeness - Excellent completeness is demonstrated overall, but individual sites may have lower completeness. In addition, the STN has very poor rural site coverage in the western U.S., where IMPROVE sites predominate.
- Comparability - Intercomparison studies recently conducted by EPA have shown good agreement with programs such as the FRM network and IMPROVE results for most of the major chemical species. Other dimensions of comparability include comparability between the four different sampler types currently in use in the STN program: MetOne SASS, Andersen RAAS, URG MASS, and the R&P 2300. In addition, the data are often intercompared with data gathered by three additional sampler types: IMPROVE, PM_{2.5} FRM, and R&P 2025 (used in Texas). All these samplers operate at a variety of different flow rates, use different modes of flow control, and utilize different particle sizing technologies.
- Representativeness - Primary site selection and field sampling operations are out of RTI's control.
- Sensitivity/Detection - The ability to quantify major species such as gravimetric mass, organic carbon, sulfate, nitrate, ammonium, iron, etc. is adequate.

However, many of the trace elements are routinely below limits of detection. Data users should carefully screen out species that are present in such low levels that their inclusion would only add noise to their analysis. Method Detection Limits (MDLs) are provided in Appendix A.

In addition to these data quality assessment criteria, there are other issues that affect data usability. The following quality-related issues and other characteristics of the data set should be taken into account in an overall assessment of the dataset:

- Lack of blank correction - The main concern is the artifact in organic carbon (OC) measurement. The IMPROVE network includes blank correction for OC in its reported data. This is a fundamental difference between the data reported by STN and IMPROVE. Since STN uses four different sampler types, the appropriate OC correction factors should be made readily available to data users.
- Intermittent media contamination issues - Equipment and media contamination issues arise from time to time. RTI makes an effort to flag data, retroactively if necessary, to invalidate or mark as suspicious any affected data items.
- Improvement of uncertainty estimates:
 - Comparability between STN and other networks - RTI is working with U.C. Davis and other experts in XRF to define an acceptable method for determining XRF uncertainty.
 - Realism of total uncertainty estimates based on statistics from sites with side-by-side collocation of samplers. Collocation results provided elsewhere in this report indicate that uncertainties reported to AQS for several major species may be overestimated by a factor of 2x or 3x. These include sulfate, nitrate, and elemental carbon. Average uncertainties currently being reported for the majority of other species appear to be in reasonable agreement with uncertainties calculated from the collocation results.

2.2 Summary of Data Completeness

Data completeness network-wide exceeded 95% for 2005. Both trends and non-trends sites exceeded 95% completeness. Completeness is defined as the number of valid measurement values divided by the potential number of values. Data records with AQS validity status codes ("suspicious" data) are included in the completeness figure, but data records with an AQS null value code are counted as missing data.

Appendix B includes more details of the sampling events and completeness for the Reporting Batches delivered in 2005. Table B.1 shows the total number of sampling events included in each Reporting Batch. Table B.2 provides the total number of records delivered by type. Table B.3 shows the percentage of routine exposure records for each delivery batch group that were valid (i.e., not invalidated with an AIRS Null Value Code) relative to the number of records for scheduled events for that batch for all trends sites. Table B.4 shows the percentage of routine exposure records for each delivery batch group that were valid (i.e., not invalidated with

an AIRS Null Value Code) relative to the number of records for scheduled events for that batch for all non-TRENDS sites. Blank cells indicate that no analyses were scheduled for a site during a particular delivery batch interval. Percentages less than 80 are usually the result of a sample being out of service or one or more exposures being missed because of problems at the site or problems with the shipping.

2.3 Corrective Actions

To ensure ongoing quality work, RTI reacts quickly and decisively to any unacceptable changes in data quality. These reactions are usually in the form of corrective actions. Most of these corrective actions have been in response to very short-term problems such that very few results were impacted negatively. What follows is a description of the major corrective actions taken to ensure the best possible PM_{2.5} data for the EPA and States.

Two formal corrective action requests (CARs) were opened and addressed during 2005.

- CAR 008 - 3/17/05 - Whatman Teflon Filter – Manufacturer’s Debris Problem. In late 2004, the number of filters with a negative net weight seemed to be increasing. A retrospective analysis of the blank data revealed that the rate of filters with net weight changes of -30 micrograms/filter or more increased during the spring of 2004, with significant acceleration in early 2005. Corrective actions were successfully taken, and the frequency of negative net mass outliers has been held to a minimum. Specific actions taken in response to this problem included replacement of approximately 6000 filters from the affected lot, and increased frequency of replicate weighing in the laboratory. See Section 3.0 of this report for further discussion.
- CAR 009 - Shipping error for 12/11/04 Paducah and Perkinstown coolers, and related data corrections. Sample coolers for Perkinstown and Paducah were mixed up and sent to the wrong sites. The site operators noted the error and wrote in the correct site names on the respective forms. RTI attempted to fix the problem by interchanging the Chain of Custody (COC) numbers in the database, but some of the field data was not changed properly the first time. After the data had been reported to AQS, RTI received notification through EPA of a data error from one of the monitoring organizations. As a result, RTI examined the original data sheets and compared them against the reported data and identified the additional changes that needed to be made.

2.3.1 Gravimetric Mass

There were several instances of facilities problems in the gravimetry laboratory during 2005. Problems included fan motors and humidification system components. These were met by actions from RTI's HVAC department and by equipment suppliers. Filter samples were generally not affected, but validity status codes may have been assigned in isolated cases when holding times, temperature, or relative humidity exceeded 2.12 guidelines. Chamber humidity and

temperature sensors were calibrated by the chamber vendor in June 2005 as part of a comprehensive system calibration and preventive maintenance check. This service was obtained to address any existing facilities issues and prevent unplanned shutdown in the future. It should be noted that weighings were not performed when chambers were not working properly.

2.3.2 Elemental Analysis

No significant corrective actions have been taken; however, RTI XRF 1 was upgraded from analog to a digital system in October 2004 and has not been used to analyze PM_{2.5} filters due to instrumental problems.

2.3.3 Ion Analysis

There were no corrective actions taken during this reporting period.

2.3.4 OC/EC Analysis

No significant corrective actions have been taken.

2.3.5 Sample Handling and Archiving Laboratory (SHAL)

Problem: Coolers arriving late at the RTI SHAL laboratory delay the processing and analysis of filters and may even cause a missed sampling event if RTI cannot repack new filters into the modules and ship them to the site in time for the next sampling event. Late arriving coolers are typically due to late returns by the site or delays in transit by the carrier. A summary and graphic of late arriving coolers for the time frame of July 1 to December 31, 2004 is presented below.

Corrective Action: Late arriving coolers are usually caused by delays in the field or by Federal Express. When a shipment is late arriving at RTI, it may not be possible for RTI to ship a set of filter modules as scheduled. When this happens, RTI will notify the EPA DOPO and any missed sampling events are flagged as “scheduled but not collected” (AF).

2.3.6 Data Processing

There were no corrective actions taken during this reporting period.

2.4 Other Quality Issues

Aside from the specific issues discussed in the CARs, there are some ongoing issues that have not been assigned CARs because there was no specific action that RTI could take, or because they required input and cooperation from others outside RTI:

- **Uncertainty harmonization for XRF instruments.** RTI has written a series of reports, presentations, and whitepapers regarding the problem of achieving harmony between the uncertainties reported by the various EDXRF instruments in use by

STN and other PM-fine networks in the U.S., including IMPROVE. Dr. Bill Gutknecht, the RTI laboratory supervisor, has been in contact with experts at U.C. Davis and Alion, Inc. to determine how certain correction factors and the associated uncertainties were estimated historically. A specific recommendation for calculating uncertainties for XRF measurements with PM-fine samples will be forthcoming in early 2006. Lessons learned in this investigation may also be applicable to speciated analysis of PM-coarse.

- **Sampler-dependent background levels for certain elements.** It has been observed since the beginning of the network that certain samplers have a higher incidence of outliers for particular species. These appear to be a function of the materials used in manufacturing the sampler downtubes, denuders, modules and other sampling components. As the analytical laboratory, RTI has no authority to make changes in the sampler technology that was chosen for the network.

3.0 Laboratory Quality Control Summaries

3.1 Gravimetric Laboratory

The Gravimetry Laboratory's two weigh chambers were used to tare 24,676 Teflon filters between January 1 and December 31, 2005. During the same time period, the laboratory performed final ("post-sampling") weighings of 22,374 Teflon filters. This number includes filters weighed for the Hurricane Katrina surveillance monitoring effort. The difference between the number of tared filters and the number of final ("post-sampling") filters is partly due to the inherent lag time between initial and final weighing sessions. Determination of PM_{2.5} mass is based on two separate weighings performed several weeks apart. The total also reflects an increase in the number of filters weighed in the last three months of the year resulting from the laboratory's support to air monitoring conducted in EPA Regions 4 and 6 after Hurricanes Katrina and Rita hit the Gulf Coast. Filter weighing totals given in this report are those recorded by the laboratory's database application.

3.1.1 Quality Issues and Corrective Actions

In March 2005, the Gravimetry Laboratory noted extraneous contaminating debris on Teflon filters purchased for the program. This extraneous debris was considered the likely cause of negative mass blanks resulting from the loss of debris from the filters between tare weighing and final weighing. In most cases, the extraneous debris was very small and matched the filter and support ring in color and texture, making it difficult to see with the naked eye in normal chamber lighting. In response to this issue, the laboratory took several actions. First, RTI replaced all the filters from this lot and received replacement filters from a different lot. Thereafter, the number of filters selected for lot stability tests was increased from six filters (two filters from each of six randomly selected boxes of filters) to 12 filters (two filters from each of six randomly selected boxes of filters) in order to get an even more representative sampling of the filter stock received from the manufacturer. Additional visual inspection of filters was also performed. The frequency of replicate QC weighings was increased from an across-the-board 10% to 100% of tared (pre-sampled) filters and at least 33% of sampled filters.

3.1.2 Description of QC Checks Applied

Internal QC checks applied in the Gravimetry Laboratory are described in Table 3-1, along with results achieved during this reporting period.

**Table 3-1. Summary of QC Checks Applied and Results
Achieved in the Gravimetry Laboratory**

| QC Check | Requirements | QC Checks Applied in RTI Laboratory | Average Value Determined by Lab | Comments |
|--|---|--|--|---|
| Working standard reference weights (mass reference standards) | Verified value \pm 3 μ g [Standard reference weights initially calibrated by Troemner and verified by North Carolina Department of Agriculture and Consumer Services (NCDA&CS) Standards Laboratory] | <u>Chamber 1</u> 100-mg S/N 12936 6/22/04 Class 1 Calibration: 99.95525 mg \pm 0.00082 Laboratory Tolerance Interval: 99.951-99.959 mg | Average = 99.956 mg Std Dev = 0.0021 for 2844 weighings | Laboratory average falls within tolerance interval. |
| | | 200-mg S/N 12935 6/22/04 Class 1 Calibration: 199.99054 mg \pm 0.00079 Laboratory Tolerance Interval: 199.987-199.994 mg | Average = 199.992 mg Std Dev = 0.0024 for 2843 weighings | Laboratory average falls within tolerance interval. |
| | | <u>Chamber 2</u> 100-mg S/N 58096 8/25/04 Class 1 Calibration: 100.00798 mg \pm 0.00082 Laboratory Tolerance Interval: 100.004-100.012 mg | Average = 100.006 mg Std Dev = 0.0017 for 664 weighings | Laboratory average falls within tolerance interval. |
| | | 100-mg S/N RTI01 6/22/04 Class 1 Calibration: 99.99279 mg \pm 0.00082 Laboratory Tolerance Interval: 99.989-99.997 mg | Average = 99.991 mg Std Dev = 0.0010 for 600 weighings | Laboratory average falls within tolerance interval. |
| | | 100-mg S/N 58097 8/25/04 Class 1 Calibration: 100.00254 mg \pm 0.00082 Laboratory Tolerance Interval: 99.999-100.006 mg | Average = 100.001 mg Std Dev = 0.0025 for 2491 weighings | Laboratory average falls within tolerance interval. |
| | | 200-mg S/N 58098 8/25/04 Class 1 Calibration: 200.00972 mg \pm 0.00079 Laboratory Tolerance Interval: 200.006-200.014 mg | Mean = 200.010 mg Std Dev = 0.0036 for 670 weighings | Laboratory average falls within tolerance interval. |

Table 3-1. (Continued)

| QC Check | Requirements | QC Checks Applied in RTI Laboratory | Average Value Determined by Lab | Comments |
|-------------------------------|--|--|---|--|
| | | 200-mg S/N 18659 6/22/04 Class 1 Calibration: 199.97943 mg \pm 0.00079 Laboratory Tolerance Interval: 199.976-199.983 mg | Mean = 199.978 mg Std Dev = 0.0009 for 602 weighings | Laboratory average falls within tolerance interval. |
| | | 200-mg S/N 58099 8/25/04 Class 1 Calibration: 200.00628 mg \pm 0.00079 Laboratory Tolerance Interval: 200.002-200.010 mg | Mean = 200.002 mg Std Dev = 0.0021 for 2489 weighings | Laboratory average falls within tolerance interval. |
| Balance Calibrations | Auto (internal) calibration daily | Daily | N/A | |
| | External calibration annually or as needed | Balance C - S/N 1118252777 (Chamber 1) and Balances D – S/N 1125430571 (replaced Balance A, which was removed from service on June 30, 2005), and B – S/N 1118311244 inspected and externally calibrated by Mettler Toledo on August 22, 2005, using NIST-traceable weight | N/A | Next inspection and external calibration scheduled for August 2006 |
| RH/T Data Logger Calibrations | Annually | Chamber 1 Data Logger S/N 01042219 and Chamber 2 Data Logger S/N 00102174 currently in service calibrated by Dickson Calibration Services February 23, 2005, using NIST-traceable standards | N/A | Chamber 1 Data Logger S/N 03082408 and Chamber 2 Data Logger S/N 03082406 to be calibrated by Dickson Calibration Services February 2006 |

Table 3-1. (Continued)

| QC Check | Requirements | QC Checks Applied in RTI Laboratory | Average Value Determined by Lab | Comments |
|----------------------------|---------------------------------|---|--|--|
| Laboratory (Filter) Blanks | Initial weight \pm 15 μ g | 2650 total replicate weighings of 349 individual laboratory blanks | Average difference between final and initial weight = 4 μ g Std Dev = 0.0048 Min wt change = 0 μ g Max wt change = 38 μ g | 71 total replicate weighings of 15 individual laboratory blanks (2.7% of the replicate weighings; 4.3% of the individual laboratory blanks) exceeded the 15 μ g criterion. Only three replicate weighings of two individual laboratory blanks exhibited excess weight changes in the positive direction. Over 95% of the replicate laboratory blank weighing failures were in the negative direction, indicating probable contamination on the filters when tared that fell off or was removed before later reweighings. |
| Replicates | Initial weight \pm 15 μ g | 20,329 Pre-sampled (Tared) Replicates 8413 Post-sampled Replicates | Average = 6 μ g Average = 4 μ g | Outliers were reweighed in order to confirm value with two weights within 5 μ g of each other. Outliers were reweighed in order to confirm value with two weights within 5 μ g of each other. |

Table 3-1. (Continued)

| QC Check | Requirements | QC Checks Applied in RTI Laboratory | Average Value Determined by Lab | Comments |
|------------------------------------|---|--|---|---------------------------------|
| Lot Blanks (Lot Stability Filters) | 24-hour weight change < $\pm 5 \mu\text{g}$ | Whatman Lot 4279001 - 6 filters weighed (2 randomly selected from each of 3 randomly selected boxes) | 24 hours = $-4 \mu\text{g}$ 48 hours = $0 \mu\text{g}$ 72 hours = $1 \mu\text{g}$ 96 hours = $2 \mu\text{g}$ | Fall well within required range |
| | | Whatman Lot 4341004 - 6 filters weighed (2 randomly selected from each of 3 randomly selected boxes) | 24 hours = $-4 \mu\text{g}$ 48 hours = $-2 \mu\text{g}$ 72 hours = $0 \mu\text{g}$ 96 hours = $0 \mu\text{g}$ | |
| | | Whatman Lot 5103003 Test 1 (Rec'd this lot twice and ran a lot stability test each time) - 12 filters weighed (2 randomly selected from each of 6 randomly selected boxes) | 24 hours = $-2 \mu\text{g}$ 48 hours = $0 \mu\text{g}$ 72 hours = $1 \mu\text{g}$ 96 hours = $0 \mu\text{g}$ | |
| | | Whatman Lot 5103003 Test 2 (Rec'd this lot twice and ran a lot stability test each time) - 12 filters weighed (2 randomly selected from each of 6 randomly selected boxes) | 24 hours = $-2 \mu\text{g}$ 48 hours = $1 \mu\text{g}$ 72 hours = $-1 \mu\text{g}$ 96 hours = $0 \mu\text{g}$ | |
| | | Whatman Lot 5103001 - 12 filters weighed (2 randomly selected from each of 6 randomly selected boxes) | 24 hours = $-2 \mu\text{g}$ 48 hours = $-1 \mu\text{g}$ 72 hours = $1 \mu\text{g}$ 96 hours = $-1 \mu\text{g}$ | |
| | | Whatman Lot 5318001 - 12 filters weighed (2 randomly selected from each of 6 randomly selected boxes) | 24 hours = $0 \mu\text{g}$ 48 hours = $-1 \mu\text{g}$ 72 hours = $1 \mu\text{g}$ 96 hours = $-1 \mu\text{g}$ | |

Table 3-1. (Continued)

| QC Check | Requirements | QC Checks Applied in RTI Laboratory | Average Value Determined by Lab | Comments |
|--|--------------|---|---------------------------------|---|
| Balance Audits <u>Chamber 1</u> Balance C - S/N 118252777 <u>Chamber 2</u> Balance D – S/N 1125430571 (replaced Balance A, which was removed from service on June 30, 2005), and Balance B – S/N 1118311244 | Annually | Performed by RTI Quality Systems Program personnel on November 29, 2005, using Class S-1 NIST-traceable weights | N/A | Included environmental evaluation, level test, scale-clarity test, zero-adjustment test, off-center (corner load) test, precision test, and accuracy test; all balances performed satisfactorily. |

3.1.3 Summary of QC Results

Internal QC values generated by the laboratory usually met the criteria shown in Table 3-1. However, some outliers were noted. Laboratory blank outliers tended to fall below the lower warning limit, indicating that the issue of debris on Teflon filters is one that must be monitored. In the case of outlier replicates, Gravimetry Laboratory analysts reweighed outliers to validate weights.

3.1.4 Assessment of Between-instrument Comparability

Beginning in 2006, the Gravimetry Laboratory will introduce an inter-instrument (intra-laboratory) round-robin program to assess both instrumental and human factors of analytical comparability. A designated filter or filters will be weighed daily by all available staff on all balances. The compilation of this data will allow for an assessment of inter-instrument comparability that has not previously been evaluated in the laboratory. Results will be assessed for evidence of bias between balances and between analysts.

3.1.5 Determination of Uncertainties and Method Detection Limits

The Laboratory's MDL calculations are based on replicate weighings of a large number of filters from filter lot acceptance batches. Because of determination of gravimetric mass requires two separate weighings, each of which contributes to the total uncertainty, MDLs reported to AQS are shown in Appendix A. All balances use the same MDLs. Multiplicative factor of 1.414 is included to account for the fact that each filter must be weighed twice to generate the final net mass.

3.1.6 Audits, Performance Evaluations, Training, and Accreditations

Table 3-2 contains information regarding audits, performance evaluations (PEs), training, and accreditations.

Table 3-2. Description of Audits, PEs, Training, and Accreditations

| Type of Evaluation | Date | Administered By | Significant Findings/Comments |
|--------------------|------------------|--|--|
| Internal Audit | January 13, 2005 | RTI FRM Project QA Officer | <p>No significant deficiency findings were reported by the FRM QAO.</p> <p>Comment: For the first time, the internal audit included an inspection of the chambers with the aid of an ultraviolet light (black light) to highlight dust and debris in the weighing area. Some material was visible, especially behind computers and under racks, with the black light that might not be visible with normal chamber lighting. The auditor recommended that staff periodically use a black light to direct their cleaning efforts.</p> |
| External | July 12, 2005 | EPA/NAREL and OAQPS Audit Team | <p>Two Teflon® filters were removed from the SHAL inventory during the audit so that NAREL could experimentally re-measure the tare mass already determined at RTI's Gravimetry Laboratory. Agreement between NAREL and RTI was excellent (within 1 µg) for one filter, but NAREL's tare mass was 30 µg smaller for the other filter described in Section 3.1.1</p> <p>Comment: It is possible that a small piece of extraneous contaminating debris was attached to the filter for measurements taken at RTI, and somehow the debris was lost from the filter before measurements were made at RTI.</p> |
| Accreditation | | Louisiana Environmental Laboratory Accreditation Program (LELAP) | RTI is accredited for the determination of fine particulates in ambient air by the Federal Reference Method (FRM) for PM _{2.5} . |

3.2 Ions Analysis Laboratory

The Ion Analysis Laboratory used four ion chromatographs to extract and analyze 20,013 cation analyses (sodium, potassium and ammonium); 21,321 nitrate analyses; and 20,072 sulfate analyses performed on the STN program during the period January 1 through December 31, 2005.

3.2.1 Quality Issues and Corrective Actions

There were no quality issues or corrective actions during the reporting period.

3.2.2 Description of QC Checks Applied

Ion chromatographic analyses were performed by personnel from RTI's Environmental Chemistry Department (ECD). Four of our six ion chromatographic systems available were used for performance of the measurements. These are described in Table 3-3. The use of these four systems was determined by the workload.

Table 3-3. Description of Ion Chromatographic Systems Used for Analysis of PM_{2.5} Filter Samples

| System No. | Dionex IC Model | Ions Measured |
|------------|-----------------|-----------------------------------|
| 3 | Model 500 (S3A) | SO ₄ , NO ₃ |
| 4 | DX-600 (D6A) | SO ₄ , NO ₃ |
| 5 | Model 500 (D5C) | Na, NH ₄ , K |
| 6 | DX-600 (D6C) | Na, NH ₄ , K |

QA/QC checks for ion analyses are summarized in Table 3-4. For ion analyses, a daily multipoint calibration (7 points for cations; 8 points for anions) is performed over the range 0.05 to 25.0 ppm for each ion (Na⁺, NH₄⁺, and K⁺ for cation analyses; NO₃⁻ and SO₄²⁻ for anion analyses) followed by QA/QC samples including (1) an RTI-prepared QC sample containing concentrations of each ion in the mid- to high-range of the calibration standard concentrations, (2) an RTI-prepared QC sample containing concentrations of each ion at the lower end of the calibration standard concentrations, and (3) a commercially-prepared, NIST-traceable QA sample containing known concentrations of each ion.

Table 3-4. Ion Analysis of PM_{2.5} – QA/QC Checks

| QA/QC Check | Frequency | Requirements |
|---|---|--|
| Calibration Regression Parameters | Daily | $r \geq 0.999$ |
| Initial QA/QC Checks: | | |
| - RTI-prepared QC sample at mid to high range concentration | Daily, immediately after calibration | Measured concentrations within 10% of known values |
| - RTI-prepared QC sample at lower end concentration | Daily, immediately after calibration | Measured concentrations within 10% of known values |
| - Commercially prepared, NIST traceable QA sample | Daily, immediately after calibration | Measured concentrations within 10% of known values |
| Periodic QA/QC Checks: | | |
| - Replicate sample† | Every 20 samples | RPD = 5% at 100x MDL* RPD = 10% at 10x MDL* RPD = 100% at MDL* |
| - QA/QC sample | Every 20 samples | Measured concentrations within 10% of known values |
| - Matrix spiked sample extract | Every 20 samples | Recoveries within 90 to 100% of target values |
| - Duplicates‡ | At least one per day | No limit set. This data gathered for comparability studies. |
| - Reagent Blanks | One reagent blank per reagent used (DI H ₂ O and/or eluent) sample set extracted | No limit set. This data gathered for comparability studies. |

* MDL = Minimum Detectable Limit

RPD = Relative% Difference

†Replicates indicate a specific sample is run twice on the same instrument.

‡Duplicates indicate a specific sample is run on two different instruments.

The regression parameters (a,b,c and correlation coefficient, r) for the standard curve for each ion are compared with those obtained in the past. Typically, a correlation coefficient of 0.999 or better is obtained for each curve. If the correlation coefficient is <0.999, the analyst carefully examines the individual chromatograms for the calibration standards and reruns any standard that is judged to be out of line with respect to the other standards or to values (peak area and/or height) obtained in the past for the same standard. Possible causes for an invalid standard run include instrumental problems such as incomplete sampling by the autosampler. If necessary, a complete recalibration is performed.

When all individual calibrations have been judged acceptable, the results for the QA/QC samples are carefully examined. If the observed value for any ion being measured differs by more than 10% from the known value, the problem is identified and corrected. Any field samples are then analyzed.

During an analysis run, a replicate sample, a QA/QC sample, and a spiked sample are analyzed at the rate of at least one every 20 field samples. Precision objectives for replicate analyses are $\pm 5\%$ for concentrations that equal or exceed 100 times the minimum detectable limit (MDL), $\pm 10\%$ for concentrations at 10 times the MDL, and $\pm 100\%$ for concentrations at the MDL. MDLs for each instrument and analyte are listed in Table 3-5. The observed value for any ion being measured must be within 10% of the known value for the QA/QC samples given in Table 3-6, and ion recoveries for the spiked samples must be within 90 to 110% of the target value. If these acceptance criteria are not met for any QA/QC or spiked sample, the problem is identified and corrected. All field samples analyzed since the last acceptable check sample are then reanalyzed.

Table 3-5. MDL * for Each Instrument and Analyte

| Instrument | Nitrate | Sulfate | Sodium | Ammonium | Potassium |
|------------|---------|---------|--------|----------|-----------|
| S3A | 0.066 | 0.074 | na | na | na |
| D6A | 0.070 | 0.100 | na | na | na |
| D5C | na | na | 0.290 | 0.160 | 0.134 |
| D6C | na | na | 0.290 | 0.160 | 0.134 |

* In $\mu\text{g}/\text{filter}$

Table 3-6. Definitions and Specifications for QA/QC Samples

| Ion | Sample ID | Description/Specification |
|---------|------------------------|--|
| Anions | QA-CPI_LOW | 0.6 ppm nitrate, 1.2 ppm sulfate |
| | QA-CPI_MED-HI | 3.0 ppm nitrate, 6.0 ppm sulfate |
| | RTI-QC-HIGH | 6.0 ppm nitrate, 12.0 ppm sulfate |
| | RTI-QC-LOW | 0.6 ppm nitrate, 1.2 ppm sulfate |
| | RTI-QC-MED | 1.5 ppm nitrate, 3.0 ppm sulfate |
| Cations | GFS 0.4 PPM QA | 0.4 ppm each sodium, ammonium, and potassium |
| | GFS 4.0 PPM QA | 4.0 ppm each sodium, ammonium, and potassium |
| | RTI 2.0 PPM QC Reg Std | 2.0 ppm each sodium, ammonium, and potassium |
| | RTI 5.0 PPM QC | 5.0 ppm each sodium, ammonium, and potassium |

3.2.3 Summary of QC Results

QC checks performed included:

- Percent recovery for QC samples (standards prepared by RTI)
- Percent recovery for QA samples (commercial standards)
- Relative percent difference (RPD) for replicates
- Spike recovery
- Reagent blank (elution solution and DI water)

Table 3-7 shows recoveries for all five analytes (nitrate, sulfate, sodium, ammonium, and potassium) with low, medium, and high QC (prepared by RTI) samples and with low and medium-high QA samples (commercially prepared and NIST-traceable) for all of the instruments used for analysis.

Table 3-7. Average Percent Recovery for QA and QC Samples

| Analyte | Sample ID | n | Conc. µg/mL | Avg% Rec * | SD | Min | Max |
|-----------|------------------------|-----|-------------|------------|------|--------|--------|
| Nitrate | QA-CPI_LOW | 490 | 0.6 | 98.6% | 1.0% | 0.564 | 0.610 |
| | QA-CPI_MED-HI | 404 | 3.0 | 101.1% | 2.8% | 2.683 | 3.145 |
| | RTI-QC-HIGH | 403 | 6.0 | 101.8% | 0.9% | 5.836 | 6.265 |
| | RTI-QC-LOW | 762 | 0.6 | 98.8% | 1.3% | 0.555 | 0.643 |
| | RTI-QC-MED | 963 | 1.5 | 99.0% | 1.1% | 1.403 | 1.531 |
| Sulfate | QA-CPI_LOW | 490 | 1.2 | 98.7% | 1.0% | 1.111 | 1.224 |
| | QA-CPI_MED-HI | 404 | 6.0 | 101.6% | 2.7% | 5.483 | 6.321 |
| | RTI-QC-HIGH | 403 | 12.0 | 102.2% | 0.9% | 11.708 | 12.905 |
| | RTI-QC-LOW | 762 | 1.2 | 99.7% | 1.3% | 1.109 | 1.280 |
| | RTI-QC-MED | 963 | 3.0 | 100.8% | 1.0% | 2.882 | 3.134 |
| Sodium | GFS 0.4 PPM QA | 607 | 0.4 | 103.1% | 3.1% | 0.386 | 0.518 |
| | GFS 4.0 PPM QA | 757 | 4.0 | 100.3% | 1.1% | 3.735 | 4.165 |
| | RTI 2.0 PPM QC Reg Std | 552 | 2.0 | 100.9% | 1.0% | 1.957 | 2.125 |
| | RTI 5.0 PPM QC | 463 | 5.0 | 100.7% | 0.9% | 4.894 | 5.182 |
| Ammonium | GFS 0.4 PPM QA | 807 | 0.4 | 102.3% | 2.4% | 0.378 | 0.441 |
| | GFS 4.0 PPM QA | 757 | 4.0 | 100.6% | 1.2% | 3.721 | 4.196 |
| | RTI 2.0 PPM QC Reg Std | 552 | 2.0 | 100.6% | 1.2% | 1.941 | 2.116 |
| | RTI 5.0 PPM QC | 463 | 5.0 | 100.8% | 1.0% | 4.853 | 5.240 |
| Potassium | GFS 0.4 PPM QA | 605 | 0.4 | 101.2% | 1.4% | 0.390 | 0.428 |
| | GFS 4.0 PPM QA | 757 | 4.0 | 99.7% | 1.0% | 3.728 | 4.104 |
| | RTI 2.0 PPM QC Reg Std | 551 | 2.0 | 100.9% | 1.0% | 1.950 | 2.099 |
| | RTI 5.0 PPM QC | 462 | 5.0 | 100.5% | 1.1% | 4.751 | 5.188 |

* Acceptance criteria for average percent recovery is $\pm 10\%$.

Average recoveries for the QC samples ranged from 98.8% to 102.2% for the year.
Average recoveries for the QA samples ranged from 98.6% to 103.1% for the year.

Table 3-8 shows percent recovery for all analyte spikes for the year. Average recoveries for the spikes ranged from 100.6% to 102.3%.

Table 3-9 presents filter blank (N BLANK) and reagent blank values for all analytes over the 12 month period. The blank data indicate that the filters supplied to the Sample Handling and Archiving Laboratory (SHAL) were of acceptable cleanliness and that the extraction tubes and the extraction procedure were not introducing contamination.

Table 3-8. Average Percent Recovery for Spikes

| Analyte | Avg Recovery * | StDev | Count | Min | Max |
|-----------|----------------|-------|-------|-------|--------|
| Nitrate | 100.6% | 1.8% | 904 | 88.3% | 108.1% |
| Sulfate | 100.7% | 1.6% | 904 | 90.2% | 106.6% |
| Sodium | 102.3% | 2.4% | 823 | 94.2% | 113.4% |
| Ammonium | 101.3% | 2.9% | 823 | 86.9% | 114.5% |
| Potassium | 101.3% | 2.5% | 823 | 93.9% | 114.7% |

*Acceptance criteria for average% recovery is $\pm 10\%$.

Table 3-9. Filter Blank and Reagent Blank Values (ppm) for all Analytes

| Analyte | Type | n | Avg | StDev | Min | Max |
|-----------|---------|-----|--------|-------|--------|-------|
| Nitrate | N QC * | 135 | 0.009 | 0.012 | 0.000 | 0.040 |
| | REAG ** | 620 | 0.001 | 0.005 | 0.000 | 0.025 |
| Sulfate | N QC | 135 | 0.002 | 0.005 | 0.000 | 0.018 |
| | REAG | 620 | 0.003 | 0.008 | 0.000 | 0.040 |
| Sodium | N QC | 151 | -0.002 | 0.006 | -0.017 | 0.011 |
| | REAG | 439 | -0.001 | 0.004 | -0.034 | 0.015 |
| Ammonium | N QC | 151 | 0.000 | 0.000 | 0.000 | 0.000 |
| | REAG | 439 | 0.000 | 0.000 | 0.000 | 0.009 |
| Potassium | N QC | 151 | 0.000 | 0.000 | 0.000 | 0.000 |
| | REAG | 439 | 0.000 | 0.001 | 0.000 | 0.015 |

* N QC is a blank filter extract analyzed to test the acceptability of the cleaned nylon filter batches. One nylon filter is tested from each bottle used for filter cleaning. If the ion loading for any ion is $> 1 \mu\text{g}$, the filters from that bottle are rejected.

** REAG is a 25-ml aliquot of either deionized water or anion eluent that has been pipetted into an extraction tube and carried through the same extraction procedure as the filters are.

3.2.4 Assessment of Between-instrument Comparability

Anion duplicates were analyzed on instruments D6A and S3A. Cation duplicates were analyzed on instruments D5C and D6C. A comparison of the ranges reported between the two instruments indicates very close results.

Cation duplicates were analyzed on instruments D5C and D6C. A comparison of the ranges reported between the two instruments indicates very close results.

Table 3-10 compares QA and QC samples run on separate instruments on the same day. Each day, both Anion instruments ran at least two QC and three QA samples. Similarly, Cation instruments ran at least two QC and two QA samples on each instrument each day. This table shows that the difference between the two instruments using the same QA or QC sample are very small. The calculated average difference and standard deviation indicate a high level of between-instrument comparability.

Table 3-10. Between-Instrument Comparability

| Analyte | QA/QC Type | Conc., µg/mL | n | Average * Difference | Standard Deviation of Diff. | Minimum Diff. | Maximum Diff. |
|-----------|----------------|-----------------|-----|-------------------------|-----------------------------------|------------------|------------------|
| Nitrate | QA-CPI_LOW | 0.6 | 235 | 0.000 | 0.004 | -0.015 | 0.012 |
| | QA-CPI_MED-HI | 3.0 | 158 | -0.004 | 0.023 | -0.104 | 0.061 |
| | RTI-QC-HIGH | 6.0 | 158 | -0.005 | 0.041 | -0.232 | 0.128 |
| | RTI-QC-LOW | 0.6 | 584 | 0.000 | 0.006 | -0.024 | 0.035 |
| | RTI-QC-MED | 1.5 | 928 | 0.000 | 0.012 | -0.070 | 0.060 |
| Sulfate | QA-CPI_LOW | 1.2 | 235 | 0.003 | 0.012 | -0.078 | 0.038 |
| | QA-CPI_MED-HI | 6.0 | 158 | -0.022 | 0.052 | -0.220 | 0.170 |
| | RTI-QC-HIGH | 12.0 | 158 | -0.023 | 0.139 | -0.514 | 0.883 |
| | RTI-QC-LOW | 1.2 | 584 | 0.001 | 0.018 | -0.098 | 0.071 |
| | RTI-QC-MED | 3.0 | 928 | 0.000 | 0.029 | -0.149 | 0.135 |
| Sodium | GFS 0.4 PPM QA | 0.4 | 701 | 0.008 | 0.019 | -0.079 | 0.113 |
| | GFS 4.0 PPM QA | 4.0 | 573 | -0.002 | 0.032 | -0.116 | 0.147 |
| | RTI 2.0 PPM QC | 2.0 | 306 | 0.010 | 0.021 | -0.065 | 0.159 |
| | RTI 5.0 PPM QC | 5.0 | 218 | 0.007 | 0.045 | -0.156 | 0.129 |
| Ammonium | GFS 0.4 PPM QA | 0.4 | 701 | 0.009 | 0.014 | -0.025 | 0.038 |
| | GFS 4.0 PPM QA | 4.0 | 573 | 0.001 | 0.043 | -0.102 | 0.178 |
| | RTI 2.0 PPM QC | 2.0 | 306 | 0.022 | 0.021 | -0.066 | 0.086 |
| | RTI 5.0 PPM QC | 5.0 | 218 | 0.010 | 0.065 | -0.209 | 0.283 |
| Potassium | GFS 0.4 PPM QA | 0.4 | 701 | 0.003 | 0.007 | -0.015 | 0.027 |
| | GFS 4.0 PPM QA | 4.0 | 573 | 0.013 | 0.032 | -0.145 | 0.119 |
| | RTI 2.0 PPM QC | 2.0 | 306 | 0.005 | 0.016 | -0.050 | 0.061 |
| | RTI 5.0 PPM QC | 5.0 | 218 | 0.030 | 0.048 | -0.218 | 0.168 |

* Differences are calculated as Concentration of D6A – Concentration of S3A for Anions and Concentration of D5C – Concentration of D6C for Cations.

3.2.5 Determination of Uncertainties and MDLs

Detection limits are determined by analyzing the lowest calibration standard 7 times and the detection limit, in µg/mL (or ppm), is calculated as 3 times the standard deviation of the 7 measurements. This detection limit is multiplied by 25 mL to determine the detection limits in µg/filter, which is the extraction volume for each filter. These calculations are performed for each instrument so that the detection limits are reported by instrument. Since most samples are not analyzed in replicate, analytical uncertainties must be estimated based on historical data and scientific judgment. A simple formula of the form $U = a \cdot C + b$ is used where U is the uncertainty and C is the concentration. The coefficients a and b vary by instrument and by analyte. The b coefficient is essentially MDL/3. The value for a is assumed to be 0.05 (5%). MDLs for the STN Program are summarized in Appendix A.

3.2.6 Audits, PEs, Training, and Accreditations

No deficiencies were found in an audit of the ion analysis laboratory performed by the U.S. Environmental Protection Agency (EPA) audit team from the National Air and Radiation Environmental Laboratory (NAREL) on July 12, 2005 (Appendix E). PE samples analyzed as a part of the audit were in good agreement with the NAREL expected values. All staff in the ion analysis laboratory have been fully trained in the extraction and analysis procedures used in the PM_{2.5} project. No additional training was needed this year.

3.3 Organic Carbon/Elemental Carbon Laboratory

The RTI OC/EC Laboratory analyzed 19,617 quartz filter samples by the STN method during the period January 1 through December 31, 2005, and reported the results of those analyses to the main STN database. Four Sunset Laboratory Carbon Aerosol Analyzers (designated by the letters R, S, T, and F) were used for STN analyses. The F analyzer was switched to other OC/EC analysis projects on March 11, 2005; while the remaining three analyzers were used for STN throughout 2005.

3.3.1 Quality Issues and Corrective Actions

No issues that affected the quality of reported data arose during the reporting period.

3.3.2 Description of QC Checks Applied

QC checks, acceptance criteria, and responses for the OC/EC Laboratory are summarized in Table 3-11.

Table 3-12 contains a list of all data flags assigned to carbon analysis data and the number of filter analysis results assigned each flag in the OC/EC Laboratory during the reporting period. Only flags assigned in OC/EC Laboratory data reports to RTI's Speciation Program Information Management System (SPIMS) are included in the table. The SHAL or the QA Officer may have assigned additional flags to the quartz filter samples based on field data or additional data validation checks.

Table 3-11. OC/EC Laboratory QC Checks, Acceptance Criteria, and Corrective Actions

| QC Element | Frequency | Acceptance Criteria | Response When Outside Criteria |
|-------------------------|---|---|--|
| Method Detection Limit | After oven replacement or annually, whichever comes first | $MDL \leq 0.5 \mu\text{g C/cm}^2$ | Investigate the source of the problem and initiate corrective action, if necessary, to correct the problem before analyzing samples. |
| Calibration Peak Area | Every analysis | Within 95% to 105% of average calibration peak area for that day | Discard the results of that analysis and, if necessary, repeat the analysis with a second punch from the same filter. |
| Instrument Blank | Daily and after about 30 samples | (1) Blank $\leq 0.3 \mu\text{g/cm}^2$, and (2) calibration peak area 90% to 110% of average for the weekly three-point calibration. | Determine if the problem is with the filter or the instrument, and, if necessary, initiate corrective action to identify and solve any instrument problem, and run an acceptable instrument blank before analyzing samples. |
| Three-Point Calibration | Weekly | (1) Correlation Coefficient (R^2) ≥ 0.998 [with force-fit through 0,0], (2) 93% to 107% recovery for all three standards, and (3) FID response factor is 90% to 110% of the average response factor for all three standards. | Determine the cause of the nonlinearity, and initiate actions that will identify and solve any problem that may have arisen. Then repeat the three-point calibration, which must yield satisfactory results before samples are analyzed. |
| Calibration Check | Daily | (1) 93% to 107% recovery, (2) calibration peak area 90% to 110% of average for the weekly three-point calibration, and (3) FID response factor is 90% to 110% of average response factor for last three-point calibration. | Initiate corrective action, if necessary, to solve the problem before analyzing samples. |
| Duplicate Analyses | 10% of all samples | (1) TC Values greater than $10 \mu\text{g C/cm}^2$ -- Less than 10% RPD, (2) TC Values $5 - 10 \mu\text{g C/cm}^2$ -- Less than 15% RPD, (3) TC Values less than $5 \mu\text{g C/cm}^2$ -- Within $\pm 0.75 \mu\text{g C/cm}^2$. | Flag analysis results for that filter with non-uniform filter deposit (LFU) flag. |

Table 3-12. OC/EC Laboratory-Assigned Data Flags

| Flag | Description | Number of Filters |
|---|---|--------------------------|
| LFA | Filter inspection flag - filter wet (Punch was dried in analyzer for 20 min, then analyzed as usual.) | 1 |
| LFU | Filter inspection flag - non-uniformity (Duplicate analysis failed applicable duplicate criterion.) | 37 |
| LFW | Filter inspection flag - sampled on wrong side of filter | 1 |
| LLI | ANALYSIS INVALID - Other (Filter was broken, and none of the pieces were large enough to get a punch for a valid analysis.) | 1 |
| Total Number of Analyses Flagged by Analysts | | 40 |
| Total Number of OC/EC Analyses Reported | | 19,934 |
| Percent of OC/EC Analyses Flagged by Analysts | | 0.201% |

3.3.3 Summary of QC Results

3.3.3.1 Instrument Blanks

Table 3-13 contains the number of instrument blanks run during the reporting period and the average, minimum, and maximum measured blank values for each of the four carbon aerosol analyzers used in the program. For all reported data, the last instrument blank run before reported samples were analyzed met the blank criterion for TC.

Table 3-13. OC/EC Instrument Blank Statistics

| Blank Statistic | OC/EC Analyzer | | | |
|---|-----------------------|-------------------|------------------|-------------------|
| | Retrofit(R) | Second (S) | Third (T) | Fourth (F) |
| Number of Instrument Blanks | 392 | 416 | 380 | 74 |
| Mean Response ($\mu\text{g C/cm}^2$) | 0.018 | 0.031 | 0.024 | 0.029 |
| Standard Deviation | 0.017 | 0.028 | 0.025 | 0.028 |
| Minimum Response ($\mu\text{g C/cm}^2$) | 0.000 | 0.000 | 0.000 | -0.014 |
| Maximum Response ($\mu\text{g C/cm}^2$) | 0.159 | 0.227 | 0.140 | 0.200 |

3.3.3.2 Calibrations

Table 3-14 provides summary statistics for full 3-point calibrations by analyzer. In addition to number of 3-point calibrations run, the table includes average, minimum, and maximum values for slope and linearity (expressed as correlation coefficient, R^2) for the calibrations and for the three percentages used as QC checks on analysis results for each individual calibration standard. The three percentages separately calculated for the low-, mid-, and high-level calibration standards include:

1. FID response to the internal standard (expressed as a percentage of the average FID response to the internal standard for the 3-point calibration),
2. Recovery (mass of carbon measured expressed as a percentage of the mass of carbon in the spiked volume of standard used), and
3. FID response factor (expressed as a percentage of the average FID response factor for the 3-point calibration).

Table 3-15 provides summary statistics for daily calibration checks by analyzer. The table gives the number of calibration checks run on each analyzer and the average, minimum, and maximum values of the three percentages used as QC checks to determine if a calibration check is acceptable. The three percentages used to evaluate the validity of each calibration check analysis include:

1. Internal standard area (as a percentage of the average internal standard area for the last 3-point calibration),
2. Recovery (mass of carbon measured expressed as a percentage of the mass of carbon in the spiked volume of standard used), and
3. FID response factor (as a percentage of the average response factor for the last 3-point calibration).

A calibration check is acceptable only if it meets all three criteria.

3.3.3.3 Duplicate Analyses

Table 3-16 gives summary statistics for all duplicate STN OC/EC analyses run on all analyzers during the reporting period. A duplicate analysis was run on the same analyzer on about every tenth filter. A total of 2,306 duplicate STN analyses were run under the laboratory support contract in 2005. OC/EC analysis results for 37 of those duplicates failed the applicable duplicate criterion and were flagged as coming from a filter with a non-uniform deposit.

Table 3-14. OC/EC Three-Point Calibration Statistics

| Variable/Statistic | | | OC/EC Analyzer | | | |
|--|------------|---------|----------------|---------|---------|---------|
| | | | R | S | T | F |
| Number of Full Calibrations Passing All Criteria | | | 50 | 52 | 48 | 10 |
| Number of Full Calibrations Failing Any Criterion | | | 0 | 0 | 0 | 0 |
| Slope (counts/μgC), forced through origin (0,0) | Average | Average | 8,584 | 5,027 | 5,898 | 10,171 |
| | | Minimum | 8,033 | 4,756 | 4,866 | 9,798 |
| | | Maximum | 9,037 | 5,395 | 6,565 | 10,474 |
| Correlation Coefficient (R ²) (Criterion: ≥0.998) | Average | Average | 0.9997 | 0.9997 | 0.9998 | 0.9995 |
| | | Minimum | 0.9987 | 0.9989 | 0.9987 | 0.9982 |
| | | Maximum | 1.0000 | 1.0000 | 1.0000 | 1.0000 |
| FID Response to Internal Standard as a percent of Average Internal Standard FID Response for 3-Point Cal (Criterion: 90% to 110%) | Low Cal | Average | 100.08% | 100.26% | 100.13% | 100.19% |
| | | Minimum | 99.34% | 99.07% | 98.86% | 98.54% |
| | | Maximum | 101.59% | 102.51% | 101.91% | 102.06% |
| | Mid Cal | Average | 99.92% | 99.96% | 99.96% | 100.52% |
| | | Minimum | 99.10% | 98.06% | 98.20% | 98.42% |
| | | Maximum | 100.37% | 102.12% | 101.28% | 102.24% |
| | High Cal | Average | 100.00% | 99.79% | 99.91% | 99.29% |
| | | Minimum | 99.28% | 97.43% | 98.55% | 96.91% |
| | | Maximum | 101.29% | 101.80% | 101.09% | 100.49% |
| Recovery: Mass of Carbon Measured as a percent of Mass of Carbon Spiked (Criterion: 93% to 107%) | Low Cal | Average | 101.59% | 101.43% | 101.18% | 102.52% |
| | | Minimum | 96.40% | 95.45% | 96.79% | 97.93% |
| | | Maximum | 104.87% | 104.78% | 105.00% | 104.81% |
| | Mid Cal | Average | 99.82% | 99.81% | 99.61% | 99.12% |
| | | Minimum | 96.03% | 97.01% | 97.40% | 97.29% |
| | | Maximum | 101.64% | 102.55% | 102.31% | 100.84% |
| | High Cal | Average | 98.45% | 98.75% | 99.19% | 98.37% |
| | | Minimum | 96.40% | 95.47% | 96.00% | 96.75% |
| | | Maximum | 101.98% | 102.39% | 101.79% | 102.14% |
| | All 3 Cals | Average | 99.95% | 100.00% | 99.99% | 100.00% |
| | | Minimum | 97.74% | 99.99% | 99.60% | 99.98% |
| | | Maximum | 100.02% | 100.01% | 100.02% | 100.02% |
| FID Response Factor as a percent of Average FID Response Factor for 3-Point Cal (Criterion: 90% to 110%) | Low Cal | Average | 101.72% | 101.69% | 101.32% | 102.71% |
| | | Minimum | 96.18% | 95.60% | 96.90% | 96.68% |
| | | Maximum | 104.74% | 106.89% | 106.95% | 105.43% |
| | Mid Cal | Average | 101.34% | 101.25% | 100.49% | 102.03% |
| | | Minimum | 96.34% | 97.67% | 94.84% | 98.09% |
| | | Maximum | 105.03% | 104.75% | 102.71% | 105.94% |
| | High Cal | Average | 98.49% | 98.55% | 99.11% | 97.67% |
| | | Minimum | 96.16% | 95.72% | 96.30% | 95.88% |
| | | Maximum | 102.12% | 102.37% | 102.95% | 102.64% |

Table 3-15. OC/EC Daily Calibration Check Statistics

| Variable/Statistic | | R | S | T | F |
|---|---------|---------|---------|---------|---------|
| Number of Cal Checks Passing All Criteria | | 238 | 248 | 224 | 46 |
| Number of Cal Checks Failing Any Criterion | | 0 | 0 | 0 | 0 |
| Internal Standard (IS) Area as a percent of Average IS Area for 3-Point Cal (Criterion: 90% to 110%) | Average | 99.98% | 99.61% | 99.75% | 100.51% |
| | Minimum | 91.89% | 91.06% | 91.20% | 90.00% |
| | Maximum | 105.30% | 106.30% | 106.74% | 105.95% |
| Recovery: Mass of Carbon Measured as a percent of Mass of Carbon Spiked (Criterion: 95% to 105%) | Average | 100.74% | 100.47% | 100.81% | 100.77% |
| | Minimum | 95.11% | 95.26% | 95.00% | 95.14% |
| | Maximum | 104.92% | 104.98% | 104.99% | 104.90% |
| FID Response Factor as a percent of Average Response Factor for 3-Point Cal (Criterion: 90% to 110%) | Average | 100.73% | 100.07% | 100.56% | 101.26% |
| | Minimum | 92.46% | 91.10% | 90.09% | 93.63% |
| | Maximum | 107.89% | 107.30% | 109.71% | 108.29% |

Table 3-16. Duplicate OC/EC Analysis Statistics

| Variable/Statistic | | Analyzer | | | |
|---|----------------|----------|--------|-------|--------|
| | | R | S | T | F |
| Total Number of Duplicate Analyses | | 756 | 776 | 704 | 124 |
| Number of Analyses Flagged as Failing Duplicate Criteria | | 10 | 13 | 12 | 2 |
| Percentage of Duplicate Analyses Failing Duplicate Criteria | | 1.32% | 1.68% | 1.70% | 1.61% |
| OC Sample/Dup Plot | Slope | 0.993 | 1.003 | 1.000 | 0.965 |
| | Intercept | 0.036 | 0.002 | 0.023 | 0.074 |
| | R ² | 0.997 | 0.997 | 0.997 | 0.996 |
| EC Sample/Dup Plot | Slope | 1.019 | 0.998 | 0.995 | 0.992 |
| | Intercept | -0.007 | 0.006 | 0.005 | 0.006 |
| | R ² | 0.993 | 0.992 | 0.993 | 0.995 |
| TC Sample/Dup Plot | Slope | 0.997 | 1.005 | 0.998 | 0.971 |
| | Intercept | 0.027 | -0.001 | 0.031 | 0.075 |
| | R ² | 0.998 | 0.997 | 0.998 | 0.997 |
| Pk1C Sample/Dup Plot | Slope | 0.993 | 0.973 | 0.993 | 0.954 |
| | Intercept | 0.009 | 0.024 | 0.007 | 0.035 |
| | R ² | 0.997 | 0.993 | 0.996 | 0.996 |
| Pk2C Sample/Dup Plot | Slope | 0.998 | 0.992 | 0.992 | 0.962 |
| | Intercept | 0.006 | 0.012 | 0.017 | 0.025 |
| | R ² | 0.990 | 0.987 | 0.989 | 0.984 |
| Pk3C Sample/Dup Plot | Slope | 1.006 | 0.985 | 0.992 | 0.940 |
| | Intercept | 0.003 | 0.010 | 0.012 | 0.024 |
| | R ² | 0.991 | 0.989 | 0.992 | 0.976 |
| Pk4C Sample/Dup Plot | Slope | 1.009 | 0.993 | 1.000 | 0.966 |
| | Intercept | -0.002 | 0.005 | 0.008 | 0.010 |
| | R ² | 0.995 | 0.993 | 0.993 | 0.989 |
| PyrolC Sample/Dup Plot | Slope | 0.921 | 1.068 | 0.931 | 1.061 |
| | Intercept | -0.001 | 0.001 | 0.000 | -0.001 |
| | R ² | 0.948 | 0.953 | 0.961 | 0.987 |

3.3.4 Assessment of Between-Instrument Comparability

While duplicate analysis results (two punches from the same filter run on the same analyzer) agree fairly well, replicate analysis results (two or more punches from the same filter run on different analyzers) for the OC Peaks do not always agree as well, especially for Pk3 C, Pk4 C and Pyrol C. The level of oxygen contamination present in the analyzer ovens during the non-oxidizing heat ramps seems to be the primary cause of the differences in OC Peak measurements between analyzers.¹ Whether the oxygen comes from diffusion through seals inside the analyzer or from some type of carry-over from the preceding analysis is not known.

Trace amounts of contaminating oxygen cause some of the carbon in thermally unstable organic species to be evolved rather than forming char during the non-oxidizing heating ramps. This early evolution of organic carbon reduces the amount of organic char formed and shifts the OC/EC split time to an earlier time in the analysis. However, the presence of oxygen does not significantly change the OC:EC mass ratio. The bad news is that the presence of oxygen shifts the evolution of OC from the later OC Peaks (especially Pyrol C) to the earlier OC Peaks.

To assess between-analyzer comparability of OC, EC, TC, and the individual OC Peaks, RTI's OC/EC Laboratory analyzed 127 filters by the STN/TOT method on three Sunset Laboratory Carbon Aerosol Analyzers over a two-year period. The results of those analyses were used to estimate uncertainties (presented below) that take into account samples collected during all seasons of the year and analyzed on different analyzers during all of the stages of oxygen contamination as analyzer ovens age and are replaced.² An F-Test analysis of the peak data indicated that the three analyzers did not give equivalent results for all analytes, but the agreement was within the same general uncertainty as the long-standing Sunset Lab-determined uncertainties for OC, EC, and TC.

RTI is continuing to run replicate analyses across all OC/EC analyzers used for the program and continuing to run duplicate analyses on the same analyzer for about 10% of samples. Replicate analysis data will be used to refine uncertainty estimates, and duplicate analysis data will continue to alert sampling personnel and data users of possible filter deposit non-uniformity issues.

3.3.5 Determination of Uncertainties and MDLs

Table 3-17 gives estimated uncertainties for OC, EC, TC, and OC Peaks measured on multiple analyzers in RTI's OC/EC Laboratory. STN/TOT OC/EC analysis results for 127 filters analyzed on three Sunset Laboratory Carbon Aerosol Analyzers over a two-year period were used to determine the estimated absolute and relative uncertainties for all reported carbon fractions.

¹The helium supply line for each RTI OC/EC analyzer is fitted with two oxygen traps: a high-capacity trap followed by an indicating trap. Only ultra-high purity (UHP) helium is used for OC/EC analysis. All OC/EC analyzers, regardless of manufacturer or model, have this problem.

²Because of the large number of samples analyzed, each STN analyzer in RTI's OC/EC Laboratory requires two to four oven replacements each year.

Table 3-17. Estimated Uncertainties for OC/EC Carbon Fractions

| Fraction | "Best Fit" Uncertainty (µgC/cm ²) |
|----------|---|
| OC | $\pm(0.20 + 0.05*OC)$ |
| EC | $\pm(0.20 + 0.05*EC)$ |
| TC | $\pm(0.30 + 0.05*TC)$ |
| Pk1 C | $\pm(0.20 + 0.05*Pk1\ C)$ |
| Pk2 C | $\pm(0.20 + 0.05*Pk2\ C)$ |
| Pk3 C | $\pm(0.30 + 0.05*Pk3\ C)$ |
| Pk4 C | $\pm(0.30 + 0.10*Pk4\ C)$ |
| Pyrol C | $\pm(0.20 + 1.40*Pyrol\ C)$ |

The estimate for each carbon fraction was obtained by determining the standard deviation(s) of the three measurements from different analyzers for each of the 127 filters in the study. A plot of average measured loading with 1-sigma error bars for each filter (y-axis) vs. averaged measured loading (x-axis) was generated for each carbon fraction. Straight trend lines of the form $y = \pm(b + mx)$, where b is absolute uncertainty, m is relative uncertainty, and x is measured loading, were then generated for points above and below the ends of the 1s error bars with values for b and m that gave trend lines that bracketed the 1s error bars for all filters except a few (1 to 3) that appeared to be outliers.³

From the table, it is obvious that Pyrol C has by far the largest relative uncertainty. Pyrol C is a measure of the pyrolyzed organic carbon remaining on the filter punch after oxygen is added at the end of the four non-oxidizing heating ramps. If the sample contains little pyrolyzable organic carbon, the trace amounts of contaminating oxygen may prevent the formation of any Pyrol C. If the sample contains sufficient pyrolyzable OC to exceed the reaction capacity of the trace amounts of contaminating oxygen, then at least some Pyrol C will be measured. Because the trace amounts of contaminating oxygen differ slightly between analyzers, the distribution of OC among the OC Peaks differs more between analyzers than it does within duplicates run on the same analyzer. Because PyrolC is formed primarily during the evolution of Pk3 C and Pk4 C, these last-evolved OC Peaks typically have the largest between-analyzer variability and, therefore, larger measurement uncertainties.

³Peterson, M.R., and M.H. Richards. 2006. *Estimation of Uncertainties for Organic Carbon Peaks Data in Thermal-Optical-Transmittance Analysis of PM_{2.5} by the Speciation Trends Network Method*. To be presented at the A&WMA Symposium on Air Quality Measurement Methods and Technology, May 9-11, 2006, Durham, NC.

Table 3-18 gives target MDLs for all reported carbon fractions. MDL values for the five OC Peaks (Pk1-Pk4 and Pyrol C) were taken from the absolute uncertainties in Table 3-18. This same approach was used to determine reasonable target MDLs for OC, EC, and TC, all of which have proven to be attainable when an analyzer is functioning properly and all operating conditions are under control.

Table 3-18. Target MDLs for OC/EC Carbon Fractions

| Fraction | Target MDL ($\mu\text{gC}/\text{cm}^2$) |
|-----------------|---|
| OC | 0.20 |
| EC | 0.20 |
| TC | 0.30 |
| Pk1 C | 0.20 |
| Pk2 C | 0.20 |
| Pk3 C | 0.30 |
| Pk4 C | 0.30 |
| Pyrol C | 0.20 |

3.3.6 Audits, PEs, Training, and Accreditations

3.3.6.1 Audits

Finding: RTI's OC/EC Laboratory was audited by EPA/NAREL's Jewell Smiley on July 17, 2005. The Technical Memorandum describing the audit and findings contained the following quote at the end of OC/EC Laboratory write-up:

The general impressions of the OC/EC laboratory developed during this audit were very positive. Only one concern was noted. Some of the routine duplicate determinations should be scheduled to collect between-instrument precision data.⁴

Response: RTI's OC/EC Laboratory has conducted evaluations of both within- and between-analyzer variability since it added a second OC/EC analyzer in the first year of the STN laboratory support program. In September 2005, RTI began analyzing replicate punches from randomly chosen filters on all STN/TOT analyzers on a regular basis.

⁴Smiley, Jewell, et al. 2005. Technical Memorandum--RTI Laboratory Audit on July 12, 2005. Memorandum dated November 5, 2005. Available on-line at <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/rti0705.pdf>.

Under the first Chemical Speciation of PM_{2.5} Laboratory Support contract, only OC, EC, and TC were reported, and Sunset Laboratory Inc. had already determined between-analyzer uncertainties for those three fractions. The validity of Sunset Laboratory's uncertainties for OC, EC, and TC have been confirmed in all RTI replicate analysis studies, including the current large study reported in Section 3.3.5. Sufficient data has now been collected to allow estimation of uncertainties for OC Peaks measurements that include between-analyzer uncertainty, which is substantially larger than within-analyzer uncertainty for the reasons cited in Section 3.3.5 above.

Discussion: Two punches from the same filter analyzed on the same analyzer (which we refer to as a duplicate analysis) can be used to evaluate the uniformity of the filter deposit. Frequent non-uniformity flags can alert field sampling personnel and data users to potential problems with a specific sampler. Two or more punches from the same filter analyzed on different analyzers (which we refer to as replicate analyses) can be used to evaluate between-analyzer variability. This data can be used to estimate uncertainties for measured carbon fractions across analyzers. In other words, duplicates and replicates serve two different purposes, both of which are very important.

To best meet the two purposes described in the preceding paragraph, RTI's OC/EC Laboratory has added regular analysis of replicates (same filter, different analyzers) to its list of QC samples without decreasing the number of duplicate analyses (same filter, same analyzer) already being performed.

Until 2004, replicate studies had consisted of periodic analysis of replicates of small numbers of filters (usually about 20) on all available analyzers over a period of a few days. These studies provided snapshots of how well analysis results agreed or disagreed across analyzers at a given time, but they were not thought to be adequate to provide estimates of uncertainties, especially for the OC Peaks, over the life of the speciation program because of changes in the distribution of OC among the OC Peaks that occur as analyzer ovens age. In 2004, 47 filters were analyzed on all RTI STN/TOT analyzers over a 1-month period. Results of that study were presented at a conference in April 2004.⁵

Currently, RTI is running replicate analyses of about four to eight filters per week on all STN analyzers used in the program. Replicate analysis data for 127 filters analyzed in 2004 and 2005 have already been evaluated to determine reasonable uncertainties and MDLs for the OC Peaks and to confirm that the uncertainties and MDLs used since the beginning of the program for OC, EC, and TC are still acceptable. Estimated uncertainties and MDLs based on the combined data are presented in the tables in Section 3.3.5 above. Future replicate analysis data will be added to this large data set to determine if uncertainties and MDLs change significantly over time.

⁵Peterson, M.R., M.H. Richards, J.L. Pritt, and C.M. Haas. 2004. *Reproducibility of Organic Carbon Peaks Data in Thermal-Optical-Transmittance Analysis of PM_{2.5} by the Speciation Trends Network Method*. Presented at the A&WMA Symposium on Air Quality Measurement Methods and Technology, Research Triangle Park, NC, April 2004.

3.3.6.2 Performance Evaluations

RTI's OC/EC Laboratory was one of four laboratories participating in the February 2005 EPA/NAREL interlaboratory comparison study. RTI's STN/TOT data, including the OC Peaks data, compared very favorably with EPA/NAREL's STN/TOT data, as shown in Appendix D.

3.3.6.3 Training

One new analyst was hired and trained during the reporting period. He has a BS in chemistry-biochemistry from a local university and came to RTI with previous chromatography and other laboratory experience. He went through intensive training in the operation of RTI's OC/EC analyzers and easily passed the analyst validation test given at the end of the training. He has been the second-shift analyst since July 2005.

3.3.6.4 Accreditations

There are no accreditation programs for OC/EC analysis.

3.4 X-ray Fluorescence Laboratories

The two X-ray fluorescence (XRF) laboratories, RTI and Chester LabNet, used two XRF instruments each to analyze a total of 19,317 filters for 48 elements during the period January 1 through December 31, 2005.

3.4.1 Chester LabNet X-Ray Fluorescence Laboratory

Chester LabNet was the original XRF contractor laboratory used for the STN program. During the period covered by this report, Chester operated two Kevex XRF instruments designated 770 and 771.

3.4.1.1 Quality Issues and Instrument Repair and Maintenance

The following repairs and maintenance were performed for XRF-770:

- 1/29/05 - replaced X-ray tube and power supply and recalibrated
- 2/8/05 - realigned X-ray tube and recalibrated
- 3/23/05 - recalibrated due to increase in excitation energy in condition 0
- 6/29/05 - replaced X-ray tube and recalibrated
- 7/19/05 - recalibrated due to decrease in excitation energy in condition 0

The following repairs and maintenance were performed for XRF-771:

- 9/12/05 - recalibrated due to decrease in excitation energy in condition 1
- 10/21/05 - replaced X-ray tube and recalibrated

3.4.1.2 Description of QC Checks Applied

QC activities for the analysis of elements by EDXRF for the RTI XRF laboratory, their frequency of application and control limits, comments and corrective actions are shown in Table 3-19.

Table 3-19. QC Procedures Performed in Support of XRF Elemental Analysis

| QC Check | QC Frequency | Control Limits | Comments/Corrective Action |
|---------------------------------------|----------------------|----------------------|----------------------------|
| Calibration | As needed | $\pm 5\%$ | Calibration |
| Calibration verification ¹ | Once per week | ± 2 sigma | Recalibrate |
| Instrument precision ² | Per 10 to 15 samples | $\pm 10\%$ | Re-analyze |
| Excitation condition check | Per 10 to 15 samples | $\pm 10\%$ | Re-analyze |
| Sample replicate precision | Per 10 samples | RPD < 2x uncertainty | Re-analyze if necessary |

1 - Using NIST SRMs

2 – Micromatter QC

3.4.1.3 Summary of QC Results

Precision

Precision was monitored by the reproducibility of the multi-element Micromatter QC sample. The QC sample has six selected elements and is analyzed with each tray of samples. The comparison of the element's values gives the measure of reproducibility or precision. The data used to monitor precision are presented in Tables 3-20A and 3-20B, for the 770 and 771 instruments, respectively. The percent coefficient of variation (%CV) for the average of all data for each of the six elements ranged between 1.82 and 4.64% for the 770 and between 2.04 and 3.05% for the 771. The Ti and Cd outliers shown for minimum values in the XRF 771 table were experienced prior to the analysis of samples. The QS standard results post-analysis were well within acceptance range, as were the “averaged” values between pre and post analytical values.

**Table 3-20A. Summary of Chester XRF 770 Laboratory QC Precision Data
1/1/2005 through 10/4/2005**

| Element | n | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|-----|------|-------|---------|---------|------|----------------|
| Si | 240 | 91.1 | 110.6 | 101.9 | 3.03 | 2.97 | 0.77 |
| Ti | 240 | 93.9 | 103.5 | 99.2 | 1.86 | 1.88 | -0.37 |
| Fe | 240 | 94.5 | 103.5 | 99.3 | 1.81 | 1.82 | 0.18 |
| Cd | 240 | 91.0 | 107.3 | 100.0 | 3.79 | 3.80 | -6.6 |
| Se | 240 | 90.3 | 107.8 | 98.7 | 4.58 | 4.64 | -9.49 |
| Pb | 240 | 90.4 | 106.9 | 99.2 | 4.37 | 4.40 | -8.2 |

**Table 3-20B. Summary of Chester XRF 771 Laboratory QC Precision Data
1/1/2005 through 10/4/2005**

| Element | n | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|-----|------|-------|---------|---------|------|----------------|
| Si | 434 | 91.2 | 105.9 | 97.1 | 2.38 | 2.45 | -0.80 |
| Ti | 434 | 85.6 | 106.0 | 96.1 | 2.79 | 2.90 | 2.56 |
| Fe | 434 | 90.4 | 103.2 | 96.8 | 1.97 | 2.04 | 0.55 |
| Cd | 434 | 89.8 | 107.7 | 98.0 | 3.11 | 3.05 | -0.29 |
| Se | 434 | 91.0 | 108.2 | 99.2 | 2.82 | 2.84 | 1.42 |
| Pb | 434 | 91.4 | 108.5 | 99.3 | 2.54 | 2.56 | 1.10 |

Accuracy

Accuracy determinations are performed with three NIST thin film SRMs, four vapor deposited Micromatter standards, and one NIST particle size standard. Recovery is calculated by dividing the measured result by the expected value. Tables 3-21A and 3-21B show recovery for 12 elements spanning the atomic mass range of the 48 elements normally measured. The min and max recovery values for all the elements ranged between 90.2 and 110.3% for the 770 and between 88.0 and 112.6% for the 771. Averages over the reporting period were within the recovery goal of twice the standard deviation for both instruments; however individual measurements were sometimes outside this criterion. Corrective actions were taken whenever a recovery was outside specifications as follows:

- If one of the elements in Tables 3-21A and 3-21B fell outside of the 2-sigma limit, a single re-analysis of the standard was performed in that excitation condition. If re-analysis resulted in failure, then recalibration of that excitation condition was necessary.
- If recalibration demonstrated that the log of the inverse of the new calibration factor (log sensitivity) –vs- atomic number (Z) for the “failed element” did not conform to a smoothly varying curve defined by the log of the sensitivity factors – vs- atomic numbers for the remaining elements, then the calibration factor was “forced” to fit the curve, with the resulting calibration factor yielding “less than optimum” recovery values.

Table 3-21A. Recovery Determined from Analysis of NIST SRMs 1832, 1833, and 2708 for Chester XRF 770 -- 1/1/2005 through 12/31/2005

| Element | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|-------|-------|---------|---------|------|----------------|
| Al | 90.2 | 106.1 | 98.6 | 4.25 | 4.31 | 3.57 |
| Si | 96.8 | 108.1 | 103.3 | 3.07 | 2.97 | 1.13 |
| Si | 93.3 | 104.5 | 99.9 | 2.70 | 2.71 | 3.14 |
| S | 92.2 | 103.4 | 97.2 | 3.13 | 3.22 | 1.17 |
| K | 94.0 | 106.7 | 100.4 | 2.85 | 2.84 | 3.29 |
| Ca | 95.1 | 106.8 | 100.2 | 2.82 | 2.82 | 2.30 |
| Ti | 94.6 | 110.3 | 100.8 | 3.80 | 3.77 | 1.68 |
| V | 93.6 | 103.6 | 98.7 | 2.38 | 2.41 | 1.86 |
| Mn | 91.9 | 110.4 | 100.2 | 4.81 | 4.80 | -7.45 |
| Fe | 96.5 | 105.6 | 101.0 | 2.55 | 2.53 | -1.64 |
| Cu | 93.7 | 109.3 | 102.3 | 3.90 | 3.81 | 0.99 |
| Zn | 94.1 | 106.0 | 100.1 | 3.21 | 3.20 | 0.99 |
| Pb | 100.0 | 108.0 | 103.8 | 1.65 | 1.59 | -0.91 |

Table 3-21B. Recovery Determined from Analysis of NIST SRMs 1832, 1833 and 2708 for Chester XRF 771 -- 1/1/2005 through 12/31/2005

| Element | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|------|-------|---------|---------|------|----------------|
| Al | 88.0 | 101.8 | 97.2 | 3.90 | 4.01 | 7.67 |
| Si | 93.5 | 104.0 | 99.2 | 2.73 | 2.76 | 6.17 |
| Si | 91.6 | 101.6 | 95.4 | 2.51 | 2.65 | 8.10 |
| S | 93.9 | 106.1 | 99.2 | 2.70 | 2.73 | -1.90 |
| K | 89.2 | 109.0 | 100.8 | 4.61 | 4.57 | 2.11 |
| Ca | 95.4 | 112.6 | 106.2 | 5.14 | 4.84 | 12.1 |
| Ti | 88.8 | 102.5 | 96.1 | 3.77 | 3.92 | 9.67 |
| V | 94.1 | 106.9 | 101.2 | 3.48 | 3.44 | 4.38 |
| Mn | 95.6 | 105.7 | 101.2 | 2.26 | 2.24 | -1.86 |
| Fe | 92.9 | 102.2 | 98.6 | 2.58 | 2.62 | 3.61 |
| Cu | 93.9 | 105.8 | 99.4 | 2.36 | 2.37 | 0.29 |
| Zn | 97.5 | 107.2 | 101.7 | 2.42 | 2.38 | 0.88 |
| Pb | 93.6 | 107.0 | 100.8 | 2.81 | 2.79 | 4.34 |

Reproducibility

Replicate analysis of field samples are used to assess reproducibility of the analytical system. Replicates were analyzed at a frequency of 5% of the filters analyzed. Six elements were selected for comparison through regression analysis. Table 3-22 shows the correlation coefficient and average RPDs for the replicate analysis. The correlation coefficients for the 770 range from 0.9987 to 0.9997, and the correlation coefficients for the 771 range from 0.9992 to 0.9999, indicating acceptable replication on both instruments.

Table 3-22. Replicate Data for Chester XRF 770 and 771

| KeveX 770 | | | | KeveX 771 | | | |
|------------------|----------|--------------------------------|--------------------|------------------|----------|--------------------------------|--------------------|
| Element | n | Correlation Coefficient | Average RPD | Element | n | Correlation Coefficient | Average RPD |
| Si | 173 | .9990 | 1.44 | Si | 141 | .9953 | -3.41 |
| S | 173 | .9992 | 0.08 | S | 141 | .9987 | -3.15 |
| K | 173 | .9987 | -0.14 | K | 141 | .9948 | -3.34 |
| Ca | 173 | .9996 | -0.14 | Ca | 141 | .9991 | -2.71 |
| Fe | 173 | .9997 | -0.05 | Fe | 141 | .9990 | -0.51 |
| Zn | 173 | .9996 | -0.42 | Zn | 141 | .9982 | -0.91 |

There are times when the distribution of a certain species across the filter is not uniform, and will not produce tight precision. This is important information for those who intend to use the data. It is Chester's position that re-analysis of particle deposits on filters received from the field represents the degree of confidence the client may expect more accurately than precision calculated from the uniformly distributed deposits from the Micromatter QC standard.

Failure of individual replicate analysis results to fall with 2x uncertainty can fall into several categories:

- The wrong sample can be re-analyzed, which is easily deduced and easily corrected by re-analyzing the correct sample.
- If one element in a sample lies outside the 2-sigma range, especially a volatile species such as Cl which can be an order of magnitude lower on subsequent analysis due to the low pressure atmosphere in the analysis chamber, no action is taken. However, if several elements in one excitation condition lie outside action levels, while other species in different excitation conditions demonstrate good precision, then the spectra for the excitation condition in question are examined for anomalies, and re-analysis of that excitation condition is performed.

3.4.1.4 Assessment of Between-instrument Comparability

For XRF, inter-instrument comparability is assessed by a round-robin filter exchange program coordinated by the RTI XRF laboratory. See Section 3.4.2.4 for comparative performance of both laboratories.

In addition, Chester has 125 samples which were analyzed by both XRF 770 and XRF 772, which will be used to gain EPA approval for use of the 772 on the STN program. Since the inception of the PM_{2.5} Speciation project, Chester has performed numerous comparisons between instruments via replicate analysis of a number of clients, but much of this data is proprietary and cannot be shared in this report.

3.4.1.5 Uncertainties and MDLs

The methods for determining uncertainties and MDLs are described in SOPs XR-002.02 and XR-006.01. MDLs were determined for the 770 and 771 instruments on 12/26/05, and are shown in Table 3-23. MDLs used during 2005 across analyzers are shown in Appendix A.

3.4.1.6 Audits, PEs, Training, and Accreditations

Chester LabNet has not received any audit visits from EPA on the STN program since the beginning of the speciation project, and would welcome any PE samples or other oversight, which the EPA might deem appropriate. No new laboratory personnel were trained during 2005, but plans for 2006 include the training of two additional analysts.

Another Chester client provides quarterly PE samples in the form of Micromatter vapor deposited standards for elements: Cr, Cu, Zn, Ga, As, Se, Cd, Te, and Pb. However, these PE samples were analyzed using instrument XRF 772, which is not currently approved for use on the STN program. These results will be provided to EPA in the revised instrument acceptance report, to be provided during the first quarter of 2006.

3.4.2 RTI International XRF Laboratory

3.4.2.1 Quality Issues and Instrument Maintenance and Repairs

No changes were made in the analytical procedures used by the RTI XRF Laboratory. However, during 2005 XRF 1 was serviced requiring a new tube and detector, which required instrument re-calibration. XRF 2 was serviced requiring a new tube, which required instrument calibration verification. Also, in October 2005 the Micromatter QC for XRF 1 was replaced with a new Micromatter QC sample. The new sample includes the same elements as the old QC sample.

3.4.2.2 Description of QC Checks Applied

QC activities for the analysis of elements by EDXRF for the RTI XRF Laboratory, their frequency of application and control limits, comments and corrective actions are shown in Table 3-24.

Table 3-23. Three-sigma MDLs^a for Chester 770 and 771 Instruments

| | 770 | 771 |
|----|-------|-------|
| Na | 0.381 | 2.124 |
| Mg | 0.174 | 0.624 |
| Al | 0.099 | 0.240 |
| Si | 0.081 | 0.126 |
| P | 0.054 | 0.090 |
| S | 0.039 | 0.072 |
| Cl | 0.108 | 0.147 |
| K | 0.039 | 0.069 |
| Ca | 0.045 | 0.048 |
| Sc | 0.030 | 0.033 |
| Ti | 0.036 | 0.030 |
| V | 0.018 | 0.021 |
| Cr | 0.018 | 0.021 |
| Mn | 0.027 | 0.024 |
| Fe | 0.033 | 0.021 |
| Co | 0.018 | 0.015 |
| Ni | 0.018 | 0.015 |
| Cu | 0.018 | 0.018 |
| Zn | 0.018 | 0.015 |
| Ga | 0.045 | 0.051 |
| Ge | 0.030 | 0.030 |
| As | 0.027 | 0.027 |
| Se | 0.024 | 0.024 |
| Br | 0.021 | 0.021 |
| Rb | 0.024 | 0.021 |
| Sr | 0.030 | 0.027 |
| Y | 0.036 | 0.033 |

| | 770 | 771 |
|----|-------|-------|
| Zr | 0.045 | 0.039 |
| Nb | 0.054 | 0.051 |
| Mo | 0.063 | 0.066 |
| Rh | 0.123 | 0.084 |
| Pd | 0.126 | 0.081 |
| Ag | 0.129 | 0.090 |
| Cd | 0.129 | 0.099 |
| In | 0.141 | 0.111 |
| Sn | 0.159 | 0.129 |
| Sb | 0.180 | 0.153 |
| Te | 0.204 | 0.189 |
| I | 0.261 | 0.243 |
| Cs | 0.443 | 0.400 |
| Ba | 0.573 | 0.531 |
| La | 0.677 | 0.705 |
| Ce | 0.838 | 0.970 |
| Sm | 0.096 | 0.063 |
| Eu | 0.108 | 0.060 |
| Tb | 0.096 | 0.063 |
| Hf | 0.081 | 0.111 |
| Ta | 0.078 | 0.180 |
| W | 0.078 | 0.120 |
| Ir | 0.075 | 0.075 |
| Au | 0.078 | 0.069 |
| Hg | 0.072 | 0.045 |
| Pb | 0.060 | 0.060 |

Notes:

a - MDLs were converted to a 3-sigma basis from the 1-sigma MDLs reported by Chester.

Table 3-24. QC Procedures Performed in RTI XRF Elemental Analysis Laboratory

| QC Check | QC Frequency | Control Limits | Comments/Corrective Action |
|---------------------------------------|--|------------------|--|
| Calibration | as needed | ----- | ----- |
| Calibration verification ¹ | weekly | 90-110% recovery | check calibration |
| Instrument precision ² | analyzed with each tray of samples (10 tray autosampler) | within 5% CV | check calibration and reanalysis of tray |
| Energy calibration | daily | ----- | ----- |
| Sample replicate precision | 5% | +/- 50 RPD | reanalysis |

1 - Using NIST SRMs

2 - Micromatter QC

3.4.2.3 Summary of QC Results

Precision was monitored by the reproducibility of the measurements of the multi-element Micromatter QC sample. The QC sample has six selected elements and is analyzed with each tray of samples. The comparison of the element's values gives the measure of reproducibility or precision. The data used to monitor precision are presented in 3-25A through 3-25C. The percent coefficient of variation (%CV) for the average of all data for each of the six elements ranged between 0.22 and 4.51% for XRF 1 and between 0.22 and 3.24% for XRF 2. Note that XRF 1 Micromatter QC was replaced with a new Micromatter QC in October 2005 and the slope percent per year calculation is calculated by the days each Micromatter QC was in use on XRF 1.

**Table 3-25A. Summary of RTI XRF 1 Laboratory QC Precision Data, ug/cm²
1/1/2005 through 10/4/2005**

| Element | n | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|-----|------|------|---------|---------|------|-------------------|
| Si | 777 | 9.00 | 9.76 | 9.26 | 0.04 | 0.42 | 0.3 |
| Ti | 777 | 8.14 | 8.27 | 8.22 | 0.02 | 0.22 | 0.2 |
| Fe | 777 | 10.1 | 10.3 | 10.2 | 0.04 | 0.37 | -0.3 |
| Cd | 777 | 5.66 | 5.81 | 5.74 | 0.03 | 0.44 | -0.1 |
| Se | 777 | 4.00 | 4.11 | 4.06 | 0.02 | 0.59 | -1.3 |
| Pb | 777 | 11.2 | 11.6 | 11.4 | 0.06 | 0.53 | -0.8 |

**Table 3-25B. Summary of RTI XRF 1 Laboratory QC Precision Data, ug/cm²
10/5/2005 through 12/31/2005**

| Element | n | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|-----|------|------|---------|---------|------|-------------------|
| Si | 135 | 4.40 | 4.54 | 4.47 | 0.04 | 0.84 | 0.8 |
| Ti | 135 | 5.70 | 6.30 | 6.06 | 0.27 | 4.51 | -4.8 |
| Fe | 135 | 6.50 | 6.74 | 6.64 | 0.09 | 1.43 | -1.5 |
| Cd | 135 | 5.41 | 5.68 | 5.52 | 0.09 | 1.58 | 1.6 |
| Se | 135 | 3.99 | 4.06 | 4.02 | 0.02 | 0.39 | 0.0 |
| Pb | 135 | 9.70 | 10.1 | 9.84 | 0.15 | 1.57 | 1.7 |

**Table 3-25C. Summary of RTI XRF 2 Laboratory QC Precision Data, ug/cm²
1/1/2005 through 12/31/2005**

| Element | n | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|------|------|------|---------|---------|------|-------------------|
| Si | 1203 | 5.19 | 5.39 | 5.30 | 0.06 | 1.10 | -0.2 |
| Ti | 1203 | 6.70 | 7.30 | 7.06 | 0.23 | 3.24 | -0.6 |
| Fe | 1203 | 6.95 | 7.09 | 7.03 | 0.02 | 0.25 | 0.0 |
| Cd | 1203 | 5.94 | 6.11 | 6.03 | 0.03 | 0.41 | 0.0 |
| Se | 1203 | 4.19 | 4.39 | 4.25 | 0.02 | 0.56 | -0.1 |
| Pb | 1203 | 8.98 | 9.13 | 9.03 | 0.02 | 0.22 | 0.0 |

n = number of observations

Min = minimum value observed

Max = maximum value observed

Std Dev = standard deviation

%CV = percent coefficient variation ((Std Dev/Average)*100)

Recovery or system accuracy was determined by the analysis of a series of NIST Standard Reference Materials (SRM) filters. Recovery is calculated by comparisons of measured and expected values. Tables 3-26A and 3-26B show recovery for 12 elements spanning the atomic mass range of the 48 elements normally measured. The recovery values for all the elements ranged between 90 and 108% for XRF 1 and between 91 and 110% for XRF 2. Note that in August 2004, NIST SRM 1833 developed a tear in the filter and was replaced with NIST SRM 2783. Even though SRM 2783 has additional analytes that were not included in SRM 1833, in being consistent with reporting, only the analytes included in SRM 1833 were reported.

Table 3-26A. Recovery Determined from Analysis of NIST SRMs 1832 and 2783 for RTI XRF 1 -- 1/1/2005 through 12/31/2005

| Element | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|-------|-------|---------|---------|------|----------------------------------|
| Al | 90.8 | 94.3 | 92.7 | 0.90 | 0.97 | -0.4 |
| Si* | 89.6 | 90.7 | 90.1 | 0.25 | 0.33 | -0.3 |
| Si** | 93.6 | 100.0 | 96.8 | 2.37 | 2.44 | -6.8 (-1.8 and 0.3) ^a |
| Co | 97.7 | 99.5 | 98.1 | 0.45 | 0.46 | 0.4 |
| K | 94.1 | 100.5 | 97.7 | 1.79 | 1.83 | -4.9 |
| Ca | 96.7 | 97.3 | 97.0 | 0.15 | 0.15 | -0.2 |
| Ti | 96.0 | 104.5 | 101.3 | 2.20 | 2.18 | -4.9 |
| V | 103.4 | 104.5 | 103.9 | 0.27 | 0.26 | 0.1 |
| Mn | 97.4 | 98.5 | 97.8 | 0.23 | 0.23 | -0.2 |
| Fe | 95.0 | 99.3 | 97.0 | 1.64 | 1.69 | -4.7 |
| Cu | 96.5 | 97.5 | 97.1 | 0.25 | 0.26 | 0.2 |
| Zn | 100.9 | 108.5 | 103.9 | 1.92 | 1.85 | -0.3 |
| Pb | 89.6 | 108.8 | 97.8 | 4.36 | 4.43 | -2.8 |

* - SRM 1832

** - SRM 2783

a – Even though the recoveries were within the control limits of 90-110%, the two values represent the slope before a slight shift in the recoveries and after the shift, which took place in 2005. The shift in recovery is evident on both XRF systems, which could indicate a possible degradation of silicon in the SRM sample.

Table 3-26B. Recovery Determined from Analysis of NIST SRMs 1832 and 2783 for RTI XRF 2 -- 1/1/2005 through 12/31/2005

| Element | Min | Max | Average | Std Dev | %CV | Slope (%/year) |
|---------|-------|-------|---------|---------|------|----------------------------------|
| Al | 94.6 | 96.5 | 95.4 | 0.39 | 0.41 | 0.2 |
| Si* | 90.5 | 91.7 | 91.0 | 0.23 | 0.25 | 0.3 |
| Si** | 92.9 | 98.7 | 95.1 | 2.23 | 2.35 | -6.4 (-1.5 and 2.5) ^a |
| Co | 98.1 | 101.3 | 100.0 | 0.61 | 0.61 | 0.4 |
| K | 91.3 | 97.0 | 94.1 | 1.49 | 1.58 | 0.8 |
| Ca | 97.9 | 99.5 | 99.0 | 0.45 | 0.46 | 1.1 |
| Ti | 92.6 | 106.0 | 99.3 | 2.62 | 2.63 | -3.2 |
| V | 99.8 | 102.5 | 101.0 | 0.62 | 0.61 | 1.5 |
| Mn | 100.4 | 101.5 | 100.9 | 0.32 | 0.32 | 0.6 |
| Fe | 94.3 | 98.1 | 96.6 | 0.98 | 1.02 | -1.7 |
| Cu | 96.7 | 98.2 | 97.6 | 0.34 | 0.34 | 0.6 |
| Zn | 97.8 | 105.8 | 102.2 | 1.95 | 1.91 | 3.9 |
| Pb | 95.9 | 110.1 | 100.9 | 3.17 | 3.13 | -2.5 |

* - SRM 1832

** - SRM 2783

a – Even though the recoveries were within the control limits of 90-110%, the two values represent the slope before a slight shift in the recoveries and after the shift, which took place in 2005. The shift in recovery is evident on both XRF systems, which could indicate a possible degradation of silicon in the SRM sample.

Replicates were analyzed at a frequency of 5% of the filters analyzed in the RTI XRF Laboratory. Six elements were selected for comparison through regression analysis. Table 3-27 shows the correlation coefficient and average RPDs for the replicate analysis. The correlation coefficients for XRF 1 range from 0.9984 to 0.9999, and the correlation coefficients for XRF 2 range from 0.9992 to 0.9999, indicating acceptable replication on both instruments. Also, for the six elements the average RPD on XRF 1 was less than seven and the average RPD for the six elements on XRF 2 was less than four.

Table 3-27. Replicates

| XRF 1 | | | | XRF 2 | | | |
|---------|-----|-------------------------|-------------|---------|-----|-------------------------|-------------|
| Element | n | Correlation Coefficient | Average RPD | Element | n | Correlation Coefficient | Average RPD |
| Si | 432 | 0.9984 | 6.49 | Si | 532 | 0.9999 | 2.90 |
| S | 432 | 0.9999 | 1.08 | S | 532 | 0.9999 | 1.27 |
| K | 432 | 0.9996 | 2.62 | K | 532 | 0.9992 | 2.56 |
| Ca | 432 | 0.9998 | 2.66 | Ca | 532 | 0.9998 | 3.32 |
| Fe | 432 | 0.9999 | 1.15 | Fe | 532 | 0.9998 | 1.92 |
| Zn | 432 | 0.9994 | 3.79 | Zn | 532 | 0.9993 | 3.73 |

3.4.2.4 Assessment of Between-Instrument Comparability

Overview of Round-Robin Samples Run During 2005

In addition to passing internal QC samples as described in the sections above, the RTI and Chester Laboratories participate in a "round-robin" filter program coordinated by the RTI XRF Laboratory. It should be emphasized that the round-robin program is only used to collect descriptive statistics about network performance; the results are not currently being used for QC purposes. The lag time between successive analyses and the potential for filter contamination and damage in transit make it impractical to use these filters for laboratory QC.

In the round-robin program, previously analyzed STN filters are recycled through all the instruments in the two laboratories. Table 3-28 summarizes the number of round robin filters analyzed during 2005.

Table 3-28. Counts of Round Robin Filter Analyses During 2005

| Laboratory | Instrument | Filters |
|---|------------|---------|
| Chester Labnet | Kevex 770 | 32 |
| Chester Labnet | Kevex 771 | 38 |
| RTI | XRF 1 | 57 |
| RTI | XRF 2 | 43 |
| Total filters common to all instruments: | | 11 |

The majority of elements reported by XRF are present in quantities at or below the detection capabilities of the instruments; therefore, it was necessary to restrict the statistical analysis of the round robin results to 15 elements that were found in sufficient quantity on a majority of the filters. A total of 11 round robin filters were analyzed by all the STN instruments during 2005. Many additional filters were analyzed by one or more of the instruments, but only 11 had been analyzed by all four instruments exclusively during 2005. The statistics to follow in this section are restricted to filters that were analyzed by all four instruments during 2005.

Assessment of Bias and Precision

The primary purpose of the round robin program is to assess bias between instruments for the various elements. Inter-laboratory precision, a component of overall network error, can also be estimated based on these statistics.

One simple way to assess potential differences in performance of the different instruments is to perform linear regression in which the individual observations for each instrument are regressed against a reference value. Tables 3-29A and 3-29B show linear regression results when the data for the 11 filters are regressed vs. the median for the four instruments for each filter. The median value is used as the reference value, since the "true"

value is unknown for these filters. Each instrument in the program reported zeros or low level detections in some of the elements (especially Ni, Cu, and Se), which can affect the calculation for slope or the correlation coefficient. The calculated uncertainty of these results for each instrument was not taken in account in doing the regression (i.e., no weighting factors were used).

Table 3-29A. Regression Results for 15 Elements - RTI XRF Instruments

| Element | RTI #1 | | | | RTI #2 | | | |
|---------|--------|-------------------------|--------|-----------|--------|-------------------------|--------|-----------|
| | n | Correlation Coefficient | Slope | Intercept | n | Correlation Coefficient | Slope | Intercept |
| Si | 11 | 0.9985 | 0.9916 | -0.158 | 11 | 0.9976 | 0.8703 | 0.004 |
| S | 11 | 0.9943 | 1.0067 | 0.128 | 11 | 0.9992 | 1.0166 | 0.037 |
| K | 11 | 0.9984 | 1.0145 | -0.019 | 11 | 0.9982 | 1.0070 | -0.038 |
| Ca | 11 | 0.9958 | 1.0940 | -0.069 | 11 | 0.9996 | 1.0410 | -0.020 |
| Mn | 11 | 0.9988 | 1.0245 | 0.003 | 11 | 0.9968 | 0.9675 | -0.004 |
| Fe | 11 | 0.9996 | 1.1140 | -0.004 | 11 | 0.9997 | 1.0310 | -0.014 |
| Ni | 11 | 0.9927 | 1.1333 | -0.000 | 11 | 0.9753 | 0.7671 | -0.000 |
| Cu | 11 | 0.9136 | 1.1280 | -0.017 | 11 | 0.9958 | 1.0750 | 0.001 |
| Zn | 11 | 0.9900 | 1.1030 | 0.007 | 11 | 0.9990 | 0.9902 | 0.003 |
| Se | 11 | 0.9914 | 0.7542 | -0.003 | 11 | 0.9905 | 1.3210 | -0.004 |
| Pb | 11 | 0.9970 | 1.1820 | -0.008 | 11 | 0.9980 | 1.0340 | -0.008 |

Note: Units for intercept are µg/filter; correlation coefficient and slope are dimensionless.

Table 3-29B. Regression Results for 15 Elements - Chester XRF Instruments

| Element | Chester 770 | | | | Chester 771 | | | |
|---------|-------------|-------------------------|--------|-----------|-------------|-------------------------|--------|-----------|
| | n | Correlation Coefficient | Slope | Intercept | n | Correlation Coefficient | Slope | Intercept |
| Si | 11 | 0.9990 | 1.0500 | 0.047 | 11 | 0.9988 | 0.9940 | 0.217 |
| S | 11 | 0.9998 | 0.9905 | 0.003 | 11 | 0.9930 | 0.9474 | 0.031 |
| K | 11 | 0.9981 | 1.0765 | -0.015 | 11 | 0.9972 | 0.9797 | 0.011 |
| Ca | 11 | 0.9974 | 0.9183 | 0.024 | 11 | 0.9998 | 0.9540 | 0.032 |
| Mn | 11 | 0.9995 | 1.1020 | -0.013 | 11 | 0.9993 | 0.9639 | 0.010 |
| Fe | 11 | 0.9996 | 0.9671 | 0.004 | 11 | 0.9985 | 0.9900 | -0.031 |
| Ni | 11 | 0.9767 | 0.9656 | -0.001 | 11 | 0.9745 | 0.9727 | 0.006 |
| Cu | 11 | 0.9907 | 0.9831 | 0.005 | 11 | 0.9902 | 0.9924 | -0.004 |
| Zn | 11 | 0.9952 | 1.0260 | -0.011 | 11 | 0.9971 | 0.9724 | -0.006 |
| Se | 11 | 0.9725 | 1.0107 | 0.002 | 11 | 0.9142 | 0.8440 | 0.012 |
| Pb | 11 | 0.9981 | 0.9814 | -0.010 | 11 | 0.9976 | 0.9316 | 0.018 |

Note: Units for intercept are µg/filter; correlation coefficient and slope are dimensionless.

Comparison of Reported Uncertainties for Round-Robin Filters

The harmonization of uncertainty calculations for PM-fine networks in the U.S., including STN and IMPROVE, is currently a matter of discussion between RTI, EPA, and other experts in XRF analysis. RTI staff have prepared a number of reports and presentations on the subject over the past several years. The 2005 round robin data show that the uncertainty values reported by Chester's and RTI's instruments differ as expected based on our research into the different uncertainty calculation methodologies used by the respective instruments.

Our assessment of the algorithms used to calculate uncertainty has uncovered the following significant areas of difference affecting the comparability of Chester vs. RTI uncertainty values:

- Chester's software appears to overestimate the attenuation uncertainty for light elements (Na - Ca). This overestimate can be up to 12% of the total concentration. In the case of sulfur, this is an overestimate of a factor of 3x or 4x.
- RTI's ThermoNoran instrument does not include an attenuation correction at all. This will result in an underestimate of the uncertainty for the same group of light elements.
- RTI's ThermoNoran instrument does not include an estimated laboratory calibration error term. Chester's software includes a factor of 5% of concentration for this term.

Table 3-30 illustrates these characteristics for the 15 elements reported here. Note that these are laboratory uncertainties reported by the instruments. RTI adds an additional 5% of concentration to account for field handling and sample volume prior to reporting the uncertainties to AQS. Only the results for the 11 filters analyzed on all four instruments were used to make Table 3-30.

Table 3-30. Average Laboratory Uncertainties

| Element | Chester 770 | Chester 771 | RTI #1 | RTI #2 |
|---------|------------------------------|------------------------------|------------------------------|------------------------------|
| | Average Relative Uncertainty | Average Relative Uncertainty | Average Relative Uncertainty | Average Relative Uncertainty |
| Si | 11.6% | 11.8% | 5.10% | 2.00% |
| S | 11.3% | 11.3% | 1.00% | 1.20% |
| K | 11.5% | 11.8% | 2.30% | 3.30% |
| Ca | 11.7% | 11.7% | 2.50% | 3.40% |
| Mn | 12.4% | 10.8% | 5.00% | 7.30% |
| Fe | 6.10% | 6.20% | 1.00% | 1.40% |
| Ni | 60.3% | 32.4% | 29.7% | 40.7% |
| Cu | 15.4% | 18.7% | 15.9% | 11.7% |
| Zn | 8.10% | 7.00% | 3.30% | 4.10% |
| Se | 44.8% | 44.8% | 57.5% | 25.6% |
| Pb | 18.8% | 20.9% | 9.80% | 10.4% |

3.4.2.5 Determination of Uncertainties and MDLs

MDLs are determined by obtaining data from the analysis of ten laboratory blanks. The MDLs are calculated as three times the average uncertainty for each element. The MDLs from XRF 1 and XRF 2 are presented in Table 3-31. Network-wide MDLs are summarized in Appendix A.

Uncertainties for each analytical result is automatically calculated by the ThermoNoran software, except for when the concentration value is zero, the software cannot calculate an uncertainty. To obtain an uncertainty value for when the concentration is zero, the following formula is used:

$$\text{Uncertainty} = \text{slope} * A * \text{sqrt}(3 * \text{sqrt}(B * t) + B * t) / t$$

Where:

t = livetime

A = scaling factor

B = background counts (cps) is incorporated during the importing of the data into the RTI XRF database.

3.4.2.6 Audits, PEs, Training, and Accreditations

In February 2005, RTI XRF laboratory received six 47mm Teflon filters from NAREL. These six samples were prepared by NAREL and were part of a PE study. During the onsite visit from NAREL in July 2005, results of the PE study showed good agreement among the participating labs as reported in the Multi-Lab Speciation PE Report by NAREL (Appendix D).

Results of the EPA systems audit conducted on 7/12/05 and reported 11/4/05 (Appendix E) included the following comments related to XRF analysis (page 12 of 13):

... the focus of the XRF audit was to discuss those samples that RTI had analyzed as part of a recent inter-laboratory comparison study sponsored by NAREL [see reference 2]. Results from this study showed aluminum to be the most controversial element reported. This study also showed that RTI generally reported uncertainties which were lower than those reported by the other participating labs. A few spectra were inspected and discussed during the audit. Two specific spectra were selected to be included in the final report for the study. Ultimately the final report included examples of the controversial spectra from all of the labs. The spectra from RTI contain a significant [diffusion peak] interference for aluminum and silicon which was not observed in the spectra from the other labs.

Comment: This observation may not be a problem for RTI's analysis since there is no standard method for calculating XRF uncertainties. However, RTI may want to take a closer look at the way uncertainties were calculated for aluminum and silicon during this study. EPA has recently initiated dialog with all of the speciation labs to learn more about the XRF analysis at each lab, and clearly there is diversity among the different labs. Any progress toward standardizing the XRF analysis is a positive step for the speciation program.

Table 3-31. MDL Values for XRF 1 and XRF 2

| Element | XRF 1 | XRF 2 | Element | XRF 1 | XRF 2 |
|---------|---------|---------|---------|---------|---------|
| Na | 1.29159 | 0.38759 | Sr | 0.02147 | 0.02373 |
| Mg | 0.15481 | 0.11639 | Y | 0.02147 | 0.02599 |
| Al | 0.34917 | 0.1356 | Zr | 0.0226 | 0.03277 |
| Si | 0.17967 | 0.1017 | Nb | 0.03842 | 0.02825 |
| P | 0.0565 | 0.17176 | Mo | 0.03277 | 0.02938 |
| S | 0.10283 | 0.11978 | Ag | 0.08023 | 0.12656 |
| Cl | 0.21018 | 0.08588 | Cd | 0.08927 | 0.1356 |
| K | 0.05085 | 0.07458 | In | 0.15594 | 0.1582 |
| Ca | 0.05537 | 0.08249 | Sn | 0.20453 | 0.20114 |
| Sc | 0.10848 | 0.11978 | Sb | 0.27007 | 0.36273 |
| Ti | 0.04859 | 0.05424 | Cs | 0.27911 | 0.11074 |
| V | 0.0339 | 0.03955 | Ba | 0.10057 | 0.1017 |
| Cr | 0.02373 | 0.02486 | La | 0.08927 | 0.08814 |
| Mn | 0.01808 | 0.02034 | Ce | 0.08136 | 0.1017 |
| Fe | 0.01582 | 0.01695 | Sm | 0.05311 | 0.05311 |
| Co | 0.01356 | 0.01356 | Eu | 0.05424 | 0.04746 |
| Ni | 0.01356 | 0.01243 | Tb | 0.04407 | 0.04181 |
| Cu | 0.01808 | 0.01695 | Hf | 0.1921 | 0.06215 |
| Zn | 0.01921 | 0.01808 | Ta | 0.17515 | 0.03616 |
| Ga | 0.01921 | 0.01017 | W | 0.06102 | 0.02486 |
| As | 0.00904 | 0.00791 | Ir | 0.04972 | 0.02373 |
| Se | 0.0113 | 0.01356 | Au | 0.03955 | 0.01808 |
| Br | 0.00791 | 0.01469 | Hg | 0.05198 | 0.03277 |
| Rb | 0.01469 | 0.01695 | Pb | 0.03503 | 0.03164 |

RTI Response: RTI is currently investigating the calculation of uncertainties by the STN and IMPROVE Laboratories in an effort to bring the uncertainties reported for PM-fine measurement programs in the U.S. into agreement ("harmonization"). Ms. Joann Rice of EPA is being kept apprised of these efforts.

3.5 Sample Handling and Archiving Laboratory

3.5.1 Quality Issues and Corrective Actions

There were no major quality issues in the SHAL laboratory during 2005. One Corrective Action was undertaken to resolve a discrepancy in data reported to the AQS database. The discrepancy in the data was the result of the SHAL sending the incorrect cooler to two different field sampling locations. The site operators noted the problem on their respective Field Sampling Data Forms. Upon arrival back at the SHAL, the error was corrected and the analytical results for the filters were assigned to the correct sampling location, however, the field sampling parameters for the two samplers were not assigned to the correct locations. After the

data for the two events was reported to AQS, RTI was notified by one of the sites that the field sampling information was not correct. RTI corrected the field sampling data for both locations and reposted the corrected data to AQS.

3.5.2 Description of QC Checks Applied



- The SHAL uses a customized database program written specifically for RTI's SHAL operation. This database has been refined over five years to incorporate many built-in QC checks. For example, RTI has assigned an inventory number to all filter modules in the network. The database will only accept allowable inventory numbers for filter modules. This avoids errors in data input for any filter module used for a sampling event. Another example is the unique number of the Teflon filters used by RTI. RTI purchases Teflon filters with a check sum digit in the numbering sequence. The database will only accept those filter numbers with the correct check sum. This prevents inadvertent entry of incorrect filter identification numbers.
- Bar-code readers are used to input identification numbers from modules, bins, containers and data forms to eliminate data transcription errors.
- A SHAL technician other than the one who prepared an outgoing shipment checks the package of outgoing filters. A checklist is used by the technician to verify that the package contents are correct before it is shipped from RTI. This check is performed on all outgoing shipments from the SHAL.
- Blank filters are taken from the SHAL refrigerator and sent unopened to the analytical laboratories for analysis. The results of the analysis of these QC filters are used to improve the overall quality of the program.
- The field site operators are provided contact information for the SHAL laboratory so they may communicate directly with personnel at RTI if any problems are discovered upon receipt of the filter modules. RTI personnel will attempt to resolve issues promptly. For example, a Field Data Form may be faxed from RTI to the site operator if necessary.

3.5.3 Summary of QC Results

During calendar year 2005, the SHAL shipped out and received back almost 20,000 coolers of filters. By employing the QC checks described in Section 3.5.2, the majority of the coolers shipped and received at RTI contained the correct filter modules and the required paperwork for completing the sampling event at the field site. The high number of correctly packaged shipments sent from RTI helped the field sampling locations meet their completion goals. (See Table B-3).

3.5.4 Summary of Scheduling Problems

Two shipping and receiving schedules are prepared for the STN. One schedule is for those sites sampling on the 1-in-3 day frequency and the other schedule is for those sites sampling on the 1-in-6 day frequency. The schedules indicate when each cooler will be sent from RTI, the scheduled sampling date for the filters, and the return ship date from the site back to RTI. The schedules are designed to allow RTI to send the sampling site clean filters allowing time for field site operators to set up and retrieve filters from the samplers. Late arriving shipments back to RTI may cause disruptions in the designated shipping schedule and could lead to missed sampling events. For instance, RTI may receive a shipment from the field sampling site, past the date that the filter modules were to be sent for a subsequent sampling event. When this happens, it may be impossible for RTI to send the filter modules to the sampling location for the next sampling event. This will mean a missed sampling event for that location. Late arriving shipments at RTI may be due to delays in transit or late return shipments from the site. Late shipments received at RTI during 2005 are summarized in Figures 3-1A and 3-1B.



Sites may also deviate from the sampling schedule and run filters on a date other than the scheduled date. Table 3-32 lists those sites with less than 95% of their filters run on the intended sampling date.

Table 3-32. Sites with Less than 95% of Filters Run on Intended Sampling Date

| Airs Code | POC | Location | Events | On Date | Pct. |
|-----------|-----|-------------------|--------|---------|------|
| 080010006 | 5 | Commerce City, CO | 102 | 81 | 79.4 |
| 080410011 | 5 | RBD, CO | 61 | 49 | 80.3 |
| 040137003 | 5 | St Johns, AZ | 47 | 39 | 83.0 |
| 471570047 | 5 | Guthrie, TN | 121 | 103 | 85.1 |
| 421010136 | 5 | Elmwood, PA | 61 | 54 | 88.5 |
| 330150014 | 5 | Portsmouth, NH | 121 | 114 | 94.2 |

3.5.4 Support Activities for Site Operators and Data Users

SHAL staff provided support to site operators and data users throughout 2005. A summary of email and phone communications with site operators and data users is presented in Table 3-33.

Table 3-33. Summary of SHAL Communications With Site Operators and Data Users

| Description | Number of Communications |
|--|--------------------------|
| Site will send cooler late | 130 |
| Site needs schedule | 26 |
| Site did not receive cooler | 43 |
| Change of operator/site information | 63 |
| Sampler problems/questions | 39 |
| Field Blank/Trip Blank ran as routine sample | 16 |
| Request change of ship date from RTI | 23 |
| Site is stopping | 27 |
| QA | 72 |
| Data questions/reporting | 173 |
| Site did not send cooler | 41 |
| Other | 134 |

3.5.5 Audits, PEs, Training, and Accreditations

- All new SHAL technicians must undergo a formal training process before they handle any filters. This process includes a Safety and Occupational Health Orientation, the viewing of a training video detailing the SHAL procedures, a review of the Standard Operating Procedure and instruction by senior staff in filter handling. A record of this training is kept on file.
- SHAL staff periodically review the Standard Operating Procedure and a record of this review is added to their training file.
- Since work in the SHAL involves lifting, during 2005 the North Carolina Department of Labor presented a course entitled “Lifting and Materials Movement and Preventing Slips in the Workplace” at RTI. All SHAL personnel attended this training.
- Throughout the year senior SHAL staff periodically observe the SHAL technicians processing the filter modules. A checklist of correct tasks has been prepared for each module type. The checklist is used during the observation of the technician. The SHAL supervisor keeps the completed checklists. Technicians are briefed following the review of any findings. A summary of the reviews for calendar year 2005 is shown in Table 3-34.

Table 3-34. Review of SHAL Technician Processing Filter Modules

| Module Type | Number Observed | Findings | Findings Reviewed with Technician |
|-------------|-----------------|----------|-----------------------------------|
| MET ONE | 28 | 6 | 6 |
| Andersen | 9 | 1 | 1 |
| URG | 2 | 0 | 0 |
| R&P Spec | 2 | 0 | 0 |

3.6 Denuder Refurbishment Laboratory

The Denuder Refurbishment Laboratory is located in RTI Building No. 3, Laboratory 220. The purpose of the laboratory is to clean and refurbish the coatings on acid-gas-removing denuders used in samplers of chemical speciation networks operated by EPA and various State and local agencies which utilize the RTI/EPA contract. The laboratory follows these protocols:

- Procedure for Coating Annular Denuders with Magnesium Oxide

- Standard Operating Procedure for Coating and Extracting Annular Denuders with Sodium Carbonate
- Procedures for Coating R & P Speciation Sampler “ChemComb” Denuders with Sodium Carbonate
- Standard Operating Procedure for Coating Annular Denuders with XAD-4 Resin.

3.6.1 Quality Issues and Corrective Actions

Ms. Constance Wall became the coordinator for the Denuder Refurbishment Laboratory. She reviewed the denuder refurbishment SOPs to ensure procedures were clearly stated and all processes were up to date. Minor revisions were made as required. Revisions mainly concerned glassware use and volumes of slurry; no revisions affected the quality of the actual denuder coating process.

The only significant problem encountered in the reporting period of operation has been the occasional receipt of broken or loose glass denuders. These were repaired by URG, Inc. and the costs were charged to the sampling site if breakage occurred there.

As personnel assignments or jobs changed, additional workers were trained in the techniques of denuder refurbishment. Hands-on training was conducted according to the several SOPs for denuder refurbishment. At present, there are four persons trained to refurbish denuders.

3.6.2 Operational Discussion

3.6.2.1 Numbers of Each Type of Denuder Serviced

Table 3-35 lists the type of denuders refurbished and the number of refurbishments completed in 2005.

**Table 3-35. Denuder Refurbishments
January 1 through December 31, 2005**

| Denuder Type | Total Refurbished |
|--------------|-------------------|
| R&P | 1,554 |
| MetOne | 708 |
| URG | 24 |
| Andersen | 56 |

3.6.2.2 Scheduling of Replacements

Denuders for the Andersen and URG speciation samplers are being cleaned and then re-coated with magnesium oxide. They are replaced at the sites at three-month intervals.

MetOne speciation sampler aluminum honeycomb denuders are also coated with magnesium oxide. Because the MetOne denuders are part of the sampling module and six sets of modules are in circulation to each site, these denuders are refurbished at 18-month intervals. RTI is able to remove MgO from denuders using a dilute hydrochloric acid solution. As needed, RTI orders uncoated aluminum honeycomb denuder substrates from MetOne, cleans them with solvent and deionized water, and then coats them with magnesium oxide. The change-out occurs whenever the MetOne denuder assembly has been in use for 18 months.

R & P ChemComb™ glass honeycomb denuders are cleaned and coated with sodium carbonate/glycerol. R & P denuders are replaced after each 24-hour sampling use.

No XAD-4 resin coated denuders (for removal of organic vapors) were ordered by EPA/OAQPS during the reporting interval.

3.6.3 Description of QC Checks Applied and Results

QC checks for coating weight are now done only occasionally. Work in earlier years of the project(s) showed that coating weights on the same types of MgO-coated denuders were usually within 10% of one another and that the amount (number of moles) of MgO applied far exceeded the expected mass (number of moles) of acidic gases that would be drawn through the denuder during the cumulative sampling period. The sodium carbonate coated R&P denuders are difficult to examine since the coating is somewhat opaque and not pure white as is MgO and the mass applied is much smaller. We depend on ensuring that all the honeycomb annuli receive the sodium carbonate uniformly during the application process.

Thickness of coating has never been evaluated. This and the uniformity of coating applied are assessed through visual examination of the interior of the denuders by holding them up to a strong light and sighting down the annuli. Examination of the occasional broken Andersen or URG denuder has also shown that the MgO coating is complete and uniform.

4.0 Data Processing

4.1 Quality Issues and Corrective Actions

No significant quality issues or corrective actions arose during the period of this report.

4.2 Operational Summary

Routine data processing activities have remained largely unchanged since the beginning of the program. These include:

- Accepting data entered from field forms
- Accepting data from the laboratories
- Backing up and maintaining the data base
- Generating data monthly for validation and review
- Posting review data monthly to the Web site for external review
- Incorporating data change requested by the States
- Uploading finalized data to AQS
- Responding to user inquiries and data requests, including support to EPA and RTI personnel.

4.3 Operational Changes and Improvements

Operational changes and improvements made during the reporting period include:

- Made minor changes to automated QA/QC review spreadsheets to make them easier to review.
- Started posting additional QA/QC spreadsheets to external web site for EPA review. Current spreadsheets include:
 - Field and trip blank levels (including gravimetric mass, organic and elemental carbon; sulfate, nitrate, and ammonium by IC; and sodium, potassium, iron, nickel, and zinc by XRF).
 - Mass balance outliers (i.e., those where the ratio of analytes (other than gravimetric mass) differs from the measured gravimetric mass by more than the control limits).
 - Mass outliers
 - Sulfur to sulfate ratio outliers (i.e., those where the ratio of sulfur differs from the measured sulfate ion by more than the control limits).

4.4 Monthly Data Postings to Web Site

Each month, RTI posts data for samples received on or before the 15th of the previous month. Table 4-1 shows monthly totals for postings and Table 4-2 shows totals for events. Sample dates may overlap between different batches due to different shipping schedules for the 1-in-3 and 1-in-6 sampling schedules. In addition, the latest date may include samples received late (i.e., after the previous report's cutoff date). Note that the number of records reported per event varies with sampler type. Thus the number of records per event will vary depending on how many of each sampler type was operating during that period.

Table 4-1. Events Posted To Web Site

| Report | | Sample Date | | Field Samples | Blanks | | Total |
|--------|------------|-------------|------------|---------------|--------|------|-------|
| Batch | Date | Earliest | Latest | | Field | Trip | |
| 60 | 1/13/2005 | 11/14/2004 | 12/14/2004 | 1,460 | 212 | 37 | 1,709 |
| 61 | 2/14/2005 | 12/17/2004 | 1/13/2005 | 1,274 | 79 | 206 | 1,559 |
| 62 | 3/11/2005 | 1/11/2005 | 2/12/2005 | 1,387 | 279 | 33 | 1,699 |
| 63 | 4/14/2005 | 12/11/2004 | 3/14/2005 | 1,312 | 211 | 48 | 1,571 |
| 64 | 5/12/2005 | 3/14/2005 | 4/13/2005 | 1,384 | 77 | 279 | 1,740 |
| 65 | 6/13/2005 | 4/16/2005 | 5/13/2005 | 1,387 | 2 | 57 | 1,446 |
| 66 | 7/15/2005 | 5/13/2005 | 6/12/2005 | 1,380 | 290 | 68 | 1,738 |
| 67 | 8/15/2005 | 6/14/2005 | 7/15/2005 | 1,479 | 72 | 107 | 1,658 |
| 68 | 9/13/2005 | 7/14/2005 | 8/11/2005 | 1,293 | 212 | 53 | 1,558 |
| 69 | 10/14/2005 | 8/14/2005 | 9/14/2005 | 1,464 | 216 | 36 | 1,716 |
| 70 | 11/14/2005 | 9/13/2005 | 10/15/2005 | 1,490 | 5 | 261 | 1,756 |
| 71 | 12/14/2005 | 10/14/2005 | 11/15/2005 | 1,385 | 282 | 29 | 1,696 |

Table 4-2. Records Posted To Web Site

| Report | | Sample Date | | Field Samples | Blanks | | Total |
|--------|------------|-------------|------------|---------------|--------|--------|---------|
| Batch | Date | Earliest | Latest | | Field | Trip | |
| 60 | 1/13/2005 | 11/14/2004 | 12/14/2004 | 166,980 | 24,312 | 4,269 | 195,561 |
| 61 | 2/14/2005 | 12/17/2004 | 1/13/2005 | 145,712 | 8,913 | 23,452 | 178,077 |
| 62 | 3/11/2005 | 1/11/2005 | 2/12/2005 | 158,716 | 31,881 | 3,764 | 194,361 |
| 63 | 4/14/2005 | 12/11/2004 | 3/14/2005 | 150,085 | 24,191 | 5,454 | 179,730 |
| 64 | 5/12/2005 | 3/14/2005 | 4/13/2005 | 157,989 | 8,681 | 31,975 | 198,645 |
| 65 | 6/13/2005 | 4/16/2005 | 5/13/2005 | 158,643 | 226 | 6,495 | 165,364 |
| 66 | 7/15/2005 | 5/13/2005 | 6/12/2005 | 157,946 | 33,162 | 7,660 | 198,768 |
| 67 | 8/15/2005 | 6/14/2005 | 7/15/2005 | 169,044 | 8,128 | 12,124 | 189,296 |
| 68 | 9/13/2005 | 7/14/2005 | 8/11/2005 | 148,011 | 24,362 | 6,037 | 178,410 |
| 69 | 10/14/2005 | 8/14/2005 | 9/14/2005 | 167,387 | 24,782 | 4,055 | 196,224 |
| 70 | 11/14/2005 | 9/13/2005 | 10/15/2005 | 170,683 | 548 | 29,967 | 201,198 |
| 71 | 12/14/2005 | 10/14/2005 | 11/15/2005 | 158,407 | 32,223 | 3,298 | 193,928 |

4.5 Postings to AQS

After data have been posted to the external website, sites have 45 days to review data and send corrections to RTI. RTI then is required to post data to AQS within 15 days. RTI met all processing deadlines for this reporting year. Table 4-3 contains totals of events posted to AQS. Table 4-4 contains totals of records posted to AQS. Note that blanks involve fewer records per event, as temperature and barometric pressure for field and trip blanks are not posted to AQS. Some data, such as results for the collocated shipping study were reported to the sites, but were not reported to AQS. In addition, the number of records posted per event varies with sampler type (with the URG posting volatile and total nitrate).

Table 4-3. Events Posted To AQS

| Report Batch | Field Samples | Blanks | | Total |
|--------------|---------------|--------|------|-------|
| | | Field | Trip | |
| 60 | 1,483 | 213 | 37 | 1,733 |
| 61 | 1,302 | 79 | 205 | 1,586 |
| 62 | 1,405 | 282 | 32 | 1,719 |
| 63 | 1,327 | 210 | 46 | 1,583 |
| 64 | 1,466 | 76 | 277 | 1,819 |
| 65 | 1,382 | 2 | 57 | 1,441 |
| 66 | 1,368 | 284 | 67 | 1,719 |
| 67 | 1,462 | 70 | 103 | 1,635 |
| 68 | 1,287 | 209 | 51 | 1,547 |
| 69 | 1,445 | 214 | 35 | 1,694 |
| 70 | 1,490 | 5 | 260 | 1,755 |

Table 4-4. Records Posted to AQS

| Report Batch | Field Samples | Blanks | | Total |
|--------------|---------------|--------|--------|---------|
| | | Field | Trip | |
| 60 | 99,499 | 14,285 | 2,487 | 116,271 |
| 61 | 87,344 | 5,307 | 13,773 | 106,424 |
| 62 | 94,267 | 18,922 | 2,148 | 115,337 |
| 63 | 89,037 | 14,086 | 3,090 | 106,213 |
| 64 | 98,374 | 5,106 | 18,587 | 122,067 |
| 65 | 92,714 | 134 | 3,831 | 96,679 |
| 66 | 91,764 | 19,054 | 4,507 | 115,325 |
| 67 | 98,092 | 4,702 | 6,917 | 109,711 |
| 68 | 86,335 | 14,017 | 3,429 | 103,781 |
| 69 | 96,953 | 14,356 | 2,355 | 113,664 |
| 70 | 99,968 | 339 | 17,446 | 117,753 |

4.6 Data User Support Activities

RTI had continuing data user support throughout the year. Most responses may be categorized into four categories; data change requests, requests for old data, support requests for the Speciation Data Validation and Analysis Tool (SDVAT), and requests from data users.

4.6.1 Data Change Requests

Sites are asked to review their data and submit any changes to RTI within 45 days. RTI then processes these changes before posting the data to AQS. Sites report changes via e-mail. Many sites do not report unless they have changes, while others send a report back indicating there are no changes to be made. Table 4-5 shows a count of the number of change requests per batch. Note that many requests represent multiple sites (often an entire state).

Table 4-5. Change Requests Per Report Batch

| Batch | 60 | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 |
|----------|----|----|----|----|----|----|----|----|----|----|----|
| Requests | 8 | 14 | 10 | 10 | 14 | 9 | 11 | 6 | 6 | 5 | 7 |

4.6.2 Requests for Old Data

RTI keeps draft data reports on its internal site for approximately 60 days. This provides enough time for sites to review their data and request changes (changes are required to be sent to RTI within 45 days of posting on the internal site). RTI makes any requested changes before posting to AQS and then removes the draft (unmodified) data from the web site. Although we recommend that all data be retrieved from AQS, as these official data incorporate any and all changes made by the sites, a few sites have found the data review format supplied by RTI to be more convenient. Such requests are often made with respect to the use of the SDVAT program (described below). Requests for old data are less frequent than in earlier years. This is likely due to AQS enhancements that allow all speciation parameters to be retrieved in a single request.

4.6.3 SDVAT Support

RTI was previously contracted by EPA to produce a software program (SDVAT) to help Speciation sites to review and approve their data. Although EPA funding for SDVAT support ended in 2002, RTI has continued to provide limited support to current STN sites. Most of these questions are from new operators who need help importing site data into the SDVAT. Introduction of new analytes (for carbon analyses) under the new contract has caused problems with the now discontinued SDVAT program (as it was not designed to handle them). RTI has produced a short note that explains a workaround for the problem, which it sends out on request.

RTI has had requests on use of the SDVAT to review data in AQS. Unfortunately, because the SDVAT was developed during the transition from the legacy AQS to the reengineered AQS, adding AQS import capabilities was not feasible. Now that the record format and procedures are available, this could be added to a future version of the SDVAT, if funding were available.

4.6.4 Data User Communications

In general, RTI's STN activity is limited to sample analysis and module preparation. Because of this, we have limited involvement with STN data users. However, the data processing staff do field a few requests each year from data users. A short summary, by topic, is below:

- Data Availability at end of calendar year - Several calls were from state or regional personnel inquiring on data availability after the end of the calendar year. We explained the process and deadlines under the current process and provided estimates of when data would be available (typically in the April 15th monthly report). The delay reflects reporting (up to 45 days), site review (45 days), and RTI posting (15 days). Thus a sample run on December 31 would be received by RTI in early January (before January 15th) and reported on by RTI on or before February 15th. The site would have until April 1st to review their data and RTI would have until April 15th to post data to AQS.
- Data Uncertainty.
 - Several state and regional contacted us about data uncertainties and how they were calculated. Most were referred to our uncertainty calculations write-up.
 - A few users also asked about backfilling uncertainty data under the old contract. We indicated that the procedures are under development.

5.0 Quality Assurance and Data Validation

5.1 QA Activities

5.1.1 QAPP Updates

RTI's QAPP for STN was updated twice during 2005 and is posted on EPA's public web site, AMTIC. Changes to the QAPP during 2005 were as follows:

- September 27, 2005
 - EPA QA Lab Director changed to E.Boswell
 - Teflon filter catalog number changed
- July 11, 2005
 - Staffing changes at EPA and RTI
 - Numerous minor edits
 - Number of sites amended
 - Correspondence files to be stored as hardcopy and electronic
 - AIRS changed to AQS throughout
 - Denuder preparation procedure corrections
 - Gravimetric analysis QC sample frequency - increase in duplicate filter weighing in response to problems with Whatman filter lot.
 - Gravimetric Laboratory disaster recovery plan updated
 - Clarification of MDL calculations
 - Update number of balances from 2 to 3

5.1.2 SOP Updates

The following SOPs were updated during 2005:

| Type | Title | Date Revised | Comment |
|------|---|--------------|---|
| SOP | Gravimetric Analysis | 7/8/2005 | Increased frequency of QC reweighings |
| SOP | Database Operations | 7/11/2005 | Maintenance updates |
| SOP | Disaster Recovery Plan--RTI CONFIDENTIAL | 7/6/2005 | Updated recovery plan for infrastructure changes at RTI |
| SOP | Procurement and Acceptance Testing of Teflon, Nylon, and Quartz Filters | 7/7/2005 | New e-procurement procedures; enhanced Teflon filter inspection requirement in response to filter debris problem. |
| SOP | Sample Handling and Archiving Laboratory (SHAL) | 7/11/2005 | Maintenance updates; revised procedures for filter module assembly and disassembly |
| SOP | Assign Field Sample Flags for the Chemical Speciation Trends Network | 7/7/2005 | Major update incorporating more details of process |
| SOP | Document Control and Storage | 7/6/2005 | Maintenance updates |

5.1.3 Internal Surveillance Activities

Internal surveillance activities during July included walkthroughs of all the laboratories to verify compliance with the SOPs prior to the EPA systems audit in July, 2005. An internal audit of the Gravimetry Laboratory was performed in January, 2005. Other inspections and investigations were prompted by issues such as the Whatman filter debris problem in the Gravimetry Laboratory. In addition, the QA Manager and Program Manager meet with laboratory supervisors on a monthly basis to discuss outstanding problems.

SHAL supervisors routinely inspect assembly of R&P model 2300 modules, which have proven to be problematic in the past. Inspection of these modules ensures that filters are fixed securely in the support rings so that bypass leaks do not occur. SHAL technicians also crosscheck each other's coolers before they are shipped to the sites.

5.1.4 Data User Support Activities

The QA Manager responded to a number of questions and requests for data during 2005. These originated from both network participants (state agency personnel and EPA) as well as data users who were not affiliated with the STN program.

- Requests for blank levels and other background data - RTI has received several requests for information such as network-wide average blank levels. When this information is readily available from monthly reports, etc., we try to provide this information. More extensive requests for data are referred to EPA staff.
- Artifacts, outliers, and poor comparison with FRM results - Reports from the state agencies regarding poor intercomparison with FRM results result in immediate investigation by the QA and technical staff. These reports are extremely important in identifying potential problems in the laboratory. However, they can also point to sampler-related issues such as leaks or sensor malfunctions.
- Data Uncertainties - Several states and regions contacted us about data uncertainties and how they were calculated. Most were referred to the uncertainty calculations write-up that was developed in 2003, when uncertainties were added to the deliverable data.
- Method Detection Limits (MDLs) - Periodic requests are received for the list of MDLs for the analyzers used by the STN program. Both MDLs and uncertainties are now included in data records uploaded to AQS. Data records prior to July 2003 lack this information, and the users are given a table of historical MDLs that was developed under RTI's previous contract.

5.2 Data Validation and Review

5.2.1 Review of Monthly Data Reports to STN Web Site

Each month, RTI reviews data completed during the previous month. The reviews include the following activities:

- Verification of data attribution to the correct site, POC, and date
- Visual review of report formats
- Investigation and corrective actions when discrepancies are found
- Automated range checks (barometric pressure, temperature)
- Level 1 checks (reconstructed mass balance, anion/cation balance, and sulfur/sulfate balance)

Tables 5-1 through 5-3 summarize the data flags attached to the data primarily through the data review process, although some of these were specified by either the field operator or one of the laboratories. Examining trends in flag percentages is a useful tool in diagnosing potential problems. For example, reporting batches 64, 65, and 71 have elevated numbers of QMB flags. These were the result of problems in the Gravimetry Laboratory which became the subject of corrective actions. Other trends in the data flags, such as the higher levels of DST flags (filter receipt temperature) in the summer, are out of the control of RTI.

5.2.2 Review of Monthly Data Packages to AQS

Approximately 60 days after initial posting on the RTI web site, the data are uploaded to the AQS database. Prior to uploading, the data processing staff prepares a QC summary report which is reviewed by the QA Manager. This summary and review includes the following main areas:

- Verification that changes requested by the state agencies have been implemented. This includes checking data flags that are different between original reporting (web site posting) and final AQS reporting.
- Verification that record counts match exactly the number of records previously reported on the RTI web site, with allowance for all records that were added and deleted during processing. Record changes include such things as elimination of duplicates, generation of aggregated nitrate values for MASS samplers, deletion of data for sites not reported to AQS (e.g., special studies).
- Scanning for unusual values such as start times other than midnight
- Scanning for formatting errors such as:
 - duplicate records
 - flags and other data in incorrect columns
 - previously delivered data (unless they are Modify records)
 - MDLs and uncertainties that do not agree between the original report and the AQS data file

Table 5-1. Summary of Validity Status Codes by Delivery Batch Number**AQS Validity Status Codes**

| Flag | Description | 60 | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 | 71 |
|------|--------------------------------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| 1 | Critical Criteria Not Met | | | | | | 0.09% | 0.04% | | | | | |
| 2 | Operational Criteria Not Met | | | 0.02% | | | | | 0.24% | | | | |
| 3 | Possible field contamination | | | | | 0.05% | | 0.06% | | | | | |
| 4 | Possible lab contamination | | 0.10% | | | | | | | | | | |
| 5 | Outlier-cause unknown | 3.87% | 3.56% | 2.45% | 4.85% | 6.68% | 7.91% | 5.87% | 5.22% | 3.70% | 4.20% | 4.97% | 8.51% |
| A | High Winds | 0.07% | 0.06% | 0.06% | 0.06% | | | 0.06% | 0.06% | 0.12% | 0.06% | 0.06% | 0.25% |
| D | Sandblasting | 0.05% | | 0.05% | | 0.10% | | 0.05% | 0.06% | 0.12% | 0.27% | | 0.11% |
| E | Forest Fire | | | | | 0.05% | 0.07% | 0.06% | 0.47% | 0.44% | | 0.06% | 0.17% |
| F | Structural Fire | 0.06% | 0.19% | | | | | | 0.06% | | | 0.07% | |
| H | Chemical Spills | | 0.08% | | | 0.05% | | | 0.06% | | | | |
| I | Unusual Traffic Congestion | | | | | | | | | | | 0.06% | |
| J | Construction/Demolition | | 0.16% | 0.22% | | | | 0.11% | 0.12% | 0.37% | 0.23% | 0.50% | 0.47% |
| K | Agricultural Tilling | | | | | | 0.07% | | 0.06% | 0.06% | | 0.06% | |
| L | Highway Construction | 0.06% | | | | | 0.07% | | | 0.06% | 0.17% | 0.06% | |
| M | Rerouting of Traffic | | | | | | | | | | 0.06% | | |
| N | Sanding/salting of Streets | | 0.12% | 0.46% | 0.31% | | | | | | | | |
| O | Infrequent Large Gatherings | | | | | | | | | 0.06% | | | |
| P | Roofing Operations | | 0.16% | 0.22% | | | | | | 0.06% | 0.13% | 0.17% | |
| Q | Prescribed Burning | | | | | | | | | 0.06% | | | |
| R | Cleanup after Major Disaster | | 0.08% | | | | | | | | 0.06% | | |
| U | Sahara Dust | | | | | | | 0.17% | | 0.12% | | | |
| W | Flow Rate Average out of specs | 0.11% | | | | 0.01% | 0.04% | 0.06% | 0.11% | 0.01% | 0.06% | 0.07% | |
| X | Filter Temperature Diff. out of spec | 0.32% | 0.19% | 0.07% | 0.13% | 0.51% | 0.47% | 0.38% | 0.80% | 0.14% | 0.13% | 0.17% | 0.23% |
| Y | Elapsed Sample Time out of specs | | 0.05% | | | | 0.13% | | | | | | |

Table 5-2. Summary of Null Value Codes by Delivery Batch Number**AQS Null Value Codes**

| Flag | Description | 60 | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 | 71 |
|------|--------------------------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| AB | Technician Unavailable | 0.35% | 0.12% | 0.23% | | | 0.34% | 0.17% | 0.29% | 0.19% | 0.06% | 0.06% | 0.06% |
| AC | Construction/Repairs in area | | | | | | | | | | 0.23% | 0.11% | |
| AF | Sched. but not collected | 0.31% | 0.55% | 0.55% | 0.14% | 0.32% | 0.20% | 0.28% | 0.27% | 0.38% | 0.51% | 0.22% | 0.17% |
| AG | Sample Time out of Limits | 0.58% | 0.72% | 0.69% | 0.43% | 0.51% | 0.66% | 0.87% | 0.62% | 0.66% | 0.28% | 0.91% | 0.48% |
| AH | Samp. Flow Rate out of Limits | 0.43% | 0.64% | 0.43% | 0.56% | 0.51% | 0.56% | 0.72% | 0.41% | 0.46% | 0.60% | 0.80% | 0.43% |
| AI | Insuff. data (can't calculate) | 0.05% | 0.12% | 0.06% | | 0.01% | 0.13% | 0.02% | 0.07% | 0.03% | | 0.06% | 0.06% |
| AJ | Filter Damage | 0.18% | 0.13% | 0.15% | 0.04% | 0.21% | 0.03% | 0.04% | 0.27% | 0.23% | 0.04% | 0.05% | 0.19% |
| AK | Filter Leak | | | | | | | | | 0.13% | | | |
| AL | Voided by Operator | 0.07% | 0.25% | 0.12% | 0.20% | 0.29% | 0.20% | 0.06% | 0.06% | 0.06% | 0.30% | 0.17% | 0.34% |
| AM | Miscellaneous void | 0.34% | 0.45% | 0.16% | 0.36% | 0.41% | 0.23% | 0.09% | 0.10% | 0.16% | 0.09% | 0.15% | 0.13% |
| AN | Machine Malfunction | 1.50% | 1.51% | 0.94% | 0.69% | 1.06% | 0.43% | 0.53% | 0.36% | 1.02% | 0.77% | 0.99% | 0.91% |
| AO | Bad Weather | 0.07% | 0.06% | 0.07% | | | 0.08% | | | | 0.63% | 0.14% | 0.19% |
| AQ | Collection Error | 0.01% | 0.14% | 0.20% | 0.17% | 0.09% | 0.20% | 0.39% | 0.08% | 0.31% | 0.09% | 0.06% | 0.35% |
| AR | Lab Error | 0.21% | 0.09% | 0.22% | 0.05% | 0.21% | 0.13% | 0.13% | 0.14% | 0.35% | 0.22% | 0.06% | 0.09% |
| AS | Poor QA Results | 0.04% | 0.06% | | | | | | 0.03% | | | | |
| AU | Monitoring Waived | | | 0.06% | 0.37% | 0.54% | | | | | | 0.06% | |
| AV | Power Failure | 0.54% | 0.57% | 0.56% | 0.34% | 0.30% | 0.60% | 0.57% | 0.80% | 0.71% | 0.43% | 0.79% | 0.48% |
| AW | Wildlife Damage | | | | | | 0.07% | 0.03% | 0.13% | 0.17% | | | |
| BA | Maint. / Routine Repairs | | | 0.16% | 0.51% | 0.12% | 0.51% | 0.18% | 0.12% | | | 0.22% | 0.18% |
| BB | Unable to Reach Site | 0.06% | 0.18% | 0.23% | | 0.05% | 0.51% | 0.11% | | 0.12% | | 0.17% | 0.12% |
| BE | Building / Site Repair | | | | | | | | | | | | 0.06% |
| BI | Lost or Damaged in Transit | 0.07% | 0.00% | | | 0.05% | 0.01% | | | | | | |
| BJ | Operator Error | | | | | 0.05% | | | | | | | |

Table 5-3. RTI-assigned Flags (not reported to AQS) by Delivery Batch Number

| Flag | Description | 60 | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 | 71 |
|------|---|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|
| ANB | Analysis not billable | 0.21% | 0.08% | 0.13% | 0.05% | 0.29% | 0.13% | 0.15% | 0.32% | 0.37% | 0.16% | 0.10% | 0.14% |
| APB | analysis partly billable | 0.74% | 0.66% | 0.48% | 0.51% | 0.85% | 0.54% | 0.47% | 0.25% | 0.28% | 0.34% | 0.15% | 0.56% |
| DFM | Filter missing | 0.01% | | | | | | | | 0.03% | | 0.01% | |
| DMA | Module assembled in correctly | | | | | | | | | 0.01% | | 0.02% | |
| DMC | Module condition invalid | | 0.01% | | | | | | | | | | |
| DST | Receipt temperature >4C | 33.10% | 26.21% | 14.79% | 21.79% | 51.83% | 82.91% | 88.89% | 86.80% | 97.55% | 97.15% | 87.07% | 51.11% |
| FBS | Field or Trip Blank appears to be actual sample | 0.06% | | | | | | | | 0.06% | | 0.07% | |
| FES | Field Environmental Substituted | 0.01% | 0.15% | 0.10% | 0.06% | 0.04% | 0.11% | 0.05% | 0.01% | 0.14% | 0.01% | 0.15% | 0.04% |
| FHT | Pickup holding time exceeded | 5.45% | 18.58% | 12.60% | 12.93% | 5.26% | 15.28% | 18.04% | 7.21% | 10.67% | 12.11% | 12.94% | 4.98% |
| FSL | Sample lost or damaged in shipment | | | | | | | | | 0.06% | | | |
| LBD | Laboratory blank duplicate outside limits | | | 0.00% | | 0.00% | | | | | | | |
| LFA | Filter inspection flags* - filter wet | 0.01% | 0.11% | 0.01% | 0.01% | 0.06% | | 0.04% | 0.01% | 0.01% | | 0.02% | 0.03% |
| LFH | Filter inspection flags* - Holes in filter | | | | | | | | | | | | |
| LFL | Filter inspection flags* -Loose Material | 0.01% | | | | | | | | 0.01% | | 0.03% | |
| LFO | Filter inspection flags* -Other | | | | | | | | | 0.07% | | | |
| LFP | Filter inspection flags* -Pinholes | | | | | | | | | | | | |
| LFT | Filter inspection flags* - Tear | | | | | | | | | 0.03% | 0.01% | | 0.01% |
| LHT | Lab holding times exceeded | | | | | | | | | | | | |
| QAC | Anion/Cation ratio out of limits | 0.08% | 0.17% | 0.11% | 0.07% | 0.11% | 0.09% | 0.13% | 0.16% | 0.08% | 0.10% | 0.13% | 0.11% |
| QL1 | Sulfur/Sulfate ratio out of limits | 0.07% | 0.07% | 0.04% | 0.04% | 0.06% | 0.07% | 0.06% | 0.09% | 0.06% | 0.05% | 0.08% | 0.07% |
| QMB | Mass balance ratio out of limits | 3.78% | 3.37% | 2.31% | 4.77% | 6.55% | 7.79% | 5.73% | 5.03% | 3.52% | 4.10% | 4.83% | 8.37% |
| SNB | Sample not billable | 0.11% | 0.24% | 0.32% | | 0.10% | 0.07% | 0.05% | | 0.19% | 0.17% | | 0.06% |
| SPB | Sample partly billable | 2.85% | 3.10% | 3.03% | 2.31% | 2.51% | 2.98% | 2.43% | 2.60% | 2.77% | 3.30% | 3.72% | 2.85% |

5.3 Analysis of Collocated Data

The STN program operated six sites with collocated samplers during 2005. The data from these sites afforded an opportunity to calculate total precision and compare the values with the uncertainty values that are currently being reported to AQS. The AQS uncertainties are only estimates based on historical QC data and scientific judgment. Table 5-4 lists the collocated sites in STN.

Table 5-4. Collocated sites in the STN

| Location Name | State | AQS Code | Sampler Type |
|----------------------------|---------------|-----------|--------------|
| Bakersfield-California Ave | California | 60290014 | MetOne SASS |
| Deer Park | Texas | 482011039 | URG MASS |
| G.T. Craig | Ohio | 390350060 | MetOne SASS |
| New Brunswick | New Jersey | 340230006 | MetOne SASS |
| Riverside-Rubidoux | California | 60658001 | MetOne SASS |
| Roxbury (Boston) | Massachusetts | 250250042 | MetOne SASS |

As indicated in the table, five of the sites use MetOne SASS samplers, and one uses a URG MASS sampler. None of the collocated sites used either the Andersen RAAS sampler or the R&P speciation sampler during 2005.

In general, the collocation data shows good or excellent agreement for the major analytes. The figures that follow (Figure 5-1) show examples of the comparisons for mass, sulfate, nitrate, sulfur, organic carbon, and elemental carbon. (This is not intended as an exhaustive list of elements -- these are presented as examples only.) The oblique line on each chart indicates perfect agreement (slope=1.000).

5.3.1 Precision

Tables 5-5 through 5-8 show the results of collocated sampling and provides a comparison with the uncertainties reported to AQS. The first column indicates the name of the chemical analyte. Columns 2-5 show the average and standard deviation of the analytical results. Note that the standard deviations reflect environmental variability of the concentration and are not determined by the laboratory uncertainties. The column titled "Average Relative Diff." is the average of the unsigned differences between the two samplers, which is calculated using the following formula:






Figure 5-1. Examples of the comparisons for mass, sulfate, nitrate, sulfur, organic carbon, and elemental carbon.




**Figure 5-1. (Continued)**

Table 5-5. Trace Elements by XRF

| Analyte | Sampler 1 | | Sampler 2 | | Average Relative Diff. ² µg/m ³ | Average AQS Uncert. ³ µg/m ³ | Ratio ⁴ AQS/Co I percent | Counts ⁵ |
|-----------|------------------------------|--|------------------------------|--|--|---|--|---------------------|
| | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | | | | |
| Aluminum | 0.084 | 0.099 | 0.081 | 0.116 | 31.4% | 17.5% | 55.6% | 125 |
| Arsenic | 0.003 | 0.002 | 0.003 | 0.002 | 23.2% | 21.6% | 93.0% | 136 |
| Barium | 0.023 | 0.026 | 0.024 | 0.019 | 26.6% | 21.2% | 79.7% | 61 |
| Bromine | 0.005 | 0.003 | 0.005 | 0.002 | 18.3% | 18.0% | 98.6% | 381 |
| Calcium | 0.083 | 0.131 | 0.079 | 0.080 | 17.8% | 9.5% | 53.2% | 470 |
| Cerium | 0.026 | 0.024 | 0.028 | 0.025 | 23.9% | 18.2% | 76.2% | 92 |
| Chlorine | 0.117 | 0.258 | 0.112 | 0.246 | 27.8% | 14.5% | 52.2% | 190 |
| Chromium | 0.008 | 0.013 | 0.006 | 0.006 | 42.2% | 22.3% | 52.8% | 123 |
| Copper | 0.011 | 0.018 | 0.009 | 0.008 | 26.8% | 15.1% | 56.4% | 386 |
| Europium | 0.013 | 0.011 | 0.014 | 0.015 | 37.7% | 22.9% | 60.7% | 30 |
| Iron | 0.157 | 0.167 | 0.157 | 0.169 | 14.4% | 5.9% | 40.9% | 479 |
| Lanthanum | 0.019 | 0.020 | 0.020 | 0.022 | 25.6% | 22.3% | 87.4% | 77 |
| Lead | 0.010 | 0.008 | 0.010 | 0.008 | 22.9% | 24.3% | 106.1% | 177 |
| Magnesium | 0.038 | 0.021 | 0.040 | 0.031 | 26.8% | 17.0% | 63.6% | 35 |
| Manganese | 0.007 | 0.007 | 0.007 | 0.007 | 23.2% | 20.5% | 88.5% | 266 |
| Mercury | 0.005 | 0.002 | 0.005 | 0.002 | 27.3% | 25.2% | 92.4% | 52 |
| Nickel | 0.004 | 0.004 | 0.004 | 0.003 | 27.3% | 22.0% | 80.4% | 243 |
| Potassium | 0.095 | 0.119 | 0.094 | 0.122 | 10.6% | 9.2% | 86.7% | 475 |
| Samarium | 0.012 | 0.010 | 0.011 | 0.010 | 27.6% | 24.0% | 87.0% | 41 |
| Selenium | 0.003 | 0.002 | 0.003 | 0.002 | 19.2% | 28.9% | 150.6% | 91 |
| Silicon | 0.166 | 0.198 | 0.159 | 0.166 | 24.6% | 11.5% | 46.6% | 338 |
| Sodium | 0.165 | 0.156 | 0.173 | 0.159 | 20.4% | 18.4% | 90.4% | 191 |
| Strontium | 0.004 | 0.004 | 0.004 | 0.003 | 20.4% | 30.5% | 149.3% | 56 |
| Sulfur | 1.196 | 1.054 | 1.186 | 1.036 | 6.3% | 6.7% | 105.1% | 477 |
| Terbium | 0.019 | 0.015 | 0.020 | 0.017 | 23.6% | 31.0% | 131.0% | 94 |
| Titanium | 0.012 | 0.010 | 0.012 | 0.009 | 24.5% | 19.7% | 80.2% | 235 |
| Vanadium | 0.006 | 0.004 | 0.006 | 0.003 | 20.6% | 22.1% | 107.4% | 219 |
| Zinc | 0.023 | 0.034 | 0.023 | 0.034 | 14.8% | 11.0% | 73.9% | 450 |
| Zirconium | 0.005 | 0.002 | 0.005 | 0.001 | 15.1% | 34.0% | 225.7% | 12 |

NOTES:

- 1 The standard deviations are a function of the natural variability of the environmental levels, and are not indicative of the analytical precision.
- 2 Calculated as the average of the absolute value of the relative difference between the two samplers' values, divided by the square root of 2.
- 3 Average value of the relative uncertainties as reported to AQS.
- 4 AQS/ARD is the ratio of reported uncertainties divided by the uncertainty determined by average relative difference of the collocated samples. Values greater than 200% are shown in bold and discussed in the text.
- 5 Counts are the number of individual observations included in the statistics. Only observations where both concentration values were above twice the uncertainty are included in the statistics.

Table 5-6. Anions and Cations by IC

| Analyte | Sampler 1 | | Sampler 2 | | Average Relative Diff. ² µg/m ³ | Average AQS Uncert. ³ µg/m ³ | Ratio ⁴ AQS/ARD percent | Counts ⁵ |
|-----------|------------------------------|---|------------------------------|---|--|---|---------------------------------------|---------------------|
| | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | | | | |
| Sulfate | 3.617 | 3.116 | 3.601 | 3.130 | 4.4% | 9.8% | 220.7% | 482 |
| Nitrate | 4.434 | 5.968 | 4.380 | 5.800 | 4.7% | 10.7% | 228.9% | 402 |
| Ammonium | 2.291 | 2.206 | 2.270 | 2.179 | 5.4% | 7.1% | 130.0% | 483 |
| Potassium | 0.114 | 0.136 | 0.115 | 0.142 | 9.1% | 7.6% | 83.3% | 258 |
| Sodium | 0.245 | 0.200 | 0.237 | 0.164 | 14.4% | 24.3% | 168.2% | 263 |

NOTES:

- 1 The standard deviations are a function of the natural variability of the environmental levels, and are not indicative of the analytical precision.
- 2 Calculated as the average of the absolute value of the relative difference between the two samplers' values, divided by the square root of 2. (ARD.)
- 3 Average value of the relative uncertainties as reported to AQS.
- 4 AQS/ARD is the ratio of reported uncertainties divided by the uncertainty determined by average relative difference of the collocated samples. Values greater than 200% are shown in bold and discussed in the text.
- 5 Counts are the number of individual observations included in the statistics. Only observations where both concentration values were above twice the uncertainty are included in the statistics.

Table 5-7. Organic and Elemental Carbon

| Analyte | Sampler 1 | | Sampler 2 | | Average Relative Diff. ² µg/m ³ | Average AQS Uncert. ³ µg/m ³ | Ratio ⁴ AQS/ARD percent | Counts ⁵ |
|------------------|------------------------------|---|------------------------------|---|--|---|---------------------------------------|---------------------|
| | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | | | | |
| Elemental carbon | 1.098 | 0.751 | 1.097 | 0.765 | 8.8% | 29.2% | 331.6% | 335 |
| Organic carbon | 4.501 | 2.370 | 4.425 | 2.388 | 9.7% | 12.3% | 127.3% | 499 |

NOTES:

- 1 The standard deviations are a function of the natural variability of the environmental levels, and are not indicative of the analytical precision.
- 2 Calculated as the average of the absolute value of the relative difference between the two samplers' values, divided by the square root of 2. (ARD.)
- 3 Average value of the relative uncertainties as reported to AQS.
- 4 AQS/ARD is the ratio of reported uncertainties divided by the uncertainty determined by average relative difference of the collocated samples. Values greater than 200% are shown in bold and discussed in the text.
- 5 Counts are the number of individual observations included in the statistics. Only observations where both concentration values were above twice the uncertainty are included in the statistics.

Table 5-8. Particulate Matter (Gravimetry)

| Analyte | Sampler 1 | | Sampler 2 | | Average Relative Diff. ² µg/m ³ | Average AQS Uncert. ³ µg/m ³ | Ratio ⁴ AQS/ARD percent | Counts ⁵ |
|------------------------|------------------------------|--|------------------------------|--|--|---|--|---------------------|
| | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | Average µg/m ³ | Standard Dev. ¹ µg/m ³ | | | | |
| PM _{2.5} mass | 16.8674 | 11.3557 | 16.7068 | 11.2434 | 9.5% | 5.8% | 60.3% | 476 |

NOTES:

- 1 The standard deviations are a function of the natural variability of the environmental levels, and are not indicative of the analytical precision.
- 2 Calculated as the average of the absolute value of the relative difference between the two samplers' values, divided by the square root of 2. (ARD.)
- 3 Average value of the relative uncertainties as reported to AQS.
- 4 AQS/ARD is the ratio of reported uncertainties divided by the uncertainty determined by average relative difference of the collocated samples. Values greater than 200% are shown in bold and discussed in the text.
- 5 Counts are the number of individual observations included in the statistics. Only observations where both concentration values were above twice the uncertainty are included in the statistics.

$$ARD = \frac{1}{\sqrt{2}} \sum \frac{|C_1 - C_2|}{(C_1 + C_2)/2}$$

Where:

- C_1 and C_2 are the concentrations from the primary and collocated samplers, respectively
- The factor of $1/\sqrt{2}$ is used to convert the difference to a single-sampler basis
- The summation is over all valid concentration values where the concentration (C_1 or C_2) is greater than twice the uncertainty reported to AQS

The column titled "Average AQS Uncert." is simply the grand average of all the relative uncertainties associated with the C_1 and C_2 values, and is calculated as follows:

$$AvAQS = \sum_i \sum_j U_{ij} / C_{ij}$$

Where

- U_{ij} and C_{ij} refer to the uncertainty and concentration for the i^{th} exposure with the j^{th} sampler ($j=1$ or 2).
- The criteria for inclusion in the average (index i) is the same as in the previous equation

The next column provides the ratio of AvAQS to ARD defined above. This is essentially the average under- or over-estimate of the uncertainty for each chemical species reported during 2005. Finally, the last column provides the number of sampling events included in the averages defined above. Only events where both concentrations were greater than twice their respective uncertainties were included.

Ratios greater than 200% or less than 50% indicate situations in which the uncertainties reported to AQS were different from the uncertainty estimated from collocation data by a factor of 2 or more. The following species disagreed by a factor of 2 or more; ratios are shown in parentheses:

- Silicon (46%) - underestimation of the uncertainty to AQS may be related to an XRF attenuation correction that is too small. Uncertainties for the XRF instruments are currently under investigation by RTI and EPA.
- Zirconium (226%) - only 12 events were included in the calculations, so this may be a statistical fluke.
- Sulfate (221%), nitrate (229%), and elemental carbon (332%) –RTI will investigate whether the MDLs used to calculate uncertainty for AQS are too large. If this is the case, the calculations will be revised.

The ratio for particulate mass (Table 5-8) is somewhat lower than expected (60%), though within a factor of 2. There was a problem with debris on the Whatman filters (described elsewhere in this report) that may have resulted in the actual uncertainty being greater than the estimated uncertainty provided to AQS

5.3.2 Bias

Biases between the primary and secondary samplers is small for all of the major analytes when data from all sites are combined. The overall averages for the primary and secondary samplers were compared using Student's t test, and only one element, copper, was found to have a significant difference between the two averages ($t = -3.77$).

5.4 Analysis of Trip and Field Blanks

In the STN program, field blanks are run at a frequency of 10% or more, while trip blanks are run at approximately 3%. Historical data has shown little difference between the two types of blanks, perhaps because the field SOPs for running them are very similar, the only difference being that the Field Blanks are mounted on the sampler for a few minutes, while the Trip Blanks are kept closed. Data from these blanks allow evaluation of contamination, which may come from a number of different sources. In addition, the Trip and Field Blank data can sometimes provide clues to problems in the analytical laboratories or with filters received from the manufacturers. Table 5-9 shows the distributions (percentiles) for trip and field blanks during 2005.

**Table 5-9. Concentration Percentiles for
Combined Trip and Field Blanks Reported During 2005**

Anions and Cations by Ion Chromatography

| ANALYTE | Percentiles | | | | | | |
|-----------------------|-------------|--------|--------|--------|--------|--------|--------|
| | 5 | 10 | 25 | 50 | 75 | 90 | 95 |
| Nitrate (SASS/nylon) | 0.0000 | 0.0000 | 0.0000 | 0.0254 | 0.0494 | 0.0775 | 0.1076 |
| Nitrate (MASS/nylon) | 0.0000 | 0.0109 | 0.0173 | 0.0236 | 0.0317 | 0.0474 | 0.0584 |
| Nitrate (MASS/Teflon) | 0.0000 | 0.0109 | 0.0161 | 0.0241 | 0.0377 | 0.0634 | 0.1197 |
| Sulfate | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0377 | 0.0637 | 0.1053 |
| Ammonium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 |
| Potassium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 |
| Sodium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0062 | 0.0343 | 0.0782 |

PM2.5 Mass by Gravimetry

| ANALYTE | Percentiles | | | | | | |
|-------------------------|-------------|----|--------|--------|--------|--------|--------|
| | 5 | 10 | 25 | 50 | 75 | 90 | 95 |
| Particulate matter 2.5u | -0.5208 | - | 0.1250 | 0.5208 | 1.0417 | 1.6667 | 2.1875 |

Organic and Elemental Carbon (OC/EC)

| ANALYTE | Percentiles | | | | | | |
|------------------|-------------|--------|--------|--------|--------|--------|--------|
| | 5 | 10 | 25 | 50 | 75 | 90 | 95 |
| Elemental carbon | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0065 | 0.0372 | 0.0840 |
| Organic carbon | 0.3920 | 0.5096 | 0.6871 | 0.9396 | 1.1864 | 1.4473 | 1.7590 |
| Pk1_OC | 0.0858 | 0.1180 | 0.1706 | 0.2398 | 0.3281 | 0.4255 | 0.4728 |
| Pk2_OC | 0.1366 | 0.1753 | 0.2517 | 0.3759 | 0.5124 | 0.6616 | 0.7936 |
| Pk3_OC | 0.0698 | 0.0959 | 0.1352 | 0.1888 | 0.2727 | 0.4003 | 0.5451 |
| Pk4_OC | 0.0080 | 0.0167 | 0.0336 | 0.0600 | 0.0999 | 0.1671 | 0.2536 |
| PyroIC | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0037 | 0.0188 | 0.0321 |

Trace Elements by XRF

| ANALYTE | Percentiles | | | | | | |
|----------|-------------|--------|--------|--------|--------|--------|--------|
| | 5 | 10 | 25 | 50 | 75 | 90 | 95 |
| Aluminum | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0041 | 0.0080 |
| Antimony | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0014 | 0.0045 | 0.0076 |
| Arsenic | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0007 | 0.0011 |
| Barium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0006 | 0.0085 | 0.0201 |
| Bromine | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0008 | 0.0012 |
| Cadmium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0005 | 0.0021 | 0.0045 |
| Calcium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0013 | 0.0026 |
| Cerium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0031 | 0.0224 |
| Cesium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0027 | 0.0096 |

Table 5-9. (Continued)

Trace Elements by XRF (continued)

| ANALYTE | Percentiles | | | | | | |
|------------|-------------|--------|--------|--------|--------|--------|--------|
| | 5 | 10 | 25 | 50 | 75 | 90 | 95 |
| Chlorine | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0001 | 0.0007 | 0.0012 |
| Chromium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0011 | 0.0015 |
| Cobalt | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0006 | 0.0008 |
| Copper | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0009 | 0.0014 |
| Europium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0005 | 0.0009 |
| Gallium | 0.0000 | 0.0000 | 0.0000 | 0.0001 | 0.0005 | 0.0010 | 0.0016 |
| Gold | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0015 | 0.0034 |
| Hafnium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0008 | 0.0015 |
| Indium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0006 | 0.0025 | 0.0055 |
| Iridium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0010 | 0.0019 |
| Iron | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0011 | 0.0032 | 0.0070 |
| Lanthanum | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0055 | 0.0194 |
| Lead | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0013 | 0.0025 |
| Magnesium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0028 | 0.0090 |
| Manganese | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0008 | 0.0012 |
| Mercury | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0011 | 0.0027 |
| Molybdenum | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0019 |
| Nickel | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0008 | 0.0011 |
| Niobium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0008 | 0.0019 |
| Phosphorus | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0012 | 0.0021 |
| Potassium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0005 | 0.0013 | 0.0021 |
| Rubidium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0008 | 0.0011 |
| Samarium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0006 | 0.0012 |
| Scandium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0007 | 0.0013 |
| Selenium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0009 | 0.0013 |
| Silicon | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0006 | 0.0025 | 0.0142 |
| Silver | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0007 | 0.0025 | 0.0053 |
| Sodium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0079 | 0.0301 |
| Strontium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0010 | 0.0013 |
| Sulfur | 0.0000 | 0.0000 | 0.0000 | 0.0001 | 0.0014 | 0.0022 | 0.0044 |
| Tantalum | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0003 | 0.0014 | 0.0041 |
| Terbium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0009 | 0.0012 |
| Tin | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0012 | 0.0044 | 0.0084 |
| Titanium | 0.0000 | 0.0000 | 0.0000 | 0.0001 | 0.0005 | 0.0013 | 0.0016 |
| Vanadium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0009 | 0.0013 |
| Wolfram | 0.0000 | 0.0000 | 0.0000 | 0.0001 | 0.0005 | 0.0023 | 0.0049 |
| Yttrium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0004 | 0.0011 | 0.0015 |
| Zinc | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0007 | 0.0011 |
| Zirconium | 0.0000 | 0.0000 | 0.0000 | 0.0000 | 0.0002 | 0.0008 | 0.0017 |

Notes: All units are micrograms per cubic meter.

Trip and Field Blanks During 2005 -- For XRF analysis, some of the largest values (95 percentile) belong to sodium, silicon, and iron. Several additional elements such as cerium, lanthanum, etc. also have large 95th percentile values, but these are unlikely to be present, and probably represent XRF instrument artifacts. Sodium may be high because it is a light element which means that accurate determination by XRF is problematic. One of the samplers, the R&P speciation sampler, uses sodium carbonate in the denuder for the nylon filter channel, which could potentially cause sodium contamination. RTI is currently running tests to gather more information on this potential problem. Iron is also a potential contaminant in some of the sampler types that use metal modules or inlet hardware. Potential contamination by sodium and iron will be discussed in the next section.

5.4.1 Outliers by Sampler Type

Table 5-10 shows 95th percentile values for EC, iron, and sodium ion for the four different sampler types that were in use during 2005. These three analytes showed outliers that are probably related to materials used in the construction of the samplers' inlets or filter modules. These are shown in bold in the Table.

- Elemental carbon for the R&P 2300 (RPSPEC) sampler is higher than for other samplers. This phenomenon has been reported previously, and is suspected to be the result of the use of silicone stopcock grease in the size-selective impactor that is built into the sampling modules. Little can be done to alleviate this problem without a significant redesign of the R&P 2300 size-selective inlets. Data users should be aware of this problem and screen their EC data for unusually high values. RTI cannot screen for high EC values in the routine (exposed) filter data because natural variability is large enough to mask most of the EC outliers.
- Iron outliers for the MetOne SASS sampler which occur occasionally may be due to stainless steel module. Although the iron outliers are easy to detect in the blank data, it would be very difficult to screen for this artifact in the routine filter data because natural variability is large enough to mask most of the outliers.
- Sodium ion levels for the R&P 2300 sampler are significantly higher than other types, possibly because of sodium carbonate denuder. Other samplers use magnesium oxide denuders.

5.4.2 Trends and Offsets in Blank Data

Other than the isolated outliers identified in the previous section, no significant trends or offsets have been observed in the trip and field data for any of the STN analytes.

Table 5-10. Outliers by Sampler Type for Selected Analytes

| Analyte | Sampler | N | MEAN | MEDIAN | 95th Pct. | Notes |
|-------------------|---------------|-------------|---------------|---------------|---------------|-------|
| EC | MASS | 226 | 0.0043 | 0.0001 | 0.0245 | |
| EC | RAAS | 348 | 0.0125 | 0.0003 | 0.0564 | |
| EC | RPSPEC | 217 | 0.0424 | 0.0008 | 0.3250 | 1 |
| EC | SASS | 2641 | 0.0147 | 0.0005 | 0.0840 | |
| Iron | MASS | 225 | 0.0004 | 0.0001 | 0.0017 | |
| Iron | RAAS | 347 | 0.0006 | 0.0000 | 0.0021 | |
| Iron | RPSPEC | 218 | 0.0006 | 0.0001 | 0.0031 | |
| Iron | SASS | 2640 | 0.0026 | 0.0003 | 0.0091 | 2 |
| Sodium Ion | MASS | 225 | 0.0047 | 0.0023 | 0.0143 | |
| Sodium Ion | RAAS | 348 | 0.0086 | 0.0000 | 0.0333 | |
| Sodium Ion | RPSPEC | 217 | 0.1413 | 0.0267 | 0.6440 | 3 |
| Sodium Ion | SASS | 2641 | 0.0137 | 0.0000 | 0.0634 | |

Notes:

- 1 EC for RPSPEC sampler is higher than other samplers
- 2 Iron for the SASS module is higher than others possibly due to stainless steel module and low flow ratio
- 3 Sodium ion for the RPSPEC sampler is significantly higher than other types, possibly because of NaCO₃ denuder.

6.0 External Audits

6.1 Performance Evaluation Audit Results

The RTI Laboratories participated in two performance audits sponsored by EPA (Appendices C and D.) These were as follows:

| Author/Organization | Report Title, Date | Description | Conclusions/Findings |
|--|--|---|---|
| Jewell Smiley - EPA/NAREL (Appendix D) | Experimental Inter-comparison of Speciation Laboratories, September 19, 2005 | Intercomparison of four laboratories currently providing PM _{2.5} speciation results for gravimetric mass, IC, OC/EC, and XRF. | RTI results were acceptable. Issue of lack of comparability of XRF uncertainty estimates was noted. |
| Steve Taylor EPA/NAREL (Appendix C) | Gravimetric Inter-Laboratory Comparison Study, November 23, 2005 | Intercomparison of five different laboratories. | RTI results satisfactory. |

6.2 System Audit Results

On July 12, 2005, a laboratory audit was conducted by National Air and Radiation Environmental Laboratory (NAREL) personnel. The audit report was received on November 4, 2005. The US EPA audit team included Eric Boswell and Jewell Smiley, with Dennis Crumpler and Joann Rice from the Office of Air Quality Planning and Standards (OAQPS). Solomon Ricks and Jeff Lance were also present during the audit as EPA observers. The audit included interviews and observations in the following areas:

- Gravimetric Laboratory
- Organic Carbon/Elemental Carbon (OC/EC) Laboratory
- X-ray Fluorescence (XRF) Laboratory
- Ion Chromatography (IC) Laboratory
- Sample Handling and Archiving Laboratory (SHAL)
- Program Management
- Quality Assurance
- Data Management

There were three major findings in the audit report (Appendix E), which are summarized in Table 6-1, along with RTI comments.

Table 6-1. Systems Audit Findings and Responses

| No. | Audit Comment | RTI Response |
|-----|---|---|
| 1 | <p>Two Teflon® filters were removed from the SHAL inventory during the audit so that NAREL could experimentally re-measure the tare mass already determined at RTI's gravimetric lab. ... NAREL's tare mass was an alarming 30 micrograms smaller for one of the filters. Comment: This finding may be an indication of serious problems like the bad filter lot that was discovered several weeks before this audit. According to the corrective action report, the bad filter lot produced negative trip and field blanks. The questionable filter would have produced this effect if it had been utilized as a trip or field blank. RTI should continue to monitor the situation and explore potential reasons for the large variability in blank filters.</p> | <p>RTI recognized that there was a problem with a certain filter lot from Whatman, which was replaced at RTI's request. New filter reweighing procedures have been implemented which require 33+% replicate initial (tare) weighings, and 100% reweighings of exposed filters (postweighing).</p> |
| 2 | <p>All of the routine OC/EC duplicates are analyzed using the same instrument that performed the original analysis. This practice was acceptable in the past when the daily sucrose spikes were able to provide evidence of acceptable between-instrument performance. Now that OC subfractions are reported, there is no daily QC that provides the necessary assurance of acceptable between-instrument precision. Recommendation. RTI should schedule some of the routine OC/EC duplicates for analysis using a different instrument. For example, half of the scheduled duplicates could be analyzed using the same instrument, and the remaining duplicates could be analyzed using one of the available instruments that did not perform the original analysis.</p> | <p>The RTI OC/EC laboratory has implemented daily between-instrument precision checks.</p> |

Table 6-1. (Continued)

| No. | Audit Comment | RTI Response |
|------------|--|--|
| 3 | <p>As stated earlier, the focus of the XRF audit was to discuss those samples that RTI had analyzed as part of a recent inter-laboratory comparison study sponsored by NAREL ... Results from this study showed aluminum to be the most controversial element reported. This study also showed that RTI generally reported uncertainties which were lower than those reported by the other participating labs. A few spectra were inspected and discussed during the audit. Two specific spectra were selected to be included in the final report for the study. Ultimately the final report included examples of the controversial spectra from all of the labs. The spectra from RTI contain a significant [diffusion peak] interference for aluminum and silicon which was not observed in the spectra from the other labs.</p> <p>Comment: This observation may not be a problem for RTI's analysis since there is no standard method for calculating XRF uncertainties. However, RTI may want to take a closer look at the way uncertainties were calculated for aluminum and silicon during this study. EPA has recently initiated dialog with all of the speciation labs to learn more about the XRF analysis at each lab, and clearly there is diversity among the different labs. Any progress toward standardizing the XRF analysis is a positive step for the speciation program.</p> | <p>RTI is currently researching the question of XRF uncertainties for fine particle analysis. We have discovered that of the several major laboratories currently conducting such analyses (including Chester Labnet, Cooper Environmental Services, the University of California at Davis, the Desert Research Institute, and RTI) there is no uniform consensus on how uncertainties are to be calculated and what components of uncertainty are included. Dr. Bill Gutknecht, director of the RTI laboratory group, is preparing an analysis report that will make recommendations for XRF uncertainty reporting.</p> |

6.3 Synoptic Summary of 2005 Speciation Trends and IMPROVE Network Audits

During 2005, EPA performed field audits at 16 sites (19 total samplers). Appendix F provides results of these audits. *Note that this is preliminary data. As of March 1, 2006, the final audit report has not been received from EPA.*

7.0 List of References

7.1 List of Current STN Documents

| Type | Title | Date Revised | Author |
|------|---|--------------|--------------|
| SOP | Gravimetric Analysis | 7/8/2005 | Greene |
| SOP | Cleaning Nylon Filters Used for Collection of PM2.5 Material | 8/14/2003 | Hardison, E. |
| SOP | XRF Analysis of PM2.5 Deposits on Teflon Filters | 8/14/2003 | McWilliams |
| SOP | R&P Speciation Sampler Chemcomb Denuders with Sodium Carbonate | 8/14/2003 | Eaton |
| SOP | Coating and Extracting Annular Denuders with Sodium Carbonate | 8/14/2003 | Eaton |
| SOP | Coating Annular Denuders with XAD-4 Resin | 8/14/2003 | Eaton |
| SOP | Coating Aluminum Honeycomb Denuders with MgO | 8/14/2003 | Eaton |
| SOP | Sample Preparation and Analysis of PM20 and PM2.5 Samples by SEM | 8/14/2003 | Crankshaw |
| SOP | Coating Annular Denuders with MgO | 8/15/2003 | Eaton |
| SOP | Database Operations | 7/11/2005 | Rickman |
| SOP | Disaster Recovery Plan--RTI CONFIDENTIAL | 7/6/2005 | Rickman |
| SOP | Anion Analysis | 8/14/2003 | Hardison, E. |
| SOP | Cation Analysis | 8/14/2003 | Hardison, E. |
| SOP | Procurement and Acceptance Testing of Teflon, Nylon, and Quartz Filters | 7/7/2005 | Hardison, E. |
| SOP | Determination of Organic, Elemental, and Total Carbon in Particulate Matter Using a Thermal/Optical-Transmittance Carbon Analyzer | 8/14/2003 | Peterson |
| SOP | Sample Handling and Archiving Laboratory (SHAL) | 7/11/2005 | O'Rourke |
| SOP | Long-Term Archiving of PM2.5 Filters and Extracts | 7/5/2002 | Haas, C. |
| SOP | Assign Field Sample Flags for the Chemical Speciation Trends Network | 7/7/2005 | Wall, C. |
| SOP | Document Control and Storage | 7/6/2005 | Haas, D. |
| SOP | Thermal/Optical Reflectance Carbon Analysis of Aerosol Filter Samples | 6/1/2000 | DRI |
| SOP | Analysis of SVOC by GC/MS | 7/1/2003 | DRI |
| SOP | Analysis of Elements in Air Particulates by XRF (Kevex 770) | 7/3/2003 | Chester |
| SOP | Kevex XRF Spectrometer Calibration | 7/3/2003 | Chester |
| SOP | Kevex XRF Spectrometer Data Generation, Interpretation and Reporting Chester Labnet Proprietary Method | 10/17/2002 | Chester |
| SOP | Analysis of Elements in Air Particulates by XRF (Kevex 771) | 8/6/2003 | Chester |
| SOP | Sample Receipt and Log In | 11/18/2002 | Chester |
| QAPP | QAPP for PM2.5 of Chemical Speciation Samples | 9/11/2005 | RTI |
| Data | Semi-Annual Data Summary Report | 5/12/2005 | RTI |
| Data | Annual Data Summary Report | 2/28/2006 | RTI |

7.2 Special Reports Issued During the Reporting Period

| Type | Title | Date Revised | Author | Document No. |
|--------|--|--------------|--------|--------------------|
| Data | Semi-Annual Data Summary Report | 5/12/2005 | RTI | RTI/08858/03QAS |
| Data | Annual Data Summary Report | 2/28/2006 | RTI | RTI/08858/04QAS |
| Report | Teflon Filter Manufacturing Defects March - April 2005 | 7/8/2005 | RTI | RTI/08858/12/03S |
| Report | Tests of Acceptance of XRF Instrument #3 Operated by RTI | 11/1/05 | RTI | RTI/0208858/02/02D |

List of Appendices

- A. *Method Detection Limits***
 - B. *Data Completeness Summary***
 - C. *Gravimetric Inter-Laboratory Comparison Study*
*(NAREL 11/23/05)***
 - D. *Experimental Inter-Comparison of Speciation Laboratories*
*(NAREL 9/19/05)***
 - E. *Systems Audit Report*
*(NAREL 11/4/05)***
 - F. *Synoptic Summary of 2005 Speciation Trends and IMPROVE*
*Network Audits***
-

Appendix A

Method Detection Limits

Maximum Detection Limits by Analysis and Analyte

| Analysis | Analyte | Mass (µg) | Concentration (ug/m ³) by Sampler Type | | | |
|---|-------------------------|-----------|--|---------|---------|--------|
| | | | MASS | RASS | R and P | SASS |
| Cations - PM2.5 (NH ₄ , Na, K) | Ammonium | 0.16 | 0.0072 | 0.017 | 0.012 | 0.019 |
| Cations - PM2.5 (NH ₄ , Na, K) | Potassium | 0.13 | 0.006 | 0.014 | 0.0097 | 0.016 |
| Cations - PM2.5 (NH ₄ , Na, K) | Sodium | 0.29 | 0.013 | 0.031 | 0.021 | 0.034 |
| Mass - PM2.5 | Particulate matter 2.5u | 7.2 | 0.32 | 0.33 | 0.31 | 0.83 |
| Nitrate - PM2.5 | Nitrate | 0.084 | | 0.0089 | 0.0061 | 0.0098 |
| Nitrate - PM2.5 (MASS/nylon) | Nitrate | 0.084 | 0.0038 | | | |
| Nitrate - PM2.5 (MASS/Teflon) | Nitrate | 0.084 | 0.0038 | | | |
| Organic and elemental carbon | Elemental carbon | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Organic and elemental carbon | Organic carbon | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Organic and elemental carbon | Pk1_OC | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Organic and elemental carbon | Pk2_OC | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Organic and elemental carbon | Pk3_OC | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Organic and elemental carbon | Pk4_OC | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Organic and elemental carbon | PyroIC | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Organic and elemental carbon | Total carbon | 2.4 | 0.10 | 0.24 | 0.17 | 0.27 |
| Sulfate - PM2.5 | Sulfate | 0.12 | 0.0054 | 0.013 | 0.0087 | 0.014 |
| Trace elements | Aluminum | 0.22 | 0.0099 | 0.0098 | 0.0094 | 0.025 |
| Trace elements | Antimony | 0.38 | 0.017 | 0.017 | 0.016 | 0.042 |
| Trace elements | Arsenic | 0.037 | 0.00065 | 0.0016 | 0.0016 | 0.0041 |
| Trace elements | Barium | 0.85 | 0.0046 | 0.039 | 0.037 | 0.097 |
| Trace elements | Bromine | 0.031 | 0.00057 | 0.0014 | 0.0013 | 0.0034 |
| Trace elements | Cadmium | 0.17 | 0.0073 | 0.0076 | 0.0071 | 0.019 |
| Trace elements | Calcium | 0.073 | 0.0032 | 0.0033 | 0.0031 | 0.0082 |
| Trace elements | Cerium | 1.2 | 0.0041 | 0.056 | 0.054 | 0.14 |
| Trace elements | Cesium | 0.53 | 0.0048 | 0.024 | 0.023 | 0.061 |
| Trace elements | Chlorine | 0.13 | 0.0035 | 0.0058 | 0.0057 | 0.015 |
| Trace elements | Chromium | 0.025 | 0.0011 | 0.0011 | 0.0011 | 0.0028 |
| Trace elements | Cobalt | 0.02 | 0.00061 | 0.00091 | 0.00087 | 0.0023 |
| Trace elements | Copper | 0.024 | 0.0011 | 0.0011 | 0.0011 | 0.0028 |
| Trace elements | Europium | 0.16 | 0.0021 | 0.0074 | 0.007 | 0.018 |
| Trace elements | Gallium | 0.071 | 0.0012 | 0.0031 | 0.003 | 0.0079 |
| Trace elements | Gold | 0.13 | 0.0017 | 0.0056 | 0.0054 | 0.014 |
| Trace elements | Hafnium | 0.38 | 0.012 | 0.017 | 0.016 | 0.043 |
| Trace elements | Indium | 0.16 | 0.0067 | 0.0074 | 0.0071 | 0.019 |
| Trace elements | Iridium | 0.17 | 0.0018 | 0.0073 | 0.0071 | 0.019 |
| Trace elements | Iron | 0.028 | 0.00092 | 0.0013 | 0.0012 | 0.0032 |
| Trace elements | Lanthanum | 1 | 0.0038 | 0.046 | 0.044 | 0.11 |
| Trace elements | Lead | 0.085 | 0.0012 | 0.0038 | 0.0037 | 0.0096 |
| Trace elements | Magnesium | 0.43 | 0.0079 | 0.019 | 0.018 | 0.048 |
| Trace elements | Manganese | 0.033 | 0.00081 | 0.0015 | 0.0014 | 0.0038 |
| Trace elements | Mercury | 0.065 | 0.0015 | 0.0029 | 0.0028 | 0.0073 |
| Trace elements | Molybdenum | 0.085 | 0.0037 | 0.0037 | 0.0036 | 0.0095 |

Maximum Detection Limits by Analysis and Analyte

| Analysis | Analyte | Mass (µg) | Concentration (ug/m ³) by Sampler Type | | | |
|----------------|------------|-----------|--|---------|---------|--------|
| | | | MASS | RASS | R and P | SASS |
| Trace elements | Nickel | 0.018 | 0.00074 | 0.00082 | 0.00078 | 0.002 |
| Trace elements | Niobium | 0.067 | 0.0015 | 0.003 | 0.0029 | 0.0075 |
| Trace elements | Phosphorus | 0.15 | 0.0068 | 0.0071 | 0.0066 | 0.017 |
| Trace elements | Potassium | 0.11 | 0.0048 | 0.0048 | 0.0046 | 0.012 |
| Trace elements | Rubidium | 0.031 | 0.00084 | 0.0014 | 0.0014 | 0.0036 |
| Trace elements | Samarium | 0.089 | 0.0021 | 0.004 | 0.0039 | 0.01 |
| Trace elements | Scandium | 0.12 | 0.0055 | 0.0055 | 0.0052 | 0.014 |
| Trace elements | Selenium | 0.033 | 0.0011 | 0.0014 | 0.0014 | 0.0037 |
| Trace elements | Silicon | 0.18 | 0.008 | 0.008 | 0.0077 | 0.021 |
| Trace elements | Silver | 0.15 | 0.0055 | 0.0069 | 0.0066 | 0.017 |
| Trace elements | Sodium | 1.6 | 0.024 | 0.07 | 0.068 | 0.18 |
| Trace elements | Strontium | 0.036 | 0.001 | 0.0016 | 0.0016 | 0.0041 |
| Trace elements | Sulfur | 0.2 | 0.0042 | 0.009 | 0.0087 | 0.023 |
| Trace elements | Tantalum | 0.28 | 0.0033 | 0.013 | 0.012 | 0.032 |
| Trace elements | Terbium | 0.11 | 0.0019 | 0.0049 | 0.0047 | 0.012 |
| Trace elements | Tin | 0.26 | 0.0087 | 0.012 | 0.011 | 0.029 |
| Trace elements | Titanium | 0.051 | 0.0022 | 0.0023 | 0.0022 | 0.0057 |
| Trace elements | Vanadium | 0.037 | 0.0016 | 0.0017 | 0.0016 | 0.0041 |
| Trace elements | Wolfram | 0.21 | 0.0027 | 0.0092 | 0.0089 | 0.023 |
| Trace elements | Yttrium | 0.044 | 0.00096 | 0.002 | 0.0019 | 0.005 |
| Trace elements | Zinc | 0.025 | 0.0011 | 0.0011 | 0.0011 | 0.0029 |
| Trace elements | Zirconium | 0.054 | 0.0014 | 0.0024 | 0.0023 | 0.006 |

1. Individual laboratory instruments used for analysis may have differing MDL values. The maximum values are shown to permit comparison of detection limits among differing species.
2. Concentration detection limits vary among sampler types due to differing sample volumes.

Appendix B

Data Completeness Summary

**Table B-1 Total Number of Sampling Events
Included in each Reporting Batch**

Sampling Events by Report Batch

| Report Batch | Samples | Blanks | | Total |
|-----------------|---------------|--------------|--------------|---------------|
| | | Field | Trip | |
| 60 | 1,460 | 212 | 37 | 1,709 |
| 61 | 1,274 | 79 | 206 | 1,559 |
| 62 | 1,387 | 279 | 33 | 1,699 |
| 63 | 1,312 | 211 | 48 | 1,571 |
| 64 | 1,384 | 77 | 279 | 1,740 |
| 65 | 1,387 | 2 | 57 | 1,446 |
| 66 | 1,380 | 290 | 68 | 1,738 |
| 67 | 1,479 | 72 | 107 | 1,658 |
| 68 | 1,293 | 212 | 53 | 1,558 |
| 69 | 1,464 | 216 | 36 | 1,716 |
| 70 | 1,490 | 5 | 261 | 1,756 |
| 71 | 1,385 | 282 | 29 | 1,696 |
| 72 | 1,342 | 74 | 81 | 1,497 |
| Total | 18,037 | 2,011 | 1,295 | 21,343 |

Table B-2 Total Number of Records Delivered by Type

Records Posted by Report Batch

| Report Batch | Samples | Blanks | | Total |
|-----------------|------------------|----------------|----------------|------------------|
| | | Field | Trip | |
| 60 | 166,980 | 24,312 | 4,269 | 195,561 |
| 61 | 145,712 | 8,913 | 23,452 | 178,077 |
| 62 | 158,716 | 31,881 | 3,764 | 194,361 |
| 63 | 150,085 | 24,191 | 5,454 | 179,730 |
| 64 | 157,989 | 8,681 | 31,975 | 198,645 |
| 65 | 158,643 | 226 | 6,495 | 165,364 |
| 66 | 157,946 | 33,162 | 7,660 | 198,768 |
| 67 | 169,044 | 8,128 | 12,124 | 189,296 |
| 68 | 148,011 | 24,362 | 6,037 | 178,410 |
| 69 | 167,387 | 24,782 | 4,055 | 196,224 |
| 70 | 170,683 | 548 | 29,967 | 201,198 |
| 71 | 158,407 | 32,223 | 3,298 | 193,928 |
| 72 | 153,303 | 8,348 | 9,396 | 171,047 |
| Total | 2,062,906 | 229,757 | 147,946 | 2,440,609 |

Table B-3. Percentage of Routine Exposure Records – STN Sites

Monthly Percent Data Completeness by Site – STN Sites

| Location | AQS Site | POC | Sampler Type | Report Batch | | | | | | | | | |
|---|-----------|-----|--------------|--------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| | | | | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 |
| 20th St. Fire Station | 120861016 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 63.6 | 100.0 | 84.3 | 100.0 |
| Allen Park | 261630001 | 5 | SASS | 100.0 | 90.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 90.0 | 90.0 |
| Bakersfield-California Ave | 060290014 | 5 | SASS | 88.9 | 90.0 | 0.0 | 71.4 | 70.0 | 88.9 | 93.4 | 75.0 | 100.0 | 88.9 |
| Bakersfield-California Ave (Collocated) | 060290014 | 6 | SASS | 88.9 | 90.0 | 0.0 | 37.5 | | | | | 100.0 | 79.0 |
| Beacon Hill | 530330080 | 6 | MASS | 99.3 | 99.6 | 99.7 | 92.6 | 80.0 | 100.0 | 100.0 | 88.9 | 100.0 | 100.0 |
| Blair Street | 295100085 | 6 | SASS | 100.0 | 100.0 | 100.0 | 93.4 | 100.0 | 100.0 | 100.0 | 99.2 | 93.1 | 90.0 |
| Burlington | 500070012 | 5 | SASS | 75.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Capitol | 220330009 | 5 | MASS | 0.0 | 100.0 | 88.8 | 97.7 | 88.8 | 76.3 | 91.8 | 100.0 | 88.9 | 79.9 |
| Chamizal | 481410044 | 5 | MASS | 90.0 | 100.0 | 80.1 | 34.8 | 55.1 | 45.1 | 100.0 | 100.0 | 100.0 | 90.0 |
| Chicopee | 250130008 | 5 | SASS | 72.7 | 100.0 | 99.2 | 93.4 | 90.0 | 100.0 | 99.9 | 99.3 | 99.7 | 100.0 |
| Com ED | 170310076 | 5 | MASS | 86.5 | 88.9 | 100.0 | 100.0 | 100.0 | 100.0 | 92.1 | 100.0 | 100.0 | 99.8 |
| Commerce City | 080010006 | 5 | SASS | 100.0 | 100.0 | 100.0 | 91.0 | 85.7 | 100.0 | 92.7 | 100.0 | 100.0 | 100.0 |
| CPW | 450190049 | 5 | SASS | 100.0 | 92.7 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 81.8 | 99.9 |
| Criscuolo Park | 090090027 | 5 | SASS | 85.7 | 99.7 | 60.1 | 25.7 | 77.8 | 77.8 | 72.7 | 66.7 | 100.0 | 88.9 |
| Deer Park | 482011039 | 6 | MASS | 100.0 | 100.0 | 83.2 | 99.1 | 100.0 | 90.9 | 98.1 | 100.0 | 100.0 | 88.9 |
| Deer Park (Collocated) | 482011039 | 7 | MASS | 42.1 | 25.1 | 69.0 | 83.7 | 100.0 | | | | 100.0 | 88.9 |
| Dover | 100010003 | 5 | SASS | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 85.4 | 100.0 | 80.0 | 100.0 |
| El Cajon | 060730003 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Elizabeth Lab | 340390004 | 5 | SASS | 84.2 | 99.3 | 75.0 | 100.0 | 99.6 | 99.1 | 99.3 | 99.6 | 100.0 | 86.2 |
| Fairbanks State Bldg | 020900010 | 6 | SASS | | | | 100.0 | 100.0 | 100.0 | 100.0 | 99.8 | | 83.3 |
| Fargo NW | 380171004 | 5 | SASS | 100.0 | 100.0 | 99.8 | 93.4 | 100.0 | 99.7 | 93.4 | 83.7 | 100.0 | 100.0 |
| Fresno - First Street | 060190008 | 5 | SASS | 85.2 | 86.2 | 99.2 | 84.3 | 100.0 | 100.0 | 100.0 | 90.7 | 100.0 | 100.0 |
| G.T. Craig | 390350060 | 5 | SASS | 100.0 | 99.8 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| G.T. Craig - Collocated | 390350060 | 6 | SASS | 83.3 | 88.9 | 87.5 | 100.0 | 89.6 | 100.0 | 100.0 | 61.0 | 100.0 | 100.0 |
| Garinger High School | 371190041 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 98.4 | 100.0 | 100.0 | 100.0 |
| Guaynabo | 720610005 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 90.0 | 100.0 | 90.4 | 85.9 | 90.9 | 90.0 |
| Gulfport | 280470008 | 5 | SASS | 100.0 | 100.0 | 88.6 | 98.0 | 99.8 | 100.0 | 99.9 | 100.0 | 44.4 | |
| Guthrie | 471570047 | 5 | RAAS | 100.0 | 100.0 | 98.8 | 100.0 | 100.0 | 100.0 | 100.0 | 99.2 | 60.5 | 99.3 |

Monthly Percent Data Completeness by Site – STN Sites

| Location | AQS Site | POC | Sampler Type | Report Batch | | | | | | | | | |
|---------------------------------|-----------|-----|--------------|--------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| | | | | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 |
| Hawthorne | 490353006 | 5 | SASS | 90.0 | 99.9 | 100.0 | 90.0 | 99.0 | 90.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Henrico Co. | 510870014 | 5 | SASS | 100.0 | 88.9 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Hinton | 481130069 | 5 | MASS | 100.0 | 99.8 | 100.0 | 100.0 | 100.0 | 87.5 | 100.0 | 100.0 | 100.0 | 100.0 |
| JFK Center | 202090021 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 81.7 | 87.5 | 100.0 | 85.7 | 100.0 | 87.5 |
| Lawrenceville | 420030008 | 6 | SASS | 100.0 | 100.0 | 87.5 | 92.7 | 87.5 | 100.0 | 100.0 | 100.0 | 90.0 | 100.0 |
| Lindon | 490494001 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 98.5 | 100.0 | 100.0 | 100.0 | 100.0 |
| McMillan Reservoir | 110010043 | 5 | RAAS | 57.1 | 88.9 | 100.0 | 100.0 | 98.7 | 100.0 | 100.0 | 85.7 | 100.0 | 75.0 |
| Missoula County Health Dept. | 300630031 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 88.9 | 100.0 | 82.7 |
| MLK | 100032004 | 5 | SASS | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 99.4 | 100.0 | 100.0 | 100.0 | 100.0 |
| Nampa NNC | 160270004 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 99.7 | 100.0 | 100.0 | 100.0 | 90.0 |
| New Brunswick | 340230006 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| New Brunswick (Collocated) | 340230006 | 6 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 40.0 | 100.0 | 100.0 |
| North Birmingham | 010730023 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 99.9 | 100.0 | 100.0 | 90.8 | 100.0 |
| NY Botanical Gardens | 360050083 | 6 | SASS | 90.0 | 99.9 | 99.8 | 99.9 | 99.9 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Peoria Site 1127 | 401431127 | 5 | SASS | 88.9 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 75.0 |
| PHILA - AMS Laboratory | 421010004 | 7 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 90.0 |
| Philips | 270530963 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 99.1 | 100.0 | 100.0 | 90.0 |
| Phoenix Supersite | 040139997 | 7 | SASS | 90.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 90.9 | 100.0 | 93.4 | 99.9 |
| Portland N. Roselawn | 410510246 | 6 | SASS | 100.0 | 99.3 | 88.9 | 100.0 | 100.0 | 99.7 | 100.0 | 100.0 | 100.0 | 100.0 |
| Portsmouth | 330150014 | 5 | RAAS | 100.0 | 99.0 | 100.0 | 90.0 | 100.0 | 97.9 | 90.9 | 88.9 | 100.0 | 100.0 |
| Reno | 320310016 | 5 | SASS | 88.9 | 100.0 | 100.0 | 100.0 | 100.0 | 99.9 | 100.0 | 100.0 | 100.0 | 100.0 |
| Riverside-Rubidoux | 060658001 | 5 | SASS | 100.0 | 90.0 | 100.0 | 90.9 | 100.0 | 100.0 | 100.0 | 87.5 | 99.8 | 99.8 |
| Riverside-Rubidoux (Collocated) | 060658001 | 6 | SASS | 100.0 | 82.7 | 100.0 | | | | | | | |
| Roxbury (Boston) | 250250042 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Roxbury (Boston) - collocated | 250250042 | 6 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Sacramento - Del Paso Manor | 060670006 | 5 | SASS | 100.0 | 100.0 | 98.8 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| San Jose - Jackson Street | 060850005 | 5 | SASS | 100.0 | 88.9 | 100.0 | 100.0 | 87.5 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| SER-DNR Headquarters | 550790026 | 5 | SASS | 100.0 | 80.0 | 100.0 | 100.0 | 99.9 | 100.0 | 100.0 | 88.9 | 100.0 | 100.0 |
| Simi Valley | 061112002 | 5 | SASS | 80.0 | 100.0 | 100.0 | 40.0 | 40.0 | 100.0 | 100.0 | 80.0 | 100.0 | 65.9 |
| South DeKalb | 130890002 | 5 | RAAS | 100.0 | 90.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 87.7 | 100.0 | 100.0 |
| Springfield Pumping Station | 170310057 | 5 | RAAS | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 99.4 | 98.5 | 100.0 | 98.5 | 96.3 |

Monthly Percent Data Completeness by Site – STN Sites

| Location | AQS Site | POC | Sampler Type | Report Batch | | | | | | | | | |
|----------------------------------|-----------|-----|--------------|--------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| | | | | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 |
| Sydney | 120573002 | 5 | SASS | 100.0 | 100.0 | 99.8 | 100.0 | 100.0 | 100.0 | 100.0 | 98.0 | 99.0 | 99.3 |
| Urban League | 440070022 | 5 | RAAS | 97.9 | 66.7 | 100.0 | 100.0 | 87.5 | 100.0 | 90.0 | 100.0 | 90.0 | 100.0 |
| Washington Park | 180970078 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 86.2 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Woolworth St | 310550019 | 5 | SASS | 96.4 | 96.0 | 96.8 | 86.2 | 97.0 | 84.9 | 96.9 | 97.0 | 97.0 | 84.9 |
| WV - Guthrie Agricultural Center | 540390011 | 5 | SASS | 100.0 | 100.0 | 100.0 | 90.9 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |

Table B-4. Percentage of Routine Exposure Records – Non-STN Sites

Monthly Percent Data Completeness by Site – Non-STN Sites

[illegible]

Monthly Percent Data Completeness by Site – Non-STN Sites

| Location | AQS Site | POC | Sampler Type | Report Batch | | | | | | | | | |
|----------------------------------|-----------|-----|--------------|--------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| | | | | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 |
| Chesterfield | 450250001 | 5 | SASS | 98.5 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 82.1 |
| Chickasaw | 010970003 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 77.9 | 100.0 | 80.0 | 100.0 |
| Children's Park | 040191028 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 |
| Chiwaukee Prairie Site | 550590019 | 5 | SASS | 100.0 | 100.0 | 100.0 | 85.4 | 100.0 | 100.0 | 100.0 | 100.0 | 77.9 | 100.0 |
| Clio | 010050002 | 5 | SASS | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Columbus | 132150011 | 5 | RAAS | 100.0 | 100.0 | 79.7 | 41.8 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Courthouse Annex-Libby | 300530018 | 5 | SASS | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Covington - University College | 211170007 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 |
| Crossett | 050030005 | 5 | SASS | 100.0 | 80.0 | 100.0 | 60.0 | 100.0 | 0.0 | | | | |
| Crown Z | 530630016 | 5 | RAAS | 100.0 | 100.0 | 60.0 | 80.0 | 100.0 | 100.0 | 100.0 | 85.4 | 100.0 | 100.0 |
| Dearborn | 261630033 | 5 | SASS | 100.0 | 80.0 | 100.0 | 100.0 | 81.8 | 79.4 | 80.0 | 100.0 | 100.0 | 100.0 |
| Decatur | 011030011 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 97.9 | 80.0 | 100.0 |
| Del Norte | 350010023 | 5 | R & P 2300 | 100.0 | 100.0 | 85.4 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 |
| Douglas | 130690002 | 5 | RAAS | 100.0 | 100.0 | 97.9 | 85.4 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Duwamish | 530330057 | 6 | RAAS | 100.0 | 85.4 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Elkhart Pierre Moran | 180390003 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 85.4 | 100.0 | 100.0 | 100.0 |
| Ellis County WMA | 400450890 | 5 | SASS | 100.0 | 100.0 | 99.7 | 100.0 | 80.0 | 60.0 | 80.0 | 100.0 | 100.0 | 100.0 |
| Ellyson | 120330004 | 6 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 84.8 |
| Elmwood | 421010136 | 5 | SASS | 77.0 | 99.7 | 60.0 | 80.0 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 |
| Erie | 420490003 | 5 | SASS | 54.3 | 100.0 | 100.0 | 100.0 | 81.8 | 85.4 | 100.0 | 100.0 | 100.0 | 100.0 |
| Essex - Met One | 240053001 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 87.5 | 100.0 | 90.0 | 100.0 | 100.0 | 100.0 |
| Evansville - Mill Road | 181630012 | 5 | SASS | 80.0 | 80.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 41.5 | 100.0 | 83.3 |
| Florence | 421255001 | 5 | SASS | 80.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 |
| Fort Wayne CAAP | 180030004 | 5 | SASS | 100.0 | | | | | | | | | |
| Freemansburg | 420950025 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Gary litri | 180890022 | 5 | SASS | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 66.7 |
| General Hospital | 390870010 | 5 | SASS | 100.0 | 100.0 | 99.7 | 100.0 | 80.0 | 100.0 | 100.0 | 9.0 | 85.4 | 83.3 |
| Grand Junction - Powell Building | 080770017 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Grand Rapids | 260810020 | 5 | SASS | 99.7 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 80.0 | 83.3 |
| Greensburg | 421290008 | 5 | SASS | 100.0 | 99.7 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |

Monthly Percent Data Completeness by Site – Non-STN Sites

| Location | AQS Site | POC | Sampler Type | Report Batch | | | | | | | | | |
|--------------------------------------|-----------|-----|--------------|--------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| | | | | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 |
| Grenada | 280430001 | 5 | SASS | 0.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 80.0 |
| Hammond Purdue | 180892004 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Harrisburg | 420430401 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 68.2 |
| Hattie Avenue | 370670022 | 5 | SASS | 100.0 | 99.7 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Hattiesburg | 280350004 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 99.7 | 99.7 | 99.7 | 99.1 | 60.0 | 99.3 |
| Haynes Pt. | 110010042 | 6 | RAAS | 100.0 | 100.0 | 98.7 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Hazard - Perry County Horse Park | 211930003 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 97.9 | 100.0 | 100.0 |
| Head Start | 390990014 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Hendersonville | 471650007 | 5 | SASS | 80.0 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 83.3 |
| Hickory | 370350004 | 5 | SASS | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 |
| Holland | 260050003 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Houghton Lake | 261130001 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 80.0 | 80.0 | 100.0 | 100.0 |
| HU-Beltsville | 240330030 | 5 | RAAS | 100.0 | 100.0 | 80.0 | 100.0 | 80.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Huntsville Old Airport | 010890014 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 81.8 | 100.0 | 100.0 |
| IL - Decatur | 171150013 | 5 | SASS | 83.3 | 97.9 | 96.4 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| IS 52 | 360050110 | 5 | R & P 2300 | 100.0 | 99.3 | 100.0 | 99.1 | 98.2 | 100.0 | 89.8 | 87.7 | 98.9 | 99.9 |
| Jackson Hinds Co. | 280490018 | 5 | SASS | 100.0 | 100.0 | 99.7 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 40.0 | 100.0 |
| Jasper Post Office | 180372001 | 5 | SASS | 50.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Jefferson Elementary (10th and Vine) | 191630015 | 5 | R & P 2300 | 100.0 | 90.0 | 100.0 | 90.9 | 100.0 | 92.7 | 100.0 | 88.9 | 45.5 | 30.0 |
| Kalamazoo | 260770008 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Kelo | 460990006 | 5 | SASS | 99.7 | 98.5 | 95.5 | 95.5 | 95.5 | 95.5 | 95.5 | 97.6 | 100.0 | 100.0 |
| Kingsport | 471631007 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 83.3 |
| Lake Forest Park | 530330024 | 6 | RAAS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 99.7 | 80.0 | 100.0 | 100.0 | 100.0 |
| Lancaster | 420710007 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Laurel | 280670002 | 5 | SASS | 100.0 | 80.0 | 100.0 | 85.4 | 100.0 | 100.0 | 100.0 | 100.0 | 60.0 | 100.0 |
| Lawrence County | 470990002 | 5 | SASS | 100.0 | 100.0 | 83.3 | 80.0 | 100.0 | 80.0 | 80.0 | 100.0 | 100.0 | 100.0 |
| Lenoir Community College | 371070004 | 5 | SASS | 25.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Lexington Health Department | 210670012 | 5 | SASS | 100.0 | 97.9 | 100.0 | 100.0 | 100.0 | 77.0 | 100.0 | 80.0 | 100.0 | 100.0 |
| Lexington (NC) | 370570002 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Liberty | 290470005 | 5 | R & P 2300 | 100.0 | 90.0 | 100.0 | 100.0 | 100.0 | 100.0 | 84.3 | 77.8 | 100.0 | 100.0 |

Monthly Percent Data Completeness by Site – Non-STN Sites

| Location | AQS Site | POC | Sampler Type | Report Batch | | | | | | | | | |
|-------------------------------|-----------|-----|--------------|--------------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| | | | | 61 | 62 | 63 | 64 | 65 | 66 | 67 | 68 | 69 | 70 |
| Pearl City | 150032004 | 5 | SASS | 100.0 | 80.0 | 40.0 | 80.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 99.5 |
| PerkinstownCASNET | 551198001 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Perry County | 420990301 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Pinnacle State Park | 361010003 | 5 | R & P 2300 | 80.0 | 100.0 | 88.9 | 81.8 | 100.0 | 88.8 | 80.9 | 99.8 | 100.0 | 99.3 |
| Platteville | 081230008 | 5 | SASS | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 99.7 | 80.0 | 100.0 | 100.0 | 100.0 |
| Pleasant Green (Central MO) | 290530001 | 5 | R & P 2300 | 80.0 | 60.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 83.3 |
| Providence | 010731009 | 5 | SASS | 100.0 | 98.5 | 100.0 | 100.0 | 100.0 | 100.0 | 97.9 | 100.0 | 80.0 | 100.0 |
| Public Health Building | 191530030 | 5 | R & P 2300 | 100.0 | 100.0 | 80.0 | 100.0 | 80.0 | 80.0 | 100.0 | 80.0 | 100.0 | 100.0 |
| Queens College | 360810124 | 6 | R & P 2300 | 90.9 | 65.4 | 83.7 | 93.4 | 97.0 | 100.0 | 89.4 | 98.7 | 81.8 | 99.9 |
| RBD | 080410011 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 99.7 | 100.0 |
| Roanoke | 517700014 | 5 | SASS | 100.0 | | | | | | | | | |
| Rochester Primary | 360551007 | 5 | R & P 2300 | 91.0 | 80.0 | 100.0 | 80.9 | 97.6 | 79.0 | 81.8 | 88.9 | 100.0 | 99.9 |
| Rockwell | 371590021 | 5 | SASS | | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 97.9 | 100.0 | 100.0 |
| Rome | 131150005 | 5 | RAAS | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Rossville | 132950002 | 5 | RAAS | | | | 69.8 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Sault Ste Marie | 260330901 | 5 | SASS | 100.0 | 89.9 | 100.0 | 100.0 | 100.0 | 100.0 | 99.8 | 100.0 | 81.8 | 100.0 |
| Savannah | 130510017 | 5 | RAAS | | | | | | | | | | |
| Scranton | 420692006 | 5 | SASS | 60.0 | 80.0 | 100.0 | 77.9 | 80.0 | 100.0 | 100.0 | 100.0 | 65.4 | 100.0 |
| Searcy | 051450001 | 5 | SASS | 100.0 | 100.0 | 100.0 | 60.0 | 100.0 | 0.0 | | | | |
| Senior Center | 040137020 | 5 | SASS | | 100.0 | 100.0 | 100.0 | 100.0 | 99.7 | 100.0 | 100.0 | 100.0 | 100.0 |
| Shenandoah High School | 180650003 | 5 | SASS | 100.0 | 100.0 | 100.0 | 80.0 | 100.0 | 100.0 | 100.0 | 85.4 | 97.9 | 83.3 |
| Shreveport Airport | 220150008 | 5 | MASS | 100.0 | 100.0 | 100.0 | 97.9 | 82.1 | 5.1 | 100.0 | 82.1 | 73.0 | 100.0 |
| Skyview | 121030026 | 5 | SASS | 98.5 | 98.0 | 100.0 | 100.0 | 87.5 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| South Bend CAAP | 181411008 | 5 | SASS | 100.0 | | | | | | | | | |
| South Charleston Library | 540391005 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 |
| Southwick Community Center | 211110043 | 5 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 83.3 |
| Spring Hill Elementary School | 470931020 | 5 | RAAS | 100.0 | 100.0 | 100.0 | 97.9 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 83.3 |
| St Johns | 040137003 | 5 | SASS | | 100.0 | 100.0 | 100.0 | 81.5 | 80.0 | 81.7 | 100.0 | 100.0 | 100.0 |
| St Theo | 390350038 | 6 | SASS | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 100.0 | 72.6 |
| St. Paul Harding | 271230871 | 5 | SASS | 80.0 | 100.0 | 100.0 | 100.0 | | | | | | |

Monthly Percent Data Completeness by Site – Non-STN Sites

[illegible]

Appendix C
***Gravimetric Inter-Laboratory
Comparison Study***

TECHNICAL MEMORANDUM



TO: Michael Papp / OAQPS
FROM: Eric Boswell / NAREL
COPY: Dennis Crumper / OAQPS
Dr. R.K.M. Jayanty / RTI
Robert Mosley / R&IE-LV
Greg Noah / Region 4
Christopher Hall/Region 10
AUTHOR: Steve Taylor
DATE: November 23, 2005
SUBJECT: Gravimetric Inter-Laboratory Comparison Study

Introduction

A gravimetric study has been conducted at the National Air and Radiation Environmental Laboratory (NAREL) to compare the performance of EPA weighing laboratories that perform PM_{2.5} mass measurements. This was the final gravimetric performance study scheduled for 2005.

Participants of this study included the Region 4 Laboratory in Athens, GA; the Region 10 contract laboratory (Manchester Laboratory) in Washington; the Radiation and Indoor Environments Laboratory (R&IE) in Las Vegas, NV; and Research Triangle Institute (RTI) in Research Triangle Park (RTP), NC. The Region 4 and Region 10 laboratories provide pre-weighing and post-weighing of filters for the PM_{2.5} Performance Evaluation Program (PEP). The R&IE Laboratory provides the PM_{2.5} gravimetric analysis for the Tribal Air Monitoring Support (TAMS) program. The RTI Laboratory facility serves as EPA's primary contractor providing laboratory services to support the PM_{2.5} Speciation air monitoring network. RTI participated in this study because of additional contract work being performed as part of the Hurricane Katrina clean-up effort. NAREL coordinated this study by supplying Performance Evaluation (PE) samples and served as the reference laboratory. All laboratories participating in this study are equipped with environmentally controlled weighing chambers and microbalances capable of mass measurements of one microgram sensitivity.

Mass determination of PM_{2.5} typically proceeds by weighing the Teflon⁷ collection filter before and after the sampling event. The amount of Particulate Matter (PM_{2.5}) captured onto the surface of the filter can be calculated by a simple subtraction of the tare weight from the loaded filter weight. In order to accurately measure particulate mass at microgram levels, the microbalance must be located in a clean, dust free environmental chamber with precise temperature and humidity control. Elimination of static from samples is also very important for accurate mass measurements.

Samples for this study were created at NAREL using Met One SASS air samplers to collect various amounts of PM_{2.5} onto Teflon⁷ filters that were previously tared by all laboratories. Blank filter samples were included as controls to provide information about filter contamination and stability of mass loading. Metallic weights were also included as samples to provide information concerning balance stability and calibration. This study compares captured mass determined by NAREL to captured mass determined by each of the participating laboratories.

Acceptance criteria for this type of comparison have not been established. There are PEP criteria established for laboratory and field blanks, and metallic standards. Laboratory and field blanks should not vary by more than 0.015 mg and 0.030 mg respectively between pre- and post-sampling. Metallic standards should not vary by more than 0.003 mg. Previous NAREL gravimetric studies have used the PEP criteria as a guideline to measure laboratory performance. As an alternative to the PEP criteria, this study uses criteria based on actual mass data compiled from gravimetric PE studies administered by NAREL.

Experimental

To begin this study, each of the four participating laboratories was provided a set of samples consisting of ten new Teflon⁷ filters and two metallic weights. Filters and weights were held in individual labeled petrislides. The metallic weights were commercially available 100 and 200 milligram stainless steel weights that were slightly altered by clipping a small corner section from each weight. Sample sets were shipped to each laboratory with instructions to equilibrate and tare the samples following their standard operating procedures for the determination of PM_{2.5} mass. The sample sets were then returned to NAREL and placed into the weighing chamber for equilibration and determination of NAREL's tare mass. After the NAREL tare masses were established for all samples, seven of the ten filters from each of the sets were loaded with PM_{2.5} collected from the ambient air at NAREL. The remaining three filters from each set were utilized as blanks.

Teflon⁷ filters were loaded with PM_{2.5} mass using two co-located Met One Super SASS air samplers. Each sampler has four flow controlled channels available to load up to eight replicate samples. To insure that mass loads were similar for each lab, filters were loaded in replicate using four different sampling events. Event one sampled for 48 hours to create eight replicates. The next two events collected air for 24 and 20 hours respectively. The fourth event, using one sampler, collected air for sixteen hours to produce four replicate samples. Sampling events are summarized in Table 7. Following sample collection, filters were returned to the weighing chamber at NAREL to equilibrate and to determine the loaded mass as well as a final mass for the remaining blank filters and the metallic weights. Several weigh sessions during the week following sample collection were conducted to insure the mass stability of the filters. The last weigh session before shipping the filters to the sites became NAREL's Aofficial® loaded mass.


Immediately after a final Aofficial® loaded mass was determined at NAREL, each sample set was placed into a cooler with frozen ice packs, a Dickson temperature logger, and a letter of instructions. The coolers were shipped to the participating laboratories by overnight Federal Express.

Instructions provided with the samples allowed laboratories two weeks from the time of receipt to equilibrate and obtain final mass measurements. All samples were then returned to NAREL, with ice packs and temperature loggers.

Gravimetric Results

Figure 1 presents the inter-laboratory capture differences for all samples with advisory limits. Inter-laboratory differences were calculated by subtracting the PM_{2.5} capture value determined at each laboratory from the capture value determined at NAREL. The advisory limits were derived from all of the PE studies administered by NAREL during the past year. The 3-sigma limits are calculated from the inter-laboratory capture differences between NAREL and the participating laboratories. Region 10 laboratories delivered results from two analysts and both sets of data are included. NAREL's capture value was calculated using the $\Lambda_{\text{official}}$ loaded mass determined immediately before the samples were shipped to the regional laboratories. Notice that a negative bar on the Figure 1 graph represents a smaller PM_{2.5} capture value determined at NAREL

Figure 1



A summary of all inter-laboratory capture differences is presented in Table 1.

| Table 1. Capture Difference Summary (mg) * | | | | | |
|--|-----------------|---------------------|---------------------|-----------------|------------|
| | Region 4 | Region 10 #1 | Region 10 #2 | R&IE | RTI |
| 48 Hour Event | 0.002 | 0.005 | -0.001 | -0.006 | 0.001 |
| 48 Hour Event | 0.003 | 0.002 | -0.001 | -0.004 | -0.001 |
| 24 Hour Event | 0.003 | 0.004 | 0.002 | -0.002 | -0.002 |
| 24 Hour Event | 0.002 | 0.005 | 0.000 | -0.007 | -0.003 |
| 20 Hour Event | 0.003 | 0.004 | 0.006 | 0.002 | -0.007 |
| 20 Hour Event | 0.003 | 0.009 | 0.007 | -0.001 | -0.004 |
| 16 Hour Event | 0.008 | 0.005 | 0.006 | -0.005 | -0.003 |
| Blank | 0.001 | 0.001 | 0.000 | -0.015 | -0.004 |
| Blank | 0.000 | -0.002 | -0.002 | -0.003 | -0.003 |
| Blank | 0.000 | 0.005 | 0.001 | -0.012 | -0.002 |
| Metallic Weight | 0.001 | 0.001 | 0.000 | 0.002 | -0.001 |
| Metallic Weight | -0.001 | -0.001 | 0.000 | 0.001 | 0.000 |
| * Capture difference = NAREL capture - Region capture A negative difference indicates a smaller capture for NAREL | | | | | |

Metallic weights were included in this study because they are more stable than a Teflon7 filter, especially a loaded Teflon7 filter. The metallic weights were weighed at each laboratory during the initial tare sessions as well as during the final loaded sessions. The difference in initial and final mass is the calculated Δ mass capture[®] for the metallic weights. Ideally, the Δ mass capture[®] for the metallic weight samples would be zero. A large difference between an initial and final mass could indicate a balance stability problem.

The temperature criteria for equilibration of Teflon7 filters is 20-23 °C, controlled to 3 2 °C for 24 hours. Data recovered from the temperature loggers assigned to each set of samples indicated that all participating laboratories were within criteria.

The PM_{2.5} mass capture for each of the four sampling events as well as the mass capture for the blank filters and metallic samples is presented graphically in figures 2 - 7 at the end of this report.

The raw data reported from all laboratories have been tabulated in Tables 2 - 6 at the end of this report. The tables include the results of all filters and the modified metallic standards weighed at each laboratory. The tables contain the filter tare mass, the final loaded mass, and the calculated PM_{2.5} capture for each filter. The tables also contain the calculated inter-laboratory difference for measuring the PM_{2.5} capture illustrated in Figure 1. A schedule of the sampling events used to load the filters is presented in Table 7.

Conclusions

Good agreement between NAREL and each participating laboratory was observed for the majority of mass measurements. Two blank filter results, illustrated in Figure 1, fell outside the lower three sigma advisory limit. Data for these samples (T05-11469 and T05-11477 listed in Table 5) indicate good between laboratory agreement for the initial tare measurements, however the post mass measurements reported by R&IE show a relatively large gain in mass for T05-11469 (0.017 mg) compared to NAREL's post measurement (0.002 mg). The R&IE captured mass for sample T05-11477 was also somewhat high for a blank filter (0.008 mg). The NAREL post measurement for T05-11477 showed a net loss of -0.004 mg.

Errors were discovered in the Region 4 laboratory's reported results. Specifically, a sample ID mix-up of the two metallic samples resulted in incorrect results reported for those samples. Also, one metallic sample result appeared to have a transcription error that indicated a three milligram change between the pre and post mass measurements. A telephone conversation with the Region 4 analyst revealed that mass results for the PE samples were not automatically recorded from the balance into a database in the same way as normal samples. The PE sample results were hand written onto a data sheet which was then manually transferred into a spreadsheet. Examination of the original raw data sheet, faxed to NAREL, showed the correct result for the transcription error, however, the sample ID mix-up occurred in the original raw data. In this case, the mistake was obvious and could easily be corrected. Once corrections were made to the data, the Region 4 laboratory results compared well with NAREL's measurements. The figures and tables in this report display the corrected results for Region 4.

Figure 2




Figure 3




Figure 4




Figure 5




Figure 6




Figure 7




Table 2. Gravimetric Data Region 4

| Sample ID | Tare Mass | | Final Mass | | Captured PM2.5 | | Inter-Lab Difference* of Captured PM2.5 (mg) |
|---|---------------|------------|---------------|------------|----------------|------------|--|
| | Region 4 (mg) | NAREL (mg) | Region 4 (mg) | NAREL (mg) | Region 4 (mg) | NAREL (mg) | |
| T05-11448 | 145.295 | 145.292 | 145.655 | 145.654 | 0.360 | 0.362 | 0.002 |
| T05-11449 | 145.101 | 145.098 | 145.458 | 145.458 | 0.357 | 0.360 | 0.003 |
| T05-11450 | 145.604 | 145.601 | 145.772 | 145.772 | 0.168 | 0.171 | 0.003 |
| T05-11451 | 143.939 | 143.936 | 144.098 | 144.097 | 0.159 | 0.161 | 0.002 |
| T05-11452 | 145.717 | 145.713 | 145.816 | 145.815 | 0.099 | 0.102 | 0.003 |
| T05-11453 | 143.732 | 143.729 | 143.831 | 143.831 | 0.099 | 0.102 | 0.003 |
| T05-11454 | 144.517 | 144.512 | 144.587 | 144.590 | 0.070 | 0.078 | 0.008 |
| T05-11455 | 145.713 | 145.709 | 145.714 | 145.711 | 0.001 | 0.002 | 0.001 |
| T05-11456 | 146.201 | 146.196 | 146.203 | 146.198 | 0.002 | 0.002 | 0.000 |
| T05-11457 | 145.503 | 145.497 | 145.506 | 145.500 | 0.003 | 0.003 | 0.000 |
| MW05-11488 | 191.060 | 191.061 | 191.059 | 191.061 | -0.001 | 0.000 | 0.001 |
| MW05-11489 | 96.351 | 96.353 | 96.352 | 96.353 | 0.001 | 0.000 | -0.001 |
| * Negative values indicate a larger capture determined by Region 4. | | | | | | | |

Table 3. Gravimetric Data Region 10 Analyst 1

| Sample ID | Tare Mass | | Final Mass | | Captured PM2.5 | | Inter-Lab Difference* of Captured PM2.5 (mg) |
|--|-------------------|------------|-------------------|------------|-------------------|------------|--|
| | Region 10 #1 (mg) | NAREL (mg) | Region 10 #1 (mg) | NAREL (mg) | Region 10 #1 (mg) | NAREL (mg) | |
| T05-11458 | 147.727 | 147.727 | 148.074 | 148.079 | 0.347 | 0.352 | 0.005 |
| T05-11459 | 149.029 | 149.032 | 149.386 | 149.391 | 0.357 | 0.359 | 0.002 |
| T05-11460 | 149.483 | 149.487 | 149.642 | 149.650 | 0.159 | 0.163 | 0.004 |
| T05-11461 | 145.360 | 145.362 | 145.518 | 145.525 | 0.158 | 0.163 | 0.005 |
| T05-11462 | 144.141 | 144.146 | 144.243 | 144.252 | 0.102 | 0.106 | 0.004 |
| T05-11463 | 144.088 | 144.090 | 144.183 | 144.194 | 0.095 | 0.104 | 0.009 |
| T05-11464 | 146.516 | 146.521 | 146.583 | 146.593 | 0.067 | 0.072 | 0.005 |
| T05-11465 | 147.334 | 147.337 | 147.333 | 147.337 | -0.001 | 0.000 | 0.001 |
| T05-11466 | 146.276 | 146.282 | 146.279 | 146.283 | 0.003 | 0.001 | -0.002 |
| T05-11467 | 146.333 | 146.334 | 146.331 | 146.337 | -0.002 | 0.003 | 0.005 |
| MW05-11490 | 193.819 | 193.822 | 193.819 | 193.822 | -0.001 | 0.000 | 0.001 |
| MW05-11491 | 92.959 | 92.960 | 92.958 | 92.959 | 0.000 | -0.001 | -0.001 |
| * Negative values indicate a larger capture determined by Region 10. | | | | | | | |

Table 4. Gravimetric Data Region 10 Analyst 2

| Sample ID | Tare Mass | | Final Mass | | Captured PM2.5 | | Inter-Lab Difference* of Captured PM2.5 (mg) |
|--|--------------------------|-------------------|--------------------------|-------------------|--------------------------|-------------------|---|
| | Region 10 #2 (mg) | NAREL (mg) | Region 10 #2 (mg) | NAREL (mg) | Region 10 #2 (mg) | NAREL (mg) | |
| T05-11458 | 147.724 | 147.727 | 148.077 | 148.079 | 0.353 | 0.352 | -0.001 |
| T05-11459 | 149.028 | 149.032 | 149.388 | 149.391 | 0.360 | 0.359 | -0.001 |
| T05-11460 | 149.482 | 149.487 | 149.643 | 149.650 | 0.161 | 0.163 | 0.002 |
| T05-11461 | 145.358 | 145.362 | 145.521 | 145.525 | 0.163 | 0.163 | 0.000 |
| T05-11462 | 144.143 | 144.146 | 144.243 | 144.252 | 0.100 | 0.106 | 0.006 |
| T05-11463 | 144.088 | 144.090 | 144.185 | 144.194 | 0.097 | 0.104 | 0.007 |
| T05-11464 | 146.517 | 146.521 | 146.583 | 146.593 | 0.066 | 0.072 | 0.006 |
| T05-11465 | 147.334 | 147.337 | 147.334 | 147.337 | 0.000 | 0.000 | 0.000 |
| T05-11466 | 146.277 | 146.282 | 146.280 | 146.283 | 0.003 | 0.001 | -0.002 |
| T05-11467 | 146.333 | 146.334 | 146.335 | 146.337 | 0.002 | 0.003 | 0.001 |
| MW05-11490 | 193.820 | 193.822 | 193.819 | 193.822 | 0.000 | 0.000 | 0.000 |
| MW05-11491 | 92.958 | 92.960 | 92.958 | 92.959 | -0.001 | -0.001 | 0.000 |
| * Negative values indicate a larger capture determined by Region 10. | | | | | | | |

Table 5. Gravimetric Data R&IE

| Sample ID | Tare Mass | | Final Mass | | Captured PM2.5 | | Inter-Lab Difference* of Captured PM2.5 (mg) |
|---|----------------------|-------------------|----------------------|-------------------|-----------------------|-------------------|---|
| | R&IE (mg) | NAREL (mg) | R&IE (mg) | NAREL (mg) | R&IE (mg) | NAREL (mg) | |
| T05-11468 | 148.505 | 148.498 | 148.870 | 148.857 | 0.365 | 0.359 | -0.006 |
| T05-11470 | 146.850 | 146.845 | 147.209 | 147.200 | 0.359 | 0.355 | -0.004 |
| T05-11471 | 144.999 | 144.994 | 145.162 | 145.155 | 0.163 | 0.161 | -0.002 |
| T05-11472 | 144.930 | 144.927 | 145.097 | 145.087 | 0.167 | 0.160 | -0.007 |
| T05-11473 | 146.365 | 146.362 | 146.462 | 146.461 | 0.097 | 0.099 | 0.002 |
| T05-11474 | 145.926 | 145.923 | 146.030 | 146.026 | 0.104 | 0.103 | -0.001 |
| T05-11475 | 145.574 | 145.569 | 145.649 | 145.639 | 0.075 | 0.070 | -0.005 |
| T05-11469 | 146.592 | 146.589 | 146.609 | 146.591 | 0.017 | 0.002 | -0.015 |
| T05-11476 | 145.384 | 145.382 | 145.388 | 145.383 | 0.004 | 0.001 | -0.003 |
| T05-11477 | 146.142 | 146.142 | 146.150 | 146.138 | 0.008 | -0.004 | -0.012 |
| MW05-11492 | 186.993 | 186.995 | 186.992 | 186.996 | -0.001 | 0.001 | 0.002 |
| MW05-11493 | 90.601 | 90.603 | 90.601 | 90.604 | 0.000 | 0.001 | 0.001 |
| * Negative values indicate a larger capture determined by R&IE-LV | | | | | | | |

Table 6. Gravimetric Data RTI

| Sample ID | Tare Mass | | Final Mass | | Captured PM2.5 | | Inter-Lab Difference* of Captured PM2.5 (mg) |
|---|---------------------|-----------------------|---------------------|-----------------------|-----------------------|-----------------------|---|
| | RTI (mg) | NAREL (mg) | RTI (mg) | NAREL (mg) | RTI (mg) | NAREL (mg) | |
| T05-11478 | 146.062 | 146.063 | 146.422 | 146.424 | 0.360 | 0.361 | 0.001 |
| T05-11479 | 146.532 | 146.535 | 146.889 | 146.891 | 0.357 | 0.356 | -0.001 |
| T05-11480 | 145.912 | 145.912 | 146.082 | 146.080 | 0.170 | 0.168 | -0.002 |
| T05-11481 | 144.168 | 144.168 | 144.336 | 144.333 | 0.168 | 0.165 | -0.003 |
| T05-11482 | 145.799 | 145.805 | 145.907 | 145.906 | 0.108 | 0.101 | -0.007 |
| T05-11483 | 145.921 | 145.924 | 146.022 | 146.021 | 0.101 | 0.097 | -0.004 |
| T05-11484 | 145.901 | 145.901 | 145.974 | 145.971 | 0.073 | 0.070 | -0.003 |
| T05-11485 | 148.313 | 148.315 | 148.317 | 148.315 | 0.004 | 0.000 | -0.004 |
| T05-11486 | 148.360 | 148.362 | 148.363 | 148.362 | 0.003 | 0.000 | -0.003 |
| T05-11487 | 147.228 | 147.229 | 147.229 | 147.228 | 0.001 | -0.001 | -0.002 |
| MW05-11494 | 181.334 | 181.336 | 181.336 | 181.337 | 0.002 | 0.001 | -0.001 |
| MW05-11495 | 88.207 | 88.208 | 88.207 | 88.208 | 0.000 | 0.000 | 0.000 |
| * Negative values indicate a larger capture determined by RTI | | | | | | | |

Table 7. Sampling Schedule

| Lab ID | Filter ID | Sample Start | Event Duration (hours) | Receiving Lab |
|---------------|------------------|---------------------|-----------------------------------|----------------------|
| T05-11448 | T2223306 | 9/29/2005 | 48 | Region 4 |
| T05-11449 | T2223307 | 9/29/2005 | 48 | Region 4 |
| T05-11450 | T2223308 | 10/3/2005 | 24 | Region 4 |
| T05-11451 | T2223309 | 10/3/2005 | 24 | Region 4 |
| T05-11452 | T2223310 | 10/4/2005 | 20 | Region 4 |
| T05-11453 | T2223311 | 10/4/2005 | 20 | Region 4 |
| T05-11454 | T2223312 | 10/5/2005 | 16 | Region 4 |
| T05-11455 | T2223314 | | 0 | Region 4 |
| T05-11456 | T2223315 | | 0 | Region 4 |
| T05-11457 | T2223316 | | 0 | Region 4 |
| T05-11458 | T2223317 | 9/29/2005 | 48 | Region 10 |
| T05-11459 | T2223318 | 9/29/2005 | 48 | Region 10 |
| T05-11460 | T2223319 | 10/3/2005 | 24 | Region 10 |
| T05-11461 | T2223320 | 10/3/2005 | 24 | Region 10 |
| T05-11462 | T2223321 | 10/4/2005 | 20 | Region 10 |
| T05-11463 | T2223323 | 10/4/2005 | 20 | Region 10 |
| T05-11464 | T2223324 | 10/5/2005 | 16 | Region 10 |
| T05-11465 | T2223327 | | 0 | Region 10 |
| T05-11466 | T2223328 | | 0 | Region 10 |
| T05-11467 | T2223329 | | 0 | Region 10 |
| T05-11468 | T2223330 | 9/29/2005 | 48 | R&IE |
| T05-11470 | T2223332 | 9/29/2005 | 48 | R&IE |
| T05-11471 | T2223333 | 10/3/2005 | 24 | R&IE |
| T05-11472 | T2223334 | 10/3/2005 | 24 | R&IE |
| T05-11473 | T2223335 | 10/4/2005 | 20 | R&IE |
| T05-11474 | T2223336 | 10/4/2005 | 20 | R&IE |
| T05-11475 | T2223337 | 10/5/2005 | 16 | R&IE |
| T05-11476 | T2223338 | | 0 | R&IE |
| T05-11477 | T2223339 | | 0 | R&IE |
| T05-11469 | T2223331 | | 0 | R&IE |
| T05-11478 | T2223340 | 9/29/2005 | 48 | RTI |
| T05-11479 | T2223341 | 9/29/2005 | 48 | RTI |
| T05-11480 | T2223342 | 10/3/2005 | 24 | RTI |
| T05-11481 | T2223343 | 10/3/2005 | 24 | RTI |
| T05-11482 | T2223344 | 10/4/2005 | 20 | RTI |
| T05-11483 | T2223345 | 10/4/2005 | 20 | RTI |
| T05-11484 | T2223346 | 10/5/2005 | 16 | RTI |
| T05-11485 | T2223347 | | 0 | RTI |
| T05-11486 | T2223348 | | 0 | RTI |
| T05-11487 | T2223349 | | 0 | RTI |

Appendix D

Experimental Inter-Comparison of Speciation Laboratories

TECHNICAL MEMORANDUM



TO: Dennis Crumpler / OAQPS
FROM: Eric Boswell / NAREL
COPY: Mike Poore / CARB
Dr. Richard Tropp / DRI
RaeAnn Haynes / ODEQ
Dr. R.K.M. Jayanty / RTI
AUTHOR: Jewell Smiley / NAREL
DATE: September 19, 2005
SUBJECT: Experimental Inter-comparison of Speciation Laboratories

Introduction

This study was conducted as part of the EPA's quality assurance oversight for the PM_{2.5} chemical speciation air monitoring networks that include the Speciation Trends Network (STN) and the Interagency Monitoring of Protected Visual Environments (IMPROVE) program. The purpose of this study was to evaluate specific laboratory performance at those laboratories that routinely analyze PM_{2.5} chemical speciation samples.

This study required each participating laboratory to analyze a set of blind Performance Evaluation (PE) filter samples. The PE samples were prepared at the National Air and Radiation Environmental Laboratory (NAREL) located in Montgomery, AL. NAREL was able to create replicate filter samples for this study by using co-located Met One speciation samplers. The co-located samplers were programmed to collect PM_{2.5} from the Montgomery air and simultaneously load several filters during each collection event. A sufficient number of replicates were prepared so that each laboratory could receive the following set of PE samples.

- Gravimetric Mass Analysis - ten Teflon® filter samples and two metallic weights
- Ion Chromatography (IC) Analysis - six Nylon® filter samples or six Teflon® filter samples
- Carbon by Thermal Optical Analysis (TOA) - six quartz filter samples
- Elemental analysis by X-Ray Fluorescence (XRF) - six Teflon® filter samples

Detailed instructions for analyzing and reporting the PE samples were provided by NAREL. This report will compare and discuss the analytical results received from all of the laboratories. Some of the laboratories received a full set of PE samples, and some received a partial set due to limitations that will be explained later in the appropriate section of this report. Table 1 identifies all of the laboratories along with their level of participation.

Table 1. List of Participating Laboratories

| Laboratory | Location | Analyses Reported |
|---|----------------------------|--|
| California Air Resources Board (CARB) | Sacramento, CA | Gravimetric mass IC analysis, Nylon® filters TOA carbon, modified STN method |
| Desert Research Institute (DRI) | Reno, NV | Gravimetric mass IC analysis, Teflon® filters IC analysis, Nylon® filters TOA carbon, STN method TOA carbon, IMPROVE method TOA carbon, IMPROVE-a method Elements by XRF |
| Oregon Dept. of Environmental Quality (ODEQ) | Portland, OR | Gravimetric mass IC analysis, Nylon® filters Elements by XRF |
| Research Triangle Institute (RTI) | Research Triangle Park, NC | Gravimetric mass IC analysis, Nylon® filters TOA carbon, STN method TOA carbon, IMPROVE method Elements by XRF |
| EPA's National Exposure Research Laboratory (NERL) | Research Triangle Park, NC | Elements by XRF |
| EPA's National Air and Radiation Environmental Laboratory (NAREL) | Montgomery, AL | Gravimetric mass IC analysis, Nylon® filters IC analysis, Teflon® filters TOA carbon, STN method TOA carbon, IMPROVE method TOA carbon, IMPROVE-a method |

Mass determination typically proceeds by weighing the Teflon® collection filter before and after the sampling event. The amount of Particulate Matter (PM_{2.5}) captured onto the surface of the filter can be calculated by a simple subtraction of the tare mass from the loaded filter mass. Each speciation laboratory routinely provides clean PRE-weighed air filters to the supported field sites. At the field site, an approved sampling device must be used to deposit the PM_{2.5} onto the collection filter. The loaded filter is returned to the originating laboratory where the gravimetric analysis is completed by POST-weighing the filter. After the gravimetric measurements are complete, the Teflon® filter is examined further using XRF to determine the elemental composition of the filter deposit. Usually XRF is the final analysis of the Teflon® filter after which the filter is placed into an archive for storage, but in some cases the filter is subjected to one more [final] analysis to determine the ions present in the filter deposit.

If the Teflon® filter is examined for ions, it must be extracted, and the extract is subsequently analyzed using Ion Chromatography.

Most of the speciation laboratories provide clean Nylon® filters to the field sites. It is usually the Nylon® filter that is used to capture PM_{2.5} for subsequent IC analysis. After the loaded filter is returned to the laboratory, the IC analysis typically proceeds by first extracting the filter using an appropriate solvent. The extract must be analyzed using an IC instrument that is optimized to determine the ions of interest. Target anions and target cations must be analyzed on separate IC instruments.

The laboratories also provide clean quartz filters to the supported field sites. The quartz filter is used to capture PM_{2.5} for subsequent carbon analysis. A thermal/optical analysis (TOA) is performed at the laboratory to determine the carbon present on the quartz filter. A carefully measured portion of the quartz filter is placed into a special oven equipped to shine a laser at the sample. The TOA technique requires heating the quartz filter material to release captured PM_{2.5}. Carbon components released from the filter are swept through the oven by a controlled purge gas. The carbon released from the filter is catalytically converted to methane and measured by a flame ionization detector (FID) positioned at the end of the sample train. A thermogram produced by the analysis contains signals from the FID and from the laser. Interpretation of the thermogram provides results for the organic carbon (OC) and the elemental carbon (EC) the sum of which represents the total carbon (TC) present in the sample. Several slightly different TOA methods were used to analyze samples during this study. A more detailed description of each TOA method will be provided later in this report.

Gravimetric Analysis

Ten new filters and two metallic transfer weights were supplied by NAREL to each laboratory for this study. These samples were placed into individual petri slides and shipped by overnight mail to the receiving lab with instructions to PRE-weigh each filter and metallic weight using the local standard procedures. After tare measurements were completed at the receiving lab, the filters and metallic weights were returned to Montgomery and immediately placed into the weighing chamber at NAREL for equilibration and determination of a stable tare mass. Shortly after NAREL's tare measurements were complete, some of the filters were loaded with PM_{2.5} captured from the Montgomery air. Co-located Met One SuperSASS air samplers were used to load seven of the filters in each sample set according to the sampling schedule presented in Table 2.

Table 2. Sampling Schedule for Gravimetric PE Filters

| Filter ID | Serial Number | Sample Start | Event Duration | Receiving Lab |
|-----------|---------------|--------------|----------------|---------------|
| T05-11285 | T2017288 | 20-Jan-05 | 24-hour | CARB |
| T05-11286 | T2017289 | 20-Jan-05 | 24-hour | CARB |
| T05-11287 | T2017290 | 21-Jan-05 | 48-hour | CARB |
| T05-11288 | T2017291 | 21-Jan-05 | 48-hour | CARB |
| T05-11289 | T2017292 | 23-Jan-05 | 12-hour | CARB |
| T05-11290 | T2017293 | 23-Jan-05 | 12-hour | CARB |
| T05-11291 | T2017310 | 24-Jan-05 | 24-hour | CARB |

Table 2. Sampling Schedule for Gravimetric PE Filters

| Filter ID | Serial Number | Sample Start | Event Duration | Receiving Lab |
|-----------|---------------|--------------|----------------|---------------|
| T05-11295 | T2017314 | 20-Jan-05 | 24-hour | DRI |
| T05-11296 | T2017315 | 20-Jan-05 | 24-hour | DRI |
| T05-11297 | T2017316 | 21-Jan-05 | 48-hour | DRI |
| T05-11298 | T2017317 | 21-Jan-05 | 48-hour | DRI |
| T05-11299 | T2017318 | 23-Jan-05 | 12-hour | DRI |
| T05-11300 | T2017319 | 23-Jan-05 | 12-hour | DRI |
| T05-11301 | T2017320 | 24-Jan-05 | 24-hour | DRI |
| T05-11305 | T2017324 | 20-Jan-05 | 24-hour | ODEQ |
| T05-11306 | T2017325 | 20-Jan-05 | 24-hour | ODEQ |
| T05-11307 | T2017326 | 21-Jan-05 | 48-hour | ODEQ |
| T05-11308 | T2017327 | 21-Jan-05 | 48-hour | ODEQ |
| T05-11309 | T2017328 | 23-Jan-05 | 12-hour | ODEQ |
| T05-11310 | T2017329 | 23-Jan-05 | 12-hour | ODEQ |
| T05-11311 | T2017330 | 24-Jan-05 | 24-hour | ODEQ |
| T05-11315 | T2017334 | 20-Jan-05 | 24-hour | RTI |
| T05-11316 | T2017335 | 20-Jan-05 | 24-hour | RTI |
| T05-11317 | T2017358 | 21-Jan-05 | 48-hour | RTI |
| T05-11318 | T2017337 | 21-Jan-05 | 48-hour | RTI |
| T05-11319 | T2017338 | 23-Jan-05 | 12-hour | RTI |
| T05-11320 | T2017339 | 23-Jan-05 | 12-hour | RTI |
| T05-11321 | T2017354 | 24-Jan-05 | 24-hour | RTI |

Table 2 shows twenty-eight filters that were loaded during four separate collection events. A sufficient number of replicates were prepared during each event such that each lab could be provided with an almost identical set of loaded filters. For example, eight replicates were created during a 24-hour collection event that started on January 20, and two of these replicates were submitted to each lab for analysis. Likewise, eight replicates were created during a 48-hour collection event that started on January 21, and two of these replicates were submitted to each lab for analysis. Table 2 does not list all of the filters that were PRE-weighed at the participating labs. Three of the ten filters that were PRE-weighed at each lab were not scheduled for loading because they were used as filter blanks for this study.

Following sample collection, the filters and the metallic weights were returned to the weighing chamber at NAREL and POST-weighed multiple times over the course of several days to demonstrate a stable final mass. Finally, the filters and metallic weights were placed into small Igloo® coolers with ice substitute and shipped back to the participating labs for POST-weighing. It is worth mentioning that the metallic weights were included in this study because they are usually less susceptible to weighing errors due to factors such as electrical static and volatility of filter constituents.

Gravimetric Results

The results of this study are summarized in Figure 1. The critical information needed by the program is the mass of PM_{2.5} deposited onto the surface of a collection filter, and therefore, PM_{2.5} capture is plotted in Figure 1 for the seven loaded filters, three travel blanks, and two metallic weights.

Figure 1





Figure 2 presents the inter-laboratory differences along with advisory limits. Inter-laboratory differences were calculated by subtracting the PM_{2.5} capture value determined at each speciation lab from the capture value determined at NAREL. Notice that a negative bar on the Figure 2 graph represents a smaller PM_{2.5} capture value determined at NAREL. The 3-sigma advisory limits were derived from all of the gravimetric PE studies administered by NAREL during the past year.

Figure 2



The raw data reported from all laboratories have been tabulated for easy viewing. At the end of this report, Table 9 contains the tare weight, the final loaded weight, and the calculated PM_{2.5} capture for each sample. Table 9 also contains the calculated inter-laboratory difference for measuring the PM_{2.5} capture which is graphed in Figure 2. RTI reported measurements made by several different analysts, and all of the results are included in Table 9. However, only the results from analyst #1 are presented in Figure 1 and Figure 2. Only one set of measurements from each lab were selected for graphical presentation because usually only one set of measurements are available for a routine sample.

All of the participating labs have an SOP for measuring the gravimetric mass of PM_{2.5} filter samples. Most of the SOP's are currently available on the web for easy viewing (see reference 1 through 5).

IC Analysis

This study included the analysis of selected ions using three slightly different IC methods. Five labs analyzed a set of Nylon® filters using the STN method, two labs analyzed a set of Teflon® filters using the STN method, and finally two labs analyzed a set of Nylon® filters using the IMPROVE method. NAREL provided each lab with a set of six filters for each method tested. Each sample set contained two blank filters and four filters that were loaded with PM_{2.5} collected from the Montgomery air. Co-located Met One SuperSASS air samplers were used to load filters and create replicates in each sample set according to the sampling schedule presented in Table 3.

Table 3. Sampling Schedule for Ion Chromatography PE Filters

| Filter ID | Filter Medium | Sample Start | Event Duration | Receiving Lab | Method |
|-----------|---------------|--------------|----------------|---------------|--------|
| N04-11197 | Nylon® | 24-Nov-04 | 116-hour | CARB | STN |
| N04-11198 | Nylon® | 24-Nov-04 | 116-hour | CARB | STN |
| N04-11208 | Nylon® | 29-Nov-04 | 159-hour | CARB | STN |
| N04-11209 | Nylon® | 29-Nov-04 | 159-hour | CARB | STN |
| N04-11199 | Nylon® | 24-Nov-04 | 116-hour | DRI | STN |
| N04-11200 | Nylon® | 24-Nov-04 | 116-hour | DRI | STN |
| N04-11210 | Nylon® | 29-Nov-04 | 159-hour | DRI | STN |
| N04-11211 | Nylon® | 29-Nov-04 | 159-hour | DRI | STN |
| N04-11201 | Nylon® | 24-Nov-04 | 116-hour | ODEQ | STN |
| N04-11202 | Nylon® | 24-Nov-04 | 116-hour | ODEQ | STN |
| N04-11212 | Nylon® | 29-Nov-04 | 159-hour | ODEQ | STN |
| N04-11213 | Nylon® | 29-Nov-04 | 159-hour | ODEQ | STN |
| N04-11203 | Nylon® | 24-Nov-04 | 116-hour | RTI | STN |
| N04-11204 | Nylon® | 24-Nov-04 | 116-hour | RTI | STN |
| N04-11214 | Nylon® | 29-Nov-04 | 159-hour | RTI | STN |
| N04-11215 | Nylon® | 29-Nov-04 | 159-hour | RTI | STN |
| N04-11205 | Nylon® | 24-Nov-04 | 116-hour | NAREL | STN |
| N04-11206 | Nylon® | 24-Nov-04 | 116-hour | NAREL | STN |
| N04-11216 | Nylon® | 29-Nov-04 | 159-hour | NAREL | STN |
| N04-11217 | Nylon® | 29-Nov-04 | 159-hour | NAREL | STN |
| T05-11333 | Teflon® | 03-Jan-05 | 144-hour | DRI | STN |
| T05-11334 | Teflon® | 03-Jan-05 | 144-hour | DRI | STN |
| T05-11337 | Teflon® | 04-Jan-05 | 216-hour | DRI | STN |
| T05-11338 | Teflon® | 04-Jan-05 | 216-hour | DRI | STN |

Table 3. Sampling Schedule for Ion Chromatography PE Filters

| Filter ID | Filter Medium | Sample Start | Event Duration | Receiving Lab | Method |
|-----------|---------------|--------------|----------------|---------------|---------|
| T05-11335 | Teflon® | 03-Jan-05 | 144-hour | NAREL | STN |
| T05-11336 | Teflon® | 03-Jan-05 | 144-hour | NAREL | STN |
| T05-11339 | Teflon® | 04-Jan-05 | 216-hour | NAREL | STN |
| T05-11340 | Teflon® | 04-Jan-05 | 216-hour | NAREL | STN |
| N04-11229 | Nylon® | 07-Dec-04 | 161-hour | RTI | IMPROVE |
| N04-11230 | Nylon® | 07-Dec-04 | 161-hour | RTI | IMPROVE |
| N04-11233 | Nylon® | 08-Dec-04 | 130-hour | RTI | IMPROVE |
| N04-11234 | Nylon® | 08-Dec-04 | 130-hour | RTI | IMPROVE |
| N04-11231 | Nylon® | 07-Dec-04 | 161-hour | NAREL | IMPROVE |
| N04-11232 | Nylon® | 07-Dec-04 | 161-hour | NAREL | IMPROVE |
| N04-11235 | Nylon® | 08-Dec-04 | 130-hour | NAREL | IMPROVE |
| N04-11236 | Nylon® | 08-Dec-04 | 130-hour | NAREL | IMPROVE |

Table 3 shows thirty-six filters that were loaded during six separate collection events. A sufficient number of replicates were prepared during each event such that each participating lab was provided with an almost identical set of loaded filters. For example, ten replicates were created during a 116-hour collection event that started on November 24, and two of these replicates were submitted to each lab for analysis. Likewise, ten replicates were created during a 159-hour collection event that started on November 29, and two of these replicates were submitted to each lab for analysis. The collection times used for this study were significantly longer than the normal 24-hours to boost the amount of PM_{2.5} collected and raise the level of most analytes to above the detection threshold. Table 3 does not list the filter blanks that were provided to each participating lab.

A filter set was provided to each participating lab with instructions to use local standard procedures, as closely as possible, for the extraction and the IC analysis. No information was given to the participating labs about the history of the individual filters. The results were reported for each sample based upon the amount of analyte present on the filter (µg/filter).

All of the participating labs have an SOP for analyzing PM_{2.5} filter samples by IC. Most of the SOP's are currently available on the web for easy viewing (see reference 6 through 13).

IC Results

Results from the analysis of Nylon® filters using the STN method are presented as a bar graph in Figure 3 and Figure 4. Ten replicates from the November 24 event are shown on the left side of the graphs, and ten replicates from the November 29 event are shown on the right side of the graphs. Nitrate, sulfate, and ammonium were the most abundant analytes captured from the Montgomery air, and these mid-level ions are plotted together in Figure 3. Sodium and potassium were present in the air at relatively low levels, and these ions are plotted in Figure 4. Since the low-level components are presented in Figure 4, an extra bar was added to this graph that represents the lowest calibration standard analyzed at NAREL. The lowest calibration standard is a good estimate of the practical quantitation limit for the analysis. Each cluster of ten bars in the graph is labeled with the ion reported, but the individual samples within each cluster are not identified. It is important to understand that the ten replicate samples within each cluster were consistently arranged, from left to right, in the same order.




Figure 3

Good precision can be seen in Figure 3 and Figure 4. The inter-laboratory precision is almost as good as the precision within each lab.

Figure 4




Figure 5 and Figure 6 show the results from replicate Teflon® filter samples that were created on January 3 and January 4. Half of the replicates were submitted to DRI for analysis using the STN method, and half were retained at NAREL for analysis using the same method. Teflon® filter samples are routinely analyzed at DRI as part of their




Figure 5

work for the Texas Commission on Environmental Quality (TCEQ). The mid-level and the low-level components are presented again as separate graphs in Figure 5 and Figure 6 respectively. It is worth noting that nitrate was not a mid-level component on the Teflon® filters even though it probably was a mid-level component in the Montgomery air. Excellent precision is observed in Figure 5 for ammonium, especially considering the non-linear response curve that ammonium offers at the IC instrument. A small consistent (eleven percent) inter-laboratory bias is observed for sulfate in Figure 5. Good precision is observed for the low-level components shown in Figure 6.

Figure 6






Figure 7



Nylon® filters are routinely analyzed at RTI using the IMPROVE method which is slightly different from the STN method with respect to the extraction procedure and the list of reported ions. Figure 7 shows good precision for all of the mid-level ions, but there is a problem in Figure 8 with the low-level components. Poor precision was reported by NAREL for the chloride analysis. After discovery of the problem, the filter extracts were re-analyzed with similar results. NAREL has not been able to explain the poor precision for chloride observed in Figure 8, but possible reasons include poor filter replication and accidental contamination of the filter extract. The variability observed for nitrite may be due to contamination which is frequently observed in blanks. Blanks were provided to all of the labs for this study even though the blank results are not presented in graphical format. The numerical results for all of the blanks and for all of the loaded filters are available in Table 10 at the end of this report.

Figure 8



Carbon Analysis

This study included the Thermal-Optical Analysis (TOA) of quartz fiber filters to determine the amount of carbon present in captured PM_{2.5}. NAREL provided each participating laboratory with a set of six 47-mm filters. Each sample set contained two blank filters and four filters that were loaded with PM_{2.5} collected from the Montgomery air. Co-located Met One SuperSASS air samplers were used to load filters and create replicates in each sample set according to the sampling schedule presented in Table 4.

Table 4. Sampling Schedule for TOA Carbon PE Filters

| Filter ID | Filter Medium | Sample Start | Event Duration | Receiving Lab | TOA Method(s) |
|-----------|---------------|--------------|----------------|---------------|----------------------------|
| Q04-11175 | quartz | 27-Apr-04 | 287-hr | CARB | STN (modified) |
| Q04-11176 | quartz | 27-Apr-04 | 287-hr | CARB | STN (modified) |
| Q04-11186 | quartz | 16-Nov-04 | 192-hr | CARB | STN (modified) |
| Q04-11187 | quartz | 16-Nov-04 | 192-hr | CARB | STN (modified) |
| Q04-11177 | quartz | 27-Apr-04 | 287-hr | DRI | STN, IMPROVE, and IMPROVEa |
| Q04-11178 | quartz | 27-Apr-04 | 287-hr | DRI | STN, IMPROVE, and IMPROVEa |
| Q04-11188 | quartz | 16-Nov-04 | 192-hr | DRI | STN, IMPROVE, and IMPROVEa |
| Q04-11189 | quartz | 16-Nov-04 | 192-hr | DRI | STN, IMPROVE, and IMPROVEa |
| Q04-11181 | quartz | 27-Apr-04 | 287-hr | RTI | STN and IMPROVE |
| Q04-11182 | quartz | 27-Apr-04 | 287-hr | RTI | STN and IMPROVE |
| Q04-11192 | quartz | 16-Nov-04 | 192-hr | RTI | STN and IMPROVE |
| Q04-11193 | quartz | 16-Nov-04 | 192-hr | RTI | STN and IMPROVE |
| Q04-11183 | quartz | 27-Apr-04 | 287-hr | NAREL | STN, IMPROVE, and IMPROVEa |
| Q04-11184 | quartz | 27-Apr-04 | 287-hr | NAREL | STN, IMPROVE, and IMPROVEa |
| Q04-11194 | quartz | 16-Nov-04 | 192-hr | NAREL | STN, IMPROVE, and IMPROVEa |
| Q04-11195 | quartz | 16-Nov-04 | 192-hr | NAREL | STN, IMPROVE, and IMPROVEa |

Table 4 shows sixteen filters that were loaded during two separate collection events. A sufficient number of replicates were prepared during each event such that each participating lab was provided with an almost identical set of loaded filters. Eight replicates were created during the 287-hour springtime event that started on April 27, and two of these replicates were submitted to each lab for analysis. Likewise, eight replicates were created during the 192-hour autumn event that started on November 16, and two of these replicates were submitted to each lab for analysis. The collection times used for this study were significantly longer than the normal 24-hours to boost the amount of elemental carbon deposited on the filter. Table 4 does not list the two filter blanks that were provided to each participating lab.

A filter set was provided to each lab with instructions to use local standard procedures, as closely as possible, for the analysis. No information was given to the participating labs about the history of the individual filters. ODEQ did not participate in this part of the study because their quartz filters are shipped to RTI for analysis. The DRI and RTI labs are set up to analyze a large volume of samples and routinely operate several TOA instruments. Both DRI and RTI were able to analyze each filter several times using more than one instrument and using more than one TOA method. The results were reported for each sample based upon the amount of carbon per square centimeter of the filter deposit ($\mu\text{g C}/\text{cm}^2$). Raw data were also supplied to NAREL so that some of the thermograms are included in this report.

This study has provided an excellent opportunity to see replicate filter samples analyzed by a variety of TOA methods. Therefore it is appropriate to ask, “what distinguishes one TOA method from another?” To answer this question we must first identify the critical elements of a TOA method. At least four different TOA methods have been identified in this report based upon the temperature protocol used during the analysis. The following table provides a brief description of each temperature protocol.

Table 5. Comparison of the Temperature Protocols for Four TOA Methods

| STN Method TOT Analysis | CARB Method (modified STN) TOT Analysis | IMPROVE Method TOR Analysis | IMPROVE-a Method TOR Analysis | Carrier Gas | Carbon Fraction* |
|--|--|--|--|------------------------|-----------------------------|
| heater off (90s) | heater off (90s) | heater off (90s) | heater off (90s) | He Purge | ----- |
| 310°C (60s) | 250°C (180s) | 120°C (150-580s) | 140°C (150-580s) | He | OC1 |
| 480°C (60s) | 400°C (150s) | 250°C (150-580s) | 279°C (150-580s) | He | OC2 |
| 615°C (60s) | 550°C (150s) | 450°C (150-580s) | 480°C (150-580s) | He | OC3 |
| 900°C (90s) | 700°C (270s) | 550°C (150-580s) | 580°C (150-580s) | He | OC4 |
| heater off (40s) | heater off (60s) | ----- | ----- | He | |
| 600°C (35s) | 550°C (100s) | 550°C (150-580s) | 580°C (150-580s) | He/O ₂ | EC1 |
| 675°C (45s) | 650°C (100s) | 700°C (150-580s) | 740°C (150-580s) | He/O ₂ | EC2 |
| 750°C (45s) | 750°C (100s) | 800°C (150-580s) | 840°C (150-580s) | He/O ₂ | EC3 |
| 825°C (45s) | 850°C (100s) | ----- | ----- | He/O ₂ | |
| 920°C (120s) | 900°C (170s) | ----- | ----- | He/O ₂ | |
| heater off (110s) | heater off (200s) | heater off (150s) | heater off (200s) | He/O ₂ +IS | |

** The carbon fractions are not consistently defined among the different methods. See text for explanation.*

Beyond the thermal protocols listed in Table 5, each TOA method is further defined by the way optical measurements are made and utilized to calculate carbon fractions. For example, the optical measurements are used to distinguish the elemental carbon (EC) from the organic carbon (OC) present in the sample. In fact as we shall see, all of the carbon fractions have a functional definition that depends upon the method of analysis.

All of the instruments used for this study are equipped with a small tubular quartz oven and a laser/diode system. The sample analysis begins by placing a carefully measured [punched] segment of the filter sample into the oven directly in the path of the laser. A purge gas removes air from the oven and surrounds the sample with a stream of pure helium before the heating and data acquisition begin. Light from the laser will interact with the sample during the analysis. Some of the light will transmit through the sample, and some light will reflect from the surface of the sample. A diode detector can be positioned to measure the light transmitted through the sample, and this configuration is needed for a TOT (thermal optical transmittance) analysis. A diode can also be positioned to measure the reflected light, and this configuration is needed for a TOR (thermal optical reflectance) analysis. As the sample segment is heated and the pure helium phase of the analysis proceeds, some of the carbon may char to form a darker pyrolyzed carbon (PyroIC). All of the methods in this study use either TOT or TOR to evaluate the PyroIC. Four different instrument configurations were used for this study. The older


Sunset [single mode] instruments are equipped with only one diode detector and are configured for the TOT analysis. The older DRI/OGC instruments are also equipped with only one diode detector and are configured for the TOR analysis. The DRI Model 2001 instruments and the Sunset Dual Mode instruments are newer designs capable of measuring the transmitted and the reflected light simultaneously. These newer instruments provide more optical information and give the user a choice of the TOT or the TOR analysis. Table 6 shows specifically how the different instruments were used for analyzing the samples in this study.

Table 6. Summary of Report Packages for the TOA Analyses

| Temperature Protocol | Optical Analysis | Instrument Model | Specific Instrument Reporting | Parameters Reported | Report Package Count |
|----------------------|------------------|--------------------|-------------------------------|--------------------------|----------------------|
| Modified STN | TOT | DRI Model 2001 | CARB Instr. #1 | OC, EC, TC | 1 |
| STN | TOT | DRI Model 2001 | DRI Instr. #7 | OC, EC, TC, OCsub, ECsub | 2 |
| | | | DRI Instr. #9 | OC, EC, TC, OCsub, ECsub | 3 |
| | | Sunset | RTI Instr. R | OC, EC, TC, OCsub | 4 |
| | | | RTI Instr. S | OC, EC, TC, OCsub | 5 |
| | | | RTI Instr. T | OC, EC, TC, OCsub | 6 |
| | | | NAREL Instr. #1 | OC, EC, TC, OCsub | 7 |
| | | Sunset (Dual Mode) | RTI Instr. F | OC, EC, TC, OCsub | 8 |
| IMPROVE | TOR | DRI/OGC | DRI Instr. #4 | OC, EC, TC, OCsub, ECsub | 9 |
| | | | DRI Instr. #5 | OC, EC, TC, OCsub, ECsub | 10 |
| | | Sunset (Dual Mode) | RTI Instr. F | OC, EC, TC, OCsub, ECsub | 11 |
| | | | NAREL Instr. #2 | OC, EC, TC, OCsub, ECsub | 12 |
| IMPROVE-a | TOR | DRI Model 2001 | DRI Instr. #7 | OC, EC, TC, OCsub, ECsub | 13 |
| | | | DRI Instr. #9 | OC, EC, TC, OCsub, ECsub | 14 |
| | | Sunset (Dual Mode) | NAREL Instr. #2 | OC, EC, TC, OCsub, ECsub | 15 |

All of the instruments in this study operate by heating a punched segment of the sample in the presence of a controlled carrier gas. Any carbonaceous material released from the quartz filter segment is swept through a series of zones that rapidly convert the released carbon to methane which is measured by a Flame Ionization Detector (FID) positioned at the end of the sample train. During the first [non-oxidizing] stage of the analysis, the carrier gas is pure helium. Oxygen is added to the carrier during the second stage of the analysis which is designed to remove any remaining carbonaceous material from the quartz residue. Most of the OC is released during the first stage of the analysis, but the EC and any PyroIC that may have formed are more difficult to oxidize, and they are expected to release during the second stage of the analysis. A known mass of methane is injected through the oven at the end of the analysis to serve as an Internal Standard (IS). Signals from the FID and from the laser may be plotted along a time axis to construct a thermogram. An example of a thermogram is shown in Figure 9. This is a thermogram of a sucrose spike which was analyzed at NAREL as a routine calibration check sample. The sucrose spike contains no EC but has a strong tendency to char and form PyroIC.

Figure 9



After the raw data acquisition is complete, the thermogram is evaluated to determine the amount of OC and the amount of EC that were present in the original sample. All of the participating labs report the Total Carbon (TC) as the sum of the OC and the EC fractions: $TC = OC + EC$. Other carbon fractions may be calculated such as the OC subfractions: $OC = OC1 + OC2 + OC3 + OC4 + PyroC$. Figure 9 shows an example of OC subfractions that were calculated by a Sunset instrument. EC subfractions may be calculated as well. For example, three EC subfractions are calculated for IMPROVE samples: $EC = EC1 + EC2 + EC3$. Unfortunately the rules [and consequently the software programs] used to determine these carbon fractions are not the same for all of the instruments. For example, we will see later that some of the instruments reported a negative PyroC, but other instruments have adopted different rules that do not allow a negative PyroC.

A “split point” must be established in each thermogram that separates the OC and the EC. The laser signal must be examined as part of determining the split point. If any of the original OC chars during the first stage of the analysis, the laser signal will decrease from its initial value, and will not recover until later in the run. The point at which the recovering laser signal reaches its initial value is usually the split point. Some samples do not form char, however, and the laser signal does not decrease and fall below its initial value. In this case, the split point is usually assigned when the oxygen valve opens for the second phase of the analysis to begin. All of the instruments follow these general rules, but there is a specific case that is controversial, and it occurs when the laser signal indicates an “early” split point. The split point is considered “early” if it is assigned during the first phase of the analysis before the oxygen valve opens. Most of the instruments were programmed to allow an early split point if the laser signal supports that assignment, but the DRI/OGC instruments did not allow early split points.

As we examine the results from all of the participating labs, it is important to understand the methods that were used, so that valid comparisons can be made. All of the results presented in this report have been identified with the instrument that performed the analysis as well as the thermal protocol and optical configuration that was used. All of the participating labs have an SOP for the TOA method(s) used at their laboratory. Most of the SOP’s are currently available on the web for easy viewing (see reference 14 through 18).

Carbon Results

Results from the analysis of replicate quartz filters using the STN method are presented below as a bar graph. Notice that each bar in the graph is labeled with the instrument number, the lab, and the last three digits of the sample number. Figure 10 shows results from replicates that were created on April 27, and Figure 11 shows the results from replicates created on November 16. The bar segments show the OC and EC components of the total carbon but do not show the more detailed fractions. The results are presented again in Figure 12 and Figure 13 with more detail.

Figure 10





Figure 11



This time in Figure 12 and Figure 13, the OC subfractions are revealed. The subfractions from CARB are not presented since CARB does not use the STN temperature protocol. As shown previously in Figure 9, some of the subfractions are directly related to the temperature set points. PyroIC, on the other hand, is related to the split point. Notice that PyroIC was negative for some of the DRI results, and the reported OC4 result was “adjusted” to maintain proper size of the stacked bar whose height represents the TC. The adjustment was performed by adding the reported OC4 value and the [negative] PyroIC value. The adjustment was performed strictly for graphical purposes.

Figure 12





Figure 13



Some of the labs were able to analyze the PE samples using the IMPROVE and the IMPROVE-a methods. The results in Figure 14 and Figure 15 show the OC and EC components of the total carbon but do not show the more detailed fractions. It can be seen in these plots that the two methods agree quite well. The IMPROVE steering committee had just approved the new IMPROVE-a method earlier this year when these PE samples were analyzed. The new IMPROVE-a method was designed to maintain as much consistency as possible with years of old data produced by the DRI/OGC instrument using the IMPROVE method. The previous results for the STN method agree quite well for TC but show EC values that are significantly smaller than those shown here.

Figure 14




Figure 15




Figure 16 and Figure 17 show the IMPROVE and the IMPROVE-a results again with more detail. Good agreement can be seen for OC subfractions when the IMPROVE-a method was used. Worse precision can be seen among the instruments when the IMPROVE method was used. The DRI/OGC instruments #4 and #5 reported consistently low values for the IMPROVE PyroIC, and this may be related to air leaks during the first stage of the analysis. The two Sunset Dual Mode instruments (RTI F and EPA #2) reported consistently low values for the IMPROVE OC1 fraction. This may be explained by a difference in the accuracy of the temperature measurements inside the sample oven. The OC1 fraction is very sensitive to the 120°C set point specified by the IMPROVE method.

Figure 16





Figure 17







Figure 18

This report includes several thermograms from all of the instruments that were used for this study, and each thermogram was derived from the analysis of a replicate PE sample that was loaded during the collection event which started on April 27, 2004. Figure 18 shows the first thermogram submitted by CARB using their modified STN method and DRI/Model 2001 #1 instrument. CARB has adopted a modified temperature protocol because about three years ago they observed symptoms of an air leak during the first [non-oxidizing] stage of their analysis using the STN method. Experiments were performed to learn more about the problem (see reference 19). Their experiments included changes to the temperature protocol. During their experiments, CARB observed the leak symptoms to become less severe as the first stage maximum temperature was reduced from the STN method value of 900°C. The thermogram shows that CARB's method currently uses a 700°C maximum temperature for the first stage of the analysis. The laser signals in Figure 18 still show some sign of a possible leak. Both laser signals decrease normally from their initial values as char forms. Unfortunately, both signals increase significantly before the oxygen valve opens at approximately 850 seconds into the run. It could be argued that the sample itself contains oxidizing compounds that cause the char to oxidize prematurely. If this were the case, we should see the same symptoms in the thermograms that follow Figure 18.

Figures 19 through 22 show individual thermograms from various instruments using the STN method.

Figure 19



Figures 19 and 20 were produced at DRI using the STN method and DRI/Model 2001 instruments.

Figure 20




Figure 21




Figure 21 is a thermogram produced at RTI using the STN method and their Sunset instrument R. Figure 22 is a thermogram produced at NAREL using the STN method and their Sunset instrument #1. Both thermograms were produced by an older model Sunset [single mode] instrument, as indicated by a single laser signal, configured to perform the TOT analysis. It should be explained that all of the Sunset thermograms were produced at NAREL from the information inside the raw data files, and the laser signal(s) presented here were not processed using the Sunset software to correct for temperature dependence of the laser/diode system.

Figure 22




Figure 23




Figure 23 is a composite of five thermograms. Two of the analyses were presented earlier as individual thermograms in Figure 21 and Figure 22. The single temperature trace was taken from the first analysis using the "R" instrument. All of the laser signals have been amplified and allowed to go off-scale during the later part of the thermogram so that critical features of each laser trace may be seen more clearly.

Figure 24 is the first IMPROVE thermogram produced at DRI using the DRI/OGC instrument #4.

Figure 24




Figure 25




Figure 25 is another IMPROVE thermogram produced at DRI using their DRI/OCG instrument #5. This thermogram shows some evidence of an air leak during the first stage of analysis as indicated by the premature rise of the laser signal before the oxygen valve opens at approximately 1800 seconds. Much less premature rise of the laser signal can be observed in the other IMPROVE thermograms presented in Figures 24, 26, and 27.

Figure 26




Figure 27




Figure 27 is the last of the IMPROVE thermograms, and Figure 28 is the first of three IMPROVE-a thermograms. Figure 28 was produced at DRI using their DRI/Model 2001 instrument #7. It is easy to interpret from this thermogram that the oxygen valve opened at approximately 1300 seconds and the split point was assigned shortly thereafter at approximately 1350 seconds. Notice that the transmitted laser signal usually supports a split point that is slightly later than the split point supported by the reflected laser signal. In this thermogram, the transmitted laser signals support a split point at approximately 1400 seconds.

Figure 28




Figure 29





Figure 29 was produced at DRI using the IMPROVE-a method and the DRI/Model 2001 instrument #9. Figure 30 is the last thermogram presented in this report, and it was produced at NAREL using their Sunset (Dual Mode) instrument which was optimized to run the IMPROVE-a method. Thirteen thermograms have been presented, and each one represents the analysis of a stable residue that was loaded onto the filter during a single collection event. Results from all of the quartz filters are presented in Table 11 at the end of this report. This table includes the uncertainty of measurement when it was available. Table 11 also contains results from the blank filters that were part of each set of PE samples.

Figure 30



XRF Analysis

NAREL provided each participating laboratory with a set of six 47-mm filters for elemental analysis using energy dispersive XRF. Each sample set contained two blank filters and four filters that were loaded with PM_{2.5} collected from the Montgomery air. Co-located Met One SuperSASS air samplers were used to load filters and create replicates in each sample set according to the sampling schedule presented in Table 7.

Table 7. Sampling Schedule for XRF PE Filters

| Filter ID | Serial Number | Sample Start | Event Duration | Receiving Lab |
|-----------|---------------|--------------|----------------|---------------|
| T04-11257 | T2017266 | 16-Dec-04 | 138-hr | DRI |
| T04-11258 | T2017268 | 16-Dec-04 | 138-hr | DRI |
| T04-11267 | T2017278 | 23-Dec-04 | 192-hr | DRI |
| T04-11268 | T2017279 | 23-Dec-04 | 192-hr | DRI |
| T04-11259 | T2017269 | 16-Dec-04 | 138-hr | ODEQ |
| T04-11260 | T2017270 | 16-Dec-04 | 138-hr | ODEQ |
| T04-11269 | T2017280 | 23-Dec-04 | 192-hr | ODEQ |
| T04-11270 | T2017281 | 23-Dec-04 | 192-hr | ODEQ |
| T04-11261 | T2017271 | 16-Dec-04 | 138-hr | RTI |
| T04-11262 | T2017272 | 16-Dec-04 | 138-hr | RTI |
| T04-11271 | T2017282 | 23-Dec-04 | 192-hr | RTI |
| T04-11272 | T2017283 | 23-Dec-04 | 192-hr | RTI |
| T04-11263 | T2017273 | 16-Dec-04 | 138-hr | EPA - NERL |
| T04-11264 | T2017274 | 16-Dec-04 | 138-hr | EPA - NERL |
| T04-11273 | T2017284 | 23-Dec-04 | 192-hr | EPA - NERL |
| T04-11274 | T2017285 | 23-Dec-04 | 192-hr | EPA - NERL |

Hidden replicate filters were present within each sample set. Table 7 shows that two of the loaded filters in each set were replicates of the same collection event. The results were reported to NAREL as mass of the element per square centimeter of deposit ($\mu\text{g}/\text{cm}^2$), and a one-sigma uncertainty was provided for each analytical result. Those results were multiplied by the total area of a filter deposit, 11.3 cm^2 , to produce final results in units of micrograms of the element per filter ($\mu\text{g}/\text{filter}$).

A request was made for each lab to provide specific information that will help us better understand how the analytical results were produced. A questionnaire was prepared and distributed to each lab. The questionnaire was designed to document those instrument conditions that were used to produce the XRF spectra. The information provided by each lab may be viewed in Tables 13 through 17 at the end of this report.

A second request was made for each lab to provide two specific XRF spectra. As requested, each lab provided the primary spectrum from which aluminum was determined for two samples. One spectrum was created during the analysis of a replicate PE sample collected on December 16, 2004 (see Table 7). The second spectrum was created during the analysis of a PE filter blank. These spectra are included in this report to serve as an example of the raw data produced at each lab.

XRF Results

A large number of XRF results were reported for this study. Forty-eight elements are routinely reported for each sample, and twenty-four samples were reported.


$$(48 \text{ elements/analysis}) \times (24 \text{ analyses}) = 1152 \text{ results}$$

CARB did not participate in this part of the study because the XRF lab was temporarily out of service due to the purchase of a new instrument. The results from all reporting laboratories are included in Table 12 at the end of this report. Table 12 also contains a median value calculated for some of the elements. A median value was calculated only when all of the reporting labs determined a concentration greater than three times the expressed uncertainty. Six of the heavy elements (Sm, Eu, Tb, Hf, Ta, and Ir) were not included in EPA's analysis, and therefore these EPA results are missing from the table of results.

All of the results have been compared to the median values by constructing a scatter plot shown in Figure 30. A log-log plot was constructed with the median values forming a straight line of unity slope. The corresponding results from all of the labs were superimposed on the median line. Most of the results were very near the median indicating good agreement among the participating labs. Even though Figure 30 gives a quick visual impression of many results that cover a wide range of concentrations, this scatter plot does not identify the element plotted nor the sample.

The more significant XRF results are presented again as stacked bar graphs in Figures 31 and 32. Each bar segment represents an individual value reported by one of the labs. Elements are identified along the horizontal axis, and the elements are arranged from left to right in order of decreasing concentration. The vertical axis of each bar graph is a linear scale, and each bar is normalized to the sum of results reported by all instruments identified in the legend. Each bar segment is color coded to identify the laboratory and labeled to show the reported concentration value. Again, the only results shown in the graphs are those that are significantly above the reported uncertainty. Those significant results can be identified in Table 12 by looking for a calculated median.

Figure 30






Figure 31

Figure 31 shows results from eight filter replicates created on December 16, 2004, and identified as samples T04-11257 through T04-11264 in Table 12. Two of these replicates were analyzed at each of the four participating laboratories. The most inconsistently reported element in Figure 31 was aluminum with values ranging from 0.56 to 2.13 µg/filter. It is worth noting that aluminum was a very small signal in the raw data spectra produced at all of the labs.

Figure 32 shows results from eight more filter replicates created on December 23, 2004, and identified as samples T04-11267 through T04-11274 in Table 12. The most inconsistent results observed in Figure 32 are for Al and Na, and both of these elements are observed as very small signals within the spectra produced at all of the labs.

Figure 32





Figure 33



A few spectra have been included in this report to give us an example of the raw data produced at each lab. Figure 33 shows two superimposed spectra that were produced at DRI. The conditions that produced these spectra are listed in column #1 of Table 13 at the end of this report. Al, Si, S, and K were detected above background in sample T04-11257 based upon these spectra.

Figure 34 shows two superimposed spectra that were produced at ODEQ. The conditions that produced these spectra are listed in column #1 of Table 14 at the end of this report. Al, Si, S, and K were detected above background in sample T04-11259 based upon these spectra. It is especially interesting to look at the signal for Al in all of the spectra, since all of the labs reported Al present [above background] in the December 16 replicates.

Figure 34




Figure 35





Figure 35 shows spectra that were produced at RTI using the conditions listed in column #1 of Table 16. Our last spectra shown in Figure 36 were produced at EPA's NERL facility using the conditions listed in column #5 of Table 17. We appreciate the effort that our participating labs made to provide us with the raw data presented here.

Figure 36







Figure 37

Figure 37 presents another view of the XRF results for aluminum which allows us to examine the uncertainty reported by each lab. Notice that the error bars represent a 3-sigma uncertainty which was used to select those results presented previously in Figures 30 through 32. Figure 37 shows results from eight filter replicates that were collected on December 16 and eight filter replicates that were collected on December 23. It is a worthy exercise to compare the spectra presented earlier with the uncertainties presented here. It is surprising that RTI consistently reported the smallest uncertainty for both collection events since RTI's spectra [in Figure 35] contain a significant interference very near aluminum.

Figure 38 presents a similar view of the XRF results for silicon. RTI reported substantially smaller uncertainties for silicon even though the spectrum shows silicon as a shoulder on the interference peak. All four of the labs actually determined silicon and aluminum from the spectra presented in this report.

Figure 38



| Table 8. Summary of XRF Results and Uncertainties (µg/filter) | | | | | | | | |
|---|----------------|----------------|----------------|----------------|-----------------|-----------------|----------------|----------------|
| | EPA Results | EPA Uncert. | RTI Results | RTI Uncert. | ODEQ Results | ODEQ Uncert. | DRI Results | DRI Uncert. |
| Replicates from Dec 16 | | | | | | | | |
| Mean | 1.475 | 0.141 | 1.169 | 0.034 | 1.164 | 0.261 | 1.272 | 0.194 |
| Max | 34.381 | 1.011 | 33.505 | 0.158 | 37.487 | 3.020 | 35.945 | 0.951 |
| Min | -0.659 | 0.014 | 0.000 | 0.004 | -1.153 | 0.008 | 0.000 | 0.006 |
| Std. Dev. | 5.266 | 0.208 | 4.823 | 0.035 | 5.399 | 0.591 | 5.025 | 0.253 |
| Count | 84 | 84 | 96 | 96 | 96 | 96 | 96 | 96 |
| Replicates from Dec 23 | | | | | | | | |
| Mean | 2.787 | 0.183 | 2.069 | 0.037 | 2.452 | 0.376 | 2.532 | 0.204 |
| Max | 84.316 | 2.287 | 74.987 | 0.237 | 88.042 | 7.077 | 85.959 | 0.993 |
| Min | -0.064 | 0.015 | 0.000 | 0.004 | -0.921 | 0.009 | 0.000 | 0.006 |
| Std. Dev. | 12.489 | 0.370 | 10.593 | 0.046 | 12.646 | 1.091 | 12.328 | 0.265 |
| Count | 84 | 84 | 96 | 96 | 96 | 96 | 96 | 96 |
| Blank Filters | | | | | | | | |
| Mean | 0.062 | 0.076 | 0.001 | 0.023 | -0.064 | 0.163 | 0.048 | 0.184 |
| Max | 1.037 | 0.554 | 0.023 | 0.131 | 0.042 | 2.258 | 1.401 | 0.942 |
| Min | -0.462 | 0.014 | 0.000 | 0.002 | -1.449 | 0.007 | 0.000 | 0.001 |
| Std. Dev. | 0.223 | 0.104 | 0.004 | 0.025 | 0.216 | 0.433 | 0.163 | 0.248 |
| Count | 84 | 84 | 96 | 96 | 96 | 96 | 96 | 96 |

Table 8 is a summary of the XRF results and the uncertainties grouped by sample type. For each sample type, two filters were analyzed at each lab. Each lab reported 96 results for each sample type, except for the EPA lab. We should remember that six of the heavy elements (Sm, Eu, Tb, Hf, Ta, and Ir) were not included in EPA's analysis, and this may skew the statistics to some extent. It is appropriate to compare these statistics as long as we fully appreciate the fact that there was no "true value" for any of the results with the possible exception of the blank filters. It is worth noting that for all three sample types, the mean uncertainty reported by RTI is considerably smaller than the mean uncertainty reported by the other labs. This may indicate a real difference in the way uncertainties are calculated at RTI, or it may indicate a real difference in the raw data itself. This report has presented only a small sample of the raw data.

Conclusions

This study was designed to evaluate the analytical performance of several PM_{2.5} speciation labs. The approach was simple. Each lab analyzed an almost identical set of blind PE samples, and the results reported from all of the labs have been compared. The scope of this study included four analytical techniques, and multiple methods were reported for IC, TOA carbon, and XRF. At least one EPA lab was able to report results for most of the methods used during this study.

Four labs analyzed a set of PE samples for gravimetric mass, and all of the labs performed well. Results for all of the samples were inside the 3-sigma advisory limits established by NAREL.

Five different labs reported IC results for at least one set of PE samples, and only one problem was observed in the IC results. NAREL reported poor analytical precision for chloride that was present in two replicates. No other problems were observed in the IC results.

Four labs analyzed a set of quartz PE filters, and all of the labs, except CARB, analyzed each filter multiple times in order to report results from more than one instrument and also report results using

more than one TOA method. Ultimately a total of fifteen data packages were used to report TOA results, and we should remember that each data package contained hidden replicates. Good precision was observed for all of the TC values reported, regardless of method and regardless of instrument. As expected, the precision was best for TC followed closely by OC. EC results for the STN method were lower than EC results reported for the IMPROVE and IMPROVE-a methods. The worst precision was observed for the DRI/OGC instruments running the IMPROVE method. There was some evidence in the raw data that a variable air leak may have contributed to the poor precision. Raw data from the CARB instrument also contained some evidence of an air leak. The thermograms included in this report help show critical information that is difficult to communicate with text.

None of the labs that reported XRF results used the same instrument. Therefore different hardware and different software were used to produce the results. XRF spectra were presented to illustrate the dramatic differences in raw data even though replicate samples were analyzed. Despite these facts, good agreement was observed for most of the elements that were significantly above the reported uncertainty. The largest disagreement in the XRF results was observed for aluminum and sodium. Both of these elements produce poor instrument response [compared to heavier elements], and larger uncertainties are expected for the lighter elements. This study has raised an important question about how uncertainties are calculated. There is no standard method for calculating the uncertainty. Each lab used a custom method to calculate the XRF uncertainty. It would be difficult to predict the outcome of using a single method at all labs since there were significant differences in the raw data.

Special effort has been made to collect information from the participating labs that help us better understand how the analytical results were produced. And that information has been included in this report. The author would like to take this opportunity to thank those individuals who answered questions, responded to the questionnaire, and provided the requested raw data. They have helped make this a better report!

References

1. CARB July 30, 2002. *Standard Operating Procedure for the Determination of PM_{2.5} Mass in Ambient Air by Gravimetric Analysis*, SOP MLD055, Northern Laboratory Branch, Monitoring and Laboratory Division, Air Resources Board, California Environmental Protection Agency. [currently available on the web] http://www.arb.ca.gov/aaqm/sop/MLD055_fin.pdf
2. DRI March 2002. *Standard Operating Procedure Gravimetric Analysis*, DRI SOP #2-114.2, Division of Atmospheric Sciences, Desert Research Institute, Reno, NV.
3. ODEQ 2003. *Standard Operating Procedure for Gravimetric Analysis of Particulate Collected with R&P Partisol Samplers and MetOne SASS Samplers*, Laboratory Division, State of Oregon Department of Environmental Quality, Portland, OR. [currently available on the web] <http://www.deq.state.or.us/lab/methods/DEQ04-LAB-0004-SOP.pdf>
4. RTI 2003. *Standard Operating Procedure for PM_{2.5} Gravimetric Analysis*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [currently available on the web] <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/gravmssop.pdf>
5. NAREL August 17, 2004. *Standard Operating Procedure, Mass Measurements for the PM_{2.5} Chemical Speciation QA Program*, National Air and Radiation Environmental Laboratory, U.S. Environmental Protection Agency, Montgomery, AL. [currently available on the web] http://epa.gov/narel/sops/sop_pm2.5_mass.pdf

6. CARB June 18, 2002. *Standard Operating Procedure for the Analysis of Anions and Cations in PM_{2.5} Speciation Samples by Ion Chromatography*, SOP MLD064, Northern Laboratory Branch, Monitoring and Laboratory Division, Air Resources Board, California Environmental Protection Agency. [currently available on the web] http://www.arb.ca.gov/aaqm/sop/MLD064_fin.pdf
7. DRI March 2002. *Standard Operating Procedure, Anion Analysis of Filter Extracts and Precipitation Samples by Ion Chromatography* DRI SOP #2-203.5, Division of Atmospheric Sciences, Desert Research Institute, Reno, NV.
8. DRI March 2002. *Standard Operating Procedure Cation Analysis of Filter Extracts and Precipitation Samples by Ion Chromatography* DRI SOP #2-208.1, Division of Atmospheric Sciences, Desert Research Institute, Reno, NV.
9. ODEQ 2003. *Standard Operating Procedure for Ion Chromatograph Analysis of Ambient Air Particulate Matter*, Laboratory Division, State of Oregon Department of Environmental Quality, Portland, OR. [currently available on the web] <http://www.deq.state.or.us/lab/methods/DEQ04-LAB-0005-SOP.pdf>
10. RTI 2003. *Standard Operating Procedure for PM_{2.5} Anion Analysis*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [currently available on the web] <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/anionsop.pdf>
11. RTI 2003. *Standard Operating Procedure for PM_{2.5} Cation Analysis*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [currently available on the web] <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/cationsop.pdf>
12. NAREL November 6, 2000. *Standard Operating Procedure, Anion Analysis for the PM_{2.5} Chemical Speciation QA Program*, National Air and Radiation Environmental Laboratory, U.S. Environmental Protection Agency, Montgomery, AL. [currently available on the web] http://www.epa.gov/narel/sops/sop_pm2.5_anion.pdf
13. NAREL July 31, 2000. *Standard Operating Procedure, Cation Analysis for the PM_{2.5} Chemical Speciation QA Program*, National Air and Radiation Environmental Laboratory, U.S. Environmental Protection Agency, Montgomery, AL. [currently available on the web] http://www.epa.gov/narel/sops/sop_pm2.5_cation.pdf
14. CARB August 1, 2002. *Standard Operating Procedure for Organic and Elemental Carbon Analysis of Exposed Quartz Microfiber Filters*, SOP MLD065, Northern Laboratory Branch, Monitoring and Laboratory Division, Air Resources Board, California Environmental Protection Agency. [currently available on the web] http://www.arb.ca.gov/aaqm/sop/MLD065_fin.pdf
15. DRI 2000. *DRI Standard Operating Procedure, Thermal/Reflectance Carbon Analysis of Aerosol Filter Samples*, Division of Atmospheric Sciences, Desert Research Institute, Reno, NV. [currently available on the web] <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/drioccc.pdf>
16. DRI March 2002. *Standard Operating Procedure DRI Model 2001 Thermal/Optical Carbon Analysis of Aerosol Filter Samples* DRI SOP #2-201.1, Division of Atmospheric Sciences, Desert Research Institute, Reno, NV.
17. RTI 2003. *Standard Operating Procedure for the Determination of Organic, Elemental, and Total Carbon in Particulate Matter Using a Thermal/Optical Transmittance Carbon Analyzer*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle

- Park, NC. [currently available on the web]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/ocecsop.pdf>
18. NAREL November 1, 2002. *Standard Operating Procedure, Carbon Analysis for the PM2.5 Chemical Speciation QA Program*, National Air and Radiation Environmental Laboratory, U.S. Environmental Protection Agency, Montgomery, AL. [currently available on the web]
http://www.epa.gov/narel/sops/sop_pm2.5_carbon.pdf
 19. EPA/NAREL. February 26, 2003. Technical Memorandum: CARB Laboratory Audit. U.S. Environmental Protection Agency. [currently available on the web]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/carbrept.pdf>
 20. DRI October 2004. *Standard Operating Procedure X-ray Fluorescence (XRF) Analysis of Aerosol Filter Samples (PANalytical Epsilon5)*, DRI SOP #2-209.1, Division of Atmospheric Sciences Desert Research Institute, Reno, NV.
 21. ODEQ 2003. *Standard Operating Procedure, Elemental Analysis of Air Particulate by Energy Dispersive X-Ray Fluorescence (EDXRF)*, Laboratory Division, State of Oregon Department of Environmental Quality, Portland, OR. [currently available on the web]
<http://www.deq.state.or.us/lab/methods/DEQ04-LAB-0006-SOP.pdf>
 22. RTI 2003. *Standard Operating Procedure for the X-Ray Fluorescence Analysis of PM2.5 Deposits on Teflon Filters*, Environmental & Industrial Measurements Division, Research Triangle Institute, Research triangle Park, NC. [currently available on the web]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/xrfsop.pdf>

Table 9. Gravimetric Mass PE Results

| Sample ID | Sample Description | Tare Mass | | Final Mass | | Captured PM _{2.5} | | Inter-Lab Difference* of Captured PM _{2.5} (mg) | Name of the Test Lab |
|------------|----------------------------------|-----------|---------|------------|---------|----------------------------|-------|---|----------------------|
| | | Test Lab | NAREL | Test Lab | NAREL | Test Lab | NAREL | | |
| | | (mg) | (mg) | (mg) | (mg) | (mg) | (mg) | | |
| T05-11285 | 24-hr collection event, 01/20/05 | 145.839 | 145.838 | 146.073 | 146.074 | 0.234 | 0.236 | 0.002 | CARB |
| T05-11286 | 24-hr collection event, 01/20/05 | 145.389 | 145.389 | 145.646 | 145.644 | 0.257 | 0.255 | -0.002 | CARB |
| T05-11287 | 48-hr collection event, 01/21/05 | 143.933 | 143.935 | 144.237 | 144.239 | 0.304 | 0.304 | 0.000 | CARB |
| T05-11288 | 48-hr collection event, 01/21/05 | 144.683 | 144.685 | 144.982 | 144.990 | 0.299 | 0.305 | 0.006 | CARB |
| T05-11289 | 12-hr collection event, 01/23/05 | 145.799 | 145.797 | 145.834 | 145.833 | 0.035 | 0.036 | 0.001 | CARB |
| T05-11290 | 12-hr collection event, 01/23/05 | 141.825 | 141.822 | 141.868 | 141.865 | 0.043 | 0.043 | 0.000 | CARB |
| T05-11291 | 24-hr collection event, 01/24/05 | 142.169 | 142.169 | 142.269 | 142.265 | 0.100 | 0.096 | -0.004 | CARB |
| T05-11292 | filter blank | 141.304 | 141.302 | 141.310 | 141.306 | 0.006 | 0.004 | -0.002 | CARB |
| T05-11293 | filter blank | 141.055 | 141.054 | 141.058 | 141.056 | 0.003 | 0.002 | -0.001 | CARB |
| T05-11294 | filter blank | 142.774 | 142.773 | 142.781 | 142.777 | 0.007 | 0.004 | -0.003 | CARB |
| MW05-11325 | metallic transfer weight | 94.833 | 94.834 | 94.831 | 94.834 | -0.002 | 0.000 | 0.002 | DRI |
| MW05-11326 | metallic transfer weight | 190.521 | 190.521 | 190.520 | 190.522 | -0.001 | 0.001 | 0.002 | DRI |
| T05-11295 | 24-hr collection event, 01/20/05 | 143.964 | 143.952 | 144.191 | 144.182 | 0.227 | 0.230 | 0.003 | DRI |
| T05-11296 | 24-hr collection event, 01/20/05 | 144.544 | 144.531 | 144.778 | 144.768 | 0.234 | 0.237 | 0.003 | DRI |
| T05-11297 | 48-hr collection event, 01/21/05 | 143.429 | 143.415 | 143.724 | 143.716 | 0.295 | 0.301 | 0.006 | DRI |
| T05-11298 | 48-hr collection event, 01/21/05 | 141.519 | 141.506 | 141.817 | 141.806 | 0.298 | 0.300 | 0.002 | DRI |
| T05-11299 | 12-hr collection event, 01/23/05 | 140.397 | 140.383 | 140.435 | 140.424 | 0.038 | 0.041 | 0.003 | DRI |
| T05-11300 | 12-hr collection event, 01/23/05 | 141.449 | 141.436 | 141.488 | 141.476 | 0.039 | 0.040 | 0.001 | DRI |
| T05-11301 | 24-hr collection event, 01/24/05 | 142.167 | 142.155 | 142.263 | 142.252 | 0.096 | 0.097 | 0.001 | DRI |
| T05-11302 | filter blank | 143.707 | 143.694 | 143.710 | 143.696 | 0.003 | 0.002 | -0.001 | DRI |
| T05-11303 | filter blank | 139.756 | 139.744 | 139.759 | 139.745 | 0.003 | 0.001 | -0.002 | DRI |
| T05-11304 | filter blank | 143.332 | 143.318 | 143.333 | 143.320 | 0.001 | 0.002 | 0.001 | DRI |
| MW05-11327 | metallic transfer weight | 97.351 | 97.356 | 97.350 | 97.356 | -0.001 | 0.000 | 0.001 | DRI |
| MW05-11328 | metallic transfer weight | 196.224 | 196.235 | 196.223 | 196.235 | -0.001 | 0.000 | 0.001 | DRI |

Table 9. Gravimetric Mass PE Results

| | | Tare Mass | | Final Mass | | Captured PM _{2.5} | | Inter-Lab Difference* | Name of the Test Lab |
|------------|----------------------------------|---------------|------------|---------------|------------|----------------------------|------------|------------------------------------|----------------------|
| Sample ID | Sample Description | Test Lab (mg) | NAREL (mg) | Test Lab (mg) | NAREL (mg) | Test Lab (mg) | NAREL (mg) | of Captured PM _{2.5} (mg) | |
| T05-11305 | 24-hr collection event, 01/20/05 | 144.336 | 144.337 | 144.559 | 144.566 | 0.223 | 0.229 | 0.006 | ODEQ |
| T05-11306 | 24-hr collection event, 01/20/05 | 142.974 | 142.972 | 143.197 | 143.205 | 0.223 | 0.233 | 0.010 | ODEQ |
| T05-11307 | 48-hr collection event, 01/21/05 | 139.657 | 139.657 | 139.951 | 139.961 | 0.294 | 0.304 | 0.010 | ODEQ |
| T05-11308 | 48-hr collection event, 01/21/05 | 141.301 | 141.301 | 141.587 | 141.594 | 0.286 | 0.293 | 0.007 | ODEQ |
| T05-11309 | 12-hr collection event, 01/23/05 | 142.031 | 142.029 | 142.062 | 142.072 | 0.031 | 0.043 | 0.012 | ODEQ |
| T05-11310 | 12-hr collection event, 01/23/05 | 141.468 | 141.468 | 141.500 | 141.505 | 0.032 | 0.037 | 0.005 | ODEQ |
| T05-11311 | 24-hr collection event, 01/24/05 | 142.483 | 142.484 | 142.575 | 142.584 | 0.092 | 0.100 | 0.008 | ODEQ |
| T05-11312 | filter blank | 141.487 | 141.487 | 141.489 | 141.490 | 0.002 | 0.003 | 0.001 | ODEQ |
| T05-11313 | filter blank | 142.208 | 142.208 | 142.209 | 142.211 | 0.001 | 0.003 | 0.002 | ODEQ |
| T05-11314 | filter blank | 140.697 | 140.698 | 140.698 | 140.703 | 0.001 | 0.005 | 0.004 | ODEQ |
| MW05-11329 | metallic transfer weight | 93.775 | 93.776 | 93.775 | 93.776 | 0.000 | 0.000 | 0.000 | ODEQ |
| MW05-11330 | metallic transfer weight | 188.879 | 188.880 | 188.878 | 188.880 | -0.001 | 0.000 | 0.001 | ODEQ |
| T05-11315 | 24-hr collection event, 01/20/05 | 139.752 | 139.751 | 139.972 | 139.978 | 0.220 | 0.227 | 0.007 | RTI analyst 1 |
| T05-11316 | 24-hr collection event, 01/20/05 | 139.527 | 139.527 | 139.749 | 139.765 | 0.222 | 0.238 | 0.016 | RTI analyst 1 |
| T05-11317 | 48-hr collection event, 01/21/05 | 142.196 | 142.196 | 142.483 | 142.491 | 0.287 | 0.295 | 0.008 | RTI analyst 1 |
| T05-11318 | 48-hr collection event, 01/21/05 | 142.533 | 142.531 | 142.829 | 142.838 | 0.296 | 0.307 | 0.011 | RTI analyst 1 |
| T05-11319 | 12-hr collection event, 01/23/05 | 141.306 | 141.306 | 141.336 | 141.344 | 0.030 | 0.038 | 0.008 | RTI analyst 1 |
| T05-11320 | 12-hr collection event, 01/23/05 | 140.549 | 140.549 | 140.584 | 140.593 | 0.035 | 0.044 | 0.009 | RTI analyst 1 |
| T05-11321 | 24-hr collection event, 01/24/05 | 140.300 | 140.297 | 140.389 | 140.400 | 0.089 | 0.103 | 0.014 | RTI analyst 1 |
| T05-11322 | filter blank | 141.646 | 141.648 | 141.648 | 141.651 | 0.002 | 0.003 | 0.001 | RTI analyst 1 |
| T05-11323 | filter blank | 145.693 | 145.692 | 145.692 | 145.696 | -0.001 | 0.004 | 0.005 | RTI analyst 1 |
| T05-11324 | filter blank | 141.573 | 141.574 | 141.573 | 141.578 | 0.000 | 0.004 | 0.004 | RTI analyst 1 |
| MW05-11331 | metallic transfer weight | 97.545 | 97.546 | 97.545 | 97.546 | 0.000 | 0.000 | 0.000 | RTI analyst 1 |
| MW05-11332 | metallic transfer weight | 192.421 | 192.422 | 192.422 | 192.421 | 0.001 | -0.001 | -0.002 | RTI analyst 1 |

Table 10. Ion Chromatography PE Results

| Sample ID | Filter | Sample Description | Lab | Method | Concentration (µg/filter) | | | | | | |
|-----------|--------|------------------------|-------|--------|---------------------------|---------|---------|---------|----------|-----------|--------|
| | Medium | | | | Chloride | Nitrate | Nitrite | Sulfate | Ammonium | Potassium | Sodium |
| N04-11197 | Nylon® | 116-hr event, 11/24/04 | CARB | STN | ---- | 22.484 | ---- | 72.523 | 27.236 | 1.898 | 1.737 |
| N04-11198 | Nylon® | 116-hr event, 11/24/04 | CARB | STN | ---- | 24.106 | ---- | 72.282 | 26.203 | 1.877 | 1.712 |
| N04-11199 | Nylon® | 116-hr event, 11/24/04 | DRI | STN | ---- | 24.039 | ---- | 77.679 | 29.991 | 1.865 | 0.948 |
| N04-11200 | Nylon® | 116-hr event, 11/24/04 | DRI | STN | ---- | 23.931 | ---- | 75.704 | 28.340 | 1.943 | 1.166 |
| N04-11201 | Nylon® | 116-hr event, 11/24/04 | ODEQ | STN | ---- | 26.700 | ---- | 79.100 | 28.600 | 2.070 | <3.6 |
| N04-11202 | Nylon® | 116-hr event, 11/24/04 | ODEQ | STN | ---- | 25.300 | ---- | 77.600 | 27.600 | 1.920 | <3.6 |
| N04-11203 | Nylon® | 116-hr event, 11/24/04 | RTI | STN | ---- | 24.496 | ---- | 76.586 | 25.559 | 2.147 | 1.744 |
| N04-11204 | Nylon® | 116-hr event, 11/24/04 | RTI | STN | ---- | 23.129 | ---- | 73.608 | 25.117 | 2.126 | 1.796 |
| N04-11205 | Nylon® | 116-hr event, 11/24/04 | NAREL | STN | ---- | 24.977 | ---- | 74.782 | 23.975 | 2.289 | 1.061 |
| N04-11206 | Nylon® | 116-hr event, 11/24/04 | NAREL | STN | ---- | 25.489 | ---- | 74.382 | 24.138 | 2.051 | 0.941 |
| N04-11208 | Nylon® | 159-hr event, 11/29/04 | CARB | STN | ---- | 64.668 | ---- | 179.290 | 79.350 | 4.506 | 3.679 |
| N04-11209 | Nylon® | 159-hr event, 11/29/04 | CARB | STN | ---- | 71.148 | ---- | 182.763 | 83.395 | 4.547 | 3.130 |
| N04-11210 | Nylon® | 159-hr event, 11/29/04 | DRI | STN | ---- | 71.388 | ---- | 172.308 | 84.330 | 4.320 | 2.445 |
| N04-11211 | Nylon® | 159-hr event, 11/29/04 | DRI | STN | ---- | 73.851 | ---- | 169.050 | 84.293 | 4.205 | 2.610 |
| N04-11212 | Nylon® | 159-hr event, 11/29/04 | ODEQ | STN | ---- | 79.700 | ---- | 178.000 | 80.700 | 3.940 | <3.6 |
| N04-11213 | Nylon® | 159-hr event, 11/29/04 | ODEQ | STN | ---- | 71.400 | ---- | 181.000 | 78.300 | 3.970 | <3.6 |
| N04-11214 | Nylon® | 159-hr event, 11/29/04 | RTI | STN | ---- | 63.602 | ---- | 173.368 | 75.905 | 3.960 | 3.397 |
| N04-11215 | Nylon® | 159-hr event, 11/29/04 | RTI | STN | ---- | 63.719 | ---- | 172.912 | 75.583 | 3.967 | 2.798 |
| N04-11216 | Nylon® | 159-hr event, 11/29/04 | NAREL | STN | ---- | 62.957 | ---- | 168.679 | 65.752 | 4.888 | 2.411 |
| N04-11217 | Nylon® | 159-hr event, 11/29/04 | NAREL | STN | ---- | 68.779 | ---- | 173.945 | 69.657 | 5.044 | 2.330 |
| N04-11219 | Nylon® | filter blank | CARB | STN | ---- | BMDL* | ---- | BMDL* | BMDL* | BMDL* | BMDL* |
| N04-11220 | Nylon® | filter blank | CARB | STN | ---- | 3.323 | ---- | BMDL* | BMDL* | BMDL* | BMDL* |
| N04-11221 | Nylon® | filter blank | DRI | STN | ---- | 0.000 | ---- | 0.000 | 0.189 | 0.000 | 0.081 |
| N04-11222 | Nylon® | filter blank | DRI | STN | ---- | 0.000 | ---- | 0.000 | 0.114 | 0.000 | 0.000 |
| N04-11223 | Nylon® | filter blank | ODEQ | STN | ---- | <1.4 | ---- | <1.4 | <0.72 | <1.1 | <3.6 |
| N04-11224 | Nylon® | filter blank | ODEQ | STN | ---- | <1.4 | ---- | <1.4 | <0.72 | <1.1 | <3.6 |
| N04-11225 | Nylon® | filter blank | RTI | STN | ---- | 1.537 | ---- | BMDL* | BMDL* | BMDL* | 0.050 |
| N04-11226 | Nylon® | filter blank | RTI | STN | ---- | 0.878 | ---- | BMDL* | BMDL* | BMDL* | 0.030 |

Table 10. Ion Chromatography PE Results

| Sample ID | Filter | Sample Description | Lab | Method | Concentration (µg/filter) | | | | | | |
|-----------|---------|------------------------|-------|---------|---------------------------|---------|---------|---------|----------|-----------|--------|
| | Medium | | | | Chloride | Nitrate | Nitrite | Sulfate | Ammonium | Potassium | Sodium |
| N04-11227 | Nylon® | filter blank | NAREL | STN | ----- | BMDL* | ----- | BMDL* | BMDL* | BMDL* | BMDL* |
| N04-11228 | Nylon® | filter blank | NAREL | STN | ----- | BMDL* | ----- | BMDL* | BMDL* | BMDL* | BMDL* |
| N04-11229 | Nylon® | 161-hr event, 12/07/04 | RTI | IMPROVE | 7.641 | 63.645 | 2.698 | 163.338 | 60.791 | ----- | ----- |
| N04-11230 | Nylon® | 161-hr event, 12/07/04 | RTI | IMPROVE | 8.273 | 65.815 | 1.533 | 161.422 | 60.242 | ----- | ----- |
| N04-11231 | Nylon® | 161-hr event, 12/07/04 | NAREL | IMPROVE | 8.285 | 64.306 | 0.602 | 156.938 | 53.281 | ----- | ----- |
| N04-11232 | Nylon® | 161-hr event, 12/07/04 | NAREL | IMPROVE | 3.506 | 53.172 | 0.633 | 158.214 | 53.945 | ----- | ----- |
| N04-11233 | Nylon® | 161-hr event, 12/07/04 | RTI | IMPROVE | 2.836 | 39.038 | 0.855 | 142.790 | 51.997 | ----- | ----- |
| N04-11234 | Nylon® | 161-hr event, 12/07/04 | RTI | IMPROVE | 2.926 | 39.349 | 0.976 | 141.670 | 50.925 | ----- | ----- |
| N04-11235 | Nylon® | 161-hr event, 12/07/04 | NAREL | IMPROVE | 3.133 | 41.645 | 0.676 | 137.541 | 45.648 | ----- | ----- |
| N04-11236 | Nylon® | 161-hr event, 12/07/04 | NAREL | IMPROVE | 5.474 | 48.054 | 0.948 | 143.390 | 47.280 | ----- | ----- |
| N04-11237 | Nylon® | filter blank | RTI | IMPROVE | 0.083 | BMDL* | 0.716 | BMDL* | BMDL* | ----- | ----- |
| N04-11238 | Nylon® | filter blank | RTI | IMPROVE | 0.062 | BMDL* | 1.048 | BMDL* | BMDL* | ----- | ----- |
| N04-11239 | Nylon® | filter blank | NAREL | IMPROVE | BMDL* | BMDL* | 0.440 | BMDL* | BMDL* | ----- | ----- |
| N04-11240 | Nylon® | filter blank | NAREL | IMPROVE | BMDL* | BMDL* | BMDL* | BMDL* | BMDL* | ----- | ----- |
| T05-11333 | Teflon® | 144-hr event, 01/03/05 | DRI | STN | ----- | 12.370 | ----- | 139.584 | 46.423 | 2.202 | 5.493 |
| T05-11334 | Teflon® | 144-hr event, 01/03/05 | DRI | STN | ----- | 12.889 | ----- | 141.185 | 46.128 | 2.289 | 5.657 |
| T05-11335 | Teflon® | 144-hr event, 01/03/05 | NAREL | STN | ----- | 11.555 | ----- | 125.824 | 42.733 | 2.882 | 5.204 |
| T05-11336 | Teflon® | 144-hr event, 01/03/05 | NAREL | STN | ----- | 10.988 | ----- | 126.785 | 43.562 | 2.546 | 4.885 |
| T05-11337 | Teflon® | 216-hr event, 01/04/05 | DRI | STN | ----- | 10.274 | ----- | 241.056 | 75.974 | 4.119 | 8.731 |
| T05-11338 | Teflon® | 216-hr event, 01/04/05 | DRI | STN | ----- | 11.050 | ----- | 245.164 | 75.126 | 4.385 | 9.891 |
| T05-11339 | Teflon® | 216-hr event, 01/04/05 | NAREL | STN | ----- | 9.125 | ----- | 217.998 | 67.936 | 5.356 | 9.503 |
| T05-11340 | Teflon® | 216-hr event, 01/04/05 | NAREL | STN | ----- | 7.224 | ----- | 217.855 | 69.121 | 5.431 | 9.136 |
| T05-11341 | Teflon® | filter blank | DRI | STN | ----- | BMDL* | ----- | BMDL* | BMDL* | BMDL* | BMDL* |
| T05-11342 | Teflon® | filter blank | DRI | STN | ----- | BMDL* | ----- | BMDL* | BMDL* | BMDL* | BMDL* |
| T05-11343 | Teflon® | filter blank | NAREL | STN | ----- | BMDL* | ----- | BMDL* | BMDL* | BMDL* | 0.180 |
| T05-11344 | Teflon® | filter blank | NAREL | STN | ----- | 0.533 | ----- | BMDL* | BMDL* | BMDL* | 0.204 |

**BMDL = Below Method Detection Limit*

Table 11. TOA Carbon PE Results

| Sample ID | Sample Description | Lab | Instrument (see text)* | Method | Concentration ($\mu\text{g C}/\text{cm}^2$) | | | | | | | |
|-----------|------------------------|-------|---------------------------|------------|---|----------------|----------------|----------------|---------------|---------------|----------------|----------------|
| | | | | | EC | OC | TC | OC1 | OC2 | OC3 | OC4 | Pyrol C |
| Q04-11175 | 287-hr event, 04/27/04 | CARB | #1 | STN (mod.) | 2.91 | 35.24 | 38.14 | ----- | ----- | ----- | ----- | ----- |
| Q04-11176 | 287-hr event, 04/27/04 | CARB | #1 | STN (mod.) | 3.22 | 35.11 | 38.33 | ----- | ----- | ----- | ----- | ----- |
| Q04-11177 | 287-hr event, 04/27/04 | DRI | #9 | STN | 4.6 ± 1.5 | 30.9 ± 3.7 | 35.5 ± 3.7 | 11.2 ± 4.8 | 5.8 ± 1.2 | 4.0 ± 0.9 | 11.2 ± 1.4 | -1.2 ± 1.0 |
| Q04-11177 | 287-hr event, 04/27/04 | DRI | #7 | STN | 1.6 ± 0.5 | 34.2 ± 4.1 | 35.9 ± 3.7 | 7.8 ± 3.3 | 6.8 ± 1.4 | 4.5 ± 1.0 | 9.2 ± 1.1 | 5.9 ± 4.9 |
| Q04-11178 | 287-hr event, 04/27/04 | DRI | #9 | STN | 1.9 ± 0.6 | 33.8 ± 4.1 | 35.7 ± 3.7 | 11.2 ± 4.8 | 4.9 ± 1.0 | 2.8 ± 0.7 | 9.0 ± 1.1 | 5.8 ± 4.8 |
| Q04-11178 | 287-hr event, 04/27/04 | DRI | #7 | STN | 1.5 ± 0.5 | 33.8 ± 4.1 | 35.3 ± 3.7 | 8.1 ± 3.4 | 5.5 ± 1.2 | 3.7 ± 0.9 | 8.5 ± 1.0 | 8.0 ± 6.6 |
| Q04-11181 | 287-hr event, 04/27/04 | RTI | R | STN | 2.7 ± 0.3 | 32.9 ± 1.8 | 35.5 ± 2.1 | 9.68 | 3.68 | 1.82 | 3.89 | 13.80 |
| Q04-11181 | 287-hr event, 04/27/04 | RTI | R | STN | 2.6 ± 0.3 | 32.7 ± 1.8 | 35.3 ± 2.1 | 9.66 | 3.66 | 1.67 | 3.12 | 14.59 |
| Q04-11181 | 287-hr event, 04/27/04 | RTI | S | STN | 2.5 ± 0.3 | 34.8 ± 1.9 | 37.3 ± 2.2 | 11.15 | 3.67 | 2.82 | 7.06 | 10.13 |
| Q04-11181 | 287-hr event, 04/27/04 | RTI | T | STN | 3.9 ± 0.4 | 33.2 ± 1.9 | 37.1 ± 2.2 | 9.80 | 3.98 | 2.67 | 5.40 | 11.34 |
| Q04-11181 | 287-hr event, 04/27/04 | RTI | F | STN | 2.6 ± 0.3 | 34.2 ± 1.9 | 36.9 ± 2.1 | 9.22 | 4.41 | 2.35 | 5.28 | 12.96 |
| Q04-11182 | 287-hr event, 04/27/04 | RTI | R | STN | 2.4 ± 0.3 | 32.6 ± 1.8 | 34.9 ± 2.0 | 9.66 | 3.59 | 1.64 | 3.12 | 14.55 |
| Q04-11182 | 287-hr event, 04/27/04 | RTI | S | STN | 3.0 ± 0.4 | 33.0 ± 1.8 | 36.0 ± 2.1 | 10.60 | 3.58 | 2.73 | 6.77 | 9.29 |
| Q04-11182 | 287-hr event, 04/27/04 | RTI | T | STN | 3.8 ± 0.4 | 31.6 ± 1.8 | 35.5 ± 2.1 | 9.03 | 4.04 | 2.46 | 5.01 | 11.11 |
| Q04-11182 | 287-hr event, 04/27/04 | RTI | T | STN | 3.9 ± 0.4 | 32.3 ± 1.8 | 36.2 ± 2.1 | 9.40 | 3.93 | 2.39 | 4.58 | 11.96 |
| Q04-11182 | 287-hr event, 04/27/04 | RTI | F | STN | 2.7 ± 0.3 | 33.2 ± 1.9 | 35.9 ± 2.1 | 9.20 | 4.19 | 1.76 | 4.81 | 13.25 |
| Q04-11183 | 287-hr event, 04/27/04 | NAREL | #1 | STN | 2.5 ± 0.3 | 32.0 ± 1.8 | 34.4 ± 2.0 | 10.68 | 3.09 | 1.65 | 3.21 | 13.32 |
| Q04-11184 | 287-hr event, 04/27/04 | NAREL | #1 | STN | 3.0 ± 0.3 | 32.9 ± 1.8 | 35.9 ± 2.1 | 11.32 | 3.29 | 1.59 | 3.34 | 13.32 |
| Q04-11186 | 192-hr event, 11/16/04 | CARB | #1 | STN (mod.) | 1.32 | 21.89 | 23.21 | ----- | ----- | ----- | ----- | ----- |
| Q04-11187 | 192-hr event, 11/16/04 | CARB | #1 | STN (mod.) | 1.44 | 22.15 | 23.59 | ----- | ----- | ----- | ----- | ----- |
| Q04-11188 | 192-hr event, 11/16/04 | DRI | #9 | STN | 3.2 ± 1.0 | 19.7 ± 2.4 | 22.9 ± 2.4 | 5.5 ± 2.4 | 4.4 ± 0.9 | 3.1 ± 0.7 | 9.5 ± 1.2 | -2.8 ± 2.3 |
| Q04-11188 | 192-hr event, 11/16/04 | DRI | #7 | STN | 3.1 ± 1.0 | 19.8 ± 2.4 | 23.0 ± 2.4 | 4.0 ± 1.7 | 5.0 ± 1.1 | 3.7 ± 0.9 | 8.6 ± 1.1 | -1.4 ± 1.2 |
| Q04-11189 | 192-hr event, 11/16/04 | DRI | #9 | STN | 3.2 ± 1.0 | 18.3 ± 2.2 | 21.5 ± 2.3 | 5.3 ± 2.2 | 3.9 ± 0.8 | 2.8 ± 0.7 | 9.1 ± 1.1 | -2.8 ± 2.3 |
| Q04-11189 | 192-hr event, 11/16/04 | DRI | #7 | STN | 3.0 ± 0.9 | 18.6 ± 2.2 | 21.5 ± 2.3 | 3.7 ± 1.6 | 4.5 ± 1.0 | 3.4 ± 0.8 | 8.0 ± 1.0 | -1.0 ± 0.8 |
| Q04-11192 | 192-hr event, 11/16/04 | RTI | R | STN | 2.0 ± 0.3 | 19.4 ± 1.2 | 21.4 ± 1.4 | 4.44 | 2.65 | 1.33 | 2.51 | 8.48 |
| Q04-11192 | 192-hr event, 11/16/04 | RTI | S | STN | 2.3 ± 0.3 | 21.2 ± 1.3 | 23.5 ± 1.5 | 5.55 | 3.00 | 2.11 | 4.65 | 5.88 |

Table 11. TOA Carbon PE Results

| Sample ID | Sample Description | Lab | Instrument (see text)* | Method | Concentration ($\mu\text{g C}/\text{cm}^2$) | | | | | | | |
|-----------|------------------------|-------|---------------------------|------------|---|----------------|----------------|---------------|---------------|---------------|---------------|---------------|
| | | | | | EC | OC | TC | OC1 | OC2 | OC3 | OC4 | Pyrol C |
| Q04-11192 | 192-hr event, 11/16/04 | RTI | T | STN | 2.8 ± 0.3 | 19.4 ± 1.2 | 22.2 ± 1.4 | 4.79 | 2.50 | 2.05 | 4.02 | 6.02 |
| Q04-11192 | 192-hr event, 11/16/04 | RTI | F | STN | 2.8 ± 0.3 | 19.0 ± 1.2 | 21.9 ± 1.4 | 3.99 | 3.29 | 1.59 | 3.68 | 6.46 |
| Q04-11192 | 192-hr event, 11/16/04 | RTI | F | STN | 2.8 ± 0.3 | 19.0 ± 1.1 | 21.8 ± 1.4 | 4.03 | 3.25 | 1.55 | 3.72 | 6.43 |
| Q04-11193 | 192-hr event, 11/16/04 | RTI | R | STN | 2.1 ± 0.3 | 19.7 ± 1.2 | 21.8 ± 1.4 | 4.62 | 2.73 | 1.33 | 2.30 | 8.68 |
| Q04-11193 | 192-hr event, 11/16/04 | RTI | S | STN | 2.5 ± 0.3 | 20.4 ± 1.2 | 23.0 ± 1.4 | 5.37 | 2.94 | 2.25 | 5.39 | 4.47 |
| Q04-11193 | 192-hr event, 11/16/04 | RTI | S | STN | 2.6 ± 0.3 | 19.8 ± 1.2 | 22.4 ± 1.4 | 5.24 | 2.79 | 2.36 | 6.01 | 3.39 |
| Q04-11193 | 192-hr event, 11/16/04 | RTI | T | STN | 2.8 ± 0.3 | 19.5 ± 1.2 | 22.3 ± 1.4 | 4.91 | 2.57 | 1.96 | 3.58 | 6.52 |
| Q04-11193 | 192-hr event, 11/16/04 | RTI | F | STN | 2.6 ± 0.3 | 20.5 ± 1.2 | 23.1 ± 1.5 | 4.62 | 3.34 | 1.63 | 3.46 | 7.46 |
| Q04-11194 | 192-hr event, 11/16/04 | NAREL | #1 | STN | 2.2 ± 0.3 | 19.3 ± 1.2 | 21.5 ± 1.4 | 5.21 | 2.57 | 1.38 | 2.31 | 7.85 |
| Q04-11195 | 192-hr event, 11/16/04 | NAREL | #1 | STN | 2.2 ± 0.3 | 19.4 ± 1.2 | 21.7 ± 1.4 | 5.19 | 2.58 | 1.41 | 2.50 | 7.75 |
| Q04-11241 | filter blank | CARB | #1 | STN (mod.) | <0.8 | <0.8 | <0.8 | ----- | ----- | ----- | ----- | ----- |
| Q04-11242 | filter blank | CARB | #1 | STN (mod.) | <0.8 | <0.8 | <0.8 | ----- | ----- | ----- | ----- | ----- |
| Q04-11243 | filter blank | DRI | #9 | STN | 0.0 ± 0.1 | 0.5 ± 0.3 | 0.5 ± 0.3 | 0.2 ± 0.1 | 0.3 ± 0.1 | 0.0 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11243 | filter blank | DRI | #7 | STN | 0.0 ± 0.1 | 0.4 ± 0.3 | 0.4 ± 0.3 | 0.2 ± 0.1 | 0.2 ± 0.1 | 0.0 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11244 | filter blank | DRI | #9 | STN | 0.0 ± 0.1 | 0.3 ± 0.3 | 0.3 ± 0.3 | 0.1 ± 0.0 | 0.2 ± 0.1 | 0.1 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11244 | filter blank | DRI | #7 | STN | 0.0 ± 0.1 | 0.5 ± 0.3 | 0.5 ± 0.3 | 0.3 ± 0.1 | 0.2 ± 0.1 | 0.0 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11247 | filter blank | RTI | R | STN | 0.0 ± 0.2 | 0.1 ± 0.2 | 0.1 ± 0.3 | 0.06 | 0.04 | 0.01 | 0.00 | 0.00 |
| Q04-11247 | filter blank | RTI | R | STN | 0.1 ± 0.2 | 0.4 ± 0.2 | 0.5 ± 0.3 | 0.06 | 0.20 | 0.12 | 0.07 | 0.00 |
| Q04-11247 | filter blank | RTI | S | STN | 0.0 ± 0.2 | 0.2 ± 0.2 | 0.2 ± 0.3 | 0.03 | 0.12 | 0.04 | 0.00 | 0.00 |
| Q04-11247 | filter blank | RTI | T | STN | 0.0 ± 0.2 | 0.1 ± 0.2 | 0.1 ± 0.3 | 0.02 | 0.05 | 0.05 | 0.00 | 0.00 |
| Q04-11247 | filter blank | RTI | F | STN | 0.0 ± 0.2 | 0.3 ± 0.2 | 0.3 ± 0.3 | 0.06 | 0.10 | 0.04 | 0.04 | 0.02 |
| Q04-11248 | filter blank | RTI | R | STN | 0.0 ± 0.2 | 0.2 ± 0.2 | 0.2 ± 0.3 | 0.05 | 0.08 | 0.03 | 0.01 | 0.00 |
| Q04-11248 | filter blank | RTI | S | STN | 0.0 ± 0.2 | 0.3 ± 0.2 | 0.3 ± 0.3 | 0.03 | 0.15 | 0.05 | 0.05 | 0.01 |
| Q04-11248 | filter blank | RTI | T | STN | 0.0 ± 0.2 | 0.2 ± 0.2 | 0.2 ± 0.3 | 0.03 | 0.04 | 0.05 | 0.03 | 0.00 |
| Q04-11248 | filter blank | RTI | T | STN | 0.0 ± 0.2 | 0.1 ± 0.2 | 0.1 ± 0.3 | 0.02 | 0.04 | 0.05 | 0.01 | 0.00 |
| Q04-11248 | filter blank | RTI | F | STN | 0.0 ± 0.2 | 0.0 ± 0.2 | 0.0 ± 0.3 | 0.00 | 0.00 | 0.00 | 0.00 | 0.00 |

Table 11. TOA Carbon PE Results

| Sample ID | Sample Description | Lab | Instrument (see text)* | Method | Concentration ($\mu\text{g C}/\text{cm}^2$) | | | | | | | |
|-----------|------------------------|-------|---------------------------|---------|---|----------------|----------------|---------------|---------------|----------------|----------------|---------------|
| | | | | | EC | OC | TC | OC1 | OC2 | OC3 | OC4 | Pyrol C |
| Q04-11249 | filter blank | NAREL | #1 | STN | 0.0 ± 0.2 | 0.5 ± 0.2 | 0.5 ± 0.3 | 0.06 | 0.22 | 0.09 | 0.07 | 0.03 |
| Q04-11250 | filter blank | NAREL | #1 | STN | 0.0 ± 0.2 | 0.4 ± 0.2 | 0.4 ± 0.3 | 0.06 | 0.19 | 0.08 | 0.06 | -0.02 |
| Q04-11177 | 287-hr event, 04/27/04 | DRI | #4 | IMPROVE | 7.1 ± 3.8 | 31.5 ± 2.1 | 38.6 ± 1.9 | 2.3 ± 0.4 | 7.5 ± 0.6 | 11.0 ± 2.8 | 9.0 ± 1.8 | 1.8 ± 2.2 |
| Q04-11177 | 287-hr event, 04/27/04 | DRI | #5 | IMPROVE | 4.4 ± 2.4 | 33.6 ± 2.2 | 37.9 ± 1.8 | 2.3 ± 0.4 | 7.2 ± 0.6 | 13.3 ± 3.4 | 10.8 ± 2.1 | 0.0 ± 0.1 |
| Q04-11178 | 287-hr event, 04/27/04 | DRI | #4 | IMPROVE | 4.3 ± 2.3 | 33.3 ± 2.2 | 37.6 ± 1.8 | 1.7 ± 0.3 | 7.7 ± 0.6 | 13.4 ± 3.4 | 10.5 ± 2.1 | 0.0 ± 0.1 |
| Q04-11178 | 287-hr event, 04/27/04 | DRI | #5 | IMPROVE | 7.2 ± 3.9 | 31.8 ± 2.1 | 39.0 ± 1.9 | 2.3 ± 0.4 | 7.5 ± 0.6 | 11.4 ± 2.9 | 9.4 ± 1.8 | 1.3 ± 1.5 |
| Q04-11181 | 287-hr event, 04/27/04 | RTI | F | IMPROVE | 6.8 ± 0.5 | 31.9 ± 1.8 | 38.8 ± 2.2 | 0.05 | 7.00 | 9.42 | 5.48 | 9.97 |
| Q04-11182 | 287-hr event, 04/27/04 | RTI | F | IMPROVE | 6.9 ± 0.5 | 31.0 ± 1.7 | 37.8 ± 2.2 | 0.06 | 6.69 | 9.19 | 4.83 | 10.18 |
| Q04-11183 | 287-hr event, 04/27/04 | NAREL | #2 | IMPROVE | 7.2 ± 0.6 | 28.9 ± 1.6 | 36.1 ± 2.1 | 0.24 | 11.69 | 4.45 | 3.77 | 8.73 |
| Q04-11184 | 287-hr event, 04/27/04 | NAREL | #2 | IMPROVE | 6.7 ± 0.5 | 29.8 ± 1.7 | 36.5 ± 2.1 | 0.88 | 11.39 | 4.65 | 3.69 | 9.21 |
| Q04-11188 | 192-hr event, 11/16/04 | DRI | #4 | IMPROVE | 3.6 ± 1.9 | 19.5 ± 1.3 | 23.0 ± 1.1 | 1.4 ± 0.3 | 3.1 ± 0.3 | 7.8 ± 2.0 | 7.2 ± 1.4 | 0.0 ± 0.1 |
| Q04-11188 | 192-hr event, 11/16/04 | DRI | #5 | IMPROVE | 6.2 ± 3.3 | 18.6 ± 1.2 | 24.8 ± 1.2 | 1.0 ± 0.2 | 3.9 ± 0.3 | 7.2 ± 1.8 | 6.5 ± 1.3 | 0.0 ± 0.1 |
| Q04-11189 | 192-hr event, 11/16/04 | DRI | #4 | IMPROVE | 3.4 ± 1.8 | 18.4 ± 1.2 | 21.9 ± 1.1 | 0.8 ± 0.1 | 3.3 ± 0.3 | 7.6 ± 1.9 | 6.8 ± 1.3 | 0.0 ± 0.1 |
| Q04-11189 | 192-hr event, 11/16/04 | DRI | #5 | IMPROVE | 6.7 ± 3.6 | 16.6 ± 1.1 | 23.3 ± 1.2 | 0.9 ± 0.2 | 3.4 ± 0.3 | 6.7 ± 1.7 | 5.7 ± 1.1 | 0.0 ± 0.1 |
| Q04-11192 | 192-hr event, 11/16/04 | RTI | F | IMPROVE | 7.1 ± 0.6 | 15.9 ± 1.0 | 23.0 ± 1.4 | 0.01 | 2.61 | 5.40 | 2.97 | 4.87 |
| Q04-11193 | 192-hr event, 11/16/04 | RTI | F | IMPROVE | 5.9 ± 0.5 | 17.7 ± 1.1 | 23.6 ± 1.5 | 0.05 | 3.50 | 6.02 | 2.80 | 5.34 |
| Q04-11194 | 192-hr event, 11/16/04 | NAREL | #2 | IMPROVE | 6.7 ± 0.5 | 16.4 ± 1.0 | 23.1 ± 1.5 | 0.19 | 5.85 | 3.71 | 2.74 | 3.93 |
| Q04-11195 | 192-hr event, 11/16/04 | NAREL | #2 | IMPROVE | 6.6 ± 0.5 | 15.7 ± 1.0 | 22.3 ± 1.4 | 0.17 | 5.58 | 3.62 | 2.44 | 3.83 |
| Q04-11243 | filter blank | DRI | #4 | IMPROVE | 0.0 ± 0.1 | 0.4 ± 0.4 | 0.4 ± 0.4 | 0.0 ± 0.0 | 0.0 ± 0.1 | 0.3 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11243 | filter blank | DRI | #5 | IMPROVE | 0.0 ± 0.1 | 0.3 ± 0.3 | 0.3 ± 0.4 | 0.0 ± 0.0 | 0.0 ± 0.1 | 0.3 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11244 | filter blank | DRI | #4 | IMPROVE | 0.0 ± 0.1 | 0.2 ± 0.3 | 0.2 ± 0.3 | 0.0 ± 0.0 | 0.0 ± 0.1 | 0.2 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11244 | filter blank | DRI | #5 | IMPROVE | 0.0 ± 0.1 | 0.5 ± 0.4 | 0.5 ± 0.4 | 0.0 ± 0.0 | 0.0 ± 0.1 | 0.4 ± 0.2 | 0.1 ± 0.1 | 0.0 ± 0.1 |
| Q04-11247 | filter blank | RTI | F | IMPROVE | 0.1 ± 0.2 | 0.5 ± 0.2 | 0.5 ± 0.3 | 0.03 | 0.06 | 0.18 | 0.12 | 0.07 |
| Q04-11248 | filter blank | RTI | F | IMPROVE | 0.0 ± 0.2 | 0.2 ± 0.2 | 0.2 ± 0.3 | 0.00 | 0.00 | 0.07 | 0.07 | 0.07 |
| Q04-11249 | filter blank | NAREL | #2 | IMPROVE | 0.0 ± 0.2 | 0.3 ± 0.2 | 0.3 ± 0.3 | 0.04 | 0.06 | 0.20 | 0.02 | 0.01 |
| Q04-11250 | filter blank | NAREL | #2 | IMPROVE | 0.0 ± 0.2 | 0.4 ± 0.2 | 0.4 ± 0.3 | 0.04 | 0.06 | 0.19 | 0.06 | 0.08 |

Table 11. TOA Carbon PE Results

| Sample ID | Sample Description | Lab | Instrument (see text)* | Method | Concentration ($\mu\text{g C/cm}^2$) | | | | | | | |
|-----------|------------------------|-------|---------------------------|-----------|--|----------------|----------------|---------------|---------------|----------------|---------------|---------------|
| | | | | | EC | OC | TC | OC1 | OC2 | OC3 | OC4 | Pyrol C |
| Q04-11177 | 287-hr event, 04/27/04 | DRI | #7 | IMPROVE-a | 7.2 ± 0.8 | 27.8 ± 1.0 | 35.0 ± 1.3 | 2.5 ± 0.3 | 5.6 ± 1.6 | 9.6 ± 4.0 | 5.7 ± 0.6 | 4.4 ± 0.6 |
| Q04-11177 | 287-hr event, 04/27/04 | DRI | #9 | IMPROVE-a | 6.8 ± 0.7 | 29.0 ± 1.1 | 35.8 ± 1.3 | 4.2 ± 0.5 | 7.2 ± 2.0 | 7.4 ± 3.1 | 6.6 ± 0.7 | 3.7 ± 0.5 |
| Q04-11178 | 287-hr event, 04/27/04 | DRI | #7 | IMPROVE-a | 7.5 ± 0.8 | 28.6 ± 1.0 | 36.1 ± 1.3 | 2.5 ± 0.3 | 5.8 ± 1.6 | 10.2 ± 4.2 | 6.0 ± 0.6 | 4.3 ± 0.5 |
| Q04-11178 | 287-hr event, 04/27/04 | DRI | #9 | IMPROVE-a | 6.4 ± 0.7 | 27.9 ± 1.0 | 34.3 ± 1.2 | 3.2 ± 0.4 | 7.7 ± 2.2 | 6.7 ± 2.8 | 6.3 ± 0.6 | 3.9 ± 0.5 |
| Q04-11183 | 287-hr event, 04/27/04 | NAREL | #2 | IMPROVE-a | 6.5 ± 0.5 | 28.7 ± 1.6 | 35.3 ± 2.1 | 3.07 | 9.33 | 5.43 | 4.72 | 6.19 |
| Q04-11184 | 287-hr event, 04/27/04 | NAREL | #2 | IMPROVE-a | 5.9 ± 0.5 | 30.1 ± 1.7 | 36.0 ± 2.1 | 3.49 | 9.50 | 5.30 | 4.82 | 7.02 |
| Q04-11188 | 192-hr event, 11/16/04 | DRI | #7 | IMPROVE-a | 6.6 ± 0.7 | 15.5 ± 0.6 | 22.2 ± 0.8 | 1.1 ± 0.1 | 3.0 ± 0.8 | 6.8 ± 2.8 | 2.5 ± 0.3 | 2.2 ± 0.3 |
| Q04-11188 | 192-hr event, 11/16/04 | DRI | #9 | IMPROVE-a | 6.1 ± 0.7 | 14.9 ± 0.6 | 21.0 ± 0.8 | 1.1 ± 0.1 | 4.3 ± 1.2 | 5.2 ± 2.2 | 2.4 ± 0.3 | 1.9 ± 0.2 |
| Q04-11189 | 192-hr event, 11/16/04 | DRI | #7 | IMPROVE-a | 6.4 ± 0.7 | 14.5 ± 0.6 | 20.8 ± 0.8 | 1.0 ± 0.1 | 2.8 ± 0.8 | 6.4 ± 2.7 | 2.5 ± 0.3 | 1.8 ± 0.2 |
| Q04-11189 | 192-hr event, 11/16/04 | DRI | #9 | IMPROVE-a | 5.6 ± 0.6 | 14.9 ± 0.6 | 20.6 ± 0.8 | 0.9 ± 0.1 | 4.4 ± 1.2 | 4.8 ± 2.0 | 2.9 ± 0.3 | 2.0 ± 0.3 |
| Q04-11194 | 192-hr event, 11/16/04 | NAREL | #2 | IMPROVE-a | 5.8 ± 0.5 | 17.1 ± 1.1 | 22.9 ± 1.5 | 1.36 | 5.63 | 3.71 | 2.83 | 3.56 |
| Q04-11195 | 192-hr event, 11/16/04 | NAREL | #2 | IMPROVE-a | 6.3 ± 0.5 | 16.1 ± 1.0 | 22.4 ± 1.4 | 1.17 | 5.66 | 3.58 | 2.33 | 3.33 |
| Q04-11243 | filter blank | DRI | #7 | IMPROVE-a | 0.0 ± 0.1 | 0.5 ± 0.5 | 0.5 ± 0.5 | 0.1 ± 0.0 | 0.1 ± 0.1 | 0.3 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11243 | filter blank | DRI | #9 | IMPROVE-a | 0.0 ± 0.1 | 0.3 ± 0.4 | 0.3 ± 0.4 | 0.0 ± 0.0 | 0.1 ± 0.1 | 0.3 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11244 | filter blank | DRI | #7 | IMPROVE-a | 0.0 ± 0.1 | 0.6 ± 0.6 | 0.6 ± 0.6 | 0.1 ± 0.0 | 0.1 ± 0.1 | 0.4 ± 0.3 | 0.1 ± 0.1 | 0.0 ± 0.1 |
| Q04-11244 | filter blank | DRI | #9 | IMPROVE-a | 0.0 ± 0.1 | 0.1 ± 0.3 | 0.1 ± 0.3 | 0.0 ± 0.0 | 0.0 ± 0.1 | 0.1 ± 0.2 | 0.0 ± 0.1 | 0.0 ± 0.1 |
| Q04-11249 | filter blank | NAREL | #2 | IMPROVE-a | ----- | ----- | ----- | ----- | ----- | ----- | ----- | ----- |
| Q04-11250 | filter blank | NAREL | #2 | IMPROVE-a | ----- | ----- | ----- | ----- | ----- | ----- | ----- | ----- |

* Instruments identified as CARB #1, DRI #7, and DRI #9 are DRI/Model 2001 instruments capable of the TOR and the TOT analysis. The DRI #4 and #5 instruments are older DRI/OGC instruments set up for the TOR analysis. RTI instruments identified as R, S, T, and the NAREL #1 instrument are early model Sunset instruments set up for the TOT analysis. The instruments identified as RTI F and NAREL #2 are newer Sunset Dual Mode instruments capable of the TOR and the TOT analysis.

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Na | 11 | 138-hr event | T04-11257 | 4.676 ± 0.859 | T04-11259 | 2.536 ± 2.291 | T04-11261 | 2.091 ± 0.124 | T04-11263 | 4.609 ± 0.805 | ----- |
| Mg | 12 | 138-hr event | T04-11257 | 0.599 ± 0.817 | T04-11259 | -0.022 ± 0.405 | T04-11261 | 0.025 ± 0.043 | T04-11263 | 1.514 ± 0.426 | ----- |
| Al | 13 | 138-hr event | T04-11257 | 1.659 ± 0.550 | T04-11259 | 0.839 ± 0.115 | T04-11261 | 0.558 ± 0.041 | T04-11263 | 1.570 ± 0.428 | 1.20 |
| Si | 14 | 138-hr event | T04-11257 | 4.800 ± 0.299 | T04-11259 | 5.637 ± 0.495 | T04-11261 | 5.196 ± 0.035 | T04-11263 | 6.428 ± 0.534 | 5.42 |
| P | 15 | 138-hr event | T04-11257 | 1.370 ± 0.054 | T04-11259 | -0.209 ± 0.085 | T04-11261 | 0.000 ± 0.130 | T04-11263 | 0.492 ± 0.242 | ----- |
| S | 16 | 138-hr event | T04-11257 | 35.945 ± 0.295 | T04-11259 | 37.487 ± 3.020 | T04-11261 | 32.499 ± 0.158 | T04-11263 | 34.381 ± 1.011 | 35.16 |
| Cl | 17 | 138-hr event | T04-11257 | 0.411 ± 0.038 | T04-11259 | 0.059 ± 0.270 | T04-11261 | 0.622 ± 0.037 | T04-11263 | 0.576 ± 0.091 | ----- |
| K | 19 | 138-hr event | T04-11257 | 5.512 ± 0.006 | T04-11259 | 5.546 ± 0.449 | T04-11261 | 5.206 ± 0.044 | T04-11263 | 5.219 ± 0.159 | 5.37 |
| Ca | 20 | 138-hr event | T04-11257 | 1.867 ± 0.047 | T04-11259 | 2.082 ± 0.174 | T04-11261 | 1.914 ± 0.029 | T04-11263 | 1.808 ± 0.065 | 1.89 |
| Sc | 21 | 138-hr event | T04-11257 | 0.000 ± 0.028 | T04-11259 | 0.002 ± 0.025 | T04-11261 | 0.000 ± 0.041 | T04-11263 | 0.016 ± 0.022 | ----- |
| Ti | 22 | 138-hr event | T04-11257 | 0.072 ± 0.026 | T04-11259 | 0.198 ± 0.049 | T04-11261 | 0.000 ± 0.025 | T04-11263 | 0.143 ± 0.045 | ----- |
| V | 23 | 138-hr event | T04-11257 | 0.009 ± 0.009 | T04-11259 | 0.029 ± 0.017 | T04-11261 | 0.078 ± 0.012 | T04-11263 | 0.013 ± 0.016 | ----- |
| Cr | 24 | 138-hr event | T04-11257 | 0.005 ± 0.040 | T04-11259 | 0.007 ± 0.008 | T04-11261 | 0.000 ± 0.010 | T04-11263 | 0.060 ± 0.015 | ----- |
| Mn | 25 | 138-hr event | T04-11257 | 0.133 ± 0.114 | T04-11259 | 0.090 ± 0.016 | T04-11261 | 0.097 ± 0.008 | T04-11263 | 0.132 ± 0.031 | ----- |
| Fe | 26 | 138-hr event | T04-11257 | 3.239 ± 0.180 | T04-11259 | 3.045 ± 0.246 | T04-11261 | 3.225 ± 0.025 | T04-11263 | 3.086 ± 0.107 | 3.16 |
| Co | 27 | 138-hr event | T04-11257 | 0.009 ± 0.019 | T04-11259 | -0.018 ± 0.019 | T04-11261 | 0.000 ± 0.009 | T04-11263 | -0.054 ± 0.026 | ----- |
| Ni | 28 | 138-hr event | T04-11257 | 0.000 ± 0.031 | T04-11259 | 0.001 ± 0.008 | T04-11261 | 0.096 ± 0.005 | T04-11263 | 0.030 ± 0.018 | ----- |
| Cu | 29 | 138-hr event | T04-11257 | 0.170 ± 0.024 | T04-11259 | 0.141 ± 0.015 | T04-11261 | 0.172 ± 0.006 | T04-11263 | 0.193 ± 0.020 | 0.17 |
| Zn | 30 | 138-hr event | T04-11257 | 0.689 ± 0.034 | T04-11259 | 0.663 ± 0.054 | T04-11261 | 0.613 ± 0.009 | T04-11263 | 0.708 ± 0.055 | 0.68 |
| Ga | 31 | 138-hr event | T04-11257 | 0.041 ± 0.097 | T04-11259 | -0.015 ± 0.054 | T04-11261 | 0.003 ± 0.004 | T04-11263 | 0.009 ± 0.028 | ----- |
| As | 33 | 138-hr event | T04-11257 | 0.066 ± 0.026 | T04-11259 | 0.079 ± 0.018 | T04-11261 | 0.124 ± 0.006 | T04-11263 | 0.065 ± 0.041 | ----- |
| Se | 34 | 138-hr event | T04-11257 | 0.000 ± 0.023 | T04-11259 | 0.048 ± 0.011 | T04-11261 | 0.067 ± 0.006 | T04-11263 | 0.119 ± 0.025 | ----- |
| Br | 35 | 138-hr event | T04-11257 | 0.199 ± 0.026 | T04-11259 | 0.188 ± 0.018 | T04-11261 | 0.216 ± 0.007 | T04-11263 | 0.220 ± 0.029 | 0.21 |
| Rb | 37 | 138-hr event | T04-11257 | 0.000 ± 0.024 | T04-11259 | -0.005 ± 0.009 | T04-11261 | 0.007 ± 0.008 | T04-11263 | 0.030 ± 0.024 | ----- |
| Sr | 38 | 138-hr event | T04-11257 | 0.007 ± 0.057 | T04-11259 | 0.007 ± 0.010 | T04-11261 | 0.008 ± 0.009 | T04-11263 | 0.021 ± 0.057 | ----- |
| Y | 39 | 138-hr event | T04-11257 | 0.066 ± 0.037 | T04-11259 | -0.006 ± 0.011 | T04-11261 | 0.014 ± 0.010 | T04-11263 | -0.010 ± 0.059 | ----- |
| Zr | 40 | 138-hr event | T04-11257 | 0.104 ± 0.078 | T04-11259 | -0.002 ± 0.012 | T04-11261 | 0.000 ± 0.011 | T04-11263 | 0.079 ± 0.053 | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Nb | 41 | 138-hr event | T04-11257 | 0.000 ± 0.062 | T04-11259 | -0.009 ± 0.014 | T04-11261 | 0.000 ± 0.009 | T04-11263 | -0.026 ± 0.055 | ----- |
| Mo | 42 | 138-hr event | T04-11257 | 0.016 ± 0.081 | T04-11259 | 0.005 ± 0.016 | T04-11261 | 0.000 ± 0.010 | T04-11263 | 0.003 ± 0.053 | ----- |
| Ag | 47 | 138-hr event | T04-11257 | 0.032 ± 0.085 | T04-11259 | -0.005 ± 0.034 | T04-11261 | 0.000 ± 0.039 | T04-11263 | 0.101 ± 0.236 | ----- |
| Cd | 48 | 138-hr event | T04-11257 | 0.000 ± 0.094 | T04-11259 | 0.009 ± 0.035 | T04-11261 | 0.000 ± 0.045 | T04-11263 | 0.046 ± 0.135 | ----- |
| In | 49 | 138-hr event | T04-11257 | 0.014 ± 0.092 | T04-11259 | -0.012 ± 0.036 | T04-11261 | 0.114 ± 0.061 | T04-11263 | 0.106 ± 0.245 | ----- |
| Sn | 50 | 138-hr event | T04-11257 | 0.000 ± 0.108 | T04-11259 | 0.040 ± 0.040 | T04-11261 | 0.000 ± 0.062 | T04-11263 | 0.047 ± 0.128 | ----- |
| Sb | 51 | 138-hr event | T04-11257 | 0.226 ± 0.098 | T04-11259 | 0.012 ± 0.044 | T04-11261 | 0.701 ± 0.136 | T04-11263 | 0.209 ± 0.095 | ----- |
| Cs | 55 | 138-hr event | T04-11257 | 0.099 ± 0.225 | T04-11259 | -0.052 ± 0.073 | T04-11261 | 0.000 ± 0.043 | T04-11263 | 0.125 ± 0.054 | ----- |
| Ba | 56 | 138-hr event | T04-11257 | 0.145 ± 0.243 | T04-11259 | 0.008 ± 0.101 | T04-11261 | 0.000 ± 0.055 | T04-11263 | 0.207 ± 0.091 | ----- |
| La | 57 | 138-hr event | T04-11257 | 0.000 ± 0.496 | T04-11259 | -0.075 ± 0.125 | T04-11261 | 0.000 ± 0.038 | T04-11263 | 0.318 ± 0.068 | ----- |
| Ce | 58 | 138-hr event | T04-11257 | 0.203 ± 0.412 | T04-11259 | -0.049 ± 0.155 | T04-11261 | 0.000 ± 0.043 | T04-11263 | 0.043 ± 0.051 | ----- |
| Sm | 62 | 138-hr event | T04-11257 | 0.000 ± 0.683 | T04-11259 | -0.223 ± 0.558 | T04-11261 | 0.000 ± 0.021 | T04-11263 | not reported | ----- |
| Eu | 63 | 138-hr event | T04-11257 | 0.735 ± 0.871 | T04-11259 | -1.153 ± 0.843 | T04-11261 | 0.000 ± 0.026 | T04-11263 | not reported | ----- |
| Tb | 65 | 138-hr event | T04-11257 | 0.052 ± 0.945 | T04-11259 | -1.086 ± 1.823 | T04-11261 | 0.147 ± 0.062 | T04-11263 | not reported | ----- |
| Hf | 72 | 138-hr event | T04-11257 | 0.000 ± 0.229 | T04-11259 | -0.047 ± 0.204 | T04-11261 | 0.049 ± 0.024 | T04-11263 | not reported | ----- |
| Ta | 73 | 138-hr event | T04-11257 | 0.000 ± 0.122 | T04-11259 | -0.100 ± 0.227 | T04-11261 | 0.000 ± 0.018 | T04-11263 | not reported | ----- |
| W | 74 | 138-hr event | T04-11257 | 0.000 ± 0.292 | T04-11259 | -0.079 ± 0.058 | T04-11261 | 0.000 ± 0.011 | T04-11263 | 0.056 ± 0.078 | ----- |
| Ir | 77 | 138-hr event | T04-11257 | 0.000 ± 0.105 | T04-11259 | -0.014 ± 0.034 | T04-11261 | 0.000 ± 0.008 | T04-11263 | not reported | ----- |
| Au | 79 | 138-hr event | T04-11257 | 0.000 ± 0.111 | T04-11259 | -0.046 ± 0.026 | T04-11261 | 0.023 ± 0.009 | T04-11263 | -0.091 ± 0.044 | ----- |
| Hg | 80 | 138-hr event | T04-11257 | 0.000 ± 0.042 | T04-11259 | -0.011 ± 0.022 | T04-11261 | 0.133 ± 0.011 | T04-11263 | -0.073 ± 0.042 | ----- |
| Pb | 82 | 138-hr event | T04-11257 | 0.084 ± 0.081 | T04-11259 | 0.199 ± 0.034 | T04-11261 | 0.255 ± 0.020 | T04-11263 | 0.352 ± 0.077 | ----- |
| Na | 11 | 138-hr event | T04-11258 | 5.015 ± 0.866 | T04-11260 | 3.153 ± 2.302 | T04-11262 | 1.774 ± 0.124 | T04-11264 | 4.488 ± 0.794 | ----- |
| Mg | 12 | 138-hr event | T04-11258 | 0.260 ± 0.814 | T04-11260 | -0.095 ± 0.406 | T04-11262 | 0.000 ± 0.057 | T04-11264 | 2.330 ± 0.460 | ----- |
| Al | 13 | 138-hr event | T04-11258 | 2.133 ± 0.553 | T04-11260 | 0.890 ± 0.119 | T04-11262 | 1.125 ± 0.042 | T04-11264 | 1.361 ± 0.424 | 1.24 |
| Si | 14 | 138-hr event | T04-11258 | 4.518 ± 0.296 | T04-11260 | 5.464 ± 0.480 | T04-11262 | 5.708 ± 0.036 | T04-11264 | 6.391 ± 0.536 | 5.59 |
| P | 15 | 138-hr event | T04-11258 | 1.268 ± 0.054 | T04-11260 | -0.213 ± 0.083 | T04-11262 | 0.000 ± 0.132 | T04-11264 | 0.015 ± 0.239 | ----- |
| S | 16 | 138-hr event | T04-11258 | 32.793 ± 0.270 | T04-11260 | 36.366 ± 2.930 | T04-11262 | 33.505 ± 0.158 | T04-11264 | 32.938 ± 0.977 | 33.22 |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Cl | 17 | 138-hr event | T04-11258 | 0.423 ± 0.038 | T04-11260 | 0.092 ± 0.263 | T04-11262 | 0.703 ± 0.038 | T04-11264 | 0.659 ± 0.095 | ----- |
| K | 19 | 138-hr event | T04-11258 | 5.105 ± 0.006 | T04-11260 | 5.353 ± 0.433 | T04-11262 | 5.429 ± 0.045 | T04-11264 | 4.876 ± 0.153 | 5.23 |
| Ca | 20 | 138-hr event | T04-11258 | 1.709 ± 0.047 | T04-11260 | 2.070 ± 0.173 | T04-11262 | 2.204 ± 0.029 | T04-11264 | 2.318 ± 0.075 | 2.14 |
| Sc | 21 | 138-hr event | T04-11258 | 0.000 ± 0.028 | T04-11260 | -0.054 ± 0.026 | T04-11262 | 0.000 ± 0.040 | T04-11264 | -0.006 ± 0.024 | ----- |
| Ti | 22 | 138-hr event | T04-11258 | 0.140 ± 0.026 | T04-11260 | 0.098 ± 0.051 | T04-11262 | 0.000 ± 0.025 | T04-11264 | 0.236 ± 0.045 | ----- |
| V | 23 | 138-hr event | T04-11258 | 0.000 ± 0.009 | T04-11260 | 0.007 ± 0.018 | T04-11262 | 0.086 ± 0.012 | T04-11264 | -0.004 ± 0.016 | ----- |
| Cr | 24 | 138-hr event | T04-11258 | 0.005 ± 0.040 | T04-11260 | 0.036 ± 0.009 | T04-11262 | 0.000 ± 0.010 | T04-11264 | 0.020 ± 0.014 | ----- |
| Mn | 25 | 138-hr event | T04-11258 | 0.111 ± 0.114 | T04-11260 | 0.084 ± 0.015 | T04-11262 | 0.081 ± 0.008 | T04-11264 | 0.175 ± 0.033 | ----- |
| Fe | 26 | 138-hr event | T04-11258 | 3.137 ± 0.180 | T04-11260 | 3.023 ± 0.244 | T04-11262 | 3.578 ± 0.026 | T04-11264 | 2.967 ± 0.105 | 3.08 |
| Co | 27 | 138-hr event | T04-11258 | 0.020 ± 0.019 | T04-11260 | -0.036 ± 0.019 | T04-11262 | 0.000 ± 0.010 | T04-11264 | -0.047 ± 0.025 | ----- |
| Ni | 28 | 138-hr event | T04-11258 | 0.023 ± 0.031 | T04-11260 | 0.015 ± 0.009 | T04-11262 | 1.597 ± 0.014 | T04-11264 | 0.048 ± 0.019 | ----- |
| Cu | 29 | 138-hr event | T04-11258 | 0.136 ± 0.024 | T04-11260 | 0.185 ± 0.018 | T04-11262 | 0.166 ± 0.007 | T04-11264 | 0.204 ± 0.021 | 0.18 |
| Zn | 30 | 138-hr event | T04-11258 | 0.622 ± 0.034 | T04-11260 | 0.654 ± 0.054 | T04-11262 | 0.637 ± 0.009 | T04-11264 | 0.746 ± 0.056 | 0.65 |
| Ga | 31 | 138-hr event | T04-11258 | 0.063 ± 0.097 | T04-11260 | -0.003 ± 0.054 | T04-11262 | 0.006 ± 0.004 | T04-11264 | -0.032 ± 0.028 | ----- |
| As | 33 | 138-hr event | T04-11258 | 0.000 ± 0.026 | T04-11260 | 0.066 ± 0.019 | T04-11262 | 0.124 ± 0.006 | T04-11264 | 0.059 ± 0.038 | ----- |
| Se | 34 | 138-hr event | T04-11258 | 0.000 ± 0.023 | T04-11260 | 0.045 ± 0.011 | T04-11262 | 0.068 ± 0.006 | T04-11264 | 0.049 ± 0.022 | ----- |
| Br | 35 | 138-hr event | T04-11258 | 0.188 ± 0.026 | T04-11260 | 0.192 ± 0.018 | T04-11262 | 0.217 ± 0.007 | T04-11264 | 0.184 ± 0.031 | 0.19 |
| Rb | 37 | 138-hr event | T04-11258 | 0.000 ± 0.024 | T04-11260 | 0.007 ± 0.010 | T04-11262 | 0.003 ± 0.008 | T04-11264 | 0.031 ± 0.024 | ----- |
| Sr | 38 | 138-hr event | T04-11258 | 0.007 ± 0.057 | T04-11260 | 0.021 ± 0.010 | T04-11262 | 0.013 ± 0.009 | T04-11264 | -0.078 ± 0.053 | ----- |
| Y | 39 | 138-hr event | T04-11258 | 0.020 ± 0.037 | T04-11260 | -0.001 ± 0.011 | T04-11262 | 0.000 ± 0.007 | T04-11264 | -0.007 ± 0.060 | ----- |
| Zr | 40 | 138-hr event | T04-11258 | 0.059 ± 0.077 | T04-11260 | 0.002 ± 0.013 | T04-11262 | 0.000 ± 0.011 | T04-11264 | 0.050 ± 0.051 | ----- |
| Nb | 41 | 138-hr event | T04-11258 | 0.099 ± 0.062 | T04-11260 | 0.001 ± 0.015 | T04-11262 | 0.000 ± 0.009 | T04-11264 | 0.133 ± 0.067 | ----- |
| Mo | 42 | 138-hr event | T04-11258 | 0.005 ± 0.081 | T04-11260 | 0.004 ± 0.018 | T04-11262 | 0.000 ± 0.009 | T04-11264 | 0.177 ± 0.066 | ----- |
| Ag | 47 | 138-hr event | T04-11258 | 0.032 ± 0.085 | T04-11260 | 0.026 ± 0.035 | T04-11262 | 0.000 ± 0.040 | T04-11264 | -0.659 ± 0.246 | ----- |
| Cd | 48 | 138-hr event | T04-11258 | 0.181 ± 0.095 | T04-11260 | 0.069 ± 0.037 | T04-11262 | 0.066 ± 0.063 | T04-11264 | 0.426 ± 0.147 | ----- |
| In | 49 | 138-hr event | T04-11258 | 0.002 ± 0.092 | T04-11260 | -0.017 ± 0.038 | T04-11262 | 0.158 ± 0.059 | T04-11264 | 0.149 ± 0.244 | ----- |
| Sn | 50 | 138-hr event | T04-11258 | 0.000 ± 0.108 | T04-11260 | 0.075 ± 0.043 | T04-11262 | 0.193 ± 0.081 | T04-11264 | -0.148 ± 0.124 | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Sb | 51 | 138-hr event | T04-11258 | 0.090 ± 0.097 | T04-11260 | 0.021 ± 0.046 | T04-11262 | 0.000 ± 0.104 | T04-11264 | 0.224 ± 0.101 | ----- |
| Cs | 55 | 138-hr event | T04-11258 | 0.000 ± 0.224 | T04-11260 | 0.087 ± 0.080 | T04-11262 | 0.000 ± 0.043 | T04-11264 | 0.074 ± 0.054 | ----- |
| Ba | 56 | 138-hr event | T04-11258 | 0.000 ± 0.240 | T04-11260 | 0.095 ± 0.110 | T04-11262 | 0.000 ± 0.055 | T04-11264 | 0.044 ± 0.092 | ----- |
| La | 57 | 138-hr event | T04-11258 | 0.000 ± 0.493 | T04-11260 | 0.010 ± 0.138 | T04-11262 | 0.000 ± 0.037 | T04-11264 | 0.200 ± 0.066 | ----- |
| Ce | 58 | 138-hr event | T04-11258 | 0.000 ± 0.410 | T04-11260 | -0.029 ± 0.170 | T04-11262 | 0.000 ± 0.043 | T04-11264 | 0.150 ± 0.051 | ----- |
| Sm | 62 | 138-hr event | T04-11258 | 0.000 ± 0.693 | T04-11260 | -0.301 ± 0.609 | T04-11262 | 0.000 ± 0.021 | T04-11264 | not reported | ----- |
| Eu | 63 | 138-hr event | T04-11258 | 0.000 ± 0.852 | T04-11260 | -1.004 ± 0.922 | T04-11262 | 0.000 ± 0.027 | T04-11264 | not reported | ----- |
| Tb | 65 | 138-hr event | T04-11258 | 0.346 ± 0.951 | T04-11260 | -0.679 ± 2.000 | T04-11262 | 0.090 ± 0.064 | T04-11264 | not reported | ----- |
| Hf | 72 | 138-hr event | T04-11258 | 0.154 ± 0.229 | T04-11260 | 0.037 ± 0.204 | T04-11262 | 0.049 ± 0.024 | T04-11264 | not reported | ----- |
| Ta | 73 | 138-hr event | T04-11258 | 0.000 ± 0.122 | T04-11260 | -0.056 ± 0.227 | T04-11262 | 0.000 ± 0.024 | T04-11264 | not reported | ----- |
| W | 74 | 138-hr event | T04-11258 | 0.007 ± 0.293 | T04-11260 | 0.011 ± 0.058 | T04-11262 | 0.000 ± 0.013 | T04-11264 | -0.026 ± 0.078 | ----- |
| Ir | 77 | 138-hr event | T04-11258 | 0.000 ± 0.104 | T04-11260 | 0.016 ± 0.034 | T04-11262 | 0.000 ± 0.008 | T04-11264 | not reported | ----- |
| Au | 79 | 138-hr event | T04-11258 | 0.000 ± 0.111 | T04-11260 | -0.007 ± 0.026 | T04-11262 | 0.015 ± 0.009 | T04-11264 | 0.013 ± 0.046 | ----- |
| Hg | 80 | 138-hr event | T04-11258 | 0.034 ± 0.042 | T04-11260 | -0.002 ± 0.022 | T04-11262 | 0.136 ± 0.012 | T04-11264 | 0.054 ± 0.055 | ----- |
| Pb | 82 | 138-hr event | T04-11258 | 0.185 ± 0.081 | T04-11260 | 0.230 ± 0.036 | T04-11262 | 0.255 ± 0.020 | T04-11264 | 0.255 ± 0.073 | ----- |
| Na | 11 | 192-hr event | T04-11267 | 3.580 ± 0.835 | T04-11269 | 8.509 ± 2.450 | T04-11271 | 2.870 ± 0.147 | T04-11273 | 5.835 ± 0.918 | 4.71 |
| Mg | 12 | 192-hr event | T04-11267 | 0.339 ± 0.815 | T04-11269 | 0.021 ± 0.416 | T04-11271 | 0.071 ± 0.050 | T04-11273 | 1.544 ± 0.450 | ----- |
| Al | 13 | 192-hr event | T04-11267 | 2.156 ± 0.553 | T04-11269 | 0.344 ± 0.100 | T04-11271 | 0.497 ± 0.044 | T04-11273 | 1.998 ± 0.522 | 1.25 |
| Si | 14 | 192-hr event | T04-11267 | 9.060 ± 0.350 | T04-11269 | 10.352 ± 0.906 | T04-11271 | 9.260 ± 0.042 | T04-11273 | 12.486 ± 0.769 | 9.81 |
| P | 15 | 192-hr event | T04-11267 | 3.607 ± 0.062 | T04-11269 | -0.500 ± 0.189 | T04-11271 | 0.000 ± 0.175 | T04-11273 | 0.417 ± 0.339 | ----- |
| S | 16 | 192-hr event | T04-11267 | 85.959 ± 0.693 | T04-11269 | 88.042 ± 7.077 | T04-11271 | 72.038 ± 0.226 | T04-11273 | 84.316 ± 2.287 | 85.14 |
| Cl | 17 | 192-hr event | T04-11267 | 0.671 ± 0.040 | T04-11269 | -0.504 ± 0.628 | T04-11271 | 0.977 ± 0.046 | T04-11273 | 0.977 ± 0.112 | ----- |
| K | 19 | 192-hr event | T04-11267 | 5.230 ± 0.006 | T04-11269 | 5.021 ± 0.407 | T04-11271 | 4.642 ± 0.043 | T04-11273 | 5.159 ± 0.159 | 5.09 |
| Ca | 20 | 192-hr event | T04-11267 | 1.516 ± 0.046 | T04-11269 | 1.610 ± 0.137 | T04-11271 | 1.523 ± 0.028 | T04-11273 | 2.039 ± 0.070 | 1.57 |
| Sc | 21 | 192-hr event | T04-11267 | 0.000 ± 0.028 | T04-11269 | -0.026 ± 0.023 | T04-11271 | 0.000 ± 0.043 | T04-11273 | -0.032 ± 0.023 | ----- |
| Ti | 22 | 192-hr event | T04-11267 | 0.174 ± 0.026 | T04-11269 | 0.177 ± 0.055 | T04-11271 | 0.000 ± 0.025 | T04-11273 | 0.060 ± 0.044 | ----- |
| V | 23 | 192-hr event | T04-11267 | 0.020 ± 0.009 | T04-11269 | 0.035 ± 0.019 | T04-11271 | 0.079 ± 0.014 | T04-11273 | 0.047 ± 0.018 | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Cr | 24 | 192-hr event | T04-11267 | 0.005 ± 0.040 | T04-11269 | -0.005 ± 0.009 | T04-11271 | 0.000 ± 0.011 | T04-11273 | 0.059 ± 0.015 | ----- |
| Mn | 25 | 192-hr event | T04-11267 | 0.088 ± 0.114 | T04-11269 | 0.105 ± 0.016 | T04-11271 | 0.103 ± 0.008 | T04-11273 | 0.092 ± 0.031 | ----- |
| Fe | 26 | 192-hr event | T04-11267 | 3.363 ± 0.180 | T04-11269 | 3.091 ± 0.249 | T04-11271 | 2.967 ± 0.024 | T04-11273 | 3.475 ± 0.118 | 3.23 |
| Co | 27 | 192-hr event | T04-11267 | 0.000 ± 0.019 | T04-11269 | -0.041 ± 0.019 | T04-11271 | 0.000 ± 0.009 | T04-11273 | -0.022 ± 0.028 | ----- |
| Ni | 28 | 192-hr event | T04-11267 | 0.011 ± 0.031 | T04-11269 | 0.030 ± 0.010 | T04-11271 | 0.032 ± 0.004 | T04-11273 | 0.483 ± 0.032 | ----- |
| Cu | 29 | 192-hr event | T04-11267 | 0.090 ± 0.024 | T04-11269 | 0.109 ± 0.013 | T04-11271 | 0.119 ± 0.006 | T04-11273 | 0.173 ± 0.020 | 0.11 |
| Zn | 30 | 192-hr event | T04-11267 | 0.802 ± 0.034 | T04-11269 | 0.671 ± 0.055 | T04-11271 | 0.611 ± 0.009 | T04-11273 | 0.782 ± 0.058 | 0.73 |
| Ga | 31 | 192-hr event | T04-11267 | 0.233 ± 0.098 | T04-11269 | -0.014 ± 0.054 | T04-11271 | 0.002 ± 0.004 | T04-11273 | -0.012 ± 0.029 | ----- |
| As | 33 | 192-hr event | T04-11267 | 0.054 ± 0.026 | T04-11269 | 0.082 ± 0.019 | T04-11271 | 0.129 ± 0.006 | T04-11273 | 0.065 ± 0.041 | ----- |
| Se | 34 | 192-hr event | T04-11267 | 0.000 ± 0.023 | T04-11269 | 0.111 ± 0.014 | T04-11271 | 0.116 ± 0.006 | T04-11273 | 0.208 ± 0.030 | ----- |
| Br | 35 | 192-hr event | T04-11267 | 0.323 ± 0.026 | T04-11269 | 0.331 ± 0.029 | T04-11271 | 0.313 ± 0.007 | T04-11273 | 0.345 ± 0.035 | 0.33 |
| Rb | 37 | 192-hr event | T04-11267 | 0.000 ± 0.024 | T04-11269 | 0.003 ± 0.011 | T04-11271 | 0.000 ± 0.006 | T04-11273 | -0.011 ± 0.024 | ----- |
| Sr | 38 | 192-hr event | T04-11267 | 0.029 ± 0.057 | T04-11269 | 0.007 ± 0.011 | T04-11271 | 0.042 ± 0.010 | T04-11273 | 0.143 ± 0.071 | ----- |
| Y | 39 | 192-hr event | T04-11267 | 0.020 ± 0.037 | T04-11269 | -0.012 ± 0.012 | T04-11271 | 0.024 ± 0.011 | T04-11273 | 0.147 ± 0.069 | ----- |
| Zr | 40 | 192-hr event | T04-11267 | 0.115 ± 0.078 | T04-11269 | -0.003 ± 0.014 | T04-11271 | 0.000 ± 0.011 | T04-11273 | 0.081 ± 0.054 | ----- |
| Nb | 41 | 192-hr event | T04-11267 | 0.000 ± 0.062 | T04-11269 | 0.014 ± 0.016 | T04-11271 | 0.000 ± 0.010 | T04-11273 | -0.006 ± 0.053 | ----- |
| Mo | 42 | 192-hr event | T04-11267 | 0.038 ± 0.082 | T04-11269 | 0.008 ± 0.019 | T04-11271 | 0.000 ± 0.010 | T04-11273 | 0.218 ± 0.068 | ----- |
| Ag | 47 | 192-hr event | T04-11267 | 0.000 ± 0.085 | T04-11269 | 0.020 ± 0.037 | T04-11271 | 0.077 ± 0.053 | T04-11273 | 0.001 ± 0.242 | ----- |
| Cd | 48 | 192-hr event | T04-11267 | 0.102 ± 0.094 | T04-11269 | -0.009 ± 0.038 | T04-11271 | 0.000 ± 0.045 | T04-11273 | -0.064 ± 0.138 | ----- |
| In | 49 | 192-hr event | T04-11267 | 0.000 ± 0.092 | T04-11269 | -0.028 ± 0.040 | T04-11271 | 0.063 ± 0.063 | T04-11273 | 0.286 ± 0.250 | ----- |
| Sn | 50 | 192-hr event | T04-11267 | 0.000 ± 0.108 | T04-11269 | -0.012 ± 0.044 | T04-11271 | 0.193 ± 0.087 | T04-11273 | -0.019 ± 0.128 | ----- |
| Sb | 51 | 192-hr event | T04-11267 | 0.079 ± 0.097 | T04-11269 | -0.011 ± 0.049 | T04-11271 | 0.000 ± 0.110 | T04-11273 | 0.151 ± 0.099 | ----- |
| Cs | 55 | 192-hr event | T04-11267 | 0.099 ± 0.225 | T04-11269 | -0.003 ± 0.085 | T04-11271 | 0.000 ± 0.045 | T04-11273 | 0.069 ± 0.055 | ----- |
| Ba | 56 | 192-hr event | T04-11267 | 0.000 ± 0.242 | T04-11269 | -0.063 ± 0.118 | T04-11271 | 0.000 ± 0.056 | T04-11273 | 0.432 ± 0.095 | ----- |
| La | 57 | 192-hr event | T04-11267 | 0.000 ± 0.495 | T04-11269 | -0.042 ± 0.148 | T04-11271 | 0.000 ± 0.038 | T04-11273 | 0.208 ± 0.068 | ----- |
| Ce | 58 | 192-hr event | T04-11267 | 0.000 ± 0.410 | T04-11269 | -0.128 ± 0.183 | T04-11271 | 0.000 ± 0.045 | T04-11273 | 0.056 ± 0.053 | ----- |
| Sm | 62 | 192-hr event | T04-11267 | 0.000 ± 0.689 | T04-11269 | -0.487 ± 0.653 | T04-11271 | 0.000 ± 0.021 | T04-11273 | not reported | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Eu | 63 | 192-hr event | T04-11267 | 0.226 ± 0.864 | T04-11269 | -0.921 ± 0.990 | T04-11271 | 0.000 ± 0.027 | T04-11273 | not reported | ----- |
| Tb | 65 | 192-hr event | T04-11267 | 0.000 ± 0.932 | T04-11269 | 1.155 ± 2.159 | T04-11271 | 0.079 ± 0.060 | T04-11273 | not reported | ----- |
| Hf | 72 | 192-hr event | T04-11267 | 0.000 ± 0.229 | T04-11269 | 0.146 ± 0.205 | T04-11271 | 0.028 ± 0.024 | T04-11273 | not reported | ----- |
| Ta | 73 | 192-hr event | T04-11267 | 0.000 ± 0.122 | T04-11269 | -0.091 ± 0.227 | T04-11271 | 0.000 ± 0.016 | T04-11273 | not reported | ----- |
| W | 74 | 192-hr event | T04-11267 | 0.041 ± 0.293 | T04-11269 | -0.026 ± 0.059 | T04-11271 | 0.000 ± 0.011 | T04-11273 | 0.080 ± 0.086 | ----- |
| Ir | 77 | 192-hr event | T04-11267 | 0.000 ± 0.105 | T04-11269 | 0.005 ± 0.034 | T04-11271 | 0.000 ± 0.008 | T04-11273 | not reported | ----- |
| Au | 79 | 192-hr event | T04-11267 | 0.000 ± 0.111 | T04-11269 | -0.048 ± 0.027 | T04-11271 | 0.035 ± 0.010 | T04-11273 | -0.054 ± 0.049 | ----- |
| Hg | 80 | 192-hr event | T04-11267 | 0.068 ± 0.042 | T04-11269 | -0.002 ± 0.022 | T04-11271 | 0.224 ± 0.012 | T04-11273 | -0.014 ± 0.053 | ----- |
| Pb | 82 | 192-hr event | T04-11267 | 0.242 ± 0.081 | T04-11269 | 0.207 ± 0.035 | T04-11271 | 0.243 ± 0.020 | T04-11273 | 0.292 ± 0.078 | ----- |
| Na | 11 | 192-hr event | T04-11268 | 6.846 ± 0.906 | T04-11270 | 7.969 ± 2.427 | T04-11272 | 2.983 ± 0.147 | T04-11274 | 5.115 ± 0.844 | 5.98 |
| Mg | 12 | 192-hr event | T04-11268 | 1.209 ± 0.825 | T04-11270 | 0.268 ± 0.415 | T04-11272 | 0.381 ± 0.050 | T04-11274 | 0.355 ± 0.418 | ----- |
| Al | 13 | 192-hr event | T04-11268 | 2.054 ± 0.551 | T04-11270 | 0.490 ± 0.104 | T04-11272 | 1.182 ± 0.043 | T04-11274 | 1.800 ± 0.515 | 1.49 |
| Si | 14 | 192-hr event | T04-11268 | 9.286 ± 0.353 | T04-11270 | 10.011 ± 0.876 | T04-11272 | 9.137 ± 0.041 | T04-11274 | 12.016 ± 0.743 | 9.65 |
| P | 15 | 192-hr event | T04-11268 | 3.437 ± 0.061 | T04-11270 | -0.587 ± 0.189 | T04-11272 | 0.000 ± 0.178 | T04-11274 | 1.023 ± 0.334 | ----- |
| S | 16 | 192-hr event | T04-11268 | 85.518 ± 0.689 | T04-11270 | 87.139 ± 7.004 | T04-11272 | 74.987 ± 0.237 | T04-11274 | 77.701 ± 2.121 | 81.61 |
| Cl | 17 | 192-hr event | T04-11268 | 0.423 ± 0.038 | T04-11270 | -0.368 ± 0.621 | T04-11272 | 0.932 ± 0.046 | T04-11274 | 0.599 ± 0.104 | ----- |
| K | 19 | 192-hr event | T04-11268 | 5.162 ± 0.006 | T04-11270 | 5.062 ± 0.410 | T04-11272 | 4.833 ± 0.043 | T04-11274 | 4.584 ± 0.149 | 4.95 |
| Ca | 20 | 192-hr event | T04-11268 | 1.584 ± 0.046 | T04-11270 | 1.723 ± 0.145 | T04-11272 | 1.533 ± 0.028 | T04-11274 | 1.728 ± 0.064 | 1.65 |
| Sc | 21 | 192-hr event | T04-11268 | 0.000 ± 0.028 | T04-11270 | -0.030 ± 0.024 | T04-11272 | 0.000 ± 0.043 | T04-11274 | -0.009 ± 0.023 | ----- |
| Ti | 22 | 192-hr event | T04-11268 | 0.106 ± 0.026 | T04-11270 | 0.191 ± 0.055 | T04-11272 | 0.000 ± 0.025 | T04-11274 | 0.121 ± 0.045 | ----- |
| V | 23 | 192-hr event | T04-11268 | 0.000 ± 0.009 | T04-11270 | 0.039 ± 0.019 | T04-11272 | 0.043 ± 0.015 | T04-11274 | 0.049 ± 0.017 | ----- |
| Cr | 24 | 192-hr event | T04-11268 | 0.027 ± 0.040 | T04-11270 | 0.000 ± 0.009 | T04-11272 | 0.000 ± 0.010 | T04-11274 | 0.052 ± 0.015 | ----- |
| Mn | 25 | 192-hr event | T04-11268 | 0.066 ± 0.114 | T04-11270 | 0.129 ± 0.018 | T04-11272 | 0.086 ± 0.008 | T04-11274 | 0.086 ± 0.030 | ----- |
| Fe | 26 | 192-hr event | T04-11268 | 3.295 ± 0.180 | T04-11270 | 3.046 ± 0.246 | T04-11272 | 3.095 ± 0.024 | T04-11274 | 2.926 ± 0.103 | 3.07 |
| Co | 27 | 192-hr event | T04-11268 | 0.000 ± 0.019 | T04-11270 | -0.031 ± 0.019 | T04-11272 | 0.000 ± 0.009 | T04-11274 | -0.014 ± 0.026 | ----- |
| Ni | 28 | 192-hr event | T04-11268 | 0.113 ± 0.031 | T04-11270 | 0.038 ± 0.010 | T04-11272 | 0.063 ± 0.005 | T04-11274 | 0.027 ± 0.018 | ----- |
| Cu | 29 | 192-hr event | T04-11268 | 0.090 ± 0.024 | T04-11270 | 0.117 ± 0.014 | T04-11272 | 0.118 ± 0.006 | T04-11274 | 0.091 ± 0.018 | 0.10 |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Zn | 30 | 192-hr event | T04-11268 | 0.757 ± 0.034 | T04-11270 | 0.695 ± 0.057 | T04-11272 | 0.618 ± 0.009 | T04-11274 | 0.627 ± 0.052 | 0.66 |
| Ga | 31 | 192-hr event | T04-11268 | 0.000 ± 0.097 | T04-11270 | -0.011 ± 0.054 | T04-11272 | 0.012 ± 0.004 | T04-11274 | -0.035 ± 0.027 | ----- |
| As | 33 | 192-hr event | T04-11268 | 0.043 ± 0.026 | T04-11270 | 0.078 ± 0.019 | T04-11272 | 0.110 ± 0.006 | T04-11274 | 0.130 ± 0.039 | ----- |
| Se | 34 | 192-hr event | T04-11268 | 0.023 ± 0.023 | T04-11270 | 0.103 ± 0.013 | T04-11272 | 0.122 ± 0.006 | T04-11274 | 0.107 ± 0.025 | ----- |
| Br | 35 | 192-hr event | T04-11268 | 0.380 ± 0.027 | T04-11270 | 0.322 ± 0.028 | T04-11272 | 0.325 ± 0.008 | T04-11274 | 0.330 ± 0.035 | 0.33 |
| Rb | 37 | 192-hr event | T04-11268 | 0.000 ± 0.024 | T04-11270 | 0.009 ± 0.010 | T04-11272 | 0.000 ± 0.006 | T04-11274 | 0.070 ± 0.025 | ----- |
| Sr | 38 | 192-hr event | T04-11268 | 0.000 ± 0.057 | T04-11270 | 0.016 ± 0.011 | T04-11272 | 0.025 ± 0.010 | T04-11274 | 0.000 ± 0.063 | ----- |
| Y | 39 | 192-hr event | T04-11268 | 0.032 ± 0.037 | T04-11270 | -0.018 ± 0.012 | T04-11272 | 0.001 ± 0.011 | T04-11274 | 0.116 ± 0.065 | ----- |
| Zr | 40 | 192-hr event | T04-11268 | 0.115 ± 0.078 | T04-11270 | -0.006 ± 0.014 | T04-11272 | 0.052 ± 0.016 | T04-11274 | 0.154 ± 0.061 | ----- |
| Nb | 41 | 192-hr event | T04-11268 | 0.032 ± 0.062 | T04-11270 | 0.019 ± 0.016 | T04-11272 | 0.000 ± 0.010 | T04-11274 | 0.088 ± 0.062 | ----- |
| Mo | 42 | 192-hr event | T04-11268 | 0.000 ± 0.081 | T04-11270 | -0.002 ± 0.019 | T04-11272 | 0.000 ± 0.010 | T04-11274 | 0.090 ± 0.062 | ----- |
| Ag | 47 | 192-hr event | T04-11268 | 0.000 ± 0.085 | T04-11270 | 0.041 ± 0.036 | T04-11272 | 0.000 ± 0.040 | T04-11274 | 0.145 ± 0.253 | ----- |
| Cd | 48 | 192-hr event | T04-11268 | 0.023 ± 0.094 | T04-11270 | 0.062 ± 0.038 | T04-11272 | 0.000 ± 0.047 | T04-11274 | 0.108 ± 0.145 | ----- |
| In | 49 | 192-hr event | T04-11268 | 0.000 ± 0.092 | T04-11270 | 0.005 ± 0.040 | T04-11272 | 0.000 ± 0.049 | T04-11274 | 0.167 ± 0.242 | ----- |
| Sn | 50 | 192-hr event | T04-11268 | 0.000 ± 0.108 | T04-11270 | 0.033 ± 0.043 | T04-11272 | 0.000 ± 0.062 | T04-11274 | 0.169 ± 0.124 | ----- |
| Sb | 51 | 192-hr event | T04-11268 | 0.045 ± 0.097 | T04-11270 | 0.078 ± 0.049 | T04-11272 | 0.045 ± 0.147 | T04-11274 | 0.039 ± 0.093 | ----- |
| Cs | 55 | 192-hr event | T04-11268 | 0.000 ± 0.224 | T04-11270 | -0.014 ± 0.084 | T04-11272 | 0.000 ± 0.045 | T04-11274 | -0.026 ± 0.053 | ----- |
| Ba | 56 | 192-hr event | T04-11268 | 1.071 ± 0.249 | T04-11270 | 0.040 ± 0.116 | T04-11272 | 0.000 ± 0.054 | T04-11274 | 0.329 ± 0.092 | ----- |
| La | 57 | 192-hr event | T04-11268 | 0.000 ± 0.496 | T04-11270 | 0.021 ± 0.147 | T04-11272 | 0.000 ± 0.038 | T04-11274 | 0.259 ± 0.067 | ----- |
| Ce | 58 | 192-hr event | T04-11268 | 0.147 ± 0.412 | T04-11270 | -0.063 ± 0.181 | T04-11272 | 0.000 ± 0.044 | T04-11274 | 0.010 ± 0.052 | ----- |
| Sm | 62 | 192-hr event | T04-11268 | 0.312 ± 0.697 | T04-11270 | 0.208 ± 0.643 | T04-11272 | 0.000 ± 0.021 | T04-11274 | not reported | ----- |
| Eu | 63 | 192-hr event | T04-11268 | 0.068 ± 0.862 | T04-11270 | 0.482 ± 0.976 | T04-11272 | 0.000 ± 0.026 | T04-11274 | not reported | ----- |
| Tb | 65 | 192-hr event | T04-11268 | 1.962 ± 0.993 | T04-11270 | 0.668 ± 2.131 | T04-11272 | 0.064 ± 0.061 | T04-11274 | not reported | ----- |
| Hf | 72 | 192-hr event | T04-11268 | 0.000 ± 0.229 | T04-11270 | 0.019 ± 0.204 | T04-11272 | 0.053 ± 0.024 | T04-11274 | not reported | ----- |
| Ta | 73 | 192-hr event | T04-11268 | 0.115 ± 0.123 | T04-11270 | -0.041 ± 0.227 | T04-11272 | 0.000 ± 0.017 | T04-11274 | not reported | ----- |
| W | 74 | 192-hr event | T04-11268 | 0.000 ± 0.290 | T04-11270 | 0.017 ± 0.059 | T04-11272 | 0.000 ± 0.011 | T04-11274 | 0.047 ± 0.075 | ----- |
| Ir | 77 | 192-hr event | T04-11268 | 0.000 ± 0.104 | T04-11270 | 0.000 ± 0.035 | T04-11272 | 0.000 ± 0.008 | T04-11274 | not reported | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Au | 79 | 192-hr event | T04-11268 | 0.192 ± 0.111 | T04-11270 | -0.031 ± 0.027 | T04-11272 | 0.034 ± 0.010 | T04-11274 | 0.074 ± 0.047 | ----- |
| Hg | 80 | 192-hr event | T04-11268 | 0.000 ± 0.042 | T04-11270 | 0.017 ± 0.022 | T04-11272 | 0.217 ± 0.011 | T04-11274 | 0.122 ± 0.053 | ----- |
| Pb | 82 | 192-hr event | T04-11268 | 0.174 ± 0.081 | T04-11270 | 0.215 ± 0.035 | T04-11272 | 0.205 ± 0.020 | T04-11274 | 0.229 ± 0.071 | ----- |
| Na | 11 | blank filter | T04-11277 | 0.000 ± 0.753 | T04-11279 | -0.258 ± 2.258 | T04-11281 | 0.000 ± 0.131 | T04-11283 | 0.174 ± 0.554 | ----- |
| Mg | 12 | blank filter | T04-11277 | 0.181 ± 0.812 | T04-11279 | -0.030 ± 0.399 | T04-11281 | 0.000 ± 0.038 | T04-11283 | 0.949 ± 0.370 | ----- |
| Al | 13 | blank filter | T04-11277 | 0.000 ± 0.539 | T04-11279 | -0.034 ± 0.080 | T04-11281 | 0.000 ± 0.049 | T04-11283 | -0.075 ± 0.239 | ----- |
| Si | 14 | blank filter | T04-11277 | 0.111 ± 0.249 | T04-11279 | -0.012 ± 0.048 | T04-11281 | 0.000 ± 0.033 | T04-11283 | 0.980 ± 0.280 | ----- |
| P | 15 | blank filter | T04-11277 | 0.000 ± 0.050 | T04-11279 | -0.023 ± 0.028 | T04-11281 | 0.000 ± 0.058 | T04-11283 | -0.212 ± 0.098 | ----- |
| S | 16 | blank filter | T04-11277 | 0.079 ± 0.014 | T04-11279 | -0.040 ± 0.047 | T04-11281 | 0.000 ± 0.036 | T04-11283 | 0.116 ± 0.077 | ----- |
| Cl | 17 | blank filter | T04-11277 | 0.005 ± 0.037 | T04-11279 | -0.001 ± 0.028 | T04-11281 | 0.000 ± 0.030 | T04-11283 | 0.025 ± 0.053 | ----- |
| K | 19 | blank filter | T04-11277 | 0.156 ± 0.001 | T04-11279 | 0.004 ± 0.017 | T04-11281 | 0.000 ± 0.026 | T04-11283 | -0.037 ± 0.040 | ----- |
| Ca | 20 | blank filter | T04-11277 | 0.002 ± 0.043 | T04-11279 | -0.029 ± 0.014 | T04-11281 | 0.000 ± 0.024 | T04-11283 | -0.027 ± 0.023 | ----- |
| Sc | 21 | blank filter | T04-11277 | 0.005 ± 0.028 | T04-11279 | -0.003 ± 0.014 | T04-11281 | 0.000 ± 0.038 | T04-11283 | 0.032 ± 0.017 | ----- |
| Ti | 22 | blank filter | T04-11277 | 0.000 ± 0.026 | T04-11279 | -0.002 ± 0.048 | T04-11281 | 0.008 ± 0.016 | T04-11283 | -0.056 ± 0.036 | ----- |
| V | 23 | blank filter | T04-11277 | 0.000 ± 0.009 | T04-11279 | -0.007 ± 0.017 | T04-11281 | 0.000 ± 0.012 | T04-11283 | -0.015 ± 0.014 | ----- |
| Cr | 24 | blank filter | T04-11277 | 0.000 ± 0.040 | T04-11279 | -0.005 ± 0.008 | T04-11281 | 0.000 ± 0.008 | T04-11283 | 0.040 ± 0.014 | ----- |
| Mn | 25 | blank filter | T04-11277 | 0.000 ± 0.113 | T04-11279 | 0.008 ± 0.013 | T04-11281 | 0.000 ± 0.006 | T04-11283 | 0.106 ± 0.027 | ----- |
| Fe | 26 | blank filter | T04-11277 | 0.041 ± 0.170 | T04-11279 | -0.005 ± 0.011 | T04-11281 | 0.000 ± 0.005 | T04-11283 | -0.008 ± 0.020 | ----- |
| Co | 27 | blank filter | T04-11277 | 0.009 ± 0.019 | T04-11279 | 0.015 ± 0.008 | T04-11281 | 0.000 ± 0.004 | T04-11283 | -0.075 ± 0.017 | ----- |
| Ni | 28 | blank filter | T04-11277 | 0.000 ± 0.031 | T04-11279 | -0.001 ± 0.008 | T04-11281 | 0.006 ± 0.003 | T04-11283 | -0.025 ± 0.015 | ----- |
| Cu | 29 | blank filter | T04-11277 | 0.000 ± 0.024 | T04-11279 | 0.008 ± 0.008 | T04-11281 | 0.000 ± 0.005 | T04-11283 | 0.011 ± 0.015 | ----- |
| Zn | 30 | blank filter | T04-11277 | 0.011 ± 0.032 | T04-11279 | -0.004 ± 0.007 | T04-11281 | 0.000 ± 0.006 | T04-11283 | -0.110 ± 0.019 | ----- |
| Ga | 31 | blank filter | T04-11277 | 0.000 ± 0.098 | T04-11279 | -0.006 ± 0.054 | T04-11281 | 0.000 ± 0.002 | T04-11283 | -0.002 ± 0.025 | ----- |
| As | 33 | blank filter | T04-11277 | 0.000 ± 0.026 | T04-11279 | 0.000 ± 0.012 | T04-11281 | 0.001 ± 0.003 | T04-11283 | -0.006 ± 0.030 | ----- |
| Se | 34 | blank filter | T04-11277 | 0.000 ± 0.023 | T04-11279 | -0.005 ± 0.009 | T04-11281 | 0.000 ± 0.003 | T04-11283 | -0.019 ± 0.018 | ----- |
| Br | 35 | blank filter | T04-11277 | 0.007 ± 0.026 | T04-11279 | -0.001 ± 0.009 | T04-11281 | 0.000 ± 0.004 | T04-11283 | -0.028 ± 0.020 | ----- |
| Rb | 37 | blank filter | T04-11277 | 0.000 ± 0.024 | T04-11279 | -0.002 ± 0.009 | T04-11281 | 0.000 ± 0.005 | T04-11283 | 0.032 ± 0.021 | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Sr | 38 | blank filter | T04-11277 | 0.000 ± 0.057 | T04-11279 | 0.000 ± 0.010 | T04-11281 | 0.000 ± 0.006 | T04-11283 | 0.080 ± 0.065 | ----- |
| Y | 39 | blank filter | T04-11277 | 0.009 ± 0.037 | T04-11279 | -0.011 ± 0.011 | T04-11281 | 0.000 ± 0.007 | T04-11283 | -0.025 ± 0.054 | ----- |
| Zr | 40 | blank filter | T04-11277 | 0.000 ± 0.077 | T04-11279 | 0.004 ± 0.013 | T04-11281 | 0.000 ± 0.010 | T04-11283 | 0.023 ± 0.044 | ----- |
| Nb | 41 | blank filter | T04-11277 | 0.043 ± 0.062 | T04-11279 | 0.004 ± 0.015 | T04-11281 | 0.000 ± 0.009 | T04-11283 | -0.028 ± 0.052 | ----- |
| Mo | 42 | blank filter | T04-11277 | 0.000 ± 0.081 | T04-11279 | -0.013 ± 0.017 | T04-11281 | 0.000 ± 0.009 | T04-11283 | -0.023 ± 0.048 | ----- |
| Ag | 47 | blank filter | T04-11277 | 0.000 ± 0.085 | T04-11279 | 0.005 ± 0.034 | T04-11281 | 0.000 ± 0.039 | T04-11283 | -0.462 ± 0.155 | ----- |
| Cd | 48 | blank filter | T04-11277 | 0.000 ± 0.094 | T04-11279 | 0.004 ± 0.036 | T04-11281 | 0.000 ± 0.044 | T04-11283 | 0.169 ± 0.085 | ----- |
| In | 49 | blank filter | T04-11277 | 0.000 ± 0.092 | T04-11279 | -0.001 ± 0.037 | T04-11281 | 0.000 ± 0.045 | T04-11283 | 0.143 ± 0.097 | ----- |
| Sn | 50 | blank filter | T04-11277 | 0.057 ± 0.108 | T04-11279 | -0.019 ± 0.040 | T04-11281 | 0.000 ± 0.054 | T04-11283 | -0.018 ± 0.050 | ----- |
| Sb | 51 | blank filter | T04-11277 | 0.124 ± 0.098 | T04-11279 | -0.002 ± 0.045 | T04-11281 | 0.000 ± 0.103 | T04-11283 | 0.065 ± 0.049 | ----- |
| Cs | 55 | blank filter | T04-11277 | 0.000 ± 0.224 | T04-11279 | -0.041 ± 0.077 | T04-11281 | 0.000 ± 0.035 | T04-11283 | 0.012 ± 0.047 | ----- |
| Ba | 56 | blank filter | T04-11277 | 0.000 ± 0.240 | T04-11279 | -0.053 ± 0.106 | T04-11281 | 0.000 ± 0.035 | T04-11283 | 0.117 ± 0.080 | ----- |
| La | 57 | blank filter | T04-11277 | 0.000 ± 0.495 | T04-11279 | -0.114 ± 0.134 | T04-11281 | 0.000 ± 0.028 | T04-11283 | 0.107 ± 0.059 | ----- |
| Ce | 58 | blank filter | T04-11277 | 0.000 ± 0.411 | T04-11279 | -0.187 ± 0.166 | T04-11281 | 0.000 ± 0.033 | T04-11283 | 0.038 ± 0.045 | ----- |
| Sm | 62 | blank filter | T04-11277 | 0.000 ± 0.684 | T04-11279 | -0.093 ± 0.591 | T04-11281 | 0.000 ± 0.017 | T04-11283 | not reported | ----- |
| Eu | 63 | blank filter | T04-11277 | 1.401 ± 0.880 | T04-11279 | -0.760 ± 0.895 | T04-11281 | 0.000 ± 0.015 | T04-11283 | not reported | ----- |
| Tb | 65 | blank filter | T04-11277 | 0.007 ± 0.942 | T04-11279 | -1.449 ± 1.938 | T04-11281 | 0.000 ± 0.014 | T04-11283 | not reported | ----- |
| Hf | 72 | blank filter | T04-11277 | 0.000 ± 0.229 | T04-11279 | 0.010 ± 0.204 | T04-11281 | 0.000 ± 0.020 | T04-11283 | not reported | ----- |
| Ta | 73 | blank filter | T04-11277 | 0.115 ± 0.123 | T04-11279 | -0.042 ± 0.226 | T04-11281 | 0.000 ± 0.011 | T04-11283 | not reported | ----- |
| W | 74 | blank filter | T04-11277 | 0.000 ± 0.292 | T04-11279 | -0.006 ± 0.052 | T04-11281 | 0.011 ± 0.009 | T04-11283 | 0.071 ± 0.070 | ----- |
| Ir | 77 | blank filter | T04-11277 | 0.000 ± 0.104 | T04-11279 | 0.010 ± 0.033 | T04-11281 | 0.000 ± 0.007 | T04-11283 | not reported | ----- |
| Au | 79 | blank filter | T04-11277 | 0.068 ± 0.111 | T04-11279 | -0.004 ± 0.025 | T04-11281 | 0.000 ± 0.005 | T04-11283 | -0.033 ± 0.040 | ----- |
| Hg | 80 | blank filter | T04-11277 | 0.023 ± 0.042 | T04-11279 | -0.009 ± 0.022 | T04-11281 | 0.000 ± 0.009 | T04-11283 | -0.107 ± 0.040 | ----- |
| Pb | 82 | blank filter | T04-11277 | 0.000 ± 0.081 | T04-11279 | 0.004 ± 0.028 | T04-11281 | 0.012 ± 0.011 | T04-11283 | 0.041 ± 0.061 | ----- |
| Na | 11 | blank filter | T04-11278 | 0.000 ± 0.759 | T04-11280 | -0.247 ± 2.258 | T04-11282 | 0.000 ± 0.129 | T04-11284 | 0.846 ± 0.546 | ----- |
| Mg | 12 | blank filter | T04-11278 | 0.418 ± 0.815 | T04-11280 | -0.170 ± 0.400 | T04-11282 | 0.000 ± 0.041 | T04-11284 | -0.145 ± 0.360 | ----- |
| Al | 13 | blank filter | T04-11278 | 0.179 ± 0.542 | T04-11280 | -0.090 ± 0.081 | T04-11282 | 0.000 ± 0.051 | T04-11284 | 0.288 ± 0.243 | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Si | 14 | blank filter | T04-11278 | 0.145 ± 0.250 | T04-11280 | -0.019 ± 0.048 | T04-11282 | 0.000 ± 0.034 | T04-11284 | 1.037 ± 0.274 | ----- |
| P | 15 | blank filter | T04-11278 | 0.000 ± 0.050 | T04-11280 | -0.011 ± 0.027 | T04-11282 | 0.000 ± 0.057 | T04-11284 | -0.065 ± 0.104 | ----- |
| S | 16 | blank filter | T04-11278 | 0.000 ± 0.012 | T04-11280 | 0.042 ± 0.045 | T04-11282 | 0.000 ± 0.033 | T04-11284 | -0.021 ± 0.077 | ----- |
| Cl | 17 | blank filter | T04-11278 | 0.000 ± 0.037 | T04-11280 | 0.039 ± 0.024 | T04-11282 | 0.000 ± 0.028 | T04-11284 | -0.092 ± 0.046 | ----- |
| K | 19 | blank filter | T04-11278 | 0.066 ± 0.001 | T04-11280 | -0.002 ± 0.016 | T04-11282 | 0.023 ± 0.018 | T04-11284 | 0.062 ± 0.042 | ----- |
| Ca | 20 | blank filter | T04-11278 | 0.000 ± 0.043 | T04-11280 | 0.004 ± 0.013 | T04-11282 | 0.000 ± 0.024 | T04-11284 | -0.059 ± 0.023 | ----- |
| Sc | 21 | blank filter | T04-11278 | 0.005 ± 0.028 | T04-11280 | 0.017 ± 0.014 | T04-11282 | 0.000 ± 0.036 | T04-11284 | 0.024 ± 0.017 | ----- |
| Ti | 22 | blank filter | T04-11278 | 0.000 ± 0.026 | T04-11280 | 0.011 ± 0.046 | T04-11282 | 0.021 ± 0.015 | T04-11284 | 0.037 ± 0.037 | ----- |
| V | 23 | blank filter | T04-11278 | 0.000 ± 0.009 | T04-11280 | 0.001 ± 0.016 | T04-11282 | 0.000 ± 0.012 | T04-11284 | 0.013 ± 0.014 | ----- |
| Cr | 24 | blank filter | T04-11278 | 0.000 ± 0.040 | T04-11280 | -0.001 ± 0.008 | T04-11282 | 0.000 ± 0.007 | T04-11284 | 0.044 ± 0.014 | ----- |
| Mn | 25 | blank filter | T04-11278 | 0.043 ± 0.114 | T04-11280 | -0.007 ± 0.012 | T04-11282 | 0.000 ± 0.006 | T04-11284 | 0.048 ± 0.025 | ----- |
| Fe | 26 | blank filter | T04-11278 | 0.063 ± 0.171 | T04-11280 | -0.006 ± 0.011 | T04-11282 | 0.000 ± 0.005 | T04-11284 | -0.035 ± 0.018 | ----- |
| Co | 27 | blank filter | T04-11278 | 0.020 ± 0.019 | T04-11280 | 0.006 ± 0.008 | T04-11282 | 0.000 ± 0.004 | T04-11284 | -0.031 ± 0.018 | ----- |
| Ni | 28 | blank filter | T04-11278 | 0.000 ± 0.031 | T04-11280 | -0.006 ± 0.008 | T04-11282 | 0.000 ± 0.004 | T04-11284 | 0.009 ± 0.016 | ----- |
| Cu | 29 | blank filter | T04-11278 | 0.000 ± 0.024 | T04-11280 | -0.014 ± 0.008 | T04-11282 | 0.008 ± 0.004 | T04-11284 | 0.013 ± 0.015 | ----- |
| Zn | 30 | blank filter | T04-11278 | 0.000 ± 0.032 | T04-11280 | -0.002 ± 0.007 | T04-11282 | 0.000 ± 0.006 | T04-11284 | -0.029 ± 0.023 | ----- |
| Ga | 31 | blank filter | T04-11278 | 0.131 ± 0.098 | T04-11280 | -0.008 ± 0.054 | T04-11282 | 0.001 ± 0.003 | T04-11284 | 0.013 ± 0.024 | ----- |
| As | 33 | blank filter | T04-11278 | 0.000 ± 0.026 | T04-11280 | 0.002 ± 0.012 | T04-11282 | 0.000 ± 0.002 | T04-11284 | -0.005 ± 0.028 | ----- |
| Se | 34 | blank filter | T04-11278 | 0.000 ± 0.023 | T04-11280 | -0.004 ± 0.009 | T04-11282 | 0.000 ± 0.003 | T04-11284 | 0.022 ± 0.019 | ----- |
| Br | 35 | blank filter | T04-11278 | 0.000 ± 0.026 | T04-11280 | -0.001 ± 0.009 | T04-11282 | 0.000 ± 0.003 | T04-11284 | 0.001 ± 0.019 | ----- |
| Rb | 37 | blank filter | T04-11278 | 0.020 ± 0.024 | T04-11280 | -0.002 ± 0.009 | T04-11282 | 0.003 ± 0.007 | T04-11284 | 0.024 ± 0.020 | ----- |
| Sr | 38 | blank filter | T04-11278 | 0.041 ± 0.057 | T04-11280 | 0.001 ± 0.010 | T04-11282 | 0.000 ± 0.006 | T04-11284 | 0.028 ± 0.054 | ----- |
| Y | 39 | blank filter | T04-11278 | 0.000 ± 0.037 | T04-11280 | -0.003 ± 0.011 | T04-11282 | 0.000 ± 0.006 | T04-11284 | 0.018 ± 0.054 | ----- |
| Zr | 40 | blank filter | T04-11278 | 0.000 ± 0.077 | T04-11280 | -0.008 ± 0.012 | T04-11282 | 0.000 ± 0.010 | T04-11284 | 0.113 ± 0.050 | ----- |
| Nb | 41 | blank filter | T04-11278 | 0.000 ± 0.062 | T04-11280 | -0.011 ± 0.014 | T04-11282 | 0.000 ± 0.008 | T04-11284 | 0.158 ± 0.062 | ----- |
| Mo | 42 | blank filter | T04-11278 | 0.016 ± 0.081 | T04-11280 | 0.017 ± 0.016 | T04-11282 | 0.000 ± 0.009 | T04-11284 | 0.188 ± 0.061 | ----- |
| Ag | 47 | blank filter | T04-11278 | 0.000 ± 0.085 | T04-11280 | 0.001 ± 0.034 | T04-11282 | 0.000 ± 0.034 | T04-11284 | -0.046 ± 0.175 | ----- |

Table 12. XRF PE Results

| Element | Z | Sample Description | DRI Sample ID | DRI (µg/filter) | ODEQ Sample ID | ODEQ (µg/filter) | RTI Sample ID | RTI (µg/filter) | EPA-NERL Sample ID | EPA-NERL (µg/filter) | Median* (µg/filter) |
|---------|----|--------------------|---------------|-----------------|----------------|------------------|---------------|-----------------|--------------------|----------------------|---------------------|
| Cd | 48 | blank filter | T04-11278 | 0.068 ± 0.094 | T04-11280 | -0.005 ± 0.035 | T04-11282 | 0.000 ± 0.046 | T04-11284 | 0.065 ± 0.096 | ----- |
| In | 49 | blank filter | T04-11278 | 0.000 ± 0.092 | T04-11280 | -0.011 ± 0.037 | T04-11282 | 0.000 ± 0.045 | T04-11284 | 0.011 ± 0.099 | ----- |
| Sn | 50 | blank filter | T04-11278 | 0.057 ± 0.108 | T04-11280 | 0.013 ± 0.039 | T04-11282 | 0.000 ± 0.056 | T04-11284 | 0.138 ± 0.054 | ----- |
| Sb | 51 | blank filter | T04-11278 | 0.034 ± 0.097 | T04-11280 | -0.012 ± 0.044 | T04-11282 | 0.000 ± 0.097 | T04-11284 | 0.113 ± 0.048 | ----- |
| Cs | 55 | blank filter | T04-11278 | 0.077 ± 0.225 | T04-11280 | -0.021 ± 0.074 | T04-11282 | 0.000 ± 0.032 | T04-11284 | 0.039 ± 0.046 | ----- |
| Ba | 56 | blank filter | T04-11278 | 0.000 ± 0.241 | T04-11280 | -0.090 ± 0.101 | T04-11282 | 0.000 ± 0.031 | T04-11284 | 0.067 ± 0.078 | ----- |
| La | 57 | blank filter | T04-11278 | 0.000 ± 0.495 | T04-11280 | -0.054 ± 0.129 | T04-11282 | 0.000 ± 0.025 | T04-11284 | 0.143 ± 0.058 | ----- |
| Ce | 58 | blank filter | T04-11278 | 0.068 ± 0.412 | T04-11280 | -0.107 ± 0.159 | T04-11282 | 0.000 ± 0.031 | T04-11284 | 0.102 ± 0.045 | ----- |
| Sm | 62 | blank filter | T04-11278 | 0.000 ± 0.683 | T04-11280 | -0.578 ± 0.568 | T04-11282 | 0.000 ± 0.016 | T04-11284 | not reported | ----- |
| Eu | 63 | blank filter | T04-11278 | 0.622 ± 0.869 | T04-11280 | -1.191 ± 0.861 | T04-11282 | 0.006 ± 0.012 | T04-11284 | not reported | ----- |
| Tb | 65 | blank filter | T04-11278 | 0.000 ± 0.940 | T04-11280 | -0.333 ± 1.864 | T04-11282 | 0.000 ± 0.014 | T04-11284 | not reported | ----- |
| Hf | 72 | blank filter | T04-11278 | 0.000 ± 0.229 | T04-11280 | -0.025 ± 0.204 | T04-11282 | 0.008 ± 0.016 | T04-11284 | not reported | ----- |
| Ta | 73 | blank filter | T04-11278 | 0.000 ± 0.122 | T04-11280 | -0.020 ± 0.227 | T04-11282 | 0.012 ± 0.015 | T04-11284 | not reported | ----- |
| W | 74 | blank filter | T04-11278 | 0.000 ± 0.292 | T04-11280 | 0.001 ± 0.053 | T04-11282 | 0.000 ± 0.007 | T04-11284 | -0.082 ± 0.066 | ----- |
| Ir | 77 | blank filter | T04-11278 | 0.047 ± 0.105 | T04-11280 | 0.007 ± 0.034 | T04-11282 | 0.000 ± 0.006 | T04-11284 | not reported | ----- |
| Au | 79 | blank filter | T04-11278 | 0.000 ± 0.111 | T04-11280 | 0.001 ± 0.025 | T04-11282 | 0.008 ± 0.006 | T04-11284 | 0.192 ± 0.045 | ----- |
| Hg | 80 | blank filter | T04-11278 | 0.000 ± 0.042 | T04-11280 | 0.001 ± 0.022 | T04-11282 | 0.000 ± 0.008 | T04-11284 | -0.031 ± 0.040 | ----- |
| Pb | 82 | blank filter | T04-11278 | 0.050 ± 0.081 | T04-11280 | -0.008 ± 0.028 | T04-11282 | 0.000 ± 0.007 | T04-11284 | 0.059 ± 0.057 | ----- |

* Median was calculated only when the result from all reporting labs was greater than three times the uncertainty.

Table 13. XRF Analysis at the DRI Laboratory

| Instrument: PanAnalytical Epsilon 5 Software: E5 Version 1.0B | | | | | | | | | | |
|--|---|--------|---------|----------------------|---|--------|----------|----------------|------------------|--------------------------------|
| | Instrument Conditions for Routine Sample Analysis | | | | | | | | | |
| Parameter | #1 | #2 | #3 | #4 | #5 | #6 | #7 | #8 | #9 | #10 |
| X-ray tube parameters: | | | | | | | | | | |
| Tube voltage (kV) | 25 | 40 | 40 | 75 | 100 | 100 | 100 | 100 | 100 | 100 |
| Tube current (mA) | 24 | 15 | 15 | 8 | 6 | 6 | 6 | 6 | 6 | 6 |
| Tube anode material | Gd | Gd | Gd | Gd | Gd | Gd | Gd | Gd | Gd | Gd |
| Direct excitation of sample: | | | | | | | | | | |
| Filter Material | | | | | | | | | | |
| Filter thickness (mm) | | | | | | | | | | |
| Secondary excitation of sample: | | | | | | | | | | |
| Secondary Fluorescor | CaF ₂ | Ti | Fe | Ge | Zr | Mo | Ag | CsI | BaF ₂ | Al ₂ O ₃ |
| Filter material | | | | | | | | | | |
| Filter thickness (mm) | | | | | | | | | | |
| Acquisition time (seconds) | 200 | 200 | 200 | 200 | 100 | 100 | 100 | 100 | 100 | 100 |
| Energy range acquired (keV) | 0-20 | 0-20 | 0-20 | 0-20 | 0-20 | 0-20 | 0-40 | 0-80 | 0-80 | 0-80 |
| Number of [MCA] channels | 2048 | 2048 | 2048 | 2048 | 2048 | 2048 | 4096 | 8192 | 8192 | 8192 |
| Sample rotation (yes/no) | yes | yes | yes | yes | yes | yes | yes | yes | yes | yes |
| Beam spot size, diameter (mm) | 20 | 20 | 20 | 20 | 20 | 20 | 20 | 20 | 20 | 20 |
| Atmosphere (vacuum, He, air) | vacuum | vacuum | vacuum | vacuum | vacuum | vacuum | vacuum | vacuum | vacuum | vacuum |
| Elements Reported | Na Mg Al Si P S Cl K | Ca Sc | Ti V Cr | Mn Fe Co Ni Cu Zn | Ga As Se Br Rb Hf Ta W Ir Au Hg Tl Pb | Sr Y | Zr Nb Mo | Pd Ag Cd In | Sn Sb | Cs Ba La Ce Sm Eu Tb |

Table 14. XRF Analysis at the ODEQ Laboratory

| Instrument: Kevex771 Software: WinXRF V2.41 | | | | | | | | | | |
|--|---|-----------|---------------|----------------------|---|--|----|----|----|-----|
| | Instrument Conditions for Routine Sample Analysis | | | | | | | | | |
| Parameter | #1 | #2 | #3 | #4 | #5 | #6 | #7 | #8 | #9 | #10 |
| X-ray tube parameters: | | | | | | | | | | |
| Tube voltage (kV) | 7.5 | 35 | 40 | 45 | 40 | 58 | | | | |
| Tube current (mA) | 0.9 | 2.1 | 2.1 | 2.1 | 0.9 | 1.5 | | | | |
| Tube anode material | Rh | Rh | Rh | Rh | Rh | Rh | | | | |
| Direct excitation of sample: | | | | | | | | | | |
| Filter Material | Whatman 41 | na | na | na | Rh | W | | | | |
| Filter thickness (mm) | 1 layer | na | na | na | 0.1 | 0.1 | | | | |
| Secondary excitation of sample: | | | | | | | | | | |
| Secondary Fluorescor | none | Ti | Fe | Ge | none | none | | | | |
| Filter material | na | none | none | none | na | na | | | | |
| Filter thickness (mm) | na | na | na | na | na | na | | | | |
| Acquisition time (seconds) | 400 | 400 | 400 | 400 | 400 | 400 | | | | |
| Energy range acquired (keV) | 10 | 10 | 10 | 10 | 20 | 80 | | | | |
| Number of [MCA] channels | 1024 | 1024 | 1024 | 1024 | 2048 | 4096 | | | | |
| Sample rotation (yes/no) | no | no | no | no | no | no | | | | |
| Beam spot size, diameter (mm) | unknown | unknown | unknown | unknown | unknown | unknown | | | | |
| Atmosphere (vacuum, He, air) | vacuum | vacuum | vacuum | vacuum | vacuum | vacuum | | | | |
| Elements Reported | Na Mg Al Si P | S Cl K Ca | Sc Ti V Cr | Mn Fe Co Ni Cu Zn | Ga As Se Br Rb Sr Y Zr Nb Mo Hf Ta W Ir Au Hg Pb | Ag Cd In Sn Sb Cs Ba La Ce Sm Eu Tb | | | | |

Table 15. XRF Analysis at the RTI Laboratory

| Instrument: ThermoNoran QuanX Software: Wintrace 3.0 Build 35 | | | | | | | | | | |
|--|---|------------------------|---|--|------------------------|----|----|----|----|-----|
| | Instrument Conditions for Routine Sample Analysis | | | | | | | | | |
| Parameter | #1 | #2 | #3 | #4 | #5 | #6 | #7 | #8 | #9 | #10 |
| X-ray tube parameters: | | | | | | | | | | |
| Tube voltage (kV) | 5 | 10 | 30 | 50 | 50 | | | | | |
| Tube current (mA) | 1.20 | 1.98 | 1.66 | 1.00 | 1.00 | | | | | |
| Tube anode material | Rh | Rh | Rh | Rh | Rh | | | | | |
| Direct excitation of sample: | | | | | | | | | | |
| Filter Material | no filter | Graphite | Pd Thin | Pd Thick | Cu Thin | | | | | |
| Filter thickness (g/cm ²) | na | 0.06 | 0.03 | 0.15 | 0.338 | | | | | |
| Secondary excitation of sample: | | | | | | | | | | |
| Secondary Fluorescor | na | na | na | na | na | | | | | |
| Filter material | na | na | na | na | na | | | | | |
| Filter thickness (mm) | na | na | na | na | na | | | | | |
| Acquisition time (seconds) | 300 | 300 | 250 | 200 | 200 | | | | | |
| Energy range acquired (keV) | 0-10 | 0-10 | 0-20 | 0-40 | 0-40 | | | | | |
| Number of [MCA] channels | 512 | 512 | 1024 | 2048 | 2048 | | | | | |
| Sample rotation (yes/no) | no | no | no | no | no | | | | | |
| Beam spot size, diameter (mm) | 9.5mm x 11mm Elipse | 9.5mm x 11mm Elipse | 9.5mm x 11mm Elipse | 9.5mm x 11mm Elipse | 9.5mm x 11mm Elipse | | | | | |
| Atmosphere (vacuum, He, air) | vacuum | vacuum | vacuum | vacuum | vacuum | | | | | |
| Elements Reported | Na Mg P | Al Si S K Ca Sc Cs | Cl Ti V Cr Mn Fe Co Ni Cu Zn La Ce Sm Eu Tb Hf | Ga As Se Br Rb Sr Y Zr Nb Mo Ta W Ir Au Hg Pb | Ag Cd In Sn Sb Ba | | | | | |

Table 16. XRF Analysis at the RTI Laboratory

| Instrument: ThermoNoran QuanX EC Software: Wintrace 3.0 Build 31 | | | | | | | | | | |
|---|---|------------------------|--|--|------------------------|----|----|----|----|-----|
| | Instrument Conditions for Routine Sample Analysis | | | | | | | | | |
| Parameter | #1 | #2 | #3 | #4 | #5 | #6 | #7 | #8 | #9 | #10 |
| X-ray tube parameters: | | | | | | | | | | |
| Tube voltage (kV) | 5 | 10 | 30 | 50 | 50 | | | | | |
| Tube current (mA) | 1.98 | 1.98 | 1.66 | 1.00 | 1.00 | | | | | |
| Tube anode material | Rh | Rh | Rh | Rh | Rh | | | | | |
| Direct excitation of sample: | | | | | | | | | | |
| Filter Material | no filter | Graphite | Pd Thin | Pd Med Thick | Cu Thin | | | | | |
| Filter thickness (g/cm ²) | na | 0.06 | 0.03 | 0.09 | 0.338 | | | | | |
| Secondary excitation of sample: | | | | | | | | | | |
| Secondary Fluorescor | na | na | na | na | na | | | | | |
| Filter material | na | na | na | na | na | | | | | |
| Filter thickness (mm) | na | na | na | na | na | | | | | |
| Acquisition time (seconds) | 300 | 300 | 300 | 300 | 300 | | | | | |
| Energy range acquired (keV) | 0-10 | 0-10 | 0-20 | 0-40 | 0-40 | | | | | |
| Number of [MCA] channels | 512 | 512 | 1024 | 2048 | 2048 | | | | | |
| Sample rotation (yes/no) | no | no | no | no | no | | | | | |
| Beam spot size, diameter (mm) | 10mm x 12mm Ellipse | 10mm x 12mm Ellipse | 10mm x 12mm Ellipse | 10mm x 12mm Ellipse | 10mm x 12mm Ellipse | | | | | |
| Atmosphere (vacuum, He, air) | vacuum | vacuum | vacuum | vacuum | vacuum | | | | | |
| Elements Reported | Na Mg Al Si | P S Cl K Ca Sc | Ti V Cr Mn Fe Co Ni Cu Zn Cs Ba La Ce Sm Eu Tb Hf | Ga As Se Br Rb Sr Y Zr Nb Mo Ta W Ir Au Hg Pb | Ag Cd In Sn Sb | | | | | |

Table 17. XRF Analysis at EPA's NERL Laboratory

| Instrument: Kevex771-EDX Software: LSQEPA v3-2004F (custom software) | | | | | | | | | | |
|---|--|---------------------|----------------------|---|---|--------|----|----|----|-----|
| | Instrument Conditions for Routine Sample Analysis | | | | | | | | | |
| Parameter | #1 | #2 | #3 | #4 | #5 | #6 | #7 | #8 | #9 | #10 |
| X-ray tube parameters: | | | | | | | | | | |
| Tube voltage (kV) | 55 | 55 | 40 | 40 | 40 | 15 | | | | |
| Tube current (mA) | 0.75 | 0.75 | 1.00 | 1.00 | 1.00 | 1.6 | | | | |
| Tube anode material | Rh | Rh | Rh | Rh | Rh | Rh | | | | |
| Direct excitation of sample: | | | | | | | | | | |
| Filter Material | none | none | none | none | none | none | | | | |
| Filter thickness (mm) | none | none | none | none | none | none | | | | |
| Secondary excitation of sample: | | | | | | | | | | |
| Secondary Fluorescor | Zr | Ag | Ge | Fe | Ti | Al | | | | |
| Filter material | none | none | none | none | none | none | | | | |
| Filter thickness (mm) | none | none | none | none | none | none | | | | |
| Acquisition time (seconds) | 200 | 100 | 100 | 200 | 200 | 200 | | | | |
| Energy range acquired (keV) | 20 | 20 | 10 | 10 | 10 | 10 | | | | |
| Number of [MCA] channels | 1024 | 1024 | 1024 | 1024 | 1024 | 1024 | | | | |
| Sample rotation (yes/no) | no | no | no | no | no | no | | | | |
| Beam spot size, diameter (mm) | ~20 | ~20 | ~20 | ~20 | ~20 | ~20 | | | | |
| Atmosphere (vacuum, He, air) | vacuum | vacuum | vacuum | vacuum | vacuum | vacuum | | | | |
| Elements Reported | Cu Zn Ga Ge As Se Br Rb W Pt Au Hg Tl Pb | Rb Sr Y Zr Nb Mo | Cr Mn Fe Co Ni Cu | K Ca Sc Ti V CrCd In Sn Sb Te I Cs Ba La Ce | Al Si P S Cl K Ca Sc Rh Pd Ag Cd In Sn Sb | Na Mg | | | | |

Appendix E

Systems Audit Report

TECHNICAL MEMORANDUM



TO: Dennis Crumpler / OAQPS
FROM: Eric Boswell / NAREL
COPY: Dr. R.K.M. Jayanty, RTI
AUTHOR: Jewell Smiley / NAREL
DATE: November 4, 2005
SUBJECT: RTI Laboratory Audit

Introduction

On July 12, 2005, a laboratory audit was conducted at the Research Triangle Institute (RTI) as part of the QA oversight for the PM_{2.5} Speciation Trends Network (STN). RTI is the prime contractor providing analytical services to support over two hundred field sites collecting speciation samples. The US EPA audit team included Eric Boswell and Jewell Smiley from the National Air and Radiation Environmental Laboratory (NAREL) with Dennis Crumpler and Joann Rice from the Office of Air Quality Planning and Standards (OAQPS). Solomon Ricks and Jeff Lance were also present during the audit as EPA observers. This audit was a routine annual inspection of the laboratory systems and operations required for acceptable contract performance.

Summary of Audit Proceedings

After a brief meeting with the RTI senior staff and supervisors, the audit team separated as necessary to complete specific assignments for the audit process. At least one member of the RTI staff was always available to escort and assist each auditor. The following specific areas on the RTI campus were visited and inspected.

- ✓ Gravimetric Laboratory - Ms. Lisa Greene
- ✓ Organic Carbon/Elemental Carbon (OC/EC) Laboratory - Dr. Max Peterson
- ✓ X-ray Fluorescence (XRF) Laboratory - Dr. William Gutknecht, Ms. Andrea McWilliams
- ✓ Ion Chromatography (IC) Laboratory - Dr. Eva Hardison
- ✓ Sample Handling and Archiving Laboratory (SHAL) - Mr. Jim O'Rourke

Besides the areas mentioned above, interviews were conducted with the following RTI staff.

- ✓ Dr. R.K.M. Jayanty - RTI Services Program Manager
- ✓ Dr. Jim Flanagan - Quality Assurance Manager

✓ Mr. Ed Rickman - Data Management Technical Supervisor

RTI has been analyzing samples from the PM_{2.5} STN since the network began in February of 2000. Members of the audit team were familiar with RTI's current Quality Assurance Project Plan (QAPP) and pertinent SOPs. A report from the previous year's on-site audit was available for reference and followup [see reference 1]. Also available was a 119-page report prepared by RTI which summarized the quality control data and corrective actions during the period July 1 through December 31, 2004. RTI was one of several laboratories to participate in a Performance Evaluation (PE) study earlier in 2005 [see reference 2], and results from that PE study were discussed with RTI staff during the audit. Several experimental activities were also performed during the course of this audit which will be described later within the appropriate section of this report.

Gravimetric Laboratory

The gravimetric laboratory is equipped with two weighing chambers located in building 11. Ms. Lisa Greene is the supervisor of this lab, and she was interviewed by Jewell Smiley and Joann Rice with Solomon Ricks observing. The interviews and inspections were performed to determine compliance with good laboratory practices, the QAPP, and the following SOPs.

- *Standard Operating Procedure for PM_{2.5} Gravimetric analysis* [see reference 3]
- *Standard Operating Procedures for Procurement and Acceptance Testing of Teflon, Nylon, and Quartz Filters* [see reference 4]

Both of the weighing chambers are configured to satisfy conditions of cleanliness, constant temperature, and constant humidity required by the program. Accurate control of climate inside the weighing chamber is important because balance calibration is very sensitive to temperature, and the equilibrated mass of an air filter sample is sensitive to humidity. Mass determination typically proceeds by weighing the Teflon® collection filter before and after the sampling event. The amount of Particulate Matter (PM) captured onto the surface of the filter can be calculated by a simple subtraction of the tare weight from the loaded filter weight.


A few items were hand-carried to the audit from NAREL so that experimental measurements could be made during the audit. Two metallic weights and four Teflon® filter samples were presented to Lisa with a request to weigh each item at least twice during the day. It should be explained that two of the filter samples were loaded with PM_{2.5} captured from the Montgomery air in January of 2005 and two filters were blank. Metallic weights were included in the sample set to represent a very stable reference material for measuring gravimetric mass. All of the test samples were placed into Chamber #1 and given approximately one hour to equilibrate before the first weighing session was performed. Mr. Maurice Gerald was the analyst selected to perform the work using microbalance "C" for all of the measurements. Results are presented in Table 1 along with mass values previously determined at NAREL. Maurice was able to weigh the test samples four times with about an hour separating each weigh session. Table 1 shows good inter-laboratory agreement for all three sample types.

Table 1. Gravimetric Mass Determinations

| Sample ID | Sample Description | NAREL Value Determined on July 7 (mg) | All RTI Values Determined on July 12 | | | |
|------------|--------------------|---------------------------------------|--------------------------------------|-------------|------------|---------------|
| | | | ~11 AM (mg) | ~12 AM (mg) | ~1 PM (mg) | ~2:30 PM (mg) |
| MW05-11331 | metallic wt. | 97.546 | 97.545 | 97.545 | 97.545 | 97.545 |
| MW05-11332 | metallic wt. | 192.422 | 192.421 | 192.421 | 192.421 | 192.420 |
| T05-11317 | loaded filter | 142.486 | 142.482 | 142.482 | 142.482 | 142.483 |
| T05-11318 | loaded filter | 142.826 | 142.823 | 142.823 | 142.824 | 142.824 |
| T05-11322 | blank filter | 141.665 | 141.663 | 141.663 | 141.664 | 141.664 |
| T05-11323 | blank filter | 145.708 | 145.705 | 145.705 | 145.705 | 145.705 |

Two Dickson data loggers were also carried to the audit from NAREL so that independent measurements of temperature and humidity could be recorded during the audit. One of the data loggers was placed into each weighing chamber immediately near RTI's device for measuring the temperature and humidity. Measurements were downloaded from all of the devices at the end of the day, and these data are presented in Figure 1 and Figure 2. Figure 1 shows good agreement between the temperature loggers placed into Chamber #1, but less agreement was observed for the humidity readings. The graph shows that humidity values measured by RTI's device were consistently lower by a small amount. The average relative humidity (RH) recorded by NAREL's device was 36.9 %, and the average RH recorded by RTI's device was 35.7 % during the same period. Both data loggers had an expected accuracy of ± 2 % RH. All of the measurement differences shown in Figure 1 are within the stated accuracy of each logging device.

Figure 1






Figure 2

Figure 2 shows the humidity and temperature data collected inside chamber #2. Both loggers show a dramatic peak in RH at about 9 AM. This surge in humidity was probably due to four extra people entering the chamber during this period. Three auditors and the supervisor entered the chamber at approximately 8:45 AM, and remained inside the chamber for about twenty minutes. The graph for chamber #2 shows humidity values measured by RTI's device consistently above those recorded by NAREL's device. The average RH recorded by NAREL's logger was 34.1 %, and the average RH recorded by RTI's device was 36.5 % during the same period.

Figure 2 shows a noticeable difference in temperature values measured inside chamber #2. The average temperature recorded by RTI's logger was 70.8 °F, and the average temperature recorded by NAREL's logger was 72.5 °F. According to RTI's QAPP, their logger is expected to have an accuracy of ± 2 °C (± 3.6 °F). NAREL's logger is expected to have an accuracy of ± 0.5 °F, and it was certified to provide this level of accuracy about one month before RTI's audit. Although difference between loggers can be seen in Figure 2, none of the temperature and humidity discrepancies are greater than RTI's stated measurement uncertainties.

Figure 3 shows one more comparison. Both of NAREL's data loggers were removed from the weighing chamber at NAREL on July 11, one day before the audit at RTI. Before they were removed, both of the loggers were located immediately near the other inside NAREL's chamber. Figure 3 shows the temperature and humidity data that were recorded by both loggers from midnight to about 8 AM at which time they were removed from the chamber and placed inside a small Igloo® container for transporting to the audit. It is important to realize that NAREL's two loggers were not exactly identical, and the most significant difference can be seen in the humidity measurements. The logger that was used to make measurements in RTI's chamber #1 shows an average RH of 35.9 % while the logger used to make measurements in RTI's chamber #2 shows an average RH of 35.5 % for the same time period. If NAREL's data loggers had been switched during the audit so that each device was placed into the opposite chamber, then the RH comparisons would have shown better agreement [by about 0.4 %] for both of RTI's chambers.




Figure 3

So how good is the temperature and humidity control at RTI? This audit has shown that both chambers were within RTI's stated control limits for temperature (68-73.4°F) and for RH (30-40%) regardless of which device was selected to provide the measurements.

Later during the audit, two Teflon® filters were removed from the SHAL inventory and traveled with the auditors back to NAREL. These two filters were placed into NAREL's weighing chamber for re-equilibration and weighing so that an independent tare mass could be determined for each filter. Those results are presented in Table 2, and excellent agreement was observed for one filter, but poor agreement was observed for the second filter.

Table 2

| Teflon® Filter ID | Filter Description | RTI Tare Mass (mg) | NAREL Tare Mass (mg) | Difference (mg) |
|-------------------|--------------------|--------------------|----------------------|-----------------|
| 12227086 | Inventory Filter 1 | 151.324 | 151.325 | 0.001 |
| 12227075 | Inventory Filter 2 | 151.099 | 151.069 | 0.030 |

NAREL's tare mass for the second filter was thirty micrograms (30 µg) lighter than the tare mass determined at RTI. Effort was made to discover a reason for this discrepancy. The filter identification was verified by checking the bar code label as well as the serial number on the filter itself. Data transcription errors were unlikely since duplicate tare measurements had been made at RTI, and both measurements agreed within one microgram. Four measurements were made at NAREL over the course of eight days, and all of NAREL's measurements agreed within two micrograms. It is possible that a small piece of extraneous contaminating debris was attached to the filter for measurements taken at RTI, and somehow the debris was lost from the filter before measurements were made at NAREL. Other explanations are also possible.

Dialog was initiated between NAREL and RTI to further investigate this significant finding. The auditors learned that corrective actions had been taken by RTI earlier in the year to deal with a defective lot of Teflon® filters. RTI had observed abnormal gravimetric mass results for filters that were supplied to the field sites during March and April of 2005. The problem revealed itself in two ways: (1) a high frequency of negative results was observed in the gravimetric mass results for the trip and field blanks and (2) a high frequency of outliers was observed in the reconstructed mass balance results for loaded filters. Several filters were examined in RTI's optical microscopy laboratory at magnifications of 3.5x to 40x under enhanced lighting. According to RTI's corrective action report, "*crumbs of filter and/or support ring material were found along the support ring. This material flaked easily from the ring with normal handling.*" RTI's report also stated that "*The negative weights may have been caused by loose debris falling off the filters between the initial and final weighings*". RTI's corrective actions included the return of 6000 unused filters to the manufacturer for replacement. RTI also increased their frequency of weighing filters in duplicate. For example, duplicate tare measurements were increased from 10 % to 100 % of the filters, and duplicate post-weighing was increased from 10 % to 30 % of the filters. It may be a coincidence that NAREL's audit finding is very similar to the blank problems described in this corrective action report.

No other deficiencies were observed as a result of this audit. The overall impression of the gravimetric lab was very positive. Earlier in the year RTI's gravimetric lab weighed several samples that were split with NAREL [see reference 2], and all of those results were within advisory limits.

Carbon Analysis Laboratory

Dr. Max Peterson is the technical supervisor of the carbon analysis laboratory located in building 3. Mr. Melville Richards and Mr. Eric Poitras were analysts working in the lab during the audit. Jewell Smiley and Joann Rice conducted this part of the audit. The interviews and inspections were performed to determine compliance with good laboratory practices, the QAPP, and the following SOP [see reference 5].

- *Standard Operating Procedure for the Determination of Organic, Elemental, and Total Carbon in Particulate Matter Using a Thermal/Optical Transmittance Carbon Analyzer.*

New quartz filters must be thermally cleaned before they are delivered to the SHAL, mounted into the appropriate sampler module, and shipped to the field for sample collection. Upon return to the laboratory, each loaded filter must be analyzed using one of the four Sunset instruments set up to run a thermal/optical method specified for all STN samples. The STN method uses a specific heating protocol to thermally remove carbon from the quartz filter material while the optical transparency of the sample is monitored by shining a laser through the sample. The STN method of carbon analysis is described in the RTI's SOP [see reference 5]. RTI currently uses the STN method to report organic carbon (OC) and elemental carbon (EC) the sum of which represents the total carbon (TC). RTI also reports five OC subfractions: OC1, OC2, OC3, OC4, and PyroC. RTI began reporting the OC subfractions in July of 2003 after a new contract was awarded.

Special attention was given to the OC subfractions during the last on-site audit because of concern that the STN thermal protocol might not provide sufficient data quality for the subfractions [see reference 1]. There was concern that the STN method might show poor precision for the

subfractions over time and between instruments. Some of the earliest evidence came from sucrose spikes which are routinely analyzed at RTI as daily calibration checks. The sucrose spikes have shown good precision for the total carbon measurement over time and between instruments, but unfortunately, sucrose shows poor precision for some of the OC subfractions. It was suggested in RTI's last audit report that we need to learn more about the data quality of the OC subfractions. Specifically, we need to learn more about the between-instrument precision. The lab routinely schedules 10 % of the filter samples for a duplicate analysis, but all of the duplicates are analyzed using the same instrument that performed the original analysis. A recommendation was made within the last audit report to change the way duplicates are scheduled so that some of the duplicates are analyzed using a different instrument. Thus far RTI has not implemented this suggestion for the OC/EC lab. As a consequence, the sucrose spikes are the only routine quality control measure of the between-instrument precision.

RTI recently participated in a study that compared OC/EC results from four different labs [see reference 2], and results from this study were discussed during the audit. A sufficient number of PM_{2.5} filter replicates were prepared at NAREL so that each participating lab received an almost identical set of samples, and each set of samples contained blind duplicates. RTI analyzed each filter sample using all four of the Sunset instruments. RTI's results from this study showed good precision for the blind duplicates and good precision among the instruments. RTI's results were virtually indistinguishable from NAREL's results, even when the OC subfractions were compared.

Later during the audit, two quartz® filters were removed from the SHAL inventory and traveled with the auditors back to NAREL. These filters were analyzed at NAREL to determine the amount of total carbon present on each filter. No significant contamination was observed on either filter.

The general impressions of the OC/EC laboratory developed during this audit were very positive. Only one concern was noted. Some of the routine duplicate determinations should be scheduled to collect between-instrument precision data.

X-Ray Fluorescence Analysis

The PM captured onto the surface of the Teflon® filter is not only weighed to determine its mass but is also analyzed to determine its elemental composition using the energy dispersive X-Ray Fluorescence (XRF) technique. The XRF analysis may not proceed before the gravimetric analysis has been completed. Historically RTI has used one of its remote subcontractor laboratories in Oregon to perform the XRF analysis, but since February of 2002, RTI has operated its own local XRF laboratory to provide a larger sample capacity. There are currently two local instruments at RTI and three remote instruments in Oregon that have been approved for analysis of STN samples.

Dr. Bill Gutknecht is responsible for the review of all XRF data, and Ms. Andrea McWilliams is the analyst responsible for operating both of local instruments. They were interviewed by Jewell Smiley and Joann Rice during this part of the audit. The interviews and inspections were performed to determine compliance with good laboratory practices, the QAPP, and the following SOP [see reference 6].

- *Standard Operating Procedure for the X-Ray Fluorescence Analysis of PM_{2.5} Deposits on Teflon Filters.*

The focus of the XRF audit was to discuss those samples that RTI had analyzed as part of a recent inter-laboratory comparison study sponsored by NAREL [see reference 2]. A sufficient number of PM_{2.5} filter replicates were prepared at NAREL so that each participating lab received an almost identical set of filter samples for XRF analysis. NAREL had received the analytical results from all of the participating labs, and had finished comparing the results reported from different labs. All of the labs reported an uncertainty along with every analytical result. Good agreement was observed among the participating labs for most of the elements that were significantly above the reported uncertainty. The most noticeable exception was aluminum. The auditors were anxious to examine some of RTI's raw data spectra, and of particular interest were the spectra from which aluminum results were derived. RTI's spectra that were used to determine the lighter elements contained a significant interference peak which Andrea described as a diffraction peak. The diffraction peak was not fully resolved from aluminum, nor was it fully resolved from silicon. One would expect an interference of this type to increase the uncertainty of aluminum and silicon results. Yet when RTI's uncertainties were compared to those reported from the other labs, RTI's uncertainties were actually smaller. This study has provided some evidence that RTI may be reporting some uncertainties that are too small. Andrea was asked to explain how the uncertainties were calculated at RTI, and she was not certain how some of the components of uncertainty were calculated by the XRF software.

Ion Chromatography (IC) Laboratory

The IC laboratory is located in building 6 where Dr. Eva Hardison is the technical supervisor, and Mr. David Hardison was the analyst on duty during the audit. Both of them were interviewed by Jewell Smiley and Joann Rice for compliance to good laboratory practices, the QAPP, and the following SOPs.

- *Standard Operating Procedures for PM_{2.5} Anion Analysis* [see reference 7]
- *Standard Operating Procedures for PM_{2.5} Cation Analysis* [see reference 8]
- *Standard Operating Procedures for Cleaning Nylon Filters Used for Collection of PM_{2.5} Material* [see reference 9]

The laboratory is equipped with multiple automated Dionex IC instruments and also has access to equipment for cleaning and extracting Nylon® filters. Four IC instruments were set up for anions and two for cations. At the instrument, multilevel calibration curves are established daily, and the calibration is checked by a second source standard. Duplicate injections have been used to evaluate precision, and post spikes have been used to evaluate accuracy. Control charts were available for recent spikes, duplicates, and laboratory blanks.

Later during the audit, two Nylon® filters were removed from the SHAL inventory and traveled with the auditors back to NAREL. These two filters were extracted and analyzed at NAREL to determine trace level ions that might be present on the filters. No ions were detected on either filter above NAREL's method detection limit.

The interviews and inspections made during this part of the audit were very satisfying, and no deficiencies associated with the IC laboratory were observed during this audit.

Sample Handling and Archiving Laboratory (SHAL)

The SHAL is currently located approximately three miles from RTI's main campus. Moving off-campus to this facility was necessary to handle the large number of samples produced by the speciation network. The network currently produces more than 5000 filter samples per month.

The SHAL is organized to be a central point for all laboratory operations. Every sample passes through the SHAL at least twice. Clean air filters are delivered to the SHAL from the analytical laboratories ready to be packaged and delivered to the field sites. Critical bookkeeping is required to insure sample integrity and to make sure that the proper equipment and information is sent to the field in a timely manner. Loaded filters returning from the field are received at the SHAL, removed from the sampler module, logged into the electronic database, and physically delivered back to the analytical laboratories where the final analysis is completed. After the final analysis is completed, each filter sample is maintained inside a refrigerated archive at RTI for up to 5.5 years, and the IC extracts are kept for six months.

The air filter is protected from the time it leaves the SHAL until it is returned from the field. Each air filter must be mounted into an appropriate sampler module to protect it from accidental contamination. Three different types of filters are required for all of the analytical fractions, and four different types of air samplers are currently operated in the field. Different samplers require different filter modules which are expensive and must be cleaned for reuse. It can be readily seen that the SHAL has a critical role for the overall operations. The correct filter must be mounted into the correct module and mailed to the correct field site on schedule. The SHAL maintains direct interaction with the field sites and with the analytical laboratories.

Eric Boswell, Jewell Smiley, Joann Rice, Dennis Crumpler, and Solomon Ricks visited the SHAL during the afternoon portion of the audit. All of the auditors were able to observe a staged demonstration of the filter assembly/disassembled process. This demonstration was planned in advance so that materials would be available. New filters which had been prepared at NAREL were used for the demonstration, and clean Met One SASS modules were supplied by RTI. SASS modules were selected for this demonstration because the majority of states use Met One air samplers at their sites. During the demonstration two Teflon® filters, two Nylon® filters, and two quartz filters were installed into six SASS modules using procedures routinely executed in the SHAL. The modules were immediately disassembled so that the filters could be recovered and placed back into their protective petri slides. Extra filters were brought from NAREL to serve as travel blanks which were not removed from their protective petri slides. All filters were carried back to NAREL for analysis.

Results from the module assembly/disassembly demonstration showed no measurable contamination transferred to the Nylon® filters and no contamination above 0.4 µg/cm² total carbon (4.7 µg/filter) was observed for the quartz filters. Results for the assembled Teflon® filters are shown in Table 3 along with the associated trip blanks and laboratory chamber blanks. No significant level of contamination was transferred to the Teflon® test filters during the demonstration.

Table 3

| Teflon® Filter ID | Filter Description | Tare Mass (mg) | Loaded Mass (mg) | Filter Residue (mg) |
|-------------------|--------------------|----------------|------------------|---------------------|
| T05-11430 | Assembled Filter 1 | 145.396 | 145.394 | -0.002 |
| T05-11431 | Assembled Filter 2 | 145.420 | 145.420 | 0.000 |
| T05-11432 | Trip Blank 1 | 144.909 | 144.907 | -0.002 |
| T05-11433 | Trip Blank 2 | 145.904 | 145.904 | 0.000 |
| T2112375 | Lab Blank 1 | 144.008 | 144.008 | 0.000 |
| T2112400 | Lab Blank 2 | 144.511 | 144.509 | -0.002 |
| T2112425 | Lab Blank 3 | 147.536 | 147.536 | 0.000 |

Other Staff Interviews

Dr. R.K.M. Jayanty, Dr. Jim Flanagan, and Mr. Ed Rickman were interviewed by Eric Boswell and Dennis Crumpler with Jeff Lantz observing. The following topics were discussed.

1. Facility and Equipment
 - a. Facility, Equipment, and Support Services
 - b. Security
 - c. Health and Safety
 - d. Waste Management
2. Organizational Structure and Management Policies
 - a. Personnel
 - b. Job Descriptions and Qualifications
 - c. Training Program and Training Records
3. Quality Assurance
 - a. Standard Operating Procedures
 - b. Performance Evaluation Results and Corrective Action Responses
 - c. Previous Audit Reports and Responses
 - d. Quality Reports to Management
 - e. Quality Control Records and Oversight
 - f. Review Process for QAPP's
 - g. Review Process for Client Data Packages

4. Procurement
 - a. Materials and Equipment
 - b. Services
5. Document Control
 - a. Controlled Document Production
 - b. Document Distribution and Tracking
 - c. Revisions to Control Documents
 - d. Retrieval and Disposal of Outdated Documents
6. Computer Management and Software Control
 - a. Personnel and Training
 - b. Facilities and Equipment
 - c. Procedures
 - d. Security
 - e. Data Entry
 - f. Records and Archives

Conclusions

Observations have been made by the audit team to determine RTI's compliance with good laboratory practices, the QAPP, and SOPs. This audit has produced the following findings, comments, and recommendations.

1. Two Teflon® filters were removed from the SHAL inventory during the audit so that NAREL could experimentally re-measure the tare mass already determined at RTI's gravimetric lab. As shown previously in Table 2, NAREL's tare mass was an alarming 30 micrograms smaller for one of the filters.

Comment: This finding may be an indication of serious problems like the bad filter lot that was discovered several weeks before this audit. According to the corrective action report, the bad filter lot produced negative trip and field blanks. The questionable filter would have produced this effect if it had been utilized as a trip or field blank. RTI should continue to monitor the situation and explore potential reasons for the large variability in blank filters.

2. All of the routine OC/EC duplicates are analyzed using the same instrument that performed the original analysis. This practice was acceptable in the past when the daily sucrose spikes were able to provide evidence of acceptable between-instrument performance. Now that OC subfractions are reported, there is no daily QC that provides the necessary assurance of acceptable between-instrument precision.

Recommendation. RTI should schedule some of the routine OC/EC duplicates for analysis using a different instrument. For example, half of the scheduled duplicates could be analyzed using the same instrument, and the remaining duplicates could be analyzed using one of the available instruments that did not perform the original analysis.

3. As stated earlier, the focus of the XRF audit was to discuss those samples that RTI had analyzed as part of a recent inter-laboratory comparison study sponsored by NAREL [see reference 2]. Results from this study showed aluminum to be the most controversial element reported. This study also showed that RTI generally reported uncertainties which were lower than those reported by the other participating labs. A few spectra were inspected and discussed during the audit. Two specific spectra were selected to be included in the final report for the study. Ultimately the final report included examples of the controversial spectra from all of the labs. The spectra from RTI contain a significant [diffusion peak] interference for aluminum and silicon which was not observed in the spectra from the other labs.

Comment: This observation may not be a problem for RTI's analysis since there is no standard method for calculating XRF uncertainties. However, RTI may want to take a closer look at the way uncertainties were calculated for aluminum and silicon during this study. EPA has recently initiated dialog with all of the speciation labs to learn more about the XRF analysis at each lab, and clearly there is diversity among the different labs. Any progress toward standardizing the XRF analysis is a positive step for the speciation program.

References

1. EPA/NAREL. October 15, 2003. Annual EPA Audit Report of the RTI Chemical Speciation Laboratory Operations. Prepared by the U.S. Environmental Protection Agency's National Air and Radiation Environmental Laboratory. [currently available on the web] <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/audit03.pdf>
2. EPA/NAREL 2005. Technical Memorandum: Experimental Inter-comparison of Speciation Laboratories. Prepared by the U.S. Environmental Protection Agency's National Air and Radiation Environmental Laboratory, Montgomery, AL.
3. RTI 2003. *Standard Operating Procedure for PM2.5 Gravimetric Analysis*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [c u r r e n t l y a v a i l a b l e o n t h e w e b] <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/gravmssop.pdf>
4. RTI 2003. *Standard Operating Procedure for Procurement and Acceptance Testing of Teflon, Nylon, and Quartz Filters*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [currently available on the web] <http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/filtersop.pdf>
5. RTI 2003. *Standard Operating Procedure for the Determination of Organic, Elemental, and Total Carbon in Particulate Matter Using a Thermal/Optical Transmittance Carbon Analyzer*, Environmental & Industrial Sciences Division, Research Triangle Institute,

- Research triangle Park, NC. [currently available on the web]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/ocecsop.pdf>
6. RTI 2003. *Standard Operating Procedure for the X-Ray Fluorescence Analysis of PM2.5 Deposits on Teflon Filters*, Environmental & Industrial Measurements Division, Research Triangle Institute, Research triangle Park, NC. [currently available on the web]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/xrfsop.pdf>
 7. RTI 2003. *Standard Operating Procedure for PM2.5 Anion Analysis*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [c u r r e n t l y a v a i l a b l e o n t h e w e b]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/anionsop.pdf>
 8. RTI 2003. *Standard Operating Procedure for PM2.5 Cation Analysis*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [c u r r e n t l y a v a i l a b l e o n t h e w e b]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/cationsop.pdf>
 9. RTI 2003. *Standard Operating Procedure for Cleaning Nylon Filters for Collection of PM2.5 Material*, Environmental & Industrial Sciences Division, Research Triangle Institute, Research triangle Park, NC. [currently available on the web]
<http://www.epa.gov/ttn/amtic/files/ambient/pm25/spec/nylonsop.pdf>

Appendix F
***Synoptic Summary of 2005 Speciation
Trends and IMPROVE Network Audits***

Draft July 17, 2006

Speciation Network information may be extracted for use in 2005 Data Summary Report

MEMORANDUM

SUBJECT: Synoptic Summary of 2005 Speciation Trends and IMPROVE Network Audits

FROM: Dennis Crumpler, EPA Quality Assurance Lead for Speciation and IMPROVE
Ambient Air Monitoring Group, AQAD (C304-06)

Jeff Lantz, Quality Assurance Lead for Monitoring Field Operations
Office of Radiation and Indoor Air, Las Vegas

TO: Addressees

The following is a digested version of the 2005 Audit Summaries of the Speciation Trends and IMPROVE networks. This version reflects further comments from UC Davis and auditors on the findings of the IMPROVE audits. The most significant change reflects the adverse findings criteria for temperature. Attached is an Excel spreadsheet that compiles and summarizes the findings, from which this synopsis is drawn. We will post the audits on AMTIC for 30 days to allow the Programs to respond to the findings (look for 2005 Speciation/IMPROVE Audit Reports). After 30 days, the audits will be posted for public viewing.

Speciation Trends and State Supplemental Sites

A total of 16 sites and 19 samplers were audited this past year by EPA-certified auditors. **Eleven out of 19 samplers passed all parameter and audit criteria.** Sites that exhibited sampler performance that fell outside of acceptance criteria for measurement quality objectives were as follows:

Significant findings:

At two sites, a total of seven channels exhibited leaks of 0.10 L/min or greater (site with four channels that initially failed corrected the problems on date of the audit):
One siting issue where samplers located < 1 meter apart and one site failed with a safety issue.

Minor findings or parameters that initially failed but were corrected by the operator on site:

- Two sites marginally failed a leak test and were corrected on date of audit.
- One barometric pressure > 10 mm from standard.
- Three clock times > 5 minutes from standard.
- Two sites with Tamb > 2 degrees from the standard.
- One site with Tfil > 2 degrees from the standard.
- Two sites with flow rates outside of 10 percent acceptance criteria, but these were corrected on site, re-audited, and passed.

Next steps:

It was noted that many of the sites were utilized for special studies underway for assessing network and programmatic questions. The goal for the future will be to audit more sites and a higher percentage of routine network sites. A newly initiated, mandatory parameter check report by network operators will be used to identify sites that are good candidates for audits. A program of certifying EPA and State/Local/Tribal auditors and audit procedures will also be initiated in order to incorporate more audit activities into the annual report. This will include a revised training class that reflects updated QAPP provisions, new audit report forms that will lead to fully electronic reporting and implementation of past and latest experiential knowledge. We will be implementing a new online auditor recertification course to save on time and travel expenses for auditors and trainers.

In a period where budgets are shrinking and under severe scrutiny, it is necessary that samplers operating outside of the acceptable parameter ranges will be corrected at the time of the audit. (Mechanical failure that must be corrected by factory service would be the exception.) This procedure will be mandatory and therefore ensure that the sampler is operating within design when the auditor departs.

IMPROVE and IMPROVE Protocol Sites

A total of 34 sites and 35 samplers were audited this past year by EPA-certified auditors. **Five sites out of the 35 passed all parameters and audit criteria.** Sampler performance that fell outside of acceptance criteria for IMPROVE measurement quality objectives were as follows:

Significant findings:

- Two clocks were >±60 minutes from reference standard.
- Two leaks where the **Vacuum** reading < 33mm Hg.
- Two sites where calibration plug was missing, thus creating zero flow through filter (**for 12 weeks**).
- Five flowrates with difference > ±10 percent of theoretical or the flow rate was questionable due to differences between calibration values and the expected design flow.

- Three sites with temperature reading $> \pm 10$ degrees Celsius from reference standard.
- Four flow rates calculated from the vacuum reading were $> \pm 10$ percent of flow measured by standard.
- One reference standard failed during audit of sampler flow rates.
- Three sites reported operator errors with respect to handling filter cassettes or reading the instrument temperature sensor.
- One had flies that were observed in one exposed filter cassette (fly eggs on the filter).
- One site was improperly positioned with respect to an adjacent shelter which caused water to drip onto one module box and then it infiltrated the sample train and sharp cut cyclone. (This has been corrected as of the writing of this memorandum.)

Minor Findings:

- Ten clocks were $> \pm 5$ minutes from reference standard but less than 60 minutes from standard.
- Sixteen sites with a temperature reading $> \pm 2$ degrees Celsius from reference standard but less than 10 degrees.
- One equipment installation-Channel D down tube severely misaligned.

Next Steps:

Several important issues have been identified as action items for the coming year.

1. We have discovered that, in some cases, the calibration may be valid but that the flow rate calculated from the magnehelic and/or vacuum gauge readings may not coincide with the expected design flow rate.
 - In most cases, a satisfactory correlation existed between the flow measured by the reference standard and the flow calculated from calibration factors and readings of the magnehelic and vacuum gauges; however,
 - There are some situations where a site's reference standard flow rate reading did not correlate well with the expected design flow rate. The following example illustrates.

At the Bronx, NY, site the Module C reference flow rate was 26.0 L/m.

The flow rate based on the magnehelic and vacuum readings: 25.1 and 25.4 L/m.

Current audit test $[(26.0-25.1)/26.0] \times 100\% = 3.4\%$; which passes.

Expected design flow rate under the condition during that audit: ~ 22.8 L/m.

Suggested additional test: $[(22.8-26.1)/22.8] \times 100\% = 14.5\%$

- We conclude that another test in the IMPROVE Audit should compare the reference standard flow rate to the design flow rate for that site and channel.

2. The procedure for implementing follow-up actions to parameters or conditions that do not meet acceptance criteria should be revised. It appeared that in several cases of sampler issues, the operator was capable and able to correct the issue. We believe that for many of the problems this will often be the case. A desirable outcome would be for the site to be functioning within all acceptance criteria prior to the conclusion of an audit. We would like to work with folks from IMPROVE to resolve, at the time of the audit, all the problems that are uncovered, so that follow-up audits, which are time consuming and expensive, may be avoided. IMPROVE has decided to include more training to encourage site operators to check and reset clocks as needed on every filter exchange date.
3. A communication tool will be implemented between IMPROVE and the EPA/State/Tribal Audit Team to provide initial audit feedback and preliminary findings. A website is being sought to host a warehouse of currently needed information. The warehouse would contain a list of currently certified auditors, the compendium of magnehelic and vacuum gauge coefficients, site operator names and contact information, and directions to the sites. From this website, certified auditors will be able to retrieve, independently of effort by IMPROVE or EPA, the information needed to audit each site. As the corps of auditors and the certification program solidifies, we will implement the construction of a contemplated audit schedule each year and post it on the website, but room will be made to make a few unscheduled site audits, if necessary, or the opportunity arises due to audits scheduled for speciation trends or (or evolving NCORE) monitoring sites. Notwithstanding all auditors will be required to provide IMPROVE (both UC Davis and the Operator) with at least a 2-weeks notice of the audit date, contingent on the availability of the operator.
4. We have a goal to audit 40-45 IMPROVE and IMPROVE Protocol sites each year. We will realize this goal this year by certifying approximately 10 additional EPA and State/Local/Tribal auditors and recertifying the current corps of 11. The training program will be combined with the speciation trends network auditor training. The training class will be updated yearly to reflect the findings and experience of the auditor corps from the previous year's audits. We will also amend QAPP and SOPs, and revise audit report forms that will lead to fully electronic reporting to reflect the lessons learned. We will develop and initiate (if time permits in 2006, but by the middle of 2007) a new online auditor recertification course to save on time and travel expenses for auditors and trainers.

Addressees:

Bill Malm, NPS

Bret Schictel, NPS

Chuck McDade, UC Davis

David Krask, MD DE (MARAMA)

David Maxwell, NPS

Dr. RKM Jayanty, RTI International

Emilio Braganza, ORIA-LV

Jed Harrison, ORIA-LV
 Joann Rice, AQAD/AAMG
 John Vimont NPSS
 Lindsey DeBelle, CIRRA, Colorado State Univ.
 Marc Pitchford, AQAD/AAMG
 Michael Papp, AQAD/AAMG
 Nicole Hyslop, UC Davis
 Phil Lorang, AQAD/AAMG
 Ray Bishop, OK DEQ (STAPPA)
 Rich Fisher, USDA FS
 Rich Poirot, VT ANR (NESCAUM)
 Robert Lebens, Westar
 Sandra Silva, FWS
 Solomon Ricks, AQAD/AAMG
 Steve Ixquiac, UC Davis
 Tom Curran, AQAD
 Tom Helms, AQAD/AQAG

cc: Brian Timin, AQAD/AQMG
 Chris St. Germaine, EPA Region 1
 Cindy Wike, Colorado Dept of Environment
 David Mintz, AQAD/AQAG
 Donald Poteat, MD Dept of the Environment
 Ed Rickman, RTI International
 Greg Noah, EPA Region 4
 Gregory Harsfield, Colorado Dept of Environment
 James O'Rourke RTI International
 Kenneth Distler, EPA Region 8
 Mark Winter, EPA Region 2
 Neil Frank, AQAD/AQAG
 Norm Beloin, EPA Region 2
 Rich Scheffe, AQAD
 Tom Gronaw MD Dept of the Environment
 Venkatesh Rao, AQAD/AQAG

2005 Summary of Audits

Audits are "Pass (P)" or "Adverse finding (A)" based on audit criteria. "(A)P" means an initial adverse finding which passed after recalibration. "Parameter of concern" box identifies the specific issue for which the audit identified and adverse finding for the site or sampler, e.g., Safety, Siting Criteria (Site), Clock, Module # Leak Check (#A-Leak), Flow rate (Flow), Filter Temperature (Tfil); Calibration Plug Missing (Plug); Magnehelic reading (Mag); Vacuum reading (Vac). Reference Standard Failure (Ref). Operator Procedural Error (OpEr) **RED** denotes failure issue; **BLUE** denotes passed parameter(s) of interest. Parameter passes but value in question, e.g. flow set point [(P)?]. Equipment installation issue (Equip). Filter Contamination (Filt C)

| Monitoring Site Location | General pass/ Adverse finding | Date | Parameters with Adverse findings | AQS Site ID No. | Monitoring Instrument | FLOW RATES | | | | | | | | | | | | Comments and other findings |
|---------------------------|-------------------------------|------------|----------------------------------|-----------------|-----------------------|------------|---------|----------|----------|---|---|---|---|---|----|----|----|--|
| | | | | | | Ch1 (A) | Ch2 (B) | Ch 3 (C) | Ch 4 (D) | 5 | 6 | 7 | 8 | 9 | 10 | 11 | 12 | |
| Atlanta | A | 7/19/2005 | Leak | 130890002 | Andersen 401 RAAS | P | P | P | P | | | | | | | | | Initial leak check failed channels 1-4; |
| Atlanta post Correction | A | 7/19/2005 | Leak | 130890002 | Andersen 401 RAAS | P | P | P | P | | | | | | | | | |
| Atlanta Follow-up Action | (A)P | 7/26/2005 | Leak | 130890002 | | NA | NA | NA | NA | | | | | | | | | Hoses and HEPA replaced on 7/26 and reported by Mactech; Region has not verified |
| Bakersfield (Primary) | P | 5/17/2005 | | 060290014 | Metone | | | | | | | | | | | | | |
| Bakersfield (collo.) | A | 5/17/2005 | Leak | 060290014 | Metone | P | P | P | | | | | | | | | | Sampler displayed leak of 0.1 L/m, but criteria is 0.08 l/m and it was not clear if the sampler really failed |
| Bakersfield (collo.) | A | | Tamb | | | P | P | P | | | | | | | | | | Tamb off by 2.5°C |
| Beacon Hill,WA | P | 3/28/2005 | None | 530330080 | URG 400 | P | | | | | | | | | | | | All paramters for both samplers passed a this site |
| | P | | None | | URG 450 | P | | | | | | | | | | | | |
| Deer Park (Primary) | P | 6/7/2005 | None | | URG 400 | P | | | | | | | | | | | | |
| | (A)P | | Tfil | | URG 450 | P | | | | | | | | | | | | Initial Tfil failure but passed after recalibration |
| Deer Park Colloc. | P | 6/7/2005 | None | 482011039 | URG 400 | P | | | | | | | | | | | | |
| | (A)P | | BP | 482011039 | URG 450 | P | | | | | | | | | | | | Initial Barometric Pressure failed but passed after recalibration |
| Fresno, CA | P | 5/20/2005 | None | 060190008 | Metone SASS | P | P | P | | | | | | | | | | |
| Phoenix | (A)P | 5/24/2005 | Clock | 040139997 | Metone SASS | P | P | P | | | | | | | | | | Initially failed Clock check but passed after recalibration |
| Rubbidoux | (A)P | 5/18/2005 | Tamb, leak, flow | 060658001 | Metone SASS | P | P | P | | | | | | | | | | Site passed but earlier audit (April 05) failed several paramters and routine operator was not there to correct issues |
| Amazon Park, Eugene, OR | A | 6/23/2005 | Siting | SLAM Site | Metone SASS | P | P | P | | | | | | | | | | Sampler was located less than 1 meter from a High volume PM-10 sampler; should be between 2-4 meters |
| NW Nazarene College, | P | 9/20/2005 | None | 160270004 | Metone SASS | P | P | P | | | | | | | | | | |
| Tacoma, WA Unit A | P | 3/31/2005 | None | SPMS | Metone SASS | P | P | P | | | | | | | | | | |
| Tacoma, WA Unit B | P | 3/31/2005 | None | SPMS | Metone SASS | P | P | P | | | | | | | | | | |
| Commerce City - 7101 E | P | 12/20/2005 | None | 080010006 | Metone SASS | P | P | P | | | | | | | | | | |
| New Brunswick, NJ | P | 8/12/2005 | Safety | 034023006 | Metone SASS | P | P | P | | | | | | | | | | Platform for single sampler need repair or Replacement |
| New Brunswick, NJ | P | 8/12/2005 | | 034023007 | Metone SASS | P | P | P | | | | | | | | | | |
| Essex MD | P | 11/29/2005 | None | 240053001 | Metone SASS | P | P | P | | | | | | | | | | |
| Hu-Beltsville, MD | A | 11/19/2005 | Clock | 240330030 | Andersen RAAS 401 | P | P | P | | | | | | | | | | Clock Failed with -6 min reading; no indication of reset; previous 05 audits indicated clock variability |
| Bismarck Residential | P | 8/30/2005 | None | 380150003 | Metone SASS | P | P | P | | | | | | | | | | Sampler Passed all Parameters |
| National Park (TRNP - NU) | p | 8/31/2005 | Leak | 380530002 | Metone SASS | P | P | P | | | | | | | | | | Sampler Failed the Leak test on all three channels and passed all other parameters. |