

PERFORMANCE SPECIFICATIONS 3
O₂ and CO₂

A. Performance Specifications

1. Instrument Zero and Span: See PSP 2c.
2. Calibration Drift: $\leq 0.5\%$ O₂ or CO₂ from the reference value of the gas, gas cell, or optical filter.
3. Relative Accuracy: (Use PDS 3-1.) $\leq 20\%$ of the mean value of the RM test data or 1.0% O₂ or CO₂, whichever is greater.

B. Test Procedure

1. Relative Accuracy Test. See PSP 2.
2. Reference Method. Unless otherwise specified in an applicable subpart of the regulations, Method 3B or any approved alternative.