

PERFORMANCE SPECIFICATIONS 7
Hydrogen Sulfide

A. Performance Specifications

1. Instrument Zero and Span: See PSP 2c.
2. Calibration Drift: $\leq 5\%$ of the established span value from the reference value of the calibration gas or reference source for 6 out of 7 test days (e.g., the established span value is 300 ppm for subpart J fuel gas combustion devices).
3. Relative Accuracy: (Use PDS 3-1) $\leq 20\%$ of the mean value of the RM test data in terms of the units of the emission standard or 10% of the applicable standard, whichever is greater.

B. Test Procedure

1. Relative Accuracy Test. See PSP 2.
2. Reference Method. Method 11, unless otherwise specified in an applicable subpart of the regulation.

